

Amendment No. 2

Date: 15/03/2014

Subject: Amendment to the Tender Enquiry Document

Ref: NIT ref.: HLL/PCD/GNCTD/02/13-14 dated 12/02/2014 read with Amendment No.1 dated 14/02/2014
Tender ID in GNCTD e-Procurement portal: 2014_HFWD_52734

The following changes have been incorporated in the referred tender .

SECTION I**Existing:**

| Description | Total Qty. | Tender Fees (Rs.) | EMD Amount (Rs.) | Date & time of Prebid meeting | Date & time of closing of online tender | Closing date & time for submission of physical Tender | Date & time of opening of tender |
|--------------------------|------------|-------------------|------------------|-------------------------------|---|---|----------------------------------|
| Gel Documentation System | 1 | 500 | 28,000 | 20-02-2014 11:00 PM | 17-03-2014 6:00 PM | 18-03-2014 1:00 PM | 18-03-2014 2:00 PM |

Amended as:

| Description | Total Qty. | Tender Fees (Rs.) | EMD Amount (Rs.) | Date & time of Prebid meeting | Date & time of closing of online tender | Closing date & time for submission of physical Tender | Date & time of opening of tender |
|--------------------------|------------|-------------------|------------------|-------------------------------|---|---|----------------------------------|
| Gel Documentation System | 1 | 500 | 28,000 | 20-02-2014 11:00 PM | 02-04-2014 6:00 PM | 03-04-2014 1:00 PM | 03-04-2014 2:00 PM |

Section – VII
Technical Specifications
Gel Documentation System

| Sl. No | Tender Page No. | Existing Specification | Read As |
|---------------|--|-------------------------------|--|
| 1 | Page No: 1 Amendment No. 1 dated 14/02/2014 Section – VII Technical Specifications | Camera: | Camera: Added Para: 16 Bits with quantum efficiency of minimum 70% should be available. |
| 2 | Page No: 1 Amendment No. 1 dated 14/02/2014 Section – VII Technical Specifications | Illumination: | Illumination: Added Para: Should have White light trasilluminator. |
| 3 | Page No: 1 Amendment No. 1 dated 14/02/2014 Section – VII Technical Specifications | | Added Para: 1. Suitable computer with monitor and printer should be included. 2. Softwares for image acquisition, 1D analysis and 2D analysis should be included. |

All other contents of the tender enquiry including terms & conditions remain unaltered.