

Amendment No: 1 Issued on dated 20/06/2014

Ref: Tender Enquiry No: HLL/PCD/NIB-15/14-15 and Pre-Bid meeting held on 09/06/2014 at HLL Lifecare Ltd., B-14A, Sector-62, Noida - 201 307

The following amendments are issued with respect to the above tender Enquiry Document.

Existing:-

SI No.	Description	Schedule
i.	Dates of sale of tender enquiry documents	31/05/14 to 30/06/14, during 1000 Hrs. to 1600 Hrs. (IST)
ii.	Place of sale of Tender Enquiry Documents	HLL Lifecare Limited, Procurement & Consultancy Services Division B-14 A, Sector-62, Noida-201 307
iii.	Cost of the Tender Enquiry Document	Rs. 2,000/-
iv.	Pre Tender Meeting Date & Time	09/06/14, 1200 Hrs. (IST)
v.	Pre Tender Meeting Venue	Same as given in 2 (ii)
vi.	Closing date & time for receipt of Tender	01/07/2014, 1400 Hrs. (IST)
vii.	Time and date of opening of Techno – Commercial tenders	01/07/14, 1430 Hrs. (IST)
viii	Venue of Opening of Techno Commercial Tender	Same as given in 2 (ii)

Read As:-

SI No.	Description	Schedule
i.	Dates of sale of tender enquiry documents	31/05/14 to 30/06/14, during 1000 Hrs. to 1600 Hrs. (IST)
ii.	Place of sale of Tender Enquiry Documents	HLL Lifecare Limited, Procurement & Consultancy Services Division B-14 A, Sector-62, Noida-201 307
iii.	Cost of the Tender Enquiry Document	Rs. 2,000/-
iv.	Pre Tender Meeting Date & Time	09/06/14, 1200 Hrs. (IST)
v.	Pre Tender Meeting Venue	Same as given in 2 (ii)
vi.	Closing date & time for receipt of Tender	04/07/2014 , 1400 Hrs. (IST)
vii.	Time and date of opening of Techno – Commercial tenders	04/07/14 , 1430 Hrs. (IST)
viii	Venue of Opening of Techno Commercial Tender	Same as given in 2 (ii)

Technical Specifications

Item: HPAEC-PAD Ion Chromatography

Existing: - Para: Autosampler

Autosampler for injection and analysis of anions & cations from same samples &/OR different sample without wasting time for changeover and equilibration.

Read as: - Autosampler for injection and analysis of carbohydrates, mono, di & poly Saccharides.

Existing: - Para: Solvent Delivery Pump – 01 no.

Pump with built-in quaternary low pressure gradient system and non-metallic PEEK base compatible for 0-14 pH & Reverse Phase (RP) compatibility for 4 solvents gradient as per the following specification. Pump should have built-in four channel degasser.

Read as: - Pump with built-in quaternary/ternary low pressure gradient system and non-metallic PEEK base compatible for 0-14 pH & Reverse Phase (RP) compatibility for 4/3 solvents gradient as per the following specification. Pump should have built-in four/three channel degasser.

Existing: - Electrochemical Detector Cell Kit – 01 no.

Para: Counter Electrode : Titanium

Read as: - Counter Electrode : Titanium/Platinum

All other terms and conditions of this tender will remain unchanged.