

AMENDMENT NO.4 Dated 29.05.2020 to Tender, Ref no: Tender No:HLL/SD/RBD/2019-20/Tender/15 Dt:05.03.2020 which has now been closed, fresh request for proposal for the supply of personal protective equipments-Goggles, N95 face mask with/Without Expiratory Valve and Nitrile Gloves floated on 11 May 2020.

The following amendment has been incorporated to the bid document for the above tender:

AMENDMENT NO.4 Dated 29.05.2020 with changes in the specification of Nitrile Gloves. All previous amendments made to the specifications of Nitrile gloves, Latex gloves, Goggles and N95 Mask vide Amendment number 1 dated 17.05.2020, Amendment number 2 dated 22.05.2020 and Amendment number 3 dated 26.05.2020 has been compiled and incorporated in this amendment number 4. The offers being submitted are to be in compliance with the below mentioned specifications for Goggles, Gloves and N95 Masks.

Specification

1. Goggles

1. Made of transparent scratch resistant material, Zero optical power, well fitting, covered from all sides with elastic band/or adjustable holder.
2. Good seal with the skin of the face.
3. Flexible frame to easily fit all face contours without too much pressure.
4. Covers the eyes and surrounding areas and accommodates for prescription glasses.
5. Fog and scratch resistant.
6. Adjustable band to secure firmly so as not to become loose during clinical activity.
7. Indirect venting to reduce fogging.
8. May be re-usable (provided appropriate arrangements for decontamination are in place) or disposable.
9. Quality compliant with the below standards, or equivalent:
 - a. EU standard directive 86/686/EEC, EN 166/2002
 - b. ANSI/SEA Z87.1-2010
 - c. IS 5983-1980, reaffirmed in 2018

2. Gloves

2.a Nitrile Gloves in pairs of Size 6.5 & 7.0 each

- Nitrile
- Non-sterile
- Powder free
- Length: minimum 240 -280 mm total length
- Different sizes in pairs of 6.5 & 7.0 each
- Quality compliant with the below standards, or equivalent:
 - a. EU standard directive 93/42/EEC Class I, EN 455
 - b. EU standard directive 89/686/EEC category III, EN 374
 - c. ANSI/SEA 105-2011
 - d. ASTM D6319-10
- Packing – The Product should be supplied in boxes of 100 pieces (The gloves of size 6.5 & 7.0 shall be individually packed).

OR

2.b Latex Medical Examination Gloves in pairs of Size 6.5 & 7.0 each

- Ambidextrous shape – fits either hand
- Size: Medium and Large in pairs
- Length: minimum 230 mm total length
- Disposable powder free latex gloves
- Single use
- Non Sterile
- Should conform to IS 15354 Part I 2018 or EN455 or ISO 11193 -1 of 2008.
- Expiry: minimum 3 years from date of manufacture
- Label on primary packaging – name and/ or trademark of manufacturer, manufacturer's product reference, expiry date by year and month

3. N-95 Masks (With / Without valve)

- Shape that will not collapse easily
- High filtration efficiency
- Good breathability, with expiratory valve
- Quality compliant with standards for surgical N95 respirator:
 - a. NIOSH N95, EN 149 FFP2, BIS or equivalent
- Fluid resistance: minimum 80 mmHg pressure based on ASTM F1862, ISO 22609, or equivalent.
- Quality Compliant with standards for particulate respirator that can be worn with full face shield

All other relevant specifications and clauses of the tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

**Vice President
Sourcing Division
HLL Lifecare Ltd.
HLL Bhavan, Poojappura,
Thiruvananthapuram.**

Offers may be submitted to:
sdcovidep.hll@lifecarehll.com / sdcovidep.hll@gmail.com
Contact: 0471 2354949 Ext: 299 (9.15 AM to 5.45 PM)