

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018



ITEM: NEEDLE FOR SURGICAL SUTURE

Tender No: HLL/AFT/MTLS/SUT/NEEDLES/2018-20

TECHNO-COMMERCIAL BID

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवनन्तपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

निविदा सूचना
TENDER NOTICE

वर्ष २०१८ - २० के लिए हमारी आक्कुलम फैक्टरी तिरुवनन्तपुरम , १६ जी ४६ एमएम नीडिल कानुल्स ब्लड बाग विनिर्माण /प्राधिकृत एजन्टों से भारत में विनिर्मित उत्पादों के लिए निर्धारितप्रपत्र में दो बोली प्रणाली (तकनीकी / वाणिज्य) के अधीन मुहरबंदित और उपरिलिखित निविदाएँ आमंत्रित किया जाता है।

Sealed and superscribed tenders under Two Bid System (Techno-Commercial) in the prescribed form are invited for products manufactured in India / abroad from Manufacturer(s)/Authorised agents for the supply of **Needle for Surgical Suture** to our Akkulam Factory for the year 2018-20.

तकनीकी विनिर्देश, अनुमानित मात्राएं और अन्य निबंधन एवं शर्तें निविदा दस्तावेजों में दिये जाते हैं, जो प्रत्येक निविदा के लिए एचएलएल लाइफ़केयर के नाम पर तिरुवनन्तपुरम में देय रूपए १०५०.०० का डीडी देकर १६.०२.२०१८ से पूर्वाह्न १०.३० बजे से अपराह्न ३.३० बजे तक हमारे कार्यालय के किसी भी कार्य दिवस पर प्राप्त किया जा सकता है। विदेशी विक्रेता अमेरिकी डालर 15.00 का डीडी देकर प्राप्त किया जा सकता है। यदि दस्तावेज डाक द्वारा अपेक्षित है तो निविदा दस्तावेजों की लागत के अतिरिक्त डाक/ कोरियर चार्ज की ओर रूपए ५०.०० का अतिरिक्त चार्ज भुगतान किया जाए। डी डी एचएलएल लाइफ़केयर के नाम पर तिरुवनन्तपुरम में देय किसी राष्ट्रीयकृत बैंक का होना चाहिए।

The technical specification, approximate quantities and other terms and conditions are given in the Tender documents, which can be had from our office on any working day from 16.02.2018 between 10.30 A.M to 3.30 P.M by paying INR.1050.00 DD (inclusive of GST) for each tender in favour of HLL Lifecare Limited, Akkulam payable at Trivandrum. In case of **overseas bidders** bid documents may be obtained on payment of USD. 15.00 in the form of Demand draft in favour of HLL payable at Trivandrum If the documents are required by Post, an additional charge of Rs.50/- may be remitted towards postal/courier charges, in addition to the cost of tender documents. The DD should be drawn on a Nationalized Bank favouring HLL Lifecare Limited, payable at Thiruvananthapuram.

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

विवरण हमारे वेबसाइट <http://www.lifecarehll.com/tender> & <https://eprocure.gov.in/cppp/> से भी मिल सकता है । निविदाकार हमारे वेबसाइट & सी पी पी पोर्टल से दस्तावेज़ डाउनलोड किया जा सकता है और ऐसी स्थिति में निविदा शुल्क, तकनीकी बोली के साथ डीडी द्वारा अदा किया जाना चाहिए।

The details can also be had from our Website at <http://www.lifecarehll.com/tender> & <https://eprocure.gov.in/cppp/> The tenderers can download the documents from our website and in such case the tender fee should be paid by way of DD along with technical bid.

तकनो-वाणिज्य बोली दस्तावेजों को जारी करने की अंतिम तारीख एवं समय ०५.०३.२०१८ मध्याह्न 1.00 बजे तक होगा। संपूर्ण निविदाएं एचएलएल लाइफ़केयर में निविदा संख्या उपरिलिखित मुहरबंदित लिफाफे में ०६.०३.२०१८, १.०० बजे तक मिलना चाहिए।

The last date & time of issue of Techno-commercial Bid documents will be upto 05.03.2018 : 1.00 PM. Completed tenders in sealed cover superscribed with Tender Nos. should reach HLL, Akkulam latest by 06.03.2018: 1.00 PM.

संक्रमण में हुई किसी विलम्ब के लिए एचएलएल लाइफ़केयर लिमिटेड उत्तरदायी नहीं होगा।

HLL will not be responsible for any delay in transit.

तकनीकी बोली में योग्यताप्राप्त निविदाकारों की मूल्यबोली मात्र खोली जाएगी।

The Price Bid of those Tenderers who qualify in the Technical Bid only will be opened.

मूल्य बोली खोलने की तारीख एवं समय अलग रूप से सूचित किया जाएगा। तकनीकी बोली मूल्यांकन में अयोग्यताप्राप्त निविदाकारों की मूल्यबोली बिना खोले वापस की जाएगी।

The date and time of opening of Price bid will be intimated separately. The Price Bids of Tenderers who do not qualify in the Technical bid evaluation will be returned unopened.

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

नये विक्रेताओं एवं विक्रेता विकास कार्यक्रम के अधीन आनेवालों की मूल्य बोली विक्रेता विकास प्रोटोकॉल में यथा निर्धारित नमूना की प्रस्तुतीकरण एवं संयंत्र परीक्षण की समाप्ति के बाद मात्र खोली जायेगी।

The Price Bid of new vendors and those under vendor development programme will be opened only after submission of sample and successful completion of plant trial as laid down in the vendor development protocol.

JT. GENERAL MANAGER (MATERIALS)

M/s.

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Dear Sirs,

Sub: Tender Forms for Technical Bid & Price Bid

We are enclosing herewith the Techno-Commercial Bid & Price Bid Tender Forms.

The Tender Forms are to be duly filled up in duplicate, signed and returned, in a sealed cover superscribed with the above Tender No., the due date and date of opening, addressed to the JOINT GENERAL MANAGER (MATERIALS), HLL LIFE CARE LIMITED, AKKULAM FACTORY, SREEKARIYAM P.O., THIRUVANANTHAPURAM-695017.

Earnest Money Deposit (EMD) – Rs 100000.00 / USD 1500.00

Tender Fee – Rs. 1050.00 / USD 15.00

The last date & time of receipt of Bids : 06.03.2018; 1.00 P.M

Date & Time of opening of Technical Bid : 06.03.2018; 3.00 P.M

The details can also be had from our Website at www.lifecarehll.com

Incomplete tender is likely to be rejected without any further correspondence.

Yours faithfully,
For HLL LIFE CARE LIMITED.

JOINT GENERAL MANAGER (MATERIALS)

Encl: As above

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

CONTENTS OF BID DOCUMENT

I. INSTRUCTION TO BIDDERS:

II. TECHNICAL BID FORMS

1. Schedule A - Specification & Quantity
2. Schedule B - Questionnaire for minimum eligibility criteria
3. Schedule C - Questionnaire - General information of the
Manufacturer / Supplier
4. Schedule D - Terms & conditions
5. Schedule E - Questionnaire for Vendor Development
6. Schedule F - Declaration accepting Terms and Conditions of
the Tender

III. PRICE BID FORMS

1. SCHEDULE I - RATE SCHEDULE FOR THIRUVANANTHAPURAM (For Domestic suppliers)
2. SCHEDULE II - RATE SCHEDULE FOR CIF THIRUVANANTHAPURAM (For Overseas Bidders)

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Page 1 of 2

INSTRUCTIONS TO BIDDERS

- a) The Bid is intended to procure material as per specification in Schedule A
- b) The Bid consist of two parts:
a. Technical Bid
b. Price Bid
- c) The Technical Bid should be accompanied by **Earnest Money Deposit (EMD)** in the form of DD drawn in favour of M/s. HLL Lifecare Limited payable at Trivandrum for a sum of Rs. 1,00,000/- and in case of foreign suppliers USD. 1500.00, failing which the Tender is likely to be summarily rejected. The EMD submitted will be converted to Security Deposit and will be released at the end of contract period. **SSI/MSME** units interested in availing exemption from payment of Tender Fee & EMD should submit a valid copy of their registration certificate issued by the concerned DIC or **NSIC / UDYOG AADHAR**. But the Party has to provide Security deposit if Tender is awarded to them.
- d) Technical & Price Bids should be in separate sealed covers. Technical and Price Bids shall be superscribed as "**Technical Bid** " and "**Price Bid**" on the respective covers in order to clearly identify between the two Bids. These two bids should be submitted in a single sealed cover with the respective Tender No. mentioned thereon, complete in all respect, addressed to the Joint General Manager (Materials) HLL Lifecare Limited , Akkulam Factory, Sreekariyam P.O., Trivandrum - 695 017, on or before the due date and time mentioned in the Tender Notification. HLL shall not be responsible for any delay, if any, in the delivery of the bidding document or non-receipt of the same.
- e) The validity of technical bid for two years (2018-20) and price bid is for one year (2018-19).
- f) **Bid submission / Bid opening dates**
The last date & time of receipt of Bids : 06.03.2018; 1.00 PM
Date & Time of opening of Technical Bid: 06.03.2018; 3.00 PM
- g) In the event of the date mentioned above being declared subsequently as holiday for the purchaser's office, the due date for submission and opening of bids will be the next working day at the same venue and time.

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Page 2 of 2

INSTRUCTIONS TO BIDDERS

- h) Bids received after the deadline for submission shall not be considered.
- i) The Bidder is expected to examine all specifications, Instructions, Forms, Terms and conditions given in the Bidding documents. Failure to furnish all information required in the Bidding documents or submission of a Bid not substantially responsive to the bidding documents in every respect will be at the Bidders risk and may result in rejection of the Bid
- j) Any clarification required will have to be obtained one week prior to the Date of opening of the Technical Bid.
- k) A certificate/Declaration as given in Schedule F stating that ALL TERMS AND CONDITIONS of this Tender is acceptable should accompany the tender failing which the tender is likely to be summarily rejected.
- l) The quantity mentioned herein are approximate annual requirement with staggered delivery and in case less or more quantity is required the suppliers should be prepared to effect supply at short notice on the same terms and conditions.
- m) **(i) The Price Bid of those Tenderers who qualify in the Technical Bid only will be opened. The date and time of opening of price bid will be intimated separately.**
- (ii) The Price Bid of new vendors (Schedule E) and those under vendor development programme will be opened only after submission of sample and successful completion of plant trial as laid down in the vendor development protocol.**
- (iii) The Price Bids of Tenderers who do not qualify will be returned unopened.**
- n) Annual purchase quantity will be shared to the suppliers @ ratio, as below: 60% to L1 party, and balance 40% shall be allotted among the vendors who matching the L1 rate

.....

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE- A

TECHNICAL BID

SPECIFICATION

<i>Material Number</i>	<i>Material Description</i>	<i>Quantity</i>
		<i>pcs</i>
1010420000	NEEDLE-36MM, 1/2 CRB (C.T), PGA 1	1200
	Specifications: Taper ratio: 1:12 (Increased taper length) USP Size: PGA 1 (Undyed) Wire diameter: 0.92+/-0.01mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420001	NEEDLE-16MM 1/2 CRB, HEAVY, PGA 2/0	2400
	USP Size: PGA 2/0 Bore diameter: 0.42+/-0.01mm Wire diameter: 0.66+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420002	NEEDLE-20MM 1/2 CRB, HEAVY, PGA 2/0	1200
	USP Size: PGA 2/0 Bore diameter: 0.42+/-0.01mm Wire diameter: 0.78+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420003	NEEDLE-26MM 1/2 CRB, PGA 2/0	1200
	USP Size: PGA 2/0 Bore diameter: 0.42+/-0.01mm Wire diameter: 0.70+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420004	NEEDLE -36MM, 1/2 CRB (C.T), PGA 2/0	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

<i>Material Number</i>	<i>Material Description</i>	<i>Quantity</i>
		<i>pcs</i>
	Specifications: USP Size: PGA 2/0 (Undyed) Taper point: 1:12 (Increased taper length) Wire diameter: 0.81+/-0.01mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420 Ribbed&siliconised	
1010420005	NEEDLE-30MM, 1/2 CRB (C.T), PGA 2/0	1200
	Specifications: USP Size: PGA 2/0 (Undyed) Taper ratio: 1:12 (Increased taper length) Wire diameter: 0.81+/-0.01mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420006	NEEDLE-30MM, 1/2 CTC, USP 2/0	2400
	Specification: USP size: PGA 2/0 (Braided) Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420007	NEEDLE-3/8 CIRCLE CUTTING 19 MM, PGA 3-0	1200
	1. 3/8 circle cutting, 19mm, Needle for PGA suture undyed braided & coated, Usp Size:3-0. 2. Bore diameter: 0.31 to 0.35 3. Wire diameter: 0.58 to 0.62mm 4. Material: High quality martensitic stainless steel grade AISI-420. Ribbed&siliconised	
1010420008	NEEDLE-16MM 1/2 CRB, HEAVY, PGA 3/0	1200
	USP Size: PGA 3/0 Bore diameter: 0.33+/-0.01mm Wire diameter: 0.61+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420009	NEEDLE-30MM 1/2 CRB, HEAVY, PGA 3/0	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	USP Size: PGA 3/0 Bore diameter: 0.33+/-0.01mm Wire diameter: 0.71+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420010	NEEDLE-40MM 1/2 CRB, HEAVY PGA 3/0	1200
	USP Size: PGA 3/0 Bore diameter: 0.33+/-0.01mm Wire diameter: 0.78+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420011	NEEDLE-12MM 5/8 CRC, PGA 4/0	1200
	USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.51+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420012	NEEDLE-13MM 3/8-CRC, PGA 4/0	1200
	USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.51+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420013	NEEDLE-13MM 5/8 CRC, PGA 4/0	1200
	USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.51+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420014	NEEDLE-16MM 3/8 CRC, PGA 4/0	1200
	USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.54+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420015	NEEDLE-16MM 5/8 CRC, PGA 4/0	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.54+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420016	NEEDLE-30MM 1/2 CRB, HEAVY PGA 4/0	1200
	USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.61+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420017	NEEDLE-40MM 1/2 CRB, HEAVY PGA 4/0	1200
	USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.70+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420018	NEEDLE-12MM 5/8 CRC, PGA 5/0	1200
	USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.46+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420019	NEEDLE-13MM 3/8 CRC, PGA 5/0	1200
	USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.46+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420020	NEEDLE-13MM 5/8 CRC, PGA 5/0	1200
	USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.46+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420021	NEEDLE-16MM, 3/8 CRC, PGA 5/0	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	NEDLE 16MM, 3/8 CIRCLE REVERSE CUTTING FOR PGA 5/0 SUTURES USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.46+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420022	NEEDLE-16MM 5/8 CRC, PGA 5/0	1200
	USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.46+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420023	NEEDLE-20MM 1/2 CRB, HEAVY, PGA 5/0	1200
	USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.54+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420 Ribbed&siliconised	
1010420024	NEEDLE-30MM 1/2 CRB, HEAVY, PGA 5/0	1200
	USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.54+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420025	NEEDLE-40MM 1/2 CRB, HEAVY, PGA 5/0	1200
	USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.72+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420 Ribbed&siliconised	
1010420026	NEEDLE-16MM 1/2 CRB, HEAVY, PGA 0	1200
	USP Size: PGA 0 Bore diameter: 0.47+/-0.01mm Wire diameter: 0.70+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420 Ribbed&siliconised	
1010420027	NEEDLE-20MM, 1/2 CRB, HEAVY PGA 0	1200
	USP Size: PGA 0 Bore diameter: 0.47+/-0.01mm Wire diameter: 0.84+/-0.02mm	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

<i>Material Number</i>	<i>Material Description</i>	<i>Quantity</i>
		<i>pcs</i>
	Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420028	NEEDLE-40MM 1/2 CRC, PGA 0 USP Size: PGA 0 Bore diameter: 0.47+/-0.01mm Wire diameter: 1+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	6000
1010420029	NEEDLE-36MM 1/2 CRC, PGA 1 USP Size: PGA 1 Bore diameter: 0.55+/-0.01mm Wire diameter: 1+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200
1010420030	NEEDLE-36MM, 1/2 CRB (C.T), PGA 0 USP Size: PGA 0 (Undyed) Taper ratio: 1:12 (Increased taper length) Wire diameter: 0.84+/-0.01mm Bore diameter: 0.47+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200
1010420031	BLUNT NEEDLE-40MM, 1/2 CRB, PGA 1 Blunt needle , 40mm, 1/2 circle round body, USP Size: PGA 1 (braided) Wire diameter: 0.92+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200
1010420032	BLUNT NEEDLE-40MM, 1/2 CRB, PGA 0. SPECIFICATION: Blunt needle 40mm, 1/2 circle round body. USP size: PGA 0 (braided) Wire diameter: 0.84+/-0.02mm Bore diameter: 0.47+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200
1010420033	NEEDLE-30 MM, 1/2 CRB, PGA 0	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	SUTURE SIZE: 0, 30mm,1/2circle,round body Bore diameter: 0.47+/-0.01mm. Wire diameter: 0.84+/-0.02mm. Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420034	NEEDLE-40 MM 1/2 CRB, PGA 0	12000
	SUTURE SIZE: 0 40mm,1/2circle,round body Wire diameter: 1+/-0.02mm Bore diameter: 0.47+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420035	NEEDLE-40 MM, 1/2 CTC, PGA 0	3000
	SUTURE SIZE:1/0 40mm,1/2circle,taper cut Wire diameter: 1+/-0.02mm Bore diameter: 0.47+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420036	NEEDLE-40 MM 1/2 CIRCLE CUTTIN	1200
	needle for PGA suture, 40mm,1/2 circle,cutting, suture size-1/0 (USP) Wire diameter: 0.92(-0.0+0.025)mm Bore diameter: 0.47 (-0.01+0.02)mm Material:high quality martensitic steel Grade:AISI-420 Ribbed&siliconised	
1010420037	NEEDLE-36MM, 1/2 CIRCLE,CUTTI	1200
	suture size:1/0 Wire diameter: 0.84(-0.0+0.025)mm Bore diameter: 0.47 (-0.01+0.02)mm Materil:High quality Martensitic steel,Grade;AISI-420 Ribbed&siliconised	
1010420038	NEEDLE-36MM, 1/2 CRC, PGA	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Wire diameter: 0.84(-0.0+0.025)mm Bore diameter: 0.47 (-0.01+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420. Ribbed&siliconised	
1010420039	NEEDLE-36MM, 1/2 CRC, OS needle,PGA 0	1200
	1. Bore diameter (mm): 0.47+/-0.02mm 2. Wire diameter (mm): 1.20+/-0.05mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420040	NEEDLE-40MM, 1/2 CRB (C.T), PGA 1	1200
	Specifications: Taper ratio: 1:12 (Increased taper length) USP Size: PGA 1 Wire diameter: 0.92 +/- 0.01mm Bore diameter: 0.55 +/- 0.01mm Material:High Quality martensitic stainless steel Grade AISI-420 Ribbed&siliconised	
1010420041	NEEDLE-40MM, 1/2CIRCLE CUTTING, USP	1200
	Wire diameter: 0.99(-0.0+0.025)mm Bore diameter: 0.55(-0.01+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420. Ribbed&siliconised	
1010420042	NEEDLE-30 MM 1/2 CRB, USP 2-0	24000
	SUTURE SIZE:2/0 30mm,1/2circle,round body Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420043	NEEDLE-40 MM 1/2 CRB, PGA 2-0	24000
	SUTURE SIZE:2/0 40mm,1/2circle round body. Wire diameter: 0.84+/-0.02mm Bore diameter: 0.42+/-0.02mm.\nMaterial:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420044	NEEDLE-40 MM 1/2 CRC, PGA 2-0	24000

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	SUTURE SIZE:2/0 40mm,1/2circle,reverse cutting Wire diameter: 0.84+/-0.01mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420045	NEEDLE-26MM, 3/8 CIRCLE,CUTTING, USP 2-0	1200
	Specification: Item name: Needle: NEEDLE-26MM, 3/8 CIRCLE CUTTING, USP 2-0-Ribbed & siliconised. 1.Wire diameter: 0.61+/-0.01mm 2. Bore diameter: 0.42+/-0.01mm 3. Suture material: Polyglycolic Acid, USP 2-0 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420046	NEEDLE-26MM 1/2 CRB, USP 2/0	1200
	Wire diameter: 0.81(-0.0+0.02)mm Bore diameter: 0.42(-0.01+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420. Ribbed&siliconised	
1010420047	NEEDLE-36MM, 1/2 CRB, USP 2-0	1200
	Specifications: USP size: PGA 2-0 Wire diameter: 0.84+/-0.02mm Bore diameter: 0.42+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420048	NEEDLE-30MM, 1/2 CRC, USP 2/0	1200
	Specification: USP size: PGA 2/0 (braided) Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420049	NEEDLE-27MM, 5/8 CTC, USP 2-0	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	SPECIFICATION: Needle: 27mm, 5/8 circle taper cutting. Suture size: PGA 2-0. Wire diameter: 0.71+/-0.02mm Bore diameter: 0.42+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420050	NEEDLE-20 MM, 1/2 CRB, USP 3-0	1200
	SUTURE SIZE:3/0 20mm,1/2circle,round body Wire diameter: 0.62+/-0.01mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420051	NEEDLE-22 MM, 3/8 CIRCLE CUTTING,USP 3-0	2400
	SUTURE SIZE:3/0 22mm,3/8circle,cutting Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420052	NEEDLE-12 MM 3/8 CRC, USP 3-0	3600
	SUTURE SIZE:3/0 12mm,3/8circle,reverse cutting Wire diameter: 0.54+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420053	NEEDLE-22MM,1/2 CIRCLE CUTTING, USP 3-0	1200
	suture size:3/0(USP) Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420054	NEEDLE-17MM,1/2 CTC, USP 3-0	1200
	SUTURE SIZE:3/0 Wire diameter: 0.60+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420055	NEEDLE-36MM, 1/2 CTC, USP 3-0	1200
	SUTURE SIZE:3/0 Wire diameter: 0.71+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420056	NEEDLE-16MM, 3/8 CIRCLE CUTTING, USP 3-0	3600
	Suture size:3/0 Wire diameter: 0.54+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High quality Martensitic grade,stainless steel Grade:AISI-420 Ribbed&siliconised	
1010420057	NEEDLE-26MM,1/2 CRB, USP 3/0	1200
	Wire diameter: 0.71(-0.01+0.02)mm Bore diameter: 0.33(-0.0+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420. Ribbed&siliconised	
1010420058	NEEDLE-36MM, 1/2 CRB, USP 3-0	1200
	Specifications: USP size: PGA 3-0 Wire diameter: 0.71+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420059	NEEDLE-12 MM, 3/8 CRC, USP 4-0	1200
	SUTURE SIZE:4/0 12mm,3/8circle,reverse cutting Wire diameter: 0.54+/-0.02mm Bore diameter: 0.28+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	
1010420060	NEEDLE-20 MM 1/2 CRB, USP 4-0	2400
	"Specification: Needle: 20mm, 1/2 circle round body Wire diameter: 0.54 +/- 0.01mm Bore diameter: 0.28 +/- 0.01mm Suture size: polyglycolic acid USP 4/0 Material: High Quality martensitic stainless steel Grade AISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Ribbed&siliconised	
1010420061	NEEDLE-16MM 1/2 CRB, USP 4/0	1200
	Specification: Suture USP Size: 4/0 Wire diameter: 0.54+/-0.02mm Bore diameter: 0.28+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420062	NEEDLE-16MM, 5/8 CIRCLE CUTTING, USP 4-0	1200
	suture size:4/0 Material:high quality Martensitic grade stainless steel Grade;AISI-420 Bore dia 0.28 +/- -0.02mm Wire dia : 0.48 to 0.51mm	
1010420063	NEEDLE-26MM, 3/8 CIRCLE,CUTTING, USP	1200
	Wire diameter: 0.56 +/- 0.01mm Bore diameter: 0.28 +/- 0.01mm suture size-4/0 material:high quality martensitic stainless steel Grade:AISI-420	
1010420064	NEEDLE-12 MM 3/8 CRC, USP 5-0	1200
	SUTURE SIZE:5/0 12mm,3/8circle,reverse cutting Wire diameter: 0.48+/-0.02mm Bore diameter: 0.22+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420065	NEEDLE-16MM,1/2 CRB, USP 5-0	2400
	suture size:5/0(USP) Wire diameter: 0.48+/-0.02mm Bore diameter: 0.22+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420066	NEEDLE-16MM, 3/8 CIRCLE,CUTTING, USP 5/0	1200
	Wire diameter: 0.48+/-0.02mm Bore diameter: 0.22+/-0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	
1010420067	NEEDLE-30mm, 1/2 CRB, USP 0	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Specification: USP Size: PGA 0 (Braided) Bore diameter: 0.47+/-0.01mm. Wire diameter: 0.84+/-0.02mm. Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420068	NEEDLE-45mm, 3/8 CRB, USP 1	1200
	Specification: USP Size: PGA 1 (Braided) Wire diameter: 1.00+/-0.02mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420069	NEEDLE-30mm, 1/2 CRB, USP 2/0	1200
	Specification: USP Size: PGA 2-0 (Braided) Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420	
1010420070	NEEDLE-40mm, 1/2 CRB USP 2/0	1200
	Specification: USP Size: PGA 2/0 (Braided) Wire diameter: 0.84+/-0.02mm Bore diameter: 0.42+/-0.02mm. Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420071	NEEDLE-40 MM, 1/2 CRB, USP 1	48000
	NEEDLE FOR PGA SUTURE(BRAIDED) 40 MM, 1/2 CIRCLE ROUND BODY SUTURE SIZE:1 (40mm,1/2circle,round body(heavy)) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420072	NEEDLE- 40 MM, 1/2 CTC, USP 1	24000
	NEEDLE FOR PGA SUTURE(BRAIDED) 40 MM, 1/2 CIRCLE TAPER CUT SUTURE SIZE:1 (40mm,1/2circle,taper cut(heavy)) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420073	NEEDLE-45 MM, 3/8 CRB, USP 1	3000

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	SUTURE SIZE:1 45mm,3/8circle,round body Wire diameter: 1+/-0.02mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420074	NEEDLE-50 MM 1/2 CRB, USP 1	3000
	Specification: Item name: Needle: 50mm, ½ circle round body, heavy. -Ribbed & siliconised. 1.Wire diameter: 1.2+/-0.01mm 2. Bore diameter: 0.55+/-0.01mm 3. Suture material: Polyglycolic Acid, USP 1 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420075	NEEDLE-50MM, 1/2 CRC,USP1	1200
	Specification: suture size:1(usp) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420076	NEEDLE-40MM,1/2 CIRCLE,CUTTING,USP 1	1200
	suture size-1, Wire diameter: 0.99(-0.0+0.025)mm Bore diameter: 0.55(-0.01+0.02)mm Material:high quality Martensitic stainless steel,Grade-AISI-420	
1010420077	NEEDLE-40MM,1/2 CRC, USP 1	24000
	1. Bore diameter (mm): 0.55+/- 0.02MM 2. Wire diameter (mm):1.15+/-0.05mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420078	NEEDLE-36mm, 1/2 CRB, USP 1	1200
	Specifications: USP size: PGA 1 Wire diameter: 1.15+/-0.02mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420	
1010420079	NEEDLE-36mm, 1/2 CRB, USP 0	1200
	Specifications: USP size: PGA 0 Wire diameter: 1.00+/-0.02mm Bore diameter: 0.47+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420080	NEEDLE- 40MM 1/2 CIRCLE CUTTING, USP 0	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Wire diameter: 0.92(-0.0+0.025)mm Bore diameter: 0.47 (-0.01+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	
1010420081	NEEDLE-40MM, 1/2 CRB, HEAVY, USP 1	1200
	Specification: Needle: 40mm, 1/2 circle round body, heavy. Suturesize: PGA Braided ,USP 1. Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality stainless steel Grade 300.	
1010420082	NEEDLE-40MM, 1/2 CTC, HEAVY, USP 1	1200
	Specification: Needle: 40mm, 1/2 circle taper cutting, heavy. Suture size: PGA (Braided), USP 1 Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality stainless steel Grade 300.	
1010420083	NEEDLE-45MM, 3/8 CIRCLE CUTTING, USP 3/0	1200
	Wire diameter: 0.71+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High quality martensitic stainless steel grade AISI-420.	
1010420084	NEEDLE-16MM, 3/8 CRB, SILK 3/0 SUTURES	1200
	Suture Size: 3/0 Wire diameter: 0.54+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420085	NEEDLE-26MM, 1/2 CRC, SILK 3/0	1200
	Suture size: 3/0 Wire diameter: 0.71+/-0.01mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420087	NEEDLE12MM, 3/8 CRB, SILK 5/0	1200
	Suture size: 5/0 Wire diameter: 0.48+/-0.02mm Bore diameter: 0.23+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420088	NEEDLE-30 MM 1/2 CRB, SILK 0	6000

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	SUTURE SIZE: 0 30mm,1/2circle,round body Wire diameter: 0.92+/-0.01mm Bore diameter: 0.48+/-0.01mm. Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420089	NEEDLE-16MM,1/2 CRB, USP 4/0 Wire diameter: 0.56+/-0.02mm Bore diameter: 0.28+/-0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	1200
1010420090	NEEDLE-45MM, 3/8 CRC, USP0 Specification: Suture size: Silk 0 Wire diameter: 1+/-0.01mm Bore diameter: 0.47+/-0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	36000
1010420091	NEEDLE-30 MM, 1/2 CRB SUTURE SIZE:2/0 30mm,1/2circle,round body Wire diameter: 0.80+/-0.01mm Bore diameter: 0.42+/-0.01mm. Material:High Quality martensitic stainless steel Grade AISI-420.	75000
1010420092	NEEDLE-45 MM 3/8 CRC, SILK 2-0 Specifications: Wire diameter: 1mm+/-0.02mm Bore diameter: 0.42+/-0.01mm SILK SUTURE: 2/0 45mm, 3/8circle,reverse cutting Material: High Quality Martensitic stainless steel Grade AISI-420	75000
1010420093	NEEDLE -25 MM 1/2 CTC, SILK 2-0 SILK SUTURE: 2/0 25mm, 1/2circle,taper cutting 1. Bore diameter (mm): 0.40+/-0.02mm 2. Wire diameter (mm): 0.76+/-0.05mm. Material: High Quality Martensitic stainless steel Grade AISI-420.	3000
1010420094	NEEDLE-50MM,1/2 CRC SILK 2-0 Bore diameter 0.40 +/- 0.02 mm Wire diameter 1.10 +/- 0.05 mm suture size:2/0 Material :High Quality martensitic stainless steel Grade AISI-420.	3000
1010420095	NEEDLE-1/2 CTC,17MM,SILK 2-0	3000

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Item name: NEEDLE-17 mm, 1/2 circle taper cutting. Silk USP 2/0 Siliconised and Ribbed. 1. Wire diameter: 0.61 (-0.0+0.025)mm 2. Bore diameter: 0.42 (-0.01+0.02)mm 3. Suture material: Silk, USP 2/0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AISI-420	
1010420096	NEEDLE-STR,RB,40MM,SILK2/0	6000
	Item name: NEEDLE-Straight, 40 MM. Silk USP 2/0 Siliconised 1. Wire diameter: 0.92 (-0.0+0.025)mm 2. Bore diameter: 0.42 (-0.01+0.02)mm 3. Suture material: Silk, USP 2/0 (Braided) 4. Material: High Quality martensitic stainless steel grade AISI-420	
1010420097	NEEDLE-36MM, 3/8CTC,SILK-2-0	1200
	Specifications: USP size: Silk 2-0 Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420098	NEEDLE-25 MM 1/2 CRB SILK	2400
	Specifications: SUTURE SIZE:3/0 Wire diameter: 0.70+/-0.01mm Bore diameter: 0.33+/-0.01mm 25mm,1/2circle,round body Material :High Quality martensitic stainless steel Grade AISI-420.	
1010420099	NEEDLE-20 MM,1/2, CTR,SILK 3/0	24000
	SUTURE SIZE:3/0 20mm,1/2circle,round body Wire diameter: 0.61 +/- 0.02mm Bore diameter: 0.33 +/- 0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420100	NEEDLE-26 MM 3/8 CRC SILK 3/0	60000
	SUTURE SIZE:3/0 26mm,3/8circle,reverse cutting Material :High Quality martensitic stainless steel Grade AISI-420. Resistance to corrosion :Corrosion resistant .Bore dia : 0.33 +/- 0.01 mm Wire dia : 0.68 to 0.70mm	
1010420101	NEEDLE-16 MM 3/8 CIRCLE CUTTING	60000

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Specifications: Wire diameter: 0.54+/-0.01mm Bore diameter: 0.33+/-0.01mm SUTURE SIZE:3/0 16mm,3/8circle,cutting Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420102	NEEDLE -1/2 CRB SILK 3/0	3000
	suture size; Silk 3/0, 35mm 1/2circle ,round body Braided Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420103	NEEDLE-36MM, 3/8 CTC SILK 3-0	1200
	Specifications: USP size: Silk 3-0 Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material :High Quality martensitic stainless steel Grade AISI-420.	
1010420104	NEEDLES -16 MM 3/8 CRB SILK4/0	2400
1010420105	NEEDLE - 16 MM 3/8 CC SILK 4/0	30000
	SUTURE SIZE:4/0 16mm,3/8circle,cutting Wire diameter: 0.56+/-0.02mm Bore diameter: 0.28+/-0.01mm. Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420106	NEEDLE -1/2 CTC17MM SILK 4/0	2400
	Wire diameter: 0.56 (-0.0+0.02)mm Bore diameter: 0.28 (-0.01+0.02) mm. Material:High Quality martensitic stainless steel Grade AISI-420	
1010420107	NEEDLE -1/2 CRB 16MM SILK 4/0	1200
	Specifications: suture size 4/0 Wire diameter:0.54 +/-0.01mm Bore diameter: 0.28+/-0.01mm 16mm, 1/2 C, Round body Material :High Quality martensitic stainless steel Grade AISI-420.	
1010420108	NEEDLE-5/0 12 MM 3/8 CRB SILK 5/0	2400
	SILK SUTURE: 5/0 12mm, 3/8circle,round body Wire diameter: 0.46+/-0.02mm Bore diameter: 0.23+/-0.01mm Material: High Quality Martens tic stainless steel Grade AISI-420.	
1010420109	NEEDLE-5/0, 12 MM 3/8 CRC 5/0	12000

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Wire diameter: 0.48 (-0.0+0.02)mm Bore diameter: 0.22 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420110	NEEDLE-3/8 CRB 16MM SILK 5/0	6000
	Wire diameter: 0.48 (-0.0+0.02)mm Bore diameter: 0.22 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420111	NEEDLE-1/2 CTC,17MM SILK 5/0	1200
	Wire diameter: 0.48 (-0.0+0.02)mm Bore diameter: 0.22 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420112	NEEDLE Straight RB 20MM SILK 5/0	1200
	Wire diameter: 0.48 (-0.0+0.02)mm Bore diameter: 0.22 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420113	NEEDLE-3/8 CRB 20MM SILK 5/0	1200
	Wire diameter: 0.48 (-0.0+0.02)mm Bore diameter: 0.22 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420114	needle-SILK 6/0,3/8 CRB 13mm	1200
	Specification: Wire diameter: 0.38+/-0.01mm Bore diameter: 0.18+/-0.01mm Material :High Quality martensitic stainless steel Grade AISI-420	
1010420115	NEEDLE - 60 MM 3/8 CC SILK 1	1200
	SUTURE SIZE:1 60mm,3/8circle,cutting 1. Bore Diameter (mm): 0.55+/-0.01mm 2. Wire diameter (mm): 1.20+/-0.05mm. Material :High Quality martensitic stainless steel Grade AISI-420.	
1010420116	NEEDLE-36MM, 3/8 CTC SILK-1	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	USP size: Silk 1 Wire diameter: 1+/-0.02mm Bore diameter: 0.55+/-0.01mm Material :High Quality martensitic stainless steel Grade AISI-420	
1010420117	Needle-3/8 CRC, 12mm PP: 0	1200
	1. Wire diameter: 0.78+/-0.02mm 2. Bore diameter: 0.47+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420118	NEEDLE - 25 MM DOUBLE 1/2 CRB PP 3/0	1200
	Wire diameter: 0.61 (-0.0+0.02)mm Bore diameter: 0.28 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420119	NEEDLE 25MM, 3/8 CC PP 3/0	1200
	USP Size: Polypropylene 3/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.61+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420120	NEEDLE-17 MM 1/2CTC PP4/0	1200
	Wire diameter: 0.56(-0.0+0.02)mm Bore diameter: 0.25(-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420	
1010420121	NEEDLE FOR PP SUTURE	1200
	Specification: SUTURE SIZE:5/0 16mm,3/8circle,round body Wire diameter: 0.48+/-0.02mm Bore diameter: 0.19+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420122	NEEDLE-12 MM 3/8 CRB PP 5/0	6000
	Specifications: SUTURE SIZE:5/0 Wire diameter:0.48 +/-0.01mm Bore diameter: 0.19+/-0.01mm 12mm 3/8 C, Round body Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420123	NEEDLE-16MM, 3/8 CRC PP 5-0	6000
	SUTURE SIZE: PP 5-0 Wire diameter: 0.48+/-0.02mm Bore diameter: 0.19+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420124	NEEDLE-3/8 CRC,12MM PP 6/0	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Suture size-6/0 Bore diameter: 0.15+/-0.01mm Wire diameter: 0.35+/-0.01mm high quality martensitic grade stainless steel-Grade-AISI-420	
1010420125	NEEDLE-PP 6/0,3/8 CRB,13MM	1200
	Suture size-6/0 (monofilament) Wire diameter: 0.36+/-0.02mm Bore diameter: 0.15+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420126	NEEDLE-PP 6/0,3/8 CC,15MM PP	2400
	Specification: Suture size-6/0 Wire diameter: 0.36+/-0.02mm Bore diameter: 0.15+/-0.01mm Material :High Quality martensitic stainless steel Grade AISI-420	
1010420127	NEEDLE - 30 MM 1/2 CRB PP:0	24000
	SUTURE SIZE: 0 30mm,1/2circle,round body Wire diameter: 0.92+/-0.02mm Bore diameter: 0.45+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420128	NEEDLE-25 MM 1/2 CTC PP:0	1200
	SUTURE SIZE:0 25mm,1/2circle,taper cut Wire diameter: 0.80+/-0.02mm Bore diameter: 0.45+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420130	NEEDLE-1/2 CRB 26MM PP-1/0	1200
	Specifications: suture size-1/0 Wire diameter: 0.78 +/-0.01mm Bore diameter: 0.45+/-0.01mm 26mm 1/2 C, Round body Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420130	NEEDLE-1/2 CRB 26MM PP-1/0	1200
	Wire diameter: 0.81 (-0.0+0.02)mm Bore diameter: 0.45 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420	
1010420131	NEEDLE-1/2CTC,30MM(HEAVY),PP1/0	2400

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Specifications: suture size-1/0 Wire diameter: 1.0 +/-0.01mm Bore diameter: 0.45+/-0.01mm 30mm 1/2 C, Taper cutting, Heavy Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420132	NEEDLE 30MM, 1/2 CC PP USP 0	1200
	USP Size: Polypropylene: 0 Bore diameter: 0.45+/-0.01mm Wire diameter: 0.92+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420133	NEEDLE 40MM, 1/2 CRB PP: 0	1200
	USP Size: Polypropylene 0 Bore diameter: 0.45+/-0.01mm Wire diameter: 0.92+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420134	NEEDLE 40MM, 1/2 CC PP: 0	1200
	USP Size: Polypropylene 0 Bore diameter: 0.45+/-0.01mm Wire diameter: 0.92+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420135	NEEDLE 45MM 1/2 CRB PP 0	1200
	USP Size: Polypropylene 0 Bore diameter: 0.45+/-0.01mm Wire diameter: 0.92+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420136	NEEDLE 45MM, 1/2 CC PP : 0	1200
	USP Size: Polypropylene 0 Bore diameter: 0.45+/-0.01mm Wire diameter: 0.92+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420	
1010420137	NEEDLE - 30 MM 1/2 CRB PP2/0	18000
	"Specification: Needle: 30mm, 1/2 circle round body Wire diameter: 0.80 +/- 0.01mm Bore diameter: 0.36 +/- 0.01mm Suture size: polypropylene USP 2/0 Material: High Quality martensitic stainless steel Grade AISI-420.	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420138	NEEDLE-25 MM 1/2 CTC PP 2/0 Needle: 25mm, 1/2 circle taper cutting Wire diameter: 0.78 +/- 0.01mm Bore diameter: 0.36 +/- 0.01mm Suture size: polypropylene USP 2/0 Material: High Quality martensitic stainless steel Grade AISI-420.	3000
1010420139	NEEDLE-2/0, 25 MM, 1/2 CTC PP2/0 Wire diameter: 0.81 (-0.0+0.02)mm Bore diameter: 0.36 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420	1200
1010420140	NEEDLE-1/2 CTC17MM PP-2/0 Wire diameter: 0.74 (-0.0+0.02)mm Bore diameter: 0.28 (-0.01+0.02)mm Material: High Quality martensitic stainless steel Grade AISI-420	2400
1010420141	NEEDLE-PP 3/0 25 MM 1/2 CRB NEEDLE-25MM, 1/2 CIRCLE ROUND BODY, POLYPROPYLENE- USP 3/0, Ribbed &Siliconised. 1. Wire diameter: 0.61 (-0.0+0.02)mm 2. Bore diameter: 0.28 (-0.01+0.02)mm 3. Suture material: Polypropylene, USP 3/0 (Monofilament) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	2400
1010420143	NEEDLE-20 MM 1/2 CRB PP 3/0 Wire diameter: 0.58 (-0.0+0.02)mm Bore diameter: 0.28 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420	1200
1010420144	NEEDLE-3/0, 22 MM, 3/8 CC PP SUTURE SIZE: Polypropylene USP3/0, 22mm, 3/8circle, cutting Bore diameter: 0.27 to 0.29mm Wire diameter: 0.57 to 0.59mm Material: High Quality Martens tic stainless steel Grade AISI-420.	3000
1010420145	NEEDLE 1/2 CTC 17MM,PP 3/0 Specifications: USP Size: 3/0 Wire diameter: 0.60+/-0.01mm Bore diameter: 0.28+/-0.01mm	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

<i>Material Number</i>	<i>Material Description</i>	<i>Quantity</i>
		<i>pcs</i>
	material;high quality martensitic stainless steel Grade;AISI-420	
1010420146	NEEDLE-PP4/0 16 MM 3/8 CC Wire diameter: 0.56 (-0.0+0.02)mm Bore diameter: 0.25 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200
1010420147	NEEDLE-PP 4/0 16 MM 1/2 CRB	6000
	SUTURE SIZE:4/0 16mm,1/2circle,round body Wire diameter: 0.54+/-0.02mm Bore diameter: 0.25+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420148	NEEDLE -3/8 CRC,12MM,PP 4/0	1200
	Specifications: suture size-4/0 Wire diameter: 0.54 +/-0.01mm Bore diameter: 0.25+/-0.01mm 12mm, 3/8 C, Reverse cutting Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420149	NEEDLE -1/2 CTC,25MM,PP 4/0	1200
	Wire diameter: 0.56 (-0.0+0.02)mm Bore diameter: 0.25 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420150	NEEDLE-16 MM 3/8 CRB	3000
	Specifications: SUTURE SIZE:5/0 16mm,3/8circle,round body Wire diameter: 0.48+/-0.02mm Bore diameter: 0.19+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420151	NEEDLE-PP 5/0 12 MM 3/8 CRC	1200
	SUTURE SIZE:5/0 12mm,3/8circle,reverse cutting Wire diameter: 0.50+/-0.01mm Bore diameter: 0.19+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420152	NEEDLE-3/8 CC,16MM PP 5/0	1200
	Wire diameter: 0.48(-0.0+0.02)mm Bore diameter: 0.19+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420153	NEEDLE -PP:1 40 MM 1/2 CRB	24000
	Specification: Item name: Needle: 40mm, ½ circle round body, heavy. -Ribbed &siliconised. 1.Wire diameter: 1.2+/-0.01mm 2. Bore diameter: 0.53+/-0.01mm 3. Suture material: Polypropylene, USP 1 (Monofilament) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420154	NEEDLE-30 MM 1/2 CRB PP:1	24000
	SUTURE SIZE:1 30mm,1/2circle,round body(heavy) Wire diameter: 1.15+/-0.05mm Bore diameter: 0.53+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420	
1010420155	NEEDLE-PP:1 45 MM HEAVY, 1/2CRC	24000
	Specifications: Wire diameter: 1.2+/-0.01mm Bore diameter: 0.53+/-0.01mm SUTURE SIZE: 1 45mm heavy,1/2circle rev cutting Material: High Quality Martens tic stainless steel Grade AISI-420.	
1010420156	NEEDLE-PP:1,1/2 CTC,30MM(HEAVY)	1200
	Specifications: suture size-1(USP) Wire diameter: 1.12 +/-0.01mm Bore diameter: 0.53+/-0.01mm 30mm 1/2 C, Taper cutting, Heavy Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420157	NEEDLE 30MM 1/2 CC,MO PP: 1	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	USP Size: Polypropylene 1 Bore diameter: 0.53+/-0.01mm Wire diameter: 1+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420158	NEEDLE 40MM 1/2 CC,MO: PP: 1	1200
	USP Size: Polypropylene 1 Bore diameter: 0.53+/-0.01mm Wire diameter: 1+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420159	NEEDLE 45MM 1/2 CC MO PP: 1	1200
	USP Size: Polypropylene 1 Bore diameter: 0.53+/-0.01mm Wire diameter: 1.2+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420160	NEEDLE 45MM 1/2 CRB,MO:PP: 1	1200
	NEEDLE 45MM 1/2 CIRCLE ROUND BODY MONOFILAMENT PP SUTURE USP 1 USP Size: Polypropylene 1 Bore diameter: 0.53+/-0.01mm Wire diameter: 1.2+/- 0.02mm Material: High Quality martensitic stainless steel Grade AISI- 420.	
1010420161	NEEDLE NYLON 1/0, 1/2 C.R.B. 40MM	24000
	"Item name: Needle: NEEDLE-40MM, 1/2 CIRCLE ROUND BODY, NYLON-USP 0, Siliconised& ribbed 1. Wire diameter: 1.00 (-0.0+0.025)mm 2. Bore diameter: 0.45 (-0.01+0.02)mm 3. Suture material: Polyamide Nylon, USP 0. (Monofilament) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420162	NEEDLE NYLON 1/0, STR C 60MM	1200
	Specification: USP Size: 0 Wire diameter: 0.84+/-0.02mm Bore diameter: 0.45+/-0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	
1010420163	NEEDLE NYLON 2/0, 3/8 CIR.REV.CU.45MM	600000
	USP size: 2-0, 45mm, 3/8 circle reverse cutting.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Bore Dia 0.36 +/- 0.01mm Wire dia : 0.99 to 1.00mm Material: High Quality martensitic Stainless Steel Grade AISI-420	
1010420164	NEEDLE 26MM, 3/8 CRC NYLON 2/0 USP Size: Nylon 2/0 Bore diameter: 0.36+/-0.01mm Wire diameter: 0.7+/-0.02mm Material:High Quality martensiticstainless steel Grade AISI-420.	1200
1010420165	NEEDLE 16MM, 3/8 CRC NYLON 2/0 USP Size: Nylon 2/0 Bore diameter: 0.36+/-0.01mm Wire diameter: 0.61+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200
1010420166	NEEDLE NYLON 3/0, 3/8 CC G 16MM Monofilament Nylon Suture 1. Bore diameter (mm): 0.28+/-0.01mm 2. Wire diameter (mm): 0.60+/-0.05mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	1200
1010420167	NEEDLE NYLON 3/0, 3/8 CIR.REV.CU. 26MM Monofilament Nylon Suture, Bore Dia 0.28 +/- 0.01mm wire dia: 0.71 to 0.72mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	300000
1010420168	NEEDLE NYLON 3/0 , 3/8 CRC 24 MM Wire diameter: 0.61 (-0.0+0.02)mm Bore diameter: 0.28 (-0.01+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	1200
1010420169	NEEDLE 45MM, 3/8 CRC NYLON 3/0 USP Size: Nylon 3/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.74+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420170	NEEDLE NYLON 4/0, 3/8 CC 16MM	75000
	Specification: Wire diameter: 0.54+/-0.02mm Bore diameter: 0.25+/-0.01mm Suture size: Monofilament Nylon Suture, USP 4-0 Material: High Quality martensitic Stainless Steel Grade AISI-420.	
1010420171	NEEDLE 16MM, 3/8 CRC NYLON 4/0	1200
	Suture size: Nylon 4/0 Wire diameter: 0.54+/-0.02mm Bore diameter: 0.25+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420172	NEEDLE 19MM, 3/8 CC NYLON 4/0	1200
	NEEDLE 19MM, 3/8 CIRCLE CUTTING FOR NYLON 4/0 SUTURES Suture size: Nylon 4/0 Wire diameter: 0.54+/-0.02mm Bore diameter: 0.25+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420	
1010420173	NEEDLE NYLON 5/0, 3/8 CC 16MM	12000
	Monofilament Nylon Suture, Wire dia : 0.46(-0.0+0.02)mm Bore Dia 0.19 ± 0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	
1010420174	NEEDLE NYLON 5/0, 3/8 CIR.REV.CU 12MM	3000
	Specification: Wire diameter: 0.46+/-0.02mm Bore diameter: 0.19+/-0.01mm Suture size: Monofilament Nylon 5-0 Material: High Quality martensitic Stainless Steel Grade AISI-420.	
1010420175	NEEDLE 13MM, 3/8 CRC NYLON 5/0	1200
	Suture size: Nylon 5/0 Wire diameter: 0.46+/-0.02mm Bore diameter: 0.19+/-0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420	
1010420176	NEEDLE NYLON 1,1/2 CRC.45MM(Heavy)	72000
	Monofilament Nylon Suture,	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	1.Bore Diameter (mm): 0.53+/-0.01mm 2. Wire diameter (mm): 1.20+/-0.05mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	
1010420177	NEEDLE NYLON 1, 1/2 Cl. R.B. 40MM Monofilament Nylon Suture, Wire dia : 1.12-0.0+0.03mm Bore Dia 0.53-0.01+0.025mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	6000
1010420178	Needle 40mm, 1/2 CRB.heavy,PGA:1 Specification: USP Size: 1 (Braided) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200
1010420179	Needle 40mm,1/2 CTC.heavy,PGA:1 Specification: USP Size: 1 (Braided) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200
1010420180	Needle 40mm, 1/2 CRC.PGA: 1 Specification: USP Size: PGA 1 (braided) Bore diameter: 0.55+/- 0.01MM Wire diameter: 1.2+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.	2400
1010420181	Needle 50mm,1/2 CRC.heavy,PGA:1 Specification: USP Size: PGA 1 (Braided) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	24000
1010420182	Needle 50mm, 1/2 CRB. heavy,PGA:1 Specification: USP Size: PGA 1 (Braided)	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420	
1010420183	NEEDLE 45 MM, 3/8 CRB.PGA 1	1200
	Specification: USP Size: PGA 1 (Braided) Wire diameter: 1.00+/-0.02mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420184	Needle 40mm, 1/2 CRC PGA 2/0	1200
	Specification: USP Size: 2/0 (Braided) Wire diameter: 0.84+/-0.02mm Bore diameter: 0.42+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420185	NEEDLE 30 mm, 1/2 CRB.PGA 2/0	1200
	Specification: USP Size: PGA 2-0 (Braided) Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420186	NEEDLE 40mm, 1/2 CRB.PGA 2/0	1200
	Specification: USP Size: PGA 2/0 (Braided) Wire diameter: 0.84+/-0.02mm Bore diameter: 0.42+/-0.02mm. Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420187	Needle 22mm, 3/8 CC.PGA: 3/0	1200
	Specification: USP Size: 3/0 (Braided) Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420188	Needle 20mm, 1/2 CCB.PGA: 3/0	1200
	Specification: USP Size: PGA 3/0 (Braided) Wire diameter: 0.62+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420189	Needle 16mm, 3/8 CC.PGA: 3/0	1200
	Specification: USP Size: PGA 3/0 (Braided) Wire diameter: 0.54+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420190	Needle 22mm, 1/2 CC.PGA: 3/0	1200
	Specification: USP Size: PGA 3/0 (Braided) Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420191	Needle 12mm, 3/8 CRC.PGA: 3/0	1200
	Specification: USP Size: PGA 3/0 (Braided) Wire diameter: 0.54+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420192	Needle 17mm, 1/2 C.TA.CU.PGA: 3/0	1200
	Specification: USP Size: PGA 3/0 (Braided) Wire diameter: 0.60+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420	
1010420193	Needle 36mm, 1/2 C.TA.CU.PGA 3/0	1200
	Specification:	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	USP Size: PGA 3/0 (Braided) Wire diameter: 0.71+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420194	Needle 20mm, 1/2 CRB.PGA: 4/0	1200
	Specification: USP Size: 4/0 (Braided) Wire diameter: 0.54+/-0.02mm Bore diameter: 0.28+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420195	Needle 16mm, 1/2 CRB.PGA: 4/0	1200
	Specification: USP Size: 4/0 (Braided) Wire diameter: 0.54+/-0.02mm Bore diameter: 0.28+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420	
1010420196	Needle16mm, 5/8 CC PGA: 4/0	1200
	Specification: USP Size: PGA 4/0 (Braided) Bore dia 0.28 +/- 0.02mm Wire dia : 0.48 to 0.51mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420197	Needle 12mm, 3/8 CRCPGA: 4/0	1200
	Specification: USP Size: PGA 4/0 (Braided) Wire diameter: 0.54+/-0.02mm Bore diameter: 0.28+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420198	NEEDLE 16mm, 1/2 CRB.PGA: 5/0	1200
	Specification: USP Size:5/0 (Braided) Wire diameter: 0.48+/-0.02mm	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Bore diameter: 0.22+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420199	Needle12mm, 3/8CRC.PGA: 5/0 Specification: USP Size: PGA 5/0 (Braided) Wire diameter: 0.48+/-0.02mm Bore diameter: 0.22+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200
1010420200	Needle 40mm, 1/2CRB.PGA: 0 Specification: USP Size: 0 (Braided) Wire diameter: 1+/-0.02mm Bore diameter: 0.47+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200
1010420201	Needle 40mm, 1/2C.TA.CPGA: 0 Specification: USP Size:PGA 0 (Braided) Wire diameter:1+/-0.02mm Bore diameter:0.47+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200
1010420202	Needle 36mm,1/2CRC(OS),PGA: 0 Specification: USP Size:PGA 0 (Braided) Bore diameter:0.47+/-0.01mm Wire diameter:1.2+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200
1010420203	NEEDLE 30 MM, 1/2 CRB.PGA: 0 Specification: USP Size:PGA 0 (Braided) Bore diameter:0.47+/-0.01mm. Wire diameter:0.84+/-0.02mm. Material:High Quality martensitic stainless steel Grade AISI-420.	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420204	NEEDLE , 40MM 1/2CRB. HEAVY,PDO: 1	50000
	Specification: Suture type:Polydioxanone(PDO) Monofilament sutureUSP1. Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420205	NEEDLE 50MM 1/2 CRB HEAVY PDO: 1.	1200
	Specifcaton: Suture type:Polydioxanone(PDO) MonofilamentUSP1. Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420206	NEEDLE 45MM 1/2CRB HEAVY PDO: 1	12000
	Specification: SUTURE SIZE:USP1 Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420207	NEEDLE 30MM, 1/2 CRB.PDO: 2-0	6000
	Specification: Suture size:USP2-0,monofilament poly(p-DIOXANONE). Wire diameter:0.78+/-0.02mm Borediameter:0.45+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420208	NEEDLE 20MM 1/2 CRB.PDO: 3/0	18000
	Specification: SUTURE SIZE:USP3/0 Wire diameter:0.78+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420209	NEEDLE 40MM 1/2 CRC-OS NEEDLE PDO	9000
	Specification SUTURE SIZE:USP1 Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420210	NEEDLE 40MM, 1/2 CRB. HEAVY, PDO: 0	1200
	Specification: Suture type:Polydioxanone(PDO) monofilament USP0. Wire diameter:1.12+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420211	NEEDLE 40MM, 1/2CRB. HEAVY,PDO: 0.	1200
	Specification: Suture type:Polydioxanone(PDO) monofilament USP0. Wire diameter:1.12+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420212	NEEDLE 30MM 1/2 CRB.PDO: 0	1200
	Specification: SUTURE SIZE:USP0 Wire diameter:1+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420213	NEEDLE 48 MM 1/2 CRC.PDO:1	12000
	Specification: SUTURE SIZE:USP1 Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420214	NEEDLE-40MM 1/2 C.TA.C.G HEAVY PDO:1	6000
	Specification: Suture material:PDOmonofilament USP1 Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420215	NEEDLE PGA-PCL 2-0,26MM 1/2 CRB.	24000
	Specification: Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] monofilament suture,USP2-0 Wire diameter:0.84+/-0.02mm Borediameter:0.45+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420216	NEEDLE PGA-PCL 2-0,26MM 3/8 CRB.	1200
	Specification Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] monofilament suture,USP2-0. Wire diameter:0.84+/-0.02mm Borediameter:0.45+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420217	NEEDLE PGA-PCL 2/0 30MM, 1/2 CRB.	18000
	Specification: SUTURE SIZE:USP2/0 Wire diameter:0.84+/-0.02mm Borediameter:0.45+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420218	NEEDLE PGA-PCL 2/0 30MM, 1/2 CRB	1200
	Specification: Suture material:PGA-PCL MonofilamentUSP2-0 Wire diameter:0.84+/-0.02mm Borediameter:0.45+/-0.01mm	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420219	NEEDLE PGA-PCL3-0 25MM 3/8 CC.	1200
	Specification: Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] MONOFILAMENTUSP3-0. Wire diameter:0.84+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420220	NEEDLE PGA-PCL 3-0 26MM 1/2 CRB.	1200
	Specification: Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] monofilament, USP3-0 Wire diameter:0.80+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420221	NEEDLE PGA-PCL3-0 26MM 3/8 CRB	1200
	Specification: Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] monofilament sutureUSP3-0 Wire diameter:0.80+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420222	NEEDLE PGA-PCL 3/0 20MM, 1/2 CRB	1200
	Specification: SUTURE SIZE:USP3/0 Wire diameter:0.78+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420223	NEEDLE PGA-PCL3/0 60MM, STR.CU.	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420224	NEEDLE PGA-PCL 3/0 26 MM 1/2 CRB.	1200
	Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] MONOFILAMENT USP3-0. Wire diameter:0.62+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420225	NEEDLE PGA-PCL 3/0 24MM 3/8CRC USP 3-0	1200
	Specification: Suture size :PGA-PCL3-0 (Monofilament) Wire diameter:0.62+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420226	NEEDLE PGA-PCL 4-0 16MM 3/8 CRC.	6000
	Specification: Suture size:PGA-PCL[Poly (glycolide-co-caprolactone)] MONOFILAMENT USP4-0. Wire diameter:0.54+/-0.02mm Borediameter:0.28+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420227	NEEDLE PGA-PLA 4/0 20MM, 1/2 CRB.	12000
	Specification: SUTURE SIZE:USP4/0 Wire diameter:0.70+/-0.02mm Borediameter:0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420228	NEEDLE 20MM 1/2 CRB.PDO:4/0	3000
	Specification SUTURE SIZE:USP4/0 Wire diameter:0.62+/-0.02mm Borediameter:0.28+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420229	NEEDLE 16 MM 3/8 CRB.PGA-PCL 4-0	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] MONOFILAMENTUSP4-0. Wire diameter:0.48+/-0.02mm Borediameter:0.28+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420230	NEEDLE30MM 1/2 CRB.PGA-PCL 0	1200
	Specification: SUTURE SIZE:USP0 Wire diameter:1+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420231	NEEDLE 40MM 1/2 C.TA.C. HEAVY PDO:0	12000
	Specification: SUTURE SIZE:USP0 Wire diameter:1.12+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420232	NEEDLE 30MM 1/2CRB HEAVY PDO:0	1200
	Specification: SUTURE SIZE:USP0 Wire diameter:1.12+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420233	NEEDLE 40MM 1/2C.TA.C.PGA-PLA: 0	1200
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP0 Wire diameter:1+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420234	NEEDLE 30MM, 1/2 CRB.PGA-PLA: 2/0.	18000
	Specification: SUTURE SIZE: poly(glycolide-co-lactide)USP2/0 Wire diameter:0.80+/-0.02mm Borediameter:0.47+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420235	NEEDLE 40MM 1/2 CRB.PGA-PLA: 2/0	18000
	Specification	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	SUTURE SIZE: Poly(glycilide-co-lactide)USP2/0 Wire diameter:0.84+/-0.02mm Borediameter:0.47+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420236	NEEDLE 40MM 1/2 CRC.PGA-PLA: 2/0	1200
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP2/0 Wire diameter:0.84+/-0.02mm Borediameter:0.47+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420237	NEEDLE 20MM 1/2 CRB.PGA-PLA: 3/0	100000
	Specification: Needle:20mm,1/2circlearoundbody Wire diameter:0.62 +/- 0.01mm Borediameter:0.36 +/- 0.01mm Suture size:poly(glycolide-co-lactide) USP3/0 Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420238	NEEDLE 22MM, 3/8 CC.PGA-PLA: 3-0	1200
	"Specification: Needle:22mm,3/8circlecutting Wire diameter:0.62 +/- 0.01mm Borediameter:0.36 +/- 0.01mm Suture size:poly(glycolide-co-lactide) USP3/0 Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420239	NEEDLE 16MM 1/2 CRB.PGA-PLA: 5/0.	3000
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP5/0 Wire diameter:0.48+/-0.02mm Borediameter:0.28+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420240	NEEDLE-40MM 1/2 CIR.REV.CU. OS	1200
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP1 Wire diameter:1.2+/-0.01mm	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420241	NEEDLE 40MM, 1/2 CRB.PGA-PLA: 0	150000
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP0 Wire diameter:1+/-0.01mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420242	NEEDLE 40MM 1/2 CRB.HEAVY PGA-PLA:1	1200
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP1 Wire diameter:1.2+/-0.01mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420243	NEEDLE - 36MM 1/2 CRC.OS NEEDLE	60000
	Specification: SUTURE SIZE: Ploy(glycolide-co-lactide)USP0 Wire diameter:1.2+/-0.01mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420244	NEEDLE40MM 1/2 C.TA.CU.HEAVY PGA-PLA	24000
	"Specification: Needle:40mm,1/2circle taper cutting, heavy Wire diameter:1.2 +/- 0.01mm Borediameter:0.60 +/- 0.01mm Suture size:poly(glycolide-co-lactide) USP1 Material:High Qualitymartensitic stainlesssteelGradeAISI-420	
1010420245	NEEDLE 30MM, 1/2 CRB. PGA-PLA: 0	24000
	Wire diameter: 0.84+/-0.02mm Bore diameter: 0.53+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420	
1010420248	NEEDLE 40MM 1/2 C.TA.CU.HEAVY PGA-PLA	1200
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide),USP1	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitystainlesssteel Grade300.	
1010420249	NEEDLE SUTURE 36MM, 1/2 CRB.PGA-PLA:1	3000
	Specification: Needle:36mm,1/2circleroundbody. Suture size:Polyglactin910, USP1 (Braided) Wire diameter:1+/-0.02mm Borediameter:0.57+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420251	NEEDLE 40MM 1/2 C.TA.CU.HEAVY PGA-PLA.	1200
	Specification: Suture size:Polyglactin910, Braided, USP1. Wire diameter:1+/-0.01mm Borediameter:0.57+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420267	NEEDLE-30MM, 3/8 CIRCLE ROUND BODY (PGA)	1200
	Specification: Suture size:PGA, USP2-0 Wire diameter:0.80 +/- 0.01mm Borediameter:0.42 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420268	NEEDLE-37MM, 1/2 CIRCLE REV CUT(PGA PLA)	9000
	Specification: Suture size:PGA-PLA,USP2-0 Wire diameter:0.84 +/- 0.01mm Borediameter:0.47 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420	
1010420269	NEEDLE-37MM,1/2 CIRCLE ROUND BODY	1200
	PGS-PLA)Specification: Suture size:PGA-PLA,USP2-0 Wire diameter:0.84 +/- 0.01mm Borediameter:0.47+/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420270	NEEDLE-26MM, 3/8 CIRCLE ROUND BODY (PGA)	1200
	Specification: Suture size:PGA, USP3-0 Wire diameter:0.70 +/- 0.01mm Borediameter:0.33 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420271	NEEDLE-30MM, 3/8 CIRCLE ROUND BODY.	1200
	Specification: Suture size:PGA, USP3-0 Wire diameter:0.70 +/- 0.01mm Borediameter:0.33 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420272	NEEDLE-37MM, 1/2 CIRCLE ROUND BODY	1200
	(PGA PLA)Specification: Suture size:PGA-PLA,USP3-0 Wire diameter:0.70 +/- 0.01mm Borediameter:0.36 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420273	NEEDLE-37MM,1/2 CIRCLE REV-CUT (PGA PLA)	1200
	Specification: Suture size:PGA-PLA,USP3-0 Wire diameter:0.70 +/- 0.01mm Borediameter:0.36 +/- 0.01mm Material:High QualitymartensiticstainlesssteelGradeAISI-420.	
1010420274	NEEDLE-13MM, 1/2 CIRCLE ROUND BODY	1200
	Suture size:PGA, USP4-0 Wire diameter:0.54 +/- 0.01mm Borediameter:0.28 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420276	NEEDLE 45MM 3/8 CIRCLE CUTTING PGA 1	1200
	SPECIFICATION: USPSIZE:PGA 1 Wire diameter:1.2mm Borediameter:0.55mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420277	NEEDLE-36MM, 3/8 CIRCLE ROUND BODY (PGA)	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Specification: Suture size:PGA, USP0 Wire diameter:0.84 +/- 0.01mm Borediameter:0.47 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420278	NEEDLE-36MM, 1/2 CIRCLE ROUND BODY (PGA)	1200
	Specification: Suture size:PGA, USP1 Wire diameter:1.00 +/- 0.01mm Borediameter:0.55 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420279	NEEDLE-26MM, 1/2 CIRCLE TAPER CUT (PGA)	1200
	Specification: Suture size:PGA, USP1 Wire diameter:0.84 +/- 0.01mm Borediameter:0.55 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420	
1010420280	NEEDLE FOR PDO SUTURE-40MM, 1/2 C R B	1200
	specification: NEEDLE :40MM, 1/2CIRCLE ROUNDBODY. Suture type:Polydioxanone(PDO) monofilament sutureUSP1. Wire diameter:1+/-0.01mm Borediameter:0.57+/-0.01mm Material:High Quality martensitic stainlesssteelGradeAISI-420.	
1010420281	NEEDLE-40MM,1/2 CIRCLE REV CUT (PGA-PLA)	3600
	Specification: Suture size:PGA-PLA,USP0 Wire diameter:1.00 +/- 0.01mm Borediameter:0.55+/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420282	NEEDLE PGA-PLA 2-0 SUTURE 30MM,1/2 C, TC	9000
	Specification: Needle:30mm,1/2circletaper cutting.	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Suture material:PGA-PLAUSP2-0 Wire diameter:0.80mm Borediameter:0.47mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420283	NEEDLE-50MM 1/2C REVERSE CUTTING, HEAVY. SPECIFICATION: NEEDLE:50MM, 1/2CIRCLE REVERSE CUTTING HEAVY. SUTURE:PGA-PLAUSP1 WIREDIAMETER:1.2+/-0.01MM BORE DIAMETER:0.60 MM Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	1200
1010420284	NEEDLE-30MM 1/2 C REVERSE CUTTING Specification: Item name:Needle:30mm,1/2circle reversecutting-siliconised. 1. Wire diameter:0.92+/-0.01mm 2. Borediameter:0.55+/-0.01mm 3. Suture material:PGA-PLA,USP0 (Braided) 4. Allneedlesmustbesiliconised 5. Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	1200
1010420285	NEEDLE-40MM 1/2 CIRCLE ROUND BODY NEEDLE-40MM 1/2CIRCLE ROUND BODY Specification: NEEDLE:40MM, 1/2 CIRCLE ROUNDBODY.SUTURE MATERIAL:PGA-PLAUSPSIZE 2. Wire diameter: 1.2MM Borediameter:0.72MM Material:High Qualitymartensitic stainless steelGradeAISI-420	75000
1010420287	NEEDLE-30MM 1/2 CIRCLE ROUND BODY NEEDLE-30MM 1/2CIRCLE ROUND BODY SPECIFICATION: NEEDLE:30MM, 1/2 CIRCLE ROUNDBODYSUTURE MATERIAL:PGA-PLAUSP2 BORE DIAMETER:0.72MM WIREDIAMETER:1MM Material: High Qualitymartensiticstainless steel GradeAISI-420	1200
1010420288	NEEDLE-40MM 1/2 CIRCLE TAPER CUTTING	24000

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	NEEDLE-40MM 1/2CIRCLE TAPER CUTTING Specification:Needle: 40mm,1/2 circletapercuttingSuturematerial: PGA-PLAUSP2 Wire diameter: 1.2mmBorediameter:0.72mm Material:High Qualitymartensitic stainless steelGradeAISI-420	
1010420289	NEEDLE-48MM 1/2 CIRCLE ROUND BODY	50000
	NEEDLE-48MM 1/2CIRCLE ROUND BODY SPECIFICATION: Needle:48mm,1/2 circleround body Suturematerial: PGA-PLAsutureUspsize 2 Wire diameter:1.3mm Borediameter: 0.72mm Material:High Quality martensitic stainlesssteelGrade AISI-420	
1010420292	NEEDLE- PGA PLA 2/0,5/8 TAPER POINT 27MM	1200
	"Specifications: SUTURE SIZE:2/0 Wire diameter:0.71+/-0.01mm Borediameter:0.47+/-0.01mm 27mm, 5/8C, TaperPoint Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	
1010420293	NEEDLE-30MM,1/2 CRB,HIDOX USP 3-0	6000
	Specification: Needle:30mm,1/2 circleroundbody Wire diameter:0. 78mm Borediameter:0.36mm Suture mtl:PolydioxanoneUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420	
1010420294	NEEDLE-50MM,1/2 CRB,HEAVY,PGA-PLA USP 1	75000
	Specification: Needle:50mm,1/2 circleround body, heavy. Wire diameter:1.2mm Borediameter:0.60mm Suture mtl:Polyglactin910 USP1 Material:High Qualitymartensitic stainlesssteel Grade AISI-420.	
1010420295	NEEDLE-30MM,1/2 CRC, PGA USP 0	1200
	Specification:	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Needle:30mm,1/2 circle reverse cutting Wire diameter:0.84mm Borediameter:0.47mm Suture mtl:PolyglycolicacidUSP0 Material:High Qualitymartensitic stainless steel Grade AISI-420.	
1010420296	NEEDLE-22MM,1/2CIRCLE CUTTING,PGA-PLA3-0	24000
	Specification: Needle:22mm,1/2circlecutting Wire diameter:0.62mm Borediameter:0.36mm Suture mtl:Polyglactin910 USP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420297	NEEDLE-45MM,1/2 CIRCLE CUTTING,PP USP-0	1200
	Specification: Needle:45mm,1/2circlecutting Wire diameter:1.1mm Borediameter:0.45mm Suture mtl:PolypropyleneUSP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420298	NEEDLE-16MM,3/8 CIRCLE RC,Silk USP 3-0	1200
	Specification: Needle:16mm,3/8circle reverse cutting Wire diameter:0.54mm Borediameter:0.33mm Suture mtl:SilkUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420299	NEEDLE-26MM,3/8CIRCLE CUTTING,PGA-PLA2-0	6000
	Specification: Needle:26mm,3/8circlecutting Wire diameter:0.78mm Borediameter:0.47mm	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Suture mtl:Polyglactin910 USP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420300	NEEDLE-22MM,1/2CIRCLE CUTTING,PGA-PLA4-0 Specification: Needle:22mm,1/2circlecutting Wire diameter:0.54mm Borediameter:0.33mm Suture mtl:Polyglactin910 USP4-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	1200
1010420301	NEEDLE-30MM,1/2 CIRCLE CUTTING,SILK 2-0 Specification: Needle:30mm,1/2circlecutting Wire diameter:0.78 +/- 0.01 mm Borediameter:0.42mm Suture mtl:SilkUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	1200
1010420302	NEEDLE-36MM,1/2CIRCLE TAPER CUT,PGA 2-0 Specification: Needle:36mm,1/2circletaper cutting Wire diameter:0.78 +/- 0.01 mm Borediameter:0.42+/- 0.01 mm Suture mtl:PolyglycolicAcidUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	6000
1010420303	NEEDLE-36MM,1/2CIR. ROUND BODY,PGA 2-0 Specification: Needle:36mm,1/2circleroundbody Wire diameter:0.78+/- 0.01 mm Borediameter:0.42+/- 0.01 mm Suture mtl:PolyglycolicAcidUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	6000
1010420304	NEEDLE-40MM,3/8CIR. ROUND BODY,PGA 1 Specification:	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Needle:40mm,3/8circleroundbody Wire diameter:1.00+/- 0.01 mm Borediameter:0.55+/- 0.01 mm Suture mtl:PolyglycolicAcidUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420305	NEEDLE-25MM,1/2CIR. ROUND BODY,PP 2/0	6000
	Needle: 25mm, 1/2 circle round body Wire diameter:0.78+/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:PolypropyleneUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420306	NEEDLE-25MM,3/8CIR. CUTTING,PP 3/0	1200
	Specification: Needle:25mm,3/8circlecutting Wire diameter:0.61+/- 0.01 mm Borediameter:0.28+/- 0.01 mm Suture mtl:PolypropyleneUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420307	NEEDLE-40MM,1/2CIRCLE REVERS CUT,PGA 3-0	1200
	Specification: Needle:40mm,1/2circlereverse cutting-ribbed&siliconised Wire diameter:0.78 +/- 0.01 mm Borediameter:0.33+/- 0.01 mm Suture mtl:PolyglycolicAcidUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420308	NEEDLE-25MM, 1/2 CRC, PGA-PLA USP 3-0	1200
	Specification: Needle:25MM, 1/2CIRCLE REVERSECUTTING-ribbed& siliconised Wire diameter:0.62 +/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:PGA-PLAUSP3-0	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420309	NEEDLE-26MM, 3/8 CRC, HEAVY,PGA-PLA USP Specification: Needle:26MM, 3/8CIRCLE REVERSECUTTING, HEAVY-ribbed&siliconised Wire diameter:0.84 +/- 0.01 mm Borediameter:0.60+/- 0.01 mm Suture mtl:PGA-PLAUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	1200
1010420310	NEEDLE-26MM, 3/8 CIRCLE CUTTING, USP 3-0 Specification: Needle:26mm,3/8circle cutting-ribbed&siliconised Wire diameter:0.62 +/- 0.01 mm Borediameter:0.33+/- 0.01 mm Suture mtl:PolyglycolicAcidUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	1200
1010420311	NEEDLE 30MM,3/8 CIRCLE CUTTING USP 3/0 Specification: Needle:30MM, 3/8CIRCLE CUTTINGNEEDLE-ribbed& siliconised Wire diameter:0.66 (+0.02 mm) Borediameter:0.28 (-0.01+0.02mm) Suture mtl: NylonUSP3/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	1200
1010420312	NEEDLE 30MM,3/8 CIRCLE CUTTING USP 2/0 Specification: Needle:30MM, 3/8CIRCLE CUTTINGNEEDLE-ribbed&siliconised Wire diameter:0.74 (+0.02mm) Borediameter:0.36 (-0.01+0.02mm) Suture mtl: NylonUSP2/0	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420313	NEEDLE 13MM,3/8 CIRCLE TAPER CUT USP 6/0 Specification: Needle:13MM, 3/8CIRCLE TAPER CUT-ribbed&siliconised Wire diameter:0.27 (+0.02 mm) Borediameter:0.15 (+/- 0.01mm) Suture mtl:PolypropeleneUSP6/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	1200
1010420314	NEEDLE-12MM, 1/2 CIRCLE CUTTING Specification: Needle:12MM, 1/2CIRCLE CUTTINGNEEDLE-ribbed&siliconised Wire diameter:0.46 (-0.0+0.02mm) Borediameter:0.22(-0.01+0.02mm) Suture mtl:PGA USP5-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	1200
1010420315	NEEDLE-30MM, 1/2C, REVERSE CUTTING Specification: Needle:30MM, 1/2CIRCLE REVERSECUTTING-ribbed&siliconised Wire diameter:0.74 (-0.0mm +0.02mm) Borediameter:0.36 (-0.01mm +0.02mm) Suture mtl:PolypropyleneUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	6000
1010420316	NEEDLE-16MM, 1/2C, REVERSE CUTTING Specification: Needle:16MM, 1/2CIRCLE REVERSECUTTING-ribbed&siliconised Wire diameter:0.56 (-0.0mm +0.02mm)	3000

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Borediameter:0.28 (-0.01mm +0.02mm) Suture mtl:PolyglycolicacidUSP5-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420317	NEEDLE-20MM, 1/2C, REVERSE CUTTING	1200
	Specification: Needle:20MM, 1/2CIRCLE REVERSECUTTING-ribbed& siliconised Wire diameter:0.74 (-0.0mm +0.02mm) Borediameter:0.36 (-0.01mm +0.02mm) Suture mtl:PolydioxanoneUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420318	NEEDLE-30 MM 1/2CIRCLE CUTTING	1200
	Specification: Needle:30MM, 1/2CIRCLE, CUTTING-ribbed&siliconised Wire diameter:0.71 (-0.0mm +0.02mm) Borediameter:0.33 (-0.01mm +0.02mm) Suture mtl:SilkUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420319	NEEDLE 13MM 3/8CIRCLE ROUND BODY	1200
	Specification: Needle:13MM, 3/8CIRCLE, ROUNDBODY-ribbed&siliconised Wire diameter:0.24 (-0.0+0.02mm) Borediameter:0.18 (± 0.01mm) Suture mtl:SilkUSP6-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420320	NEEDLE 23MM 3/8CIRCLE REV.CUTTING HEAVY	1200
	Specification: Needle:23MM, 3/8CIRCLE, REVERSECUTTING, HEAVY-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm)	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420321	NEEDLE 23MM 3/8CIRCLE REV.CUTTING HEAVY	1200
	Specification: Needle:23MM, 3/8CIRCLE REVERSECUTTING, HEAVY-ribbed&siliconised Wire diameter:0.84 +/- 0.01 mm Borediameter:0.60+/- 0.01 mm Suture mtl:PGA-PLAUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420322	NEEDLE-26MM 3/8CIR. REVERSE CUTTING,2/0	1200
	Specification: Needle:26MM, 3/8CIRCLE, REVERSECUTTING- ribbed& siliconised Wire diameter:0.71 (-0.0+0.02mm) Borediameter:0.36 (-0.01+0.02mm) Suture mtl:NYLONUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420323	NEEDLE-19MM 3/8CIR. REVERSE CUTTING,3/0	1200
	Specification: Needle:19MM, 3/8CIRCLE, REVERSECUTTING- ribbed& siliconised Wire diameter:0.61 (-0.0+0.02mm) Borediameter:0.28 (-0.01+0.02mm) Suture mtl:NYLONUSP3/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420324	NEEDLE-22MM 3/8CIR. REVERSE CUTTING,3/0	1200
	Specification: Needle:22MM, 3/8CIRCLE, REVERSECUTTING-ribbed& siliconised	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Wire diameter:0.61 (-0.0+0.02mm) Borediameter:0.28 (-0.01+0.02mm) Suture mtl: NYLON USP3/0 Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420325	NEEDLE-19MM 3/8CIR. REVERSE CUTTING,4/0	1200
	Specification: Needle:19MM, 3/8CIRCLE, REVERSE CUTTING- ribbed & siliconised Wire diameter:0.56 (-0.0+0.02mm) Borediameter:0.25 (-0.01+0.02mm) Suture mtl: NYLON USP4/0 Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420326	NEEDLE-22MM 3/8CIR. REVERSE CUTTING,4/0	3000
	Specification: Needle:22MM, 3/8CIRCLE, REVERSE CUTTING- ribbed & siliconised Wire diameter:0.56 (-0.0+0.02mm) Borediameter:0.25 (-0.01+0.02mm) Suture mtl: NYLON USP4/0 Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420327	NEEDLE-19MM 3/8CIR. REVERSE CUTTING,3/0	6000
	Specification: Needle:19MM, 3/8CIRCLE, REVERSE CUTTING- ribbed & siliconised Wire diameter:0.61 (-0.0+0.02mm) Borediameter:0.33 (-0.01+0.02mm) Suture mtl: SILK USP3/0 Material:High Quality martensitic stainless steel Grade AISI-420.	
1010420328	NEEDLE-19MM 3/8CIR. REVERSE CUTTING,4/0	6000
	Specification: Needle:19MM, 3/8CIRCLE, REVERSE CUTTING- ribbed & siliconised Wire diameter:0.56 (-0.0+0.02mm) Borediameter:0.28 (-0.01+0.025mm)	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Suture mtl: SILKUSP4/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420329	NEEDLE-23MM 1/2 CIR.REV.CUT,HEAVY,USP O	1200
	Specification: NEEDLE-23MM 1/2CIRCLE REVERSECUTTING, HEAVY, USP O-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420330	NEEDLE-23MM 1/2 CIR.REV.CUT,HEAVY,USP 1	1200
	Specification: Needle:NEEDLE-23MM 1/2CIRCLE REVERSECUTTING,HEAVY,USP 1- ribbed&siliconised Wire diameter:0.84 (-0.0+0.02mm) Borediameter:0.60 (-0.01+0.02mm) Suture mtl:PGA-PLAUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420331	NEEDLE-23MM 1/2 CIR.REV.CUT,HEAVY,USP O	75000
	Specification: NEEDLE-23MM 1/2CIRCLE REVERSECUTTING,HEAVY,USP O-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420332	NEEDLE-23MM 1/2 CIR.REV.CUT,HEAVY,USP 1	75000
	Specification:	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	NEEDLE-23MM 1/2CIRCLE REVERSE CUTTING,HEAVY,USP 1-ribbed&siliconised Wire diameter:0.84 (-0.0+0.02mm) Borediameter:0.60 (-0.01+0.02mm) Suture mtl:PGA-PLAUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420333	NEEDLE-36MM 1/2 CIR. REVERSE CUTTING, 2/ Specification: Needle:NEEDLE-36MM 1/2CIR. REVERSE CUTTING,2/0-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.47 (-0.01+0.02mm) Suture mtl:Polyglactin910Rapid USP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	75000
1010420334	NEEDLE-36MM 1/2 CIR. ROUND BODY, 2/0 Specification: Needle:NEEDLE-36MM 1/2CIR. ROUNDBODY,2/0-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.47 (-0.01+0.02mm) Suture mtl:Polyglactin910Rapid USP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	75000
1010420335	NEEDLE-20MM 3/8 CIR. ROUND BODY, 4/0 Specification: Needle:NEEDLE-20MM 3/8CIR. ROUNDBODY,4/0-ribbed&siliconised Wire diameter:0.58 (-0.0+0.02mm) Borediameter:0.33 (-0.01+0.02mm)	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Suture mtl:Polyglactin910 USP4-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420336	NEEDLE-45MM 3/8 CRB, PGA USP 1 Specification: Needle:NEEDLE-45MM 3/8CIRCLE ROUNDBODY,USP1-ribbed&siliconised Wire diameter:0.86 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:PolyglycolicacidUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420	6000
1010420337	NEEDLE-40MM 1/2 CRB, PGA-PLA USP 0 Specification: Needle:NEEDLE-40MM 1/2CIRCLE ROUNDBODY,USP0-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	1200
1010420338	NEEDLE-40MM 1/2 CRB, HEAVY PGA-PLA USP1 Specification: Needle:NEEDLE-40MM 1/2CIRCLE ROUNDBODY,HEAVY,USP 1-ribbed&siliconised Wire diameter:1.00 (-0.0+0.02mm) Borediameter:0.60 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420	1200
1010420339	NEEDLE-40MM 1/2 CTC, PGA-PLA USP0 Specification: Needle:NEEDLE-40MM 1/2CIRCLE TAPERCUTTING,USP0-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm)	6000

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420340	NEEDLE 40MM 1/2 CRC.PGA-PLA: 2/0	30000
	Specification: Needle:NEEDLE-40MM 1/2CIRCLE REVERSECUTTING,USP 2/0-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.47 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420341	NEEDLE-40MM, 3/8CRC,NYLON1	1200
	Specification: Needle:NEEDLE-40MM 3/8CIR. REVERSECUTTING,ribbed& siliconised Wire diameter:1 (-0.0+0.025mm) Borediameter:0.53 (-0.01+0.025mm) Suture mtl: Nylon1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420342	NEEDLE-40MM,1/2CRB,HEAVY,AISI 302,USP1	1500000
	Specification: Item name:Needle:NEEDLE-40MM, 1/2CIRCLE ROUND BODY, HEAVY-USP 1- Siliconised. 1.Wire diameter:1.00 (-0.0+ 0.025)mm 2. Borediameter:0.60 (-0.01+ 0.025)mm 3. Suture material:PGA-PLA,USP1 (Braided)	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

<i>Material Number</i>	<i>Material Description</i>	<i>Quantity</i>
		<i>pcs</i>
	4. Allneedlesmustbesiliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.	
1010420343	NEEDLE-40MM,1/2CTC,HEAVY,AISI 302,USP1 Specification: Item name:Needle:NEEDLE-40MM, 1/2CIRCLE TAPERCUTTING, HEAVY-USP 1 Siliconised. 1. Wire diameter:1.00 (-0.0+ 0.025)mm 2. Borediameter:0.60 (-0.01+ 0.025)mm 3. Suture material:PGA-PLA,USP1 (Braided) 4. Allneedlesmustbesiliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.	1000000
1010420344	NEEDLE-45MM,1/2CRC,PGA-PLA USP 2 Specification: Item name:Needle:NEEDLE-45MM, 1/2CIRCLE REVERSE CUTTING,-USP2 Ribbed &siliconised. 1. Wire diameter:1.20 (-0.0+0.03)mm 2. Borediameter:0.72 (-0.01+0.025)mm 3. Suture material:PGA-PLA,USP2 (Braided) 4. Allneedlesmustbesiliconised& ribbed. 5. Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	1200
1010420345	NEEDLE-19MM,1/2 CIR.CUT.,PGA-PLA USP 5/0 Specification: Item name:Needle:NEEDLE-19MM, 1/2CIRCLE CUTTING-USP5/0 Ribbed &siliconised. 1. Wire diameter:0.56 (-0.0+0.02)mm 2. Borediameter:0.28 (-0.01+0.02)mm	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	3. Suture material:PGA-PLA,USP 5/0(Braided) 4. Allneedlesmustbesiliconised& ribbed. 5. Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420346	NEEDLE-30MM,1/2 CRB,AISI 302,PGA-PLA 3/0	12000
	Specification: Item name:Needle:NEEDLE-30MM, 1/2CIRCLE ROUND BODY,USP 3-0 Siliconised&RIBBED. 1.Wire diameter:0.56 (-0.0+0.02)mm 2. Borediameter:0.36 (-0.01+ 0.02)mm 3. Suture material:PGA-PLA,USP 3-0(Braided) 4. Allneedlesmustbesiliconised& ribbed. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.	
1010420347	NEEDLE-25MM,3/8 CIRCLE PRIME CUTTING	120000
	Specification: Item name:Needle: NEEDLE-25MM,3/8 CIRCLE PRIME CUTTING, AISI302,USP3-0 Siliconised 1.Wire diameter:0.58 (-0.0+0.02)mm 2. Borediameter:0.36 (-0.01+ 0.02)mm 3. Suture material:PGA-PCL,USP 3-0 4. Allneedlesmustbesiliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.	
1010420348	NEEDLE-19MM, 3/8 CRC, PGA-PLA 3-0	3000
	Needle:19MM, 3/8CIRCLE, REVERSE CUTTING-ribbed&siliconised Wire diameter:0.61 (-0.0+0.02mm) Borediameter:0.36 (-0.01+0.02mm) Suture mtl: SILK USP3/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420349	NEEDLE-40MM 1/2 CRB,AISI 302,PGLA USP 0	300000
	Needle:NEEDLE-40MM 1/2CIRCLE	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	ROUNDBODY,USP0- ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High QualityAustenitic stainlesssteelGrade AISI-302.	
1010420350	NEEDLE-40MM 1/2 CTC,AISI 302,PGLA USP0	120000
	Needle:NEEDLE-40MM 1/2CIRCLE TAPERCUTTING,USP0- ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High QualityAustenitic stainlesssteelGrade AISI-302.	
1010420351	NEEDLE 30MM, 1/2 CRB,AISI 302,PGLA 2/0.	240000
	NEEDLE: 30MM, 1/2CIRCLE ROUNDBODY- Ribbed&siliconised, PGA-PLA: 2/0. Wire diameter:0.74(-0.0+0.02)mm Borediameter:0.47(-0.01+0.02)mm Material:High QualityAustenitic stainlesssteelGrade AISI-302.	
1010420352	NEEDLE-40MM 1/2 CIR.REV.CU. OS, AISI-302	240000
	Specification: NEEDLE:40MM, 1/2CIRCLE REVERSECUTTING,OS NEEDLE-ribbed&siliconised, pga-plaUSP1 Wire diameter:1.2(-0.0+0.03)mm Borediameter:0.60(-0.01+0.025)mm Material:High QualityAustenitic stainlesssteelGrade AISI-302.	
1010420353	NEEDLE -60 MM 3/8 CC SILK 1 BD053	120000
	Specification: Needle:60mm,3/8circle,cutting-ribbed &siliconised, silkusp 1 1. BoreDiameter(mm): 0.53(-0.01+0.025) 2. Wire diameter(mm): 1.20(-0.0+0.03)	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420354	NEEDLE-36MM 1/2 CRC,PGA-PLA USP 0 Specification: Needle:NEEDLE-36MM 1/2CIRCLE REVERSECUTTING,USP0-ribbed &siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420	1200
1010420355	NEEDLE-22MM 1/2 CRC, PGA-PLA USP2-0 Specification: NEEDLE: 22MM, 1/2CIRCLE REVERSECUTTING-Ribbed&siliconised, PGA-PLA: 2/0. Wire diameter:0.71(-0.0+0.02)mm Borediameter:0.47(-0.01+0.02)mm Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	1200
1010420356	NEEDLE-40MM 1/2 CRC, PGA-PLA USP2 Specification: Needle:NEEDLE-40MM 1/2CIRCLE REVERSECUTTING,USP2-ribbed &siliconised Wire diameter:1.2 (-0.0+0.03mm) Borediameter:0.72 (-0.01+0.025mm) Suture mtl:Polyglactin910 USP2 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	18000
1010420357	NEEDLE-22MM 1/2 CRC, PGA-PLA USP3-0 Specification: NEEDLE: 22MM, 1/2CIRCLE REVERSECUTTING-Ribbed& siliconised, PGA-PLA: 3/0. Wire diameter:0.62(-0.0+0.02)mm	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

<i>Material Number</i>	<i>Material Description</i>	<i>Quantity</i>
		<i>pcs</i>
	Borediameter:0.36(-0.01+0.02)mm Material:High Qualitymartensitic stainlesssteelGrade AISI-420	
1010420358	NEEDLE-60MM STRAIGHT CUTTING,USP 2-0	1200
	Specification: NEEDLE: 60MM, STRAIGHT CUTTING-Ribbed&siliconised, NYLON2/0. Wire diameter:0.74(-0.0+0.02)mm Borediameter:0.36(-0.01+0.02)mm Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420359	NEEDLE-50MM 1/2 CRB, PGA 0	24000
	Specification: Needle:NEEDLE-50MM 1/2CIRCLE ROUNDBODY,USP0-ribbed& siliconised Wire diameter:1 (-0.0+0.025)mm Borediameter:0.47 (-0.01+0.02)mm Suture mtl:PGA USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420360	NEEDLE-50MM, 1/2 CRC-PGA 2/0	1200
	Specification: Needle:NEEDLE-50MM 1/2CIRCLE REVERSECUTTING,USP 2/0-ribbed&siliconised Wire diameter:0.88 (-0.0+0.025mm) Borediameter:0.42 (-0.01+0.02mm) Suture mtl:PGA USP2/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420361	NEEDLE-50MM, 1/2 CRB-PGA 2/0	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Specification: Needle:NEEDLE-50MM 1/2CIRCLE ROUNDBODY,USP2/0-ribbed& siliconised Wire diameter:0.88 (-0.0+0.025)mm Borediameter:0.42 (-0.01+0.02)mm Suture mtl:PGA USP2/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420	
1010420362	NEEDLE-50MM, 1/2 CRB, PGA-PLA 1,AISI 302	240000
	"Item name: Needle: NEEDLE-50MM, 1/2CIRCLE ROUNDBODY,HEAVY-USP 1 Siliconised, AISI-302 1.Wire diameter:1.08 (-0.0+ 0.025)mm 2. Borediameter:0.60 (-0.01+ 0.025)mm 3. Suture material:PGA-PLA,USP1 (Braided) 4. Allneedlesmustberibbedand siliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.	
1010420363	NEEDLE-45MM 1/2 CTC, HEAVY, PET USP 2	75000
	Item name: Needle: NEEDLE-45MM, 1/2CIRCLE TAPERCUTTING,HEAVY-USP 2 SiliconisedandRibbed. 1. Wire diameter:1.20 (-0.0+ 0.03)mm 2. Borediameter:0.65 (-0.01+ 0.025)mm 3. Suture material:Polyester,USP2 (Braided) 4. Allneedlesmustbesiliconised and ribbed. 5. Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420364	NEEDLE-26MM, 1/2 CC, PGA-PLA USP 3/0	1200
	Specification: Item name:NEEDLE-26MM, ½ CIRCLE CUTTING,Siliconised. 1.Wire diameter:0.56 (-0.0+0.02)mm	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	2. Borediameter:0.36 (-0.01+ 0.02)mm 3. Suture material:PGA-PLA,USP3-0 (Braided) 4. Allneedlesmustbesiliconised& ribbed. 5. Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010420365	NEEDLE 45MM,1/2CRB 4-0	1200
	"Item name: Needle: NEEDLE-45 MM, 1/2CIRCLE ROUNDBODY, USP4-0, AISI- 302 1. Wire diameter:0.81(-0.0+0.02)mm 2. Borediameter:0.28 (-0.01+ 0.02)mm 3. Suture material: PLA,USP 4-0 (Braided) 4. Allneedlesmustberibbedand siliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.	
1010420366	NEEDLE 45MM,1/2CRB 5-0	1200
	Item name: Needle: NEEDLE-45 MM, 1/2CIRCLE ROUNDBODY, USP5-0, AISI- 302 1. Wire diameter:0.74 (-0.0+0.02)mm 2. Borediameter:0.22 (-0.01+ 0.02)mm 3. Suture material: PLA,USP 5-0 (Braided) 4. Allneedlesmustberibbedand siliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.	
1010420367	NEEDLE-40MM 1/2 CTC, PGA, USP 2	6000
	Item name: Needle: NEEDLE-40MM, 1/2CIRCLE TAPERCUTTING- USP2 SiliconisedandRibbed. 1. Wire diameter:1.12 (-0.0+ 0.03)mm 2. Borediameter:0.60 (-0.01+ 0.025)mm 3. Suture material:PGA, USP2 (Braided) 4. Allneedlesmustbesiliconised and ribbed. 5. Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420368	NEEDLE-45 MM,3/8 CRC, NYLON-USP 0	75000
	1. Wire diameter: 1(-0.0+0.025)mm 2. Bore diameter: 0.45(-0.01+0.02)mm 3. Suture material: Nylon USP 0 (Monofilament) 4. All needles must be siliconised & ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	
1010420369	NEEDLE-48 MM,1/2 CTC, PGA-PLA USP2	1200
	Specification: Item: 48 mm, 1/2 Circle Taper Cutting, Ribbed & siliconised. SUTURE SIZE: Polyglactin 910 USP 2 Wire diameter: 1.20 (-0.0+0.03) mm Bore diameter: 0.72 (-0.01+0.025) mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420370	NEEDLE 48MM 1/2 CRC-PGA PLA 2	1200
	Item name: Needle: NEEDLE-48mm, 1/2 circle round body. PGA-PLA USP 0 Siliconised and Ribbed. 1. Wire diameter: 1.12 (-0.0 + 0.03)mm 2. Bore diameter: 0.55 (-0.01 + 0.025)mm 3. Suture material: PGLA, USP 0(Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420371	NEEDLE 48MM 1/2 CRB-PGA PLA 0	1200
	Item name: Needle: NEEDLE-48mm, 1/2 circle round body. PGA-PLA USP 0 Siliconised and Ribbed. 1. Wire diameter: 1.12 (-0.0 + 0.03)mm 2. Bore diameter: 0.55 (-0.01 + 0.025)mm 3. Suture material: PGLA, USP 0(Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420372	NEEDLE 45MM ½ CRC PGA PLA-1	1200
	Item name: Needle: NEEDLE-45mm, 1/2 circle reverse cutting. PGA-PLA USP 1 Siliconised and Ribbed. 1. Wire diameter: 1.12 (-0.0+0.03)mm 2. Bore diameter: 0.60 (-0.01+0.025)mm 3. Suture material: PGLA, USP 1 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AIST-420	
1010420373	NEEDLE 16MM ½ CRC PGA PLA-4-0	1200
	Item name: Needle: NEEDLE-16mm, 1/2 circle reverse cutting. PGA-PLA USP 4/0 Siliconised and Ribbed. 1. Wire diameter: 0.56 (-0.0+0.02)mm 2. Bore diameter: 0.33 (-0.01+0.02)mm 3. Suture material: PGLA, USP 4/0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	
1010420374	NEEDLE 31 MM ½ CRB PGA PLA-2-0	2400
	Item name: Needle: NEEDLE-31mm, 1/2 circle round body. PGA-PLA USP 2/0 Siliconised and Ribbed. 1. Wire diameter: 0.80 (-0.0+0.02)mm 2. Bore diameter: 0.47 (-0.01+0.02)mm 3. Suture material: PGLA, USP 2/0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	
1010420375	NEEDLE 31MM ½ CRC PGA PLA-2-0	12000
	Item name: Needle: NEEDLE-31mm, 1/2 circle reverse cutting. PGA-PLA USP 2/0 Siliconised and Ribbed. 1. Wire diameter: 0.80 (-0.0+0.02)mm 2. Bore diameter: 0.47 (-0.01+0.02)mm 3. Suture material: PGLA, USP 2/0 (Braided)	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	
1010420376	NEEDLE 26 MM ½ CRB PGA PLA 3-0 Item name: Needle: NEEDLE-26mm, 1/2 circle round body. PGA-PLA USP 3/0 Siliconised and Ribbed. 1. Wire diameter: 0.66 (-0.0+0.02)mm 2. Bore diameter: 0.36 (-0.01+0.02)mm 3. Suture material: PGLA, USP 3/0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	12000
1010420377	NEEDLE 50 MM ½ CRC PGA PLA 0 Item name: Needle: NEEDLE-50mm, 1/2 circle reverse cutting. PGA-PLA USP 0 Siliconised and Ribbed. 1. Wire diameter:1.12 (-0.0+0.03)mm 2. Bore diameter: 0.55 (-0.01+0.025)mm 3. Suture material: PGLA, USP 0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	1200
1010420378	NEEDLE 45MM ½ CRB PGA PLA -1 Item name: Needle: NEEDLE-45mm, 1/2 circle round body. PGA-PLA USP 1 Siliconised and Ribbed. 1. Wire diameter: 1.12 (-0.0+0.03)mm 2. Bore diameter: 0.60 (-0.01+0.025)mm 3. Suture material: PGLA, USP 1 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	1200
1010420379	NEEDLE-35 MM,3/8 CRC,RAPID 0,B55W1 "Needle: NEEDLE-35MM 3/8 CIRCLE REVERSE CUTTING, USP 0-ribbed & siliconised Wire diameter: 1.00 (-0.0+0.025mm)	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Bore diameter: 0.55 (-0.01+0.02mm) Suture mtl: PGA-PLA Rapid USP 0 Material: High Quality martensitic stainless steel grade AIST-420	
1010420380	NEEDLE-23 MM,1/2 CTC,PDO, USP 0	1200
	"Needle: NEEDLE-23MM 1/2 CIRCLE TAPER CUTTING, USP 0-ribbed & siliconised Wire diameter: 0.81 (-0.0+0.02mm) Bore diameter: 0.55 (-0.01+0.02mm) Suture mtl: Polydioxanone USP 0 Material: High Quality martensitic stainless steel grade AIST-420	
1010420381	NEEDLE-55MM,1/2 CTC,HEAVY,POLYESTER- 5	60000
	Needle: NEEDLE-55MM 1/2 CIRCLE TAPER CUTTING- POLYESTER USP 5-ribbed & siliconised Wire diameter: 1.3 (-0.0+0.03mm) Bore diameter: 0.90 (-0.01+0.02mm) Suture mtl: Polyester USP 5 Material: High Quality martensitic stainless steel grade AIST-420	
1010420382	NEEDLE-90 MM, 3/8 CCC,NYLON USP 1	1200
	"Item name: Needle: NEEDLE-90 mm, 3/8 Circle Conventional Cutting,Nylon USP 1 Siliconised and Ribbed. 1. Wire diameter: 1.20 (-0.0+0.03)mm 2. Bore diameter: 0.53 (-0.01+0.025)mm 3. Suture material: Nylon, USP 1 (Monofilament) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	
1010420383	NEEDLE-65MM,1/2CRB,PGA-1	1200
	NEEDLE FOR PGA SUTURE(BRAIDED) 65 MM, 1/2 CIRCLE ROUND BODY SIZE:1 (65mm,1/2circle,round body(heavy))Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.02mm Material:High Quality	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

<i>Material Number</i>	<i>Material Description</i>	<i>Quantity</i>
		<i>pcs</i>
	martensitic stainless steel Grade AISI-420.	
1010420384	NEEDLE-30MM,1/2CRB,PGA-4/0	1200
	Needle 30 MM,1/2 CIRCLE ROUND BODY Suture size:4/0 PGA Material:high quality Martensitic grade stainless steel Grade;AISI-420 CORROSION RESISTANT,Free from dirt,grease,fibresetc Needle type:Drilled end suitable for above said applications Bore dia 0.28 +/-0.02mm Wire dia : 0.48 to 0.51mm	
1010420385	NEEDLE-30MM,1/2CRC,NYLON 4/0	1200
	Needle 30 MM ,1/2 CIRCLE REVERSE CUTTING Wire diameter: 0.54+/-0.02mm Bore diameter: 0.25+/-0.01mm Suture size: Monofilament Nylon Suture, USP 4-0 Material: High Quality martensitic Stainless Steel Grade AIST-420	
1010420386	NEEDLE-48MM,1/2 CRB,PGAPLA2-0	24000
	"Item name: Needle: NEEDLE- 48MM, 1/2 CRB,PGLA 2-0 Siliconised and Ribbed. 1. Wire diameter: 1.12 (-0.0+0.03) mm 2. Bore diameter: 0.47 (-0.01+0.02) mm 3. Suture material: PGLA, USP 2-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AIST-420	
1010420387	NEEDLE-16MM,1/2 CRB,PGAPLA3-0	12000
	Item name: Needle: NEEDLE- 16 MM, 1/2 CRB,PGLA 3-0 Siliconised and Ribbed.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	1. Wire diameter: 0.61 (-0.0+0.02) mm 2. Bore diameter: 0.36 (-0.01+0.02) mm 3. Suture material: PGLA, USP 3-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420388	NEEDLE- 45MM, 1/2 CRB,PGA3-0	60000
	Item name: Needle: NEEDLE- 45MM, 1/2 CRB,PGA3-0 Siliconised and Ribbed. 1. Wire diameter: 0.84 (-0.0+0.025) mm 2. Bore diameter: 0.33 (-0.01+0.02) mm 3. Suture material: PGA, USP 3-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AIST-420	
1010420389	NEEDLE-35MM, 3/8 CRC,HISIL 2-0	3000
	tem name: Needle: NEEDLE-35MM, 3/8 CIRCLE REVERSE CUTTING-USP 2-0 Siliconised and Ribbed. 1. Wire diameter: 0.84 (-0.0+0.025)mm 2. Bore diameter: 0.42 (-0.01+0.02)mm 3. Suture material: Silk, USP 2-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	
1010420390	NEEDLE-35MM,3/8 CRC,HISIL 0	3000
	Item name: Needle: NEEDLE-35MM, 3/8 CIRCLE REVERSE CUTTING-USP 0 Siliconised and Ribbed. 1. Wire diameter: 0.92 (-0.0+0.025)mm 2. Bore diameter: 0.47 (-0.01+0.02)mm 3. Suture material: Silk, USP 0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	
1010420391	NEEDLE 30MM-1/2 CC(PGCL) 2/0	6000

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Specification: Needle: 30mm, 1/2 circle cutting Wire diameter: 0.84(-0.0+0.025)mm Bore diameter: 0.45(-0.01+0.02)mm Suture size: PGA-PCL, USP 2-0 Material: High Quality martensitic stainless steel Grde AIST-420	
1010420392	NEEDLE 30MM-1/2 CRC(PGCL) 2/0	6000
	Specification: Needle: 30mm, 1/2 circle reverse cutting Wire diameter: 0.84(-0.0+0.025)mm Bore diameter: 0.45(-0.01+0.02)mm Suture size: PGA-PCL, USP 2-0 Material: High Quality martensitic stainless steel Grade AIST-420	
1010420393	NEEDLE 20MM-1/2 CC(PGCL) 4/0	2400
	Specification: Needle: 20mm, 1/2 circle cutting Wire diameter: 0.66(-0.0+0.02)mm Bore diameter: 0.28(-0.01+0.02)mm Suture size: PGA-PCL, USP 4-0 Material: High Quality martensitic stainless steel Grde AIST-420	
1010420394	NEEDLE 30MM-1/2CRB(PGCL)3/0	6000
	Specification: Needle: 30mm, 1/2 circle round body Wire diameter: 0.74(-0.0+0.02)mm Bore diameter: 0.36(-0.01+0.02)mm Suture size: PGA-PCL, USP 3-0 Material: High Quality martensitic stainless steel Grade AIST-420	
1010420396	NEEDLE 20MM-1/2 CRC(PGCL) 4/0	6000
	Specification: Needle: 20mm, 1/2 circle reverse cutting Wire diameter: 0.66(-0.0+0.02)mm Bore diameter: 0.28(-0.01+0.02)mm Suture size: PGA-PCL, USP 4-0	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Material: High Quality martensitic stainless steel Grade AIST-420	
1010420397	NEEDLE 30MM-1/2 CRB(PGCL) 4/0	6000
	pecification: Needle: 30mm, 1/2 circle round body Wire diameter: 0.71(-0.0+0.02)mm Bore diameter: 0.28(-0.01+0.02)mm Suture size: PGA-PCL, USP 4-0 Material: High Quality martensitic stainless steel Grade AIST-420	
1010420398	NEEDLE 45MM-1/2 CRB(HIDOX)	1200
	Specification: Needle: 45mm, 1/2 circle round body Wire diameter: 1.12 (-0.0+0.03)mm Bore diameter: 0.60 (-0.01+0.025)mm Suture size: PDO USP 1 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420399	NEEDLE 45MM-1/2 CRB(HIDOX) 2	12000
	pecification: Needle: 45mm, 1/2 circle round body Wire diameter: 1.12 (-0.0+0.03)mm Bore diameter: 0.72 (-0.01+0.025)mm Suture size: PDO USP 2 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420400	NEEDLE 20MM-1/2 CRB(HIDOX)2/0	6000
	Specification: Needle: 20mm, 1/2 circle round body Wire diameter: 0.71(-0.0+0.02)mm Bore diameter: 0.45(-0.01+0.02)mm Suture size: PDO USP 2/0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420401	NEEDLE 26MM-1/2CRC(PP)2/0	6000

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Needle: NEEDLE-26MM 1/2 CIRCLE REVERSE CUTTING, USP 2-0-ribbed &siliconised Wire diameter: 0.71 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: Polypropylene USP 2-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420402	NEEDLE 22MM-1/2CRB(PP)3/0	1200
	eedle: NEEDLE-22MM 1/2 CIRCLE ROUND BODY, USP 3-0-ribbed &siliconised Wire diameter: 0.61 (-0.0+0.02mm) Bore diameter: 0.28 (-0.01+0.02mm) Suture mtl: Polypropylene USP 3-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420403	NEEDLE 22MM-1/2CRC(PP)3/0	1200
	Needle: NEEDLE-22MM 1/2 CIRCLE REVERSE CUTTING, USP 3-0-ribbed &siliconised Wire diameter: 0.61 (-0.0+0.02mm) Bore diameter: 0.28 (-0.01+0.02mm) Suture mtl: Polydioxanone USP 3-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420404	NEEDLE 20MM-1/2 CRC(HIDOX)3/0	6000
	E-20MM 1/2 CIRCLE REVERSE CUTTING, USP 3-0-ribbed &siliconised Wire diameter: 0.61 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: Polydioxanone USP 3-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420405	NEEDLE 16MM-1/2 CRC(HIDOX)4/0	3000
	Needle: NEEDLE-16MM 1/2 CIRCLE REVERSE CUTTING, USP 4-0-ribbed &siliconised	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Wire diameter: 0.46 (-0.0+0.02mm) Bore diameter: 0.28 (-0.01+0.02mm) Suture mtl: Polydioxanone USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420406	NEEDLE 12MM-1/2 CRC(RAPID)5/0	1200
	Needle: NEEDLE-12MM 1/2 CIRCLE REVERSE CUTTING, USP 5-0-ribbed &siliconised Wire diameter: 0.38 (-0.0+0.02mm) Bore diameter: 0.28 (-0.01+0.02mm) Suture mtl: PGA-PLA RAPID USP 5-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420407	NEEDLE 13MM-1/2CRC(PP)5/0	18000
	Needle: NEEDLE-13MM 1/2 CIRCLE REVERSE CUTTING, USP 5-0-ribbed &siliconised Wire diameter: 0.38 (-0.0+0.02mm) Bore diameter: 0.19 (-0.01+0.02mm) Suture mtl: Polypropylene USP 5-0 Material: High Quality martensitic stainless steel grade AISI-420.	
1010420408	NEEDLE 31MM-1/2CRB(PP)2/0	6000
	Needle: NEEDLE-31MM 1/2 CIRCLE ROUND BODY, USP 2-0-ribbed &siliconised Wire diameter: 0.74 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: Polypropylene USP 2-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420409	NEEDLE 26MM-1/2CRB(PP)2/0	6000
	Needle: NEEDLE-26MM 1/2 CIRCLE ROUND BODY, USP 2-0-ribbed &siliconised Wire diameter: 0.71 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: Polypropylene USP2-0	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420410	NEEDLE 16MM-1/2 CRB(PP)4/0 Needle: NEEDLE-16MM 1/2 CIRCLE ROUND BODY, USP 4-0-ribbed &siliconised Wire diameter: 0.56 (-0.0+0.02mm) Bore diameter: 0.25 (-0.01+0.02mm) Suture mtl: Polypropylene USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.	6000
1010420411	NEEDLE 13MM-1/2CRB(PP)5/0 Needle: NEEDLE-13MM 1/2 CIRCLE ROUND BODY, USP 5-0-ribbed &siliconised Wire diameter: 0.38 (-0.0+0.02mm) Bore diameter: 0.19 (-0.01+0.01mm) Suture mtl: Polypropylene USP 5-0 Material: High Quality martensitic stainless steel Grade AISI-420.	12000
1010420412	NEEDLE 30MM-1/2CRC(PGCL)3/0 Needle: NEEDLE-30MM 1/2 CIRCLE REVERSE CUTTING, USP 3-0-ribbed &siliconised Wire diameter: 0.71 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: Polyglecaprone25 USP 3-0 Material: High Quality martensitic stainless steel Grade AISI-420.	1200
1010420413	NEEDLE 25MM-1/2CRC (PP)4/0 Needle: NEEDLE-25MM 1/2 CIRCLE REVERSE CUTTING, USP 4-0-ribbed &siliconised Wire diameter: 0.51 (-0.0+0.02mm) Bore diameter: 0.25 (-0.01+0.02mm) Suture mtl: Polypropylene USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.	1200

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420414	NEEDLE 25 MM-1/2CRB(PP)4/0	6000
	Needle: NEEDLE-25MM 1/2 CIRCLE ROUND BODY, USP 4-0-ribbed &siliconised Wire diameter: 0.51 (-0.0+0.02mm) Bore diameter: 0.25 (-0.01+0.02mm) Suture mtl: Polypropylene USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420415	NEEDLE 26MM-1/2CRC(PGLA)3/0	1200
	Needle: NEEDLE-26MM 1/2 CIRCLE REVERSE CUTTING, USP 3-0-ribbed &siliconised Wire diameter: 0.71 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: PGA-PLA USP 3-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420416	NEEDLE- 26MM, 1/2 C TROCAR,PGLA2-0	24000
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide) USP 2-0 Wire diameter: 0.84 (-0.0+0.025)mm Bore diameter: 0.47 (-0.01+0.02)mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420417	NEEDLE-40MM ,1/2C TROCAR,PGLA1	24000
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide) USP 1 Wire diameter: 1.00 (-0.0+0.025)mm Bore diameter: 0.60 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420418	NEEDLE-30MM,1/2C TROCAR,PGLA0	24000
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide) USP 0	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Wire diameter: 0.92 (-0.0+0.025)mm Bore diameter: 0.55 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420419	NEEDLE-40MM,1/2CTC,SHORT TIP,SS302,USP1	100000
	""Specification: Item name: Needle: NEEDLE-40MM, 1/2 CIRCLE TAPER CUTTING, HEAVY-USP 1 with short taper cutting edge 1.Wire diameter: 1.00 (-0.0 + 0.025)mm 2. Bore diameter: 0.60 (-0.01 + 0.025)mm 3. Suture material: PGA-PLA, USP 1 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel G	
1010420420	NEEDLE-35MM,1/2CTC,SS302,USP2-0	1200
	pecification: Item name: Needle: NEEDLE-35MM, 1/2 CIRCLE TAPER CUTTING-USP 2-0 1.Wire diameter: 0.88 (-0.0 + 0.02)mm 2. Bore diameter: 0.47 (-0.01 + 0.025)mm 3. Suture material: PGA-PLA, USP 2-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-302.	
1010420421	NEEDLE-31MM,1/2CRB,PGLA-0	6000
	Specification: Item name: Needle: NEEDLE-31MM, 1/2 CIRCLE ROUND BODY-USP 1-0 1.Wire diameter: 0.84 (-0.0 + 0.025)mm 2. Bore diameter: 0.55 (-0.01 + 0.025)mm 3. Suture material: PGA-PLA, USP 1-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420422	NEEDLE-17MM,1/2CRB,PGLA5-0	1200
	Specification: Item name: Needle: NEEDLE-17MM, 1/2 CIRCLE ROUND BODY -USP 5-0 1.Wire diameter: 0.51 (-0.0 + 0.02)mm 2. Bore diameter: 0.28 (-0.01 + 0.02)mm 3. Suture material: PGA-PLA, USP 5-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420423	NEEDLE-20MM,3/8CC,PGLA4-0	2400
	Specification: Item name: Needle: NEEDLE-20MM, 3/8 CIRCLE CUTTING-USP 4-0 1.Wire diameter: 0.58 (-0.0 + 0.02)mm 2. Bore diameter: 0.33(-0.01 + 0.02)mm 3. Suture material: PGA-PLA, USP 4-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420424	NEEDLE-36MM,1/2CRB,PGLA-0	1200
	Specification: Item name: Needle: NEEDLE-36MM, 1/2 CIRCLE ROUND BODY -USP 1-0 1.Wire diameter: 1.00(-0.0 + 0.025)mm 2. Bore diameter: 0.55 (-0.01 + 0.025)mm 3. Suture material: PGA-PLA, USP 1-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420425	NEEDLE-40MM,1/2CC,PGLA-2	3000
	Specification: Item name: Needle: NEEDLE-40MM, 1/2 CIRCLE CUTTING-USP 2 1.Wire diameter: 1.20 (-0.0 + 0.03)mm	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	2. Bore diameter: 0.72 (-0.01 + 0.025)mm 3. Suture material: PGA-PLA, USP 2 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420426	NEEDLE-19MM,3/8CC,PGLA3-0	1200
	Specification: Item name: Needle: NEEDLE-19MM, 3/8 CIRCLE CUTTING-USP 3-0 1.Wire diameter: 0.62 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PGA-PLA, USP 3-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420427	NEEDLE-40MM,1/2CRB,HILON2-0	12000
	Specification: Item name: Needle: NEEDLE-40MM, 1/2 CIRCLE ROUND BODY-USP 2-0 1.Wire diameter: 1.00 (-0.0 + 0.025)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 2-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420428	NEEDLE-40MM,1/2CRC,HILON2-0	12000
	Specification: Item name: Needle: NEEDLE-40 MM, 1/2 CIRCLE REVERSE CUTTING-USP 2-0 1.Wire diameter: 1.00 (-0.0 + 0.025)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 2-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420429	NEEDLE-36MM,1/2CRB,HILON3-0	6000
	Specification: Item name: Needle: NEEDLE-36MM, 1/2 CIRCLE ROUND BODY-USP 3-0 1.Wire diameter: 0.86 (-0.0 + 0.025)mm 2. Bore diameter: 0.28 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 3-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420430	NEEDLE-36MM,1/2CRC,HILON3-0	6000
	Specification: Item name: Needle: NEEDLE-36MM, 1/2 CIRCLE REVERSE CUTTING-USP 3-0 1.Wire diameter: 0.86 (-0.0 + 0.025)mm 2. Bore diameter: 0.28 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 3-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420431	NEEDLE-20MM,1/2CRB,HILON4-0	6000
	Specification: Item name: Needle: NEEDLE-20MM, 1/2 CIRCLE ROUND BODY -USP 4-0 1.Wire diameter: 0.56 (-0.0 + 0.02)mm 2. Bore diameter: 0.25 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 4-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420432	NEEDLE-20MM,1/2CRC,HILON4-0	1200
	Specification: Item name: Needle: NEEDLE-20MM, 1/2 CIRCLE	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	REVERSE CUTTING-USP 4-0 1.Wire diameter: 0.56 (-0.0 + 0.02)mm 2. Bore diameter: 0.25 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 4-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420433	NEEDLE-20MM,1/2CRB,HILON5-0	2400
	Specification: Item name: Needle: NEEDLE-20 MM, 1/2 CIRCLE ROUND BODY -USP 5-0 1.Wire diameter: 0.51 (-0.0 + 0.02)mm 2. Bore diameter: 0.19 (+/-0.01)mm 3. Suture material: Nylon, USP 5-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420434	NEEDLE-20MM,1/2CRC,HILON5-0	6000
	Specification: Item name: Needle: NEEDLE-20 MM, 1/2 CIRCLE REVERSE CUTTING-USP 5-0 1.Wire diameter: 0.51 (-0.0 + 0.02)mm 2. Bore diameter: 0.19 (+/-0.01)mm 3. Suture material: Nylon, USP 5-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420435	NEEDLE-38MM,1/2CRC,HILON2-0	1200
	Specification: Item name: Needle: NEEDLE-38MM, 1/2 CIRCLE REVERSE CUTTING-USP 2-0 1.Wire diameter: 1.00 (-0.0 + 0.025)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 2-0 4. All needles must be siliconised and ribbed.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	5. Material: High Quality austenitic stainless steel Grade AISI-420.	
1010420436	NEEDLE-38MM,1/2CRC,HILON1-0 Specification: Item name: Needle: NEEDLE-38MM, 1/2 CIRCLE REVERSE CUTTING-USP 1-0 1.Wire diameter: 1.00 (-0.0 + 0.025)mm 2. Bore diameter: 0.45 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 1-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AIST-420	1200
1010420437	NEEDLE-38MM,1/2CRC,HILON1 Specification: Item name: Needle: NEEDLE-38MM, 1/2 CIRCLE REVERSE CUTTING-USP 1 1.Wire diameter: 1.12 (-0.0 + 0.025)mm 2. Bore diameter: 0.53 (-0.01 + 0.025)mm 3. Suture material: Nylon, USP 1 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	1200
1010420438	NEEDLE-40MM,1/2CRC,HILON2 Specification: Item: 40mm, 1/2 Circle Reverse Cutting, Ribbed &siliconised. SUTURE SIZE: Polyamide Nylon USP 2 Wire diameter: 1.12 (-0.0+0.03) mm Bore diameter: 0.60 (-0.01+0.025) mm Material: High Quality martensitic stainless steel Grade AISI-420.	2400
1010420439	NEEDLE-45MM,1/2CRB,HILON2 Specification: Item: 45MM, 1/2 Circle Round Body, Ribbed &siliconised. SUTURE SIZE: Polyamide Nylon USP 2	1200

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Wire diameter: 1.20 (-0.0+0.03) mm Bore diameter: 0.60 (-0.01+0.025) mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420440	NEEDLE-45MM,1/2CRC,HILON2	1200
	Specification: Item: 45MM, 1/2 Circle Reverse cutting, Ribbed &siliconised. SUTURE SIZE: Polyamide Nylon USP 2 Wire diameter: 1.20 (-0.0+0.03) mm Bore diameter: 0.60 (-0.01+0.025) mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420441	NEEDLE-45MM,1/2CRC,PGLA -0	6000
	Specification: Item: 45mm, 1/2 circle Reverse cutting. SUTURE SIZE: Polyglactin 910 USP 0 Wire diameter: 1.00 (-0.0+0.025) mm Bore diameter: 0.55 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420442	NEEDLE-30MM,1/2CRB,PGCL-0	1200
	Specification: Item: 30mm, 1/2 circle round body. SUTURE SIZE: Polyglecaprone 25 USP 0 Wire diameter: 0.92 (-0.0+0.025) mm Bore diameter: 0.55 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420443	NEEDLE-36MM,1/2CRB,HILON2-0	3000
	Specification: Item: 36mm, 1/2 circle round body. SUTURE SIZE: Nylon USP 2-0 Wire diameter: 0.88 (-0.0+0.025) mm	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	Bore diameter: 0.36 (-0.01+0.02)mm Material: High Quality martensitic stainless steel Grade AIST-420	
1010420444	NEEDLE-36MM,1/2CRC,HILON2-0	3000
	Specification: Item: 36mm, ½ circle reverse cutting. SUTURE SIZE: Nylon USP 2-0 Wire diameter: 0.88 (-0.0+0.025) mm Bore diameter: 0.36 (-0.01+0.02)mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420448	NEEDLE-48MM-1/2 CRB,RAPID-1	6000
	Specification: Item: 48mm, ½ circle Round Body. SUTURE SIZE: Polyglactin 910 Undyed USP 1 Wire diameter: 1.20 (-0.0+0.025) mm Bore diameter: 0.60 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420449	NEEDLE-48MM-1/2 CRC,RAPID-1	3000
	Specification: Item: 48mm, ½ circle Reverse cutting. SUTURE SIZE: Polyglactin 910 Undyed USP 1 Wire diameter: 1.20 (-0.0+0.025) mm Bore diameter: 0.60 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420450	NEEDLE-48MM-1/2 CRB,PGCL-1	1200
	Specification: Item: 48mm, ½ circle Round body. SUTURE SIZE: Polyglecaprone 25 Undyed USP 1 Wire diameter: 1.20 (-0.0+0.025) mm Bore diameter: 0.60 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010420451	NEEDLE-48MM-1/2 CRB,PGCL-2	3000
	Specification: Item: 48mm, ½ circle Round body. SUTURE SIZE: Polyglecaprone 25 Undyed USP 2 Wire diameter: 1.20 (-0.0+0.025) mm Bore diameter: 0.72 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420452	NEEDLE_3/8 CIRCLE RC 22MM PGLA 4-0	1200
	""Specification: Needle: 22mm, 3/8 circle reverse cutting Wire diameter: 0.61 (-0.0+0.02mm) Bore diameter: 0.33mm (-0.01+0.02) Suture mtl: Polyglactin 910 USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420453	NEEDLE 13MM 1/2 CRB VISI BLACK PP 5-0	1200
	Specification: Needle: NEEDLE 13MM1/2 CRB VISI BLACK Wire diameter: 0.46mm (-0.0+0.02) Bore diameter: 0.19mm (-0.01+0.02) Suture mtl: Polyglactin 910 USP 2-0 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010420454	NEEDLE 30MM-3/8CRC PA2-0	1200
	Specification: Needle: 30MM-3/8 circle reverse cutting Wire diameter: 0.99mm (-0.0+0.02) Bore diameter: 0.36mm (-0.01+0.02) Suture mtl: Polyamide USP 2-0 Material: High Quality martensitic stainless steel Grade AISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010430008	36MM, 3/8CIRCLE REVERSE CUTTING	1200
	Specification: Needle:36MM, 3/8CIRCLE REVERSE CUTTING-ribbed& siliconised Wire diameter:1.00 +/- 0.01 mm Borediameter:0.60+/- 0.01 mm Suture mtl:Polyglactin910 USP1 Material:High Quality martensitic stainless steel Grade AISI-420.	
1010430009	36MM, 1/2CIRCLE TAPER CUTTING	3000
	Specification: Needle:36MM, 1/2CIRCLE TAPER CUTTING-ribbed& siliconised Wire diameter:0.84 +/- 0.01 mm Borediameter:0.55+/- 0.01 mm Suture mtl:Polyglactin910 USP0 Material:High Quality martensitic stainless steel Grade AISI-420.	
1010430010	35MM, 1/2CIRCLE TAPER CUTTING	75000
	Specification: Needle:35MM, 1/2CIRCLE TAPER CUTTING-ribbed& siliconised Wire diameter:0.80 +/- 0.01 mm Borediameter:0.47+/- 0.01 mm Suture mtl:Polyglactin910 USP2-0 Material:High Quality martensitic stainless steel Grade AISI-420	
1010430011	16MM, 3/8CIRCLE CUTTING	6000
	Specification: Needle:16MM, 3/8CIRCLE CUTTING-ribbed& siliconised Wire diameter:0.54 +/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:Polyglactin910 USP3-0 Material:High Quality martensitic stainless steel Grade AISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010430012	22MM, 1/2CIRCLE CUTTING	6000
	Specification: Needle:22MM, 1/2CIRCLE CUTTING-ribbed&siliconised Wire diameter:0.62 +/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:Polyglactin910 USP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010430013	26MM, 3/8CIRCLE REVERSE CUTTING	24000
	Specification: Needle:26MM, 3/8CIRCLE REVERSECUTTING-ribbed& siliconised Wire diameter:0.62 +/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:Polyglactin910 USP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010430014	26MM, 3/8CIRCLE ROUND BODY	12000
	Specification: Needle:26MM, 3/8CIRCLE ROUND BODY-ribbed&siliconised Wire diameter:0.62 +/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:Polyglactin910 USP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	
1010430015	16MM, 3/8CIRCLE CUTTING	24000
	Specification: Needle: 16MM, 3/8CIRCLE CUTTING-ribbed &siliconised Wire diameter: 0.48 +/- 0.01 mm Bore diameter: 0.33+/- 0.01 mm Suture mtl: Polyglactin 910 USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
1010430017	45MM, 1/2CIRCLE, ROUND BODY	3000
	Specification: Needle: 45MM, 1/2CIRCLE ROUND BODY-ribbed &siliconised Wire diameter: 1.2 (-0.0mm +0.03mm) Bore diameter: 0.60 (-0.01 mm +0.02mm) Suture mtl: Polyglycolicacid USP 2 Material: High Quality martensitic stainless steel Grade AISI-420.	
1010430028	NEEDLE 45MM,1/2 CRB PGA1	60000
	Specification: Item name: Needle: NEEDLE-45MM, 1/2 CIRCLE ROUND BODY, PGA-USP 1 Ribbed &siliconised. 1. Wire diameter: 1.12(-0.0 + 0.03)mm 2. Bore diameter: 0.55(-0.01 + 0.025)mm 3. Suture material: PGA USP 1 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010430029	NEEDLE 45MM,1/2 CRC PGA2-0	24000
	Specification: Item name: Needle: NEEDLE-45MM, 1/2 CIRCLE REVERSE CUTTING, PGA-USP 2-0 Ribbed &siliconised. 1. Wire diameter: 0.92(-0.0+0.025)mm 2. Bore diameter: 0.42(-0.01+0.02)mm 3. Suture material: PGA USP 2-0 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010430030	NEEDLE 45MM,1/2 CRB PGA2-0	60000
	Specification: Item name: Needle: NEEDLE-45MM, 1/2 CIRCLE ROUND BODY, PGA-USP 2-0 Ribbed &siliconised. 1. Wire diameter: 0.92(-0.0+0.025)mm 2. Bore diameter: 0.42(-0.01+0.02)mm 3. Suture material: PGA USP 2-0 (Braided) 4. All needles must be siliconised& ribbed.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010430031	NEEDLE 48MM,1/2 CRB PGA1	1200
	Specification: Item name: Needle: NEEDLE-48MM, 1/2 CIRCLE ROUND BODY, PGA-USP 1 Ribbed &siliconised. 1.Wire diameter: 1.12(-0.0 + 0.03)mm 2. Bore diameter: 0.55(-0.01 +0.025)mm 3. Suture material: PGA USP 1 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010430032	NEEDLE 48MM,1/2 CRB PGA2	1200
	Specification: Item name: Needle: NEEDLE-48MM, 1/2 CIRCLE ROUND BODY, PGA-USP 2 Ribbed &siliconised. 1.Wire diameter: 1.2(-0.0+0.03)mm 2. Bore diameter: 0.60(-0.01 + 0.025)mm 3. Suture material: PGA USP 2 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010430033	NEEDLE 48MM,3/8 CRC PGA2	1200
	Specification: Item name: Needle: NEEDLE-48MM, 3/8 CIRCLE REVERSE CUTTING, PGA-USP 2 Ribbed &siliconised. 1.Wire diameter: 1.2(-0.0+0.03)mm 2. Bore diameter: 0.60(-0.01 + 0.025)mm 3. Suture material: PGA USP 2 (Braided) 4. All needles must be siliconised&ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010430035	NEEDLE-45MM,1/2CRB,PGLA-2	60000
	Specification: Item name: Needle: NEEDLE-45MM, 1/2 CIRCLE ROUND BODY,-USP 2 Ribbed &siliconised.	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	1.Wire diameter: 1.20 (-0.0+0.03)mm 2. Bore diameter: 0.72 (-0.01+0.025)mm 3. Suture material: PGA-PLA, USP 2 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	
1010430037	NEEDLE 26MM-1/2CRC(PGLA)3/0	1200
	Specification: Item name:NEEDLE-26MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.74 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PGLA Braided 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
1010430041	NEEDLE 16MM-1/2 CRC(HIDOX)4/0	1200
	Specification: Item name:NEEDLE-16 MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.56 (-0.0 + 0.02)mm 2. Bore diameter: 0.28 (-0.01 + 0.02)mm 3. Suture material: PDO Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
1010430042	NEEDLE 12MM-1/2 CRC(RAPID)5/0	1200
	Specification: Item name:NEEDLE-12 MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.48 (-0.0 + 0.02)mm 2. Bore diameter: 0.28 (-0.01 + 0.02)mm 3. Suture material: PGLA RAPID Braided 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
1010430043	NEEDLE 13MM-1/2CRC(PP)5/0	1200
	Specification: Item name:NEEDLE-13 MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	1.Wire diameter: 0.48 (-0.0 + 0.02)mm 2. Bore diameter: 0.19 (-0.01 + 0.02)mm 3. Suture material: PP Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
1010430044	NEEDLE 31MM-1/2CRB(PP)2/0	1200
	Specification: Item name:NEEDLE-31MM, 1/2 CIRCLE ROUND BODY, Siliconised. 1.Wire diameter: 0.81 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PP Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
1010430045	NEEDLE 26MM-1/2CRB(PP)2/0	1200
	Specification: Item name:NEEDLE-26MM, 1/2 CIRCLE ROUND BODY, Siliconised. 1.Wire diameter: 0.74 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PP Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
1010430048	NEEDLE 30MM-1/2CRC(PGCL)3/0	1200
	Specification: Item name:NEEDLE-30MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.81 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PGCL Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
1010430049	NEEDLE 25MM-1/2CRC (PP)4/0	1200
	Specification: Item name:NEEDLE-25MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.60 (-0.0 + 0.02)mm 2. Bore diameter: 0.25 (-0.01 + 0.02)mm	

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity
		pcs
	3. Suture material: PP Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
1010430050	NEEDLE 25 MM-1/2CRB(PP)4/0	1200
	Specification: Item name:NEEDLE-25MM, 1/2 CIRCLE ROUND BODY, Siliconised. 1. Wire diameter: 0.60 (-0.0 + 0.02)mm 2. Bore diameter: 0.25 (-0.01 + 0.02)mm 3. Suture material: PP Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
1010430051	NEEDLE-3/8 CIRCLE RC 26MM PGLA2-0	1200
	Specification: Item name:NEEDLE-26MM, 3/8 CIRCLE REVERSE CUTTING, Siliconised. 1. Wire diameter: 0.71 (-0.0 + 0.02)mm 2. Bore diameter: 0.47 (-0.01 + 0.02)mm 3. Suture material: PGA-PLA,USP 2-0 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
1010430052	NEEDLE-3/8 CIRCLE RC 22MM PGLA3-0	1200
	Specification: Item name:NEEDLE-22M, 3/8 CIRCLE REVERSE CUTTING, Siliconised. 1. Wire diameter: 0.62 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PGA-PLA,USP 3-0 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420	
10104203499	NEEDLE-26MM, 1/2 CC, PGA-PLA USP 3/0	1200
	Specification: Item name:NEEDLE-26MM, 1/2 CIRCLE CUTTING, Siliconised. 1. Wire diameter: 0.56 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PGA-PLA,USP 3-0 (Braided) 4. All needles must be siliconised& ribbed.	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

<i>Material Number</i>	<i>Material Description</i>	<i>Quantity</i>
		<i>pcs</i>
	5. Material: High Quality martensitic stainless steel Grade AISI-420.	
	<p>General Specifications:</p> <p>All needles must be siliconised and ribbed. Resistance to corrosion: Corrosion resistant Cleanliness: Freefrom dirt,grease,fibersetc. Needle type: Drilled end suitable for above applications. Needlesfor monofilament sutures should have tapped hole.</p> <p>General Conditions:</p> <ol style="list-style-type: none"> 1.The surface of the suture needle shall be smooth and shall be free from pitting. 2.The needle shall be free from grinding marks, polishing dirt, grease, fibers or other materials, which could necessitate cleaning. 3.The needle should meet the respective national and international standards. 4.The needle should be able to carry suture material through tissues with minimal trauma. 5.Needle should be with silicon or similar lubricant to improve the needle penetration. 6.The needle should be sharpened enough to penetrate tissue with minimal resistance. 7.Needle should be rigid enough to resist bending, yet ductile enough to resist breaking during surgery. 8.The needle should be suitable for ETO/Gamma sterilization. 9.The needle offered should conform to the standard stipulated / quality specifications. 10. The supplier must also provide internal test certificates. 11.Packing: The needle shall be properly packed and labeled as per international standards. The packing shall be such that there won't be any damage to the needles during transit. Supply in standard & uniform packets. 12.The supplier should supply suitable dies for crimping the needles free of cost 	

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE B
Page 1 of 1

TECHNICAL BID
QUESTIONNAIRE TO BE FILLED FOR MINIMUM ELIGIBILITY CRITERIA
Material: Needle for Surgical Suture

1.	Are you a manufacturer?	YES/NO
2.	Are you an authorised agent? In case of authorised agent following documents from the Principal supplier should have to be enclosed Letter confirming the agency from the Manufacturer valid for 2016-18 Letter from the manufacturer that they also agree to abide by all the terms and conditions of this tender	YES/NO
3.	What is your annual production / process capacity? Out of the above how much can you supply to HLL per month?	
4.	Do you have testing facility to check HLL's parameters? If yes, will you be issuing test certificates for each Sample / batch /lot	YES/NO
5.	Do you have minimum one year experience in supplying this item?	YES/NO
6.	Are you an existing supplier to HLL? And if yes, Since when	YES/NO
7.	Are you an approved vendor of Akkulam Factory? (If not please fill the Schedule E and submit with at least 200 Nos. of samples)	YES/NO
8.	EMD furnished or not	YES/NO
9.	Tender Fee furnished or not	YES/NO
10.	Declaration as per Schedule G	YES/NO
11.	What is your Annual Turn over for the last three years ?	
12.	Do the facility ISO 9000 / ISO 13485 certification	YES / NO

All the information provided herein are true & correct

PLACE:
DATE:

NAME & SIGNATURE OF THE APPLICANT
(WITH OFFICE SEAL)

NOTE: Forms must be completed in all respect and supported by relevant documents.

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE -C
Page 1 of 2

TECHNICAL BID

QUESTIONNAIRE

(General information of the manufacturer / supplier)

1. Name & Address of the Supplier with :
(a) Telephone No.
(b) Fax No.
(c) E-mail Address
(d) Name of contact person
(e) Whether proprietary/partnership/
Limited company.
(f) Specify whether SSI / MSME unit / UDYOG AADHAR
2. How many years have you been in the :
business of manufacturing/selling?
3. What is your annual production capacity :
4. Indicate the quantity you can supply HLL :
per month
5. What would be the minimum period :
required to deliver the ordered quantity
from the date of confirmed purchase
order?
6. Have you been a supplier to any Blood bag :
Manufacturer, if so give details of the
name, address, quantity and values of
orders received and executed during the
last three year? (Attach separate sheet)
7. Kindly furnish the name and address of :
the Transporter through whom the rejected
material is to be returned on freight to pay
basis (Ref. clause d of Schedule D)

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE - C
Page 2 of 2

TECHNICAL BID

QUESTIONNAIRE

8. What is your Annual Turn Over during the :
last 3 years?
(Copy of P&L and Balance sheet to be attached.)
9. Details of tax registration :-
a) GSTIN. :
b) PAN :
10. Name & Address of your Banker(s) :
11. A/C no. & Swift Code :
12. Any other details :

All the information provided herein is true & correct.

PLACE:
DATE :

NAME & SIGNATURE OF THE APPLICANT
(WITH OFFICE SEAL)

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE - D
Page 1 of 5

TERMS & CONDITIONS

1. The Tender should be completed in all respects and submitted along **Tender Fees (non-refundable)** and **EMD**. Incomplete tenders are liable to be rejected.
 2. Unsealed Tenders received are liable to be rejected and this will be at the sole risk of the Tenderer.
 3. The EMD of successful bidders will be converted to Security Deposit and will be refunded at the end of contract period.
 4. (a) The quoted rate shall be valid for a period of one year during which period no price escalation and no change in terms and conditions will be allowed on any ground.
(b) In the event of failure on the part of the supplier to supply the material as per the delivery schedule or supply material at the agreed rate during the contract period, HLL would procure the material from alternate source at the risk and cost of the supplier. **In such cases, the additional cost incurred will be recovered from the supplier.**
(c) Failure to supply the material for any of the above reasons will also entail forfeiture of Security Deposit and even blacklisting of the supplier for a minimum period of two years.
- i. Conditional tenders / tenders with deviation will not be considered.
 - ii. The quantity mentioned herein is approximate annual requirement and in case the company requires less/more quantity, the supplier should be prepared to effect supply at short notice at the agreed rate, terms and conditions.
 - iii. HLL reserves the right to split up the quantity and place the order on more than a supplier and also reserves the right to accept or reject the offer without assigning any reason.

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE - D
Page 2 of 5

- iv. The tender is liable to be suspended or cancelled at anytime at the discretion of the company without assigning any reason.

In the event of placing orders: -

- a. Sample should be submitted and got approved before effecting bulk supply.
- b. The remaining period of shelf life of material on arrival at our stores should be minimum 5/6th of the total shelf life.
- c. Test Report of the product is to be attached with each consignment batch wise specifying therein the readings of the final test.
- d. In case of rejection the material should be taken back and replaced at supplier's risk and cost within 7 days of intimation from HLL. The procedure in this regard would be as follows:-

Material shall be returned through the Transporter indicated in Schedule C or through our approved Transporter on freight to-pay basis. In the case of overseas bidders material shall be returned through their specified freight forwarder.

Alternatively the rejected material shall be destroyed at our end.

The cost incurred for disposal of the rejected material shall be recovered from supplier.

The supplier is required to replace the rejected material failing which HLL reserves the right to procure the material from alternate source(s) at supplier's risk and cost.

- e. A certificate confirming that the product offered is environment friendly i.e., non-hazardous to the environment is to be enclosed with the Techno-Commercial Bid.
- f. The supplier agrees to supply strictly as per the order terms in respect of quantity and quality. HLL reserves the right not to receive material beyond the delivery date given in the order.

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE - D
Page 3 of 5

- g. Acceptance of the delayed supplies and excess quantity is solely at the discretion of HLL. In the case of excess supply the excess quantity shall be returned back through the Transporter indicated in Schedule C (clause 5) or our authorized transporter on freight to pay basis.
- h. The parties have to abide by delivery schedule given in the supply order strictly. Penalty @ 0.5% value of the materials per week of delay subject to a maximum of 7.5% of the value of the supply defaulted will be imposed if material is accepted by the company after the stipulated delivery period
- i. HLL reserves the right to inspect the facilities for GMP by their officials or by an authorized independent third party agency for compliance.
- j. All dispatch documents like delivery note/challan, packing list and invoice should contain the following details
- i. **Supply Order No.& date**
 - ii. **Description of items as contained in the supply Order.**
 - iii. **Quantity dispatched**
 - iv. **Manufacturing date.**
 - v. **Expiry date.**
 - vi. **Total Number of packages/ serially numbered**
 - vii. **Test certificate**
 - viii. **HLL's CST / TIN No.**

The L.R No / AWB /RR No..... should be intimated immediately after dispatch of the material by e-mail/fax

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE - D
Page 4 of 5

- k. The following information shall be stenciled or labelled on the exterior of the packing in bold letters, clearly visible, at least 50mm high with waterproof ink.
- a. Name of Product (as given in supply Order)
 - b. Date of Manufacture (Monthly/Yearly)
 - c. Date of expiry.
 - d. Quantity
 - e. Instruction off storage and handling
 - f. Name & Address of Manufacturer
 - g. Company's address in full
 - h. All packages should be separately numbered and it should appear on top of the packages with proper labelling.
- l. The following documents should accompany the consignment
- a. Invoice in triplicate
 - b. Delivery Note/Challan
 - c. Packing List
 - d. LR/AWB/----
 - e. Material Safety Data Sheet(MSDS)
 - f. Environment clearance certificate
- m. Mode of Transport shall be by road, unless otherwise specified. In case of overseas bidders mode of freight will be By Air unless otherwise specified. The materials shall be sent through the Transporter who has been granted a Certificate of Registration for transportation of goods as per Carriage by Road Act 2007.
- n. Payment shall be made by cheque/RTGS with in 30 days of receipt, inspection and acceptance of the material. In the case of overseas bidders the payment term will be LC at Sight / Wire Transfer within 30 days of receipt.
- o. The jurisdiction of any dispute, suits and proceedings arising out of this tender shall be only in the court of Thiruvananthapuram as the case may be.
- p. Annual purchase quantity will be shared to the suppliers @ ratio, as below: 60% to L1 party, and balance 40% shall be allotted among the vendors who matching the L1 rate.

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE - D
Page 5 of 5

INDEMINITY CLAUSE:

If the supplier fails to execute the order within the time prescribed for the delivery of goods ordered or violates or infringes the existing rates as agreed to as mentioned in the supply order, the supplier shall and will indemnify the company against all losses or damages whatsoever to be incurred or sustained including the legal cost or expenses incurred by the company by reason of non-delivery of goods at agreed quantity and rate with in the time specified in the supply order. The company will initiate legal action if the supplier fails to execute the supply order as per the schedule in the supply order for the actual loss suffered. No quantity tolerance will be permitted in this regard unless otherwise prior approval is taken by the company before dispatching any excess quantity supplied which shall be returned back on freight to pay basis at the risk of the supplier. Responsiveness of the Bid shall be at the discretion of HLL.

The supplier shall have no right to change the quantity stipulated in the supply order.

Bid pronounced Non Responsive by HLL shall be summarily rejected.

The decision of HLL will be final and no correspondence on this shall be entertained.

We have read and understood the above conditions and agree to abide by the same.

PLACE:
DATE:

NAME AND SIGNATURE OF THE APPLICANT
(WITH OFFICE SEAL)

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE - E
Page 1 of 1

QUESTIONNAIRE FOR VENDOR DEVELOPMENT
(TO BE FILLED BY NEW VENDORS)

NAME OF PRODUCT:

1) Name & Address of the Manufacturer :

Telephone No.

Fax No.

E-mail Address

Name of contact person

(i) Whether proprietary/partnership/
Limited company.

(ii) Furnish Name/'s, address,
phone nos. of Proprietor/
Managing partner/ M.D / Director's
(Attach separate sheet)

Specify whether SSI / MSME unit / UDYOG AADHAR

If Yes, pl specify if your unit is owned by :
SC/ST entrepreneur

2) Details of tax registration :- :

c) GSTIN.

d) PAN.

3) Name & Address of your Banker(s) :

Account no.

Swift Code

4) Annual Turn over

5) Do you have a Quality Control department. If yes, give the
details of the facilities. : Yes/No

6) a) Do you have inspection for incoming materials : Yes/No

b) Do you have in process inspection facilities : Yes/No

c) Do you have final inspection of the product : Yes/No

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE - E
Page 2 of 2

- 7) Do you have any accreditation : Yes/No
(a)ISO 9000 / ISO 13485
(b)Any National/International Laboratory/Authority
Certification
If 'Yes', give details along with a copy of certificate issued
by them.
- 8) Do you have an environmental policy / ISO 14001 certification? : Yes / No
If yes, please give the details.
- 9) Furnish the details of testing equipments.
- 10) Have you been assessed previously by HLL : Yes/No
- 11) List of customers
a) Government Dept. / PSU's
b) Private Sector
- 12) Have you furnished the sample along with this format : Yes/No

Place :

NAME AND SIGNATURE OF THE VENDOR

Date :

(Office Seal)

This is to be filled up by HLL LIFECARE LIMITED, THIRUVANANTHAPURAM

Recommendation/Remarks of the committee

Committee Members:

1.Head of Quality Control/Head of Technical Services and Material Testing

2.Head of User Department

3.Head of Purchase Department

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE - F
Page 1 of 1

TECHNICAL BID

Item: Needle for Surgical Suture

DECLARATION

We confirm having read and understood all the specifications, instructions, forms, terms and conditions and other requirements of the above tender (both expressed and implied) in full and that we agree to abide by all without any deviation.

SIGNATURE SEAL OF THE APPLICANT

NAME AND ADDRESS OF APPLICANT

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

CONTENTS OF PRICE BID FORMS

For

Needle for Surgical Suture

1. SCHEDULE I - RATE SCHEDULE FOR THIRUVANANTHAPURAM (For Domestic suppliers)
2. SCHEDULE II - RATE SCHEDULE FOR CIF THIRUVANANTHAPURAM (For Overseas Bidders)

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

SCHEDULE - I
PRICE BID - RATE SCHEDULE FOR SUPPLY TO AKKULAM FACTORY, THIRUVANANTHAPURAM,
KERALA (For Domestic Bidders)

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
1010420000	NEEDLE-36MM, 1/2 CRB (C.T), PGA 1 Specifications: Taper ratio: 1:12 (Increased taper length) USP Size: PGA 1 (Undyed) Wire diameter: 0.92+/-0.01mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420001	NEEDLE-16MM 1/2 CRB, HEAVY, PGA 2/0 USP Size: PGA 2/0 Bore diameter: 0.42+/-0.01mm Wire diameter: 0.66+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	2400		
1010420002	NEEDLE-20MM 1/2 CRB, HEAVY, PGA 2/0 USP Size: PGA 2/0 Bore diameter: 0.42+/-0.01mm Wire diameter: 0.78+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420003	NEEDLE-26MM 1/2 CRB, PGA 2/0 USP Size: PGA 2/0 Bore diameter: 0.42+/-0.01mm Wire diameter: 0.70+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Ribbed&siliconised			
1010420004	NEEDLE -36MM, 1/2 CRB (C.T), PGA 2/0 Specifications: USP Size: PGA 2/0 (Undyed) Taper point: 1:12 (Increased taper length) Wire diameter: 0.81+/-0.01mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420 Ribbed&siliconised	1200		
1010420005	NEEDLE-30MM, 1/2 CRB (C.T), PGA 2/0 Specifications: USP Size: PGA 2/0 (Undyed) Taper ratio: 1:12 (Increased taper length) Wire diameter: 0.81+/-0.01mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420006	NEEDLE-30MM, 1/2 CTC, USP 2/0 Specification: USP size: PGA 2/0 (Braided) Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	2400		
1010420007	NEEDLE-3/8 CIRCLE CUTTING 19 MM, PGA 3-0 1. 3/8 circle cutting, 19mm, Needle for PGA suture undyed braided & coated, Usp Size:3-0. 2. Bore diameter: 0.31 to 0.35 3. Wire diameter: 0.58 to 0.62mm 4. Material: High quality martensitic stainless steel grade AISI-420. Ribbed&siliconised	1200		
1010420008	NEEDLE-16MM 1/2 CRB, HEAVY, PGA 3/0	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	USP Size: PGA 3/0 Bore diameter: 0.33+/-0.01mm Wire diameter: 0.61+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420009	NEEDLE-30MM 1/2 CRB, HEAVY, PGA 3/0	1200		
	USP Size: PGA 3/0 Bore diameter: 0.33+/-0.01mm Wire diameter: 0.71+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420010	NEEDLE-40MM 1/2 CRB, HEAVY PGA 3/0	1200		
	USP Size: PGA 3/0 Bore diameter: 0.33+/-0.01mm Wire diameter: 0.78+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420011	NEEDLE-12MM 5/8 CRC, PGA 4/0	1200		
	USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.51+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420012	NEEDLE-13MM 3/8-CRC, PGA 4/0	1200		
	USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.51+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420013	NEEDLE-13MM 5/8 CRC, PGA 4/0	1200		
	USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.51+/-0.02mm			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420014	NEEDLE-16MM 3/8 CRC, PGA 4/0 USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.54+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420015	NEEDLE-16MM 5/8 CRC, PGA 4/0 USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.54+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420016	NEEDLE-30MM 1/2 CRB, HEAVY PGA 4/0 USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.61+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420017	NEEDLE-40MM 1/2 CRB, HEAVY PGA 4/0 USP Size: PGA 4/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.70+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420018	NEEDLE-12MM 5/8 CRC, PGA 5/0 USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.46+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Ribbed&siliconised			
1010420019	NEEDLE-13MM 3/8 CRC, PGA 5/0 USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.46+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420020	NEEDLE-13MM 5/8 CRC, PGA 5/0 USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.46+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420021	NEEDLE-16MM, 3/8 CRC, PGA 5/0 NEDLE 16MM, 3/8 CIRCLE REVERSE CUTTING FOR PGA 5/0 SUTURES USP Size: PGA 5/0 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.46+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420022	NEEDLE-16MM 5/8 CRC, PGA 5/0 USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.46+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420023	NEEDLE-20MM 1/2 CRB, HEAVY, PGA 5/0 USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.54+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Ribbed&siliconised			
1010420024	NEEDLE-30MM 1/2 CRB, HEAVY, PGA 5/0 USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.54+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420025	NEEDLE-40MM 1/2 CRB, HEAVY, PGA 5/0 USP Size: PGA 5/0 Bore diameter: 0.22+/-0.01mm Wire diameter: 0.72+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420 Ribbed&siliconised	1200		
1010420026	NEEDLE-16MM 1/2 CRB, HEAVY, PGA 0 USP Size: PGA 0 Bore diameter: 0.47+/-0.01mm Wire diameter: 0.70+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420 Ribbed&siliconised	1200		
1010420027	NEEDLE-20MM, 1/2 CRB, HEAVY PGA 0 USP Size: PGA 0 Bore diameter: 0.47+/-0.01mm Wire diameter: 0.84+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420028	NEEDLE-40MM 1/2 CRC, PGA 0 USP Size: PGA 0 Bore diameter: 0.47+/-0.01mm Wire diameter: 1+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	6000		
1010420029	NEEDLE-36MM 1/2 CRC, PGA 1	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	USP Size: PGA 1 Bore diameter: 0.55+/-0.01mm Wire diameter: 1+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420030	NEEDLE-36MM, 1/2 CRB (C.T), PGA 0	1200		
	USP Size: PGA 0 (Undyed) Taper ratio: 1:12 (Increased taper length) Wire diameter: 0.84+/-0.01mm Bore diameter: 0.47+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420031	BLUNT NEEDLE-40MM, 1/2 CRB, PGA 1	1200		
	Blunt needle , 40mm, 1/2 circle round body, USP Size: PGA 1 (braided) Wire diameter: 0.92+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420032	BLUNT NEEDLE-40MM, 1/2 CRB, PGA 0.	1200		
	SPECIFICATION: Blunt needle 40mm, 1/2 circle round body. USP size: PGA 0 (braided) Wire diameter: 0.84+/-0.02mm Bore diameter: 0.47+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420033	NEEDLE-30 MM, 1/2 CRB, PGA 0	1200		
	SUTURE SIZE: 0, 30mm,1/2circle,round body Bore diameter: 0.47+/-0.01mm. Wire diameter: 0.84+/-0.02mm. Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420034	NEEDLE-40 MM 1/2 CRB, PGA 0	12000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	SUTURE SIZE: 0 40mm, 1/2circle, round body Wire diameter: 1+/-0.02mm Bore diameter: 0.47+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420035	NEEDLE-40 MM, 1/2 CTC, PGA 0	3000		
	SUTURE SIZE: 1/0 40mm, 1/2circle, taper cut Wire diameter: 1+/-0.02mm Bore diameter: 0.47+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420036	NEEDLE-40 MM 1/2 CIRCLE CUTTIN	1200		
	needle for PGA suture, 40mm, 1/2 circle, cutting, suture size-1/0 (USP) Wire diameter: 0.92(-0.0+0.025)mm Bore diameter: 0.47 (-0.01+0.02)mm Material: high quality martensitic steel Grade: AISI-420 Ribbed&siliconised			
1010420037	NEEDLE-36MM, 1/2 CIRCLE, CUTTI	1200		
	suture size: 1/0 Wire diameter: 0.84(-0.0+0.025)mm Bore diameter: 0.47 (-0.01+0.02)mm Material: High quality Martensitic steel, Grade; AISI-420 Ribbed&siliconised			
1010420038	NEEDLE-36MM, 1/2 CRC, PGA	1200		
	Wire diameter: 0.84(-0.0+0.025)mm Bore diameter: 0.47 (-0.01+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420. Ribbed&siliconised			
1010420039	NEEDLE-36MM, 1/2 CRC, OS needle, PGA 0	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	1. Bore diameter (mm): 0.47+/-0.02mm 2. Wire diameter (mm): 1.20+/-0.05mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420040	NEEDLE-40MM, 1/2 CRB (C.T), PGA 1 Specifications: Taper ratio: 1:12 (Increased taper length) USP Size: PGA 1 Wire diameter: 0.92 +/- 0.01mm Bore diameter: 0.55 +/- 0.01mm Material:High Quality martensitic stainless steel Grade AISI-420 Ribbed&siliconised	1200		
1010420041	NEEDLE-40MM, 1/2CIRCLE CUTTING, USP Wire diameter: 0.99(-0.0+0.025)mm Bore diameter: 0.55(-0.01+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420. Ribbed&siliconised	1200		
1010420042	NEEDLE-30 MM 1/2 CRB, USP 2-0 SUTURE SIZE:2/0 30mm,1/2circle,round body Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	24000		
1010420043	NEEDLE-40 MM 1/2 CRB, PGA 2-0 SUTURE SIZE:2/0 40mm,1/2circle round body. Wire diameter: 0.84+/-0.02mm Bore diameter: 0.42+/-0.02mm.\nMaterial:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	24000		
1010420044	NEEDLE-40 MM 1/2 CRC, PGA 2-0 SUTURE SIZE:2/0 40mm,1/2circle,reverse cutting Wire diameter: 0.84+/-0.01mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	24000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420045	NEEDLE-26MM, 3/8 CIRCLE,CUTTING, USP 2-0	1200		
	Specification: Item name: Needle: NEEDLE-26MM, 3/8 CIRCLE CUTTING, USP 2-0-Ribbed & siliconised. 1.Wire diameter: 0.61+/-0.01mm 2. Bore diameter: 0.42+/-0.01mm 3. Suture material: Polyglycolic Acid, USP 2-0 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420046	NEEDLE-26MM 1/2 CRB, USP 2/0	1200		
	Wire diameter: 0.81(-0.0+0.02)mm Bore diameter: 0.42(-0.01+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420. Ribbed&siliconised			
1010420047	NEEDLE-36MM, 1/2 CRB, USP 2-0	1200		
	Specifications: USP size: PGA 2-0 Wire diameter: 0.84+/-0.02mm Bore diameter: 0.42+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420048	NEEDLE-30MM, 1/2 CRC, USP 2/0	1200		
	Specification: USP size: PGA 2/0 (braided) Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420049	NEEDLE-27MM, 5/8 CTC, USP 2-0	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	SPECIFICATION: Needle: 27mm, 5/8 circle taper cutting. Suture size: PGA 2-0. Wire diameter: 0.71+/-0.02mm Bore diameter: 0.42+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420050	NEEDLE-20 MM, 1/2 CRB, USP 3-0	1200		
	SUTURE SIZE:3/0 20mm,1/2circle,round body Wire diameter: 0.62+/-0.01mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420051	NEEDLE-22 MM, 3/8 CIRCLE CUTTING,USP 3-0	2400		
	SUTURE SIZE:3/0 22mm,3/8circle,cutting Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420052	NEEDLE-12 MM 3/8 CRC, USP 3-0	3600		
	SUTURE SIZE:3/0 12mm,3/8circle,reverse cutting Wire diameter: 0.54+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420053	NEEDLE-22MM,1/2 CIRCLE CUTTING, USP 3-0	1200		
	suture size:3/0(USP) Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420054	NEEDLE-17MM,1/2 CTC, USP 3-0	1200		
	SUTURE SIZE:3/0 Wire diameter: 0.60+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Ribbed&siliconised			
1010420055	NEEDLE-36MM, 1/2 CTC, USP 3-0 SUTURE SIZE:3/0 Wire diameter: 0.71+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420056	NEEDLE-16MM, 3/8 CIRCLE CUTTING, USP 3-0 Suture size:3/0 Wire diameter: 0.54+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High quality Martensitic grade,stainless steel Grade:AISI-420 Ribbed&siliconised	3600		
1010420057	NEEDLE-26MM,1/2 CRB, USP 3/0 Wire diameter: 0.71(-0.01+0.02)mm Bore diameter: 0.33(-0.0+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420. Ribbed&siliconised	1200		
1010420058	NEEDLE-36MM, 1/2 CRB, USP 3-0 Specifications: USP size: PGA 3-0 Wire diameter: 0.71+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420059	NEEDLE-12 MM, 3/8 CRC, USP 4-0 SUTURE SIZE:4/0 12mm,3/8circle,reverse cutting Wire diameter: 0.54+/-0.02mm Bore diameter: 0.28+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised	1200		
1010420060	NEEDLE-20 MM 1/2 CRB, USP 4-0	2400		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	"Specification: Needle: 20mm, 1/2 circle round body Wire diameter: 0.54 +/- 0.01mm Bore diameter: 0.28 +/- 0.01mm Suture size: polyglycolic acid USP 4/0 Material: High Quality martensitic stainless steel Grade AISI-420. Ribbed&siliconised			
1010420061	NEEDLE-16MM 1/2 CRB, USP 4/0	1200		
	Specification: Suture USP Size: 4/0 Wire diameter: 0.54+/-0.02mm Bore diameter: 0.28+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420062	NEEDLE-16MM, 5/8 CIRCLE CUTTING, USP 4-0	1200		
	suture size:4/0 Material:high quality Martensitic grade stainless steel Grade;AISI-420 Bore dia 0.28 +/- -0.02mm Wire dia : 0.48 to 0.51mm			
1010420063	NEEDLE-26MM, 3/8 CIRCLE,CUTTING, USP	1200		
	Wire diameter: 0.56 +/- 0.01mm Bore diameter: 0.28 +/- 0.01mm suture size-4/0 material:high quality martensitic stainless steel Grade:AISI-420			
1010420064	NEEDLE-12 MM 3/8 CRC, USP 5-0	1200		
	SUTURE SIZE:5/0 12mm,3/8circle,reverse cutting Wire diameter: 0.48+/-0.02mm Bore diameter: 0.22+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420065	NEEDLE-16MM,1/2 CRB, USP 5-0	2400		
	suture size:5/0(USP) Wire diameter: 0.48+/-0.02mm Bore diameter: 0.22+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420066	NEEDLE-16MM, 3/8 CIRCLE,CUTTING, USP 5/0	1200		
	Wire diameter: 0.48+/-0.02mm			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Bore diameter: 0.22+/-0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420.			
1010420067	NEEDLE-30mm, 1/2 CRB, USP 0 Specification: USP Size: PGA 0 (Braided) Bore diameter: 0.47+/-0.01mm. Wire diameter: 0.84+/-0.02mm. Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420068	NEEDLE-45mm, 3/8 CRB, USP 1 Specification: USP Size: PGA 1 (Braided) Wire diameter: 1.00+/-0.02mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420069	NEEDLE-30mm, 1/2 CRB, USP 2/0 Specification: USP Size: PGA 2-0 (Braided) Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420	1200		
1010420070	NEEDLE-40mm, 1/2 CRB USP 2/0 Specification: USP Size: PGA 2/0 (Braided) Wire diameter: 0.84+/-0.02mm Bore diameter: 0.42+/-0.02mm. Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420071	NEEDLE-40 MM, 1/2 CRB, USP 1 NEEDLE FOR PGA SUTURE(BRAIDED) 40 MM, 1/2 CIRCLE ROUND BODY SUTURE SIZE:1 (40mm,1/2circle,round body(heavy)) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	48000		
1010420072	NEEDLE- 40 MM, 1/2 CTC, USP 1	24000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	NEEDLE FOR PGA SUTURE(BRAIDED) 40 MM, 1/2 CIRCLE TAPER CUT SUTURE SIZE:1 (40mm,1/2circle,taper cut(heavy)) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420073	NEEDLE-45 MM, 3/8 CRB, USP 1 SUTURE SIZE:1 45mm,3/8circle,round body Wire diameter: 1+/-0.02mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	3000		
1010420074	NEEDLE-50 MM 1/2 CRB, USP 1 Specification: Item name: Needle: 50mm, ½ circle round body, heavy. -Ribbed &siliconised. 1.Wire diameter: 1.2+/-0.01mm 2. Bore diameter: 0.55+/-0.01mm 3. Suture material: Polyglycolic Acid, USP 1 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	3000		
1010420075	NEEDLE-50MM, 1/2 CRC,USP1 Specification: suture size:1(usp) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420076	NEEDLE-40MM,1/2 CIRCLE,CUTTING,USP 1 suture size-1, Wire diameter: 0.99(-0.0+0.025)mm Bore diameter: 0.55(-0.01+0.02)mm Material:high quality Martensitic stainless steel,Grade-AISI-420	1200		
1010420077	NEEDLE-40MM,1/2 CRC, USP 1 1. Bore diameter (mm): 0.55+/- 0.02MM 2. Wire diameter (mm):1.15+/-0.05mm Material:High Quality martensitic stainless steel Grade AISI-420.	24000		
1010420078	NEEDLE-36mm, 1/2 CRB, USP 1	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Specifications: USP size: PGA 1 Wire diameter: 1.15+/-0.02mm Bore diameter: 0.55+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420			
1010420079	NEEDLE-36mm, 1/2 CRB, USP 0	1200		
	Specifications: USP size: PGA 0 Wire diameter: 1.00+/-0.02mm Bore diameter: 0.47+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420080	NEEDLE- 40MM 1/2 CIRCLE CUTTING, USP 0	1200		
	Wire diameter: 0.92(-0.0+0.025)mm Bore diameter: 0.47 (-0.01+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420.			
1010420081	NEEDLE-40MM, 1/2 CRB, HEAVY, USP 1	1200		
	Specification: Needle: 40mm, 1/2 circle round body, heavy. Suturesize: PGA Braided ,USP 1. Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality stainless steelGrade 300.			
1010420082	NEEDLE-40MM, 1/2 CTC, HEAVY, USP 1	1200		
	Specification: Needle: 40mm, 1/2 circle taper cutting, heavy. Suture size: PGA (Braided), USP 1 Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality stainless steel Grade 300.			
1010420083	NEEDLE-45MM, 3/8 CIRCLE CUTTING, USP 3/0	1200		
	Wire diameter: 0.71+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High quality martensitic stainless steel grade AISI-420.			
1010420084	NEEDLE-16MM, 3/8 CRB, SILK 3/0 SUTURES	1200		
	Suture Size: 3/0 Wire diameter: 0.54+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420085	NEEDLE-26MM, 1/2 CRC, SILK 3/0 Suture size: 3/0 Wire diameter: 0.71+/-0.01mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420087	NEEDLE12MM, 3/8 CRB, SILK 5/0 Suture size: 5/0 Wire diameter: 0.48+/-0.02mm Bore diameter: 0.23+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420088	NEEDLE-30 MM 1/2 CRB, SILK 0 SUTURE SIZE: 0 30mm,1/2circle,round body Wire diameter: 0.92+/-0.01mm Bore diameter: 0.48+/-0.01mm. Material:High Quality martensitic stainless steel Grade AISI-420.	6000		
1010420089	NEEDLE-16MM,1/2 CRB, USP 4/0 Wire diameter: 0.56+/-0.02mm Bore diameter: 0.28+/-0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	1200		
1010420090	NEEDLE-45MM, 3/8 CRC, USPO Specification: Suture size: Silk, USP 0 Wire diameter: 1+/-0.01mm Bore diameter: 0.47+/-0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	36000		
1010420091	NEEDLE-30 MM, 1/2 CRB SUTURE SIZE:2/0 30mm,1/2circle,round body Wire diameter: 0.80+/-0.01mm Bore diameter: 0.42+/-0.01mm. Material:High Quality martensitic stainless steel Grade AISI-420.	75000		
1010420092	NEEDLE-45 MM 3/8 CRC, SILK 2-0	75000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Specifications: Wire diameter: 1mm+/-0.02mm Bore diameter: 0.42+/-0.01mm SILK SUTURE: 2/0 45mm, 3/8circle,reverse cutting Material: High Quality Martensitic stainless steel Grade AISI-420			
1010420093	NEEDLE -25 MM 1/2 CTC, SILK 2-0 SILK SUTURE: 2/0 25mm, 1/2circle,taper cutting 1. Bore diameter (mm): 0.40+/-0.02mm 2. Wire diameter (mm): 0.76+/-0.05mm. Material: High Quality Martens tic stainless steel Grade AISI-420.	3000		
1010420094	NEEDLE-50MM,1/2 CRC SILK 2-0 Bore diameter 0.40 +/- 0.02 mm Wire diameter 1.10 +/- 0.05 mm suture size:2/0 Material :High Quality martensitic stainless steel Grade AISI-420.	3000		
1010420095	NEEDLE-1/2 CTC,17MM,SILK 2-0 Item name: NEEDLE-17 mm, 1/2 circle taper cutting. Silk USP 2/0 Siliconised and Ribbed. 1. Wire diameter: 0.61 (-0.0+0.025)mm 2. Bore diameter: 0.42 (-0.01+0.02)mm 3. Suture material: Silk, USP 2/0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AISI-420	3000		
1010420096	NEEDLE-STR,RB,40MM,SILK2/0 Item name: NEEDLE-Straight, 40 MM. Silk USP 2/0 Siliconised 1. Wire diameter: 0.92 (-0.0+0.025)mm 2. Bore diameter: 0.42 (-0.01+0.02)mm	6000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	3. Suture material: Silk, USP 2/0 (Braided) 4. Material: High Quality martensitic stainless steel grade AISI-420			
1010420097	NEEDLE-36MM, 3/8CTC,SILK-2-0 Specifications: USP size: Silk 2-0 Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420098	NEEDLE-25 MM 1/2 CRB SILK Specifications: SUTURE SIZE:3/0 Wire diameter: 0.70+/-0.01mm Bore diameter: 0.33+/-0.01mm 25mm,1/2circle,round body Material :High Quality martensitic stainless steel Grade AISI-420.	2400		
1010420099	NEEDLE-20 MM,1/2, CTR,SILK 3/0 SUTURE SIZE:3/0 20mm,1/2circle,round body Wire diameter: 0.61 +/- 0.02mm Bore diameter: 0.33 +/- 0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	24000		
1010420100	NEEDLE-26 MM 3/8 CRC SILK 3/0 SUTURE SIZE:3/0 26mm,3/8circle,reverse cutting Material :High Quality martensitic stainless steel Grade AISI-420. Resistance to corrosion :Corrosion resistant .Bore dia : 0.33 +/- 0.01 mm Wire dia : 0.68 to 0.70mm	60000		
1010420101	NEEDLE-16 MM 3/8 CIRCLE CUTTING Specifications: Wire diameter: 0.54+/-0.01mm Bore diameter: 0.33+/-0.01mm SUTURE SIZE:3/0 16mm,3/8circle,cutting Material:High Quality martensitic stainless steel Grade AISI-420.	60000		
1010420102	NEEDLE -1/2 CRB SILK 3/0 suture size; Silk 3/0, 35mm 1/2circle ,round body Braided Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	3000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
1010420103	NEEDLE-36MM, 3/8 CTC SILK 3-0 Specifications: USP size: Silk 3-0 Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material :High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420104	NEEDLES -16 MM 3/8 CRB SILK4/0 Wire diameter: 0.56 (-0.0+0.02)mm Bore diameter: 0.28 (-0.01+0.02) mm. Material:High Quality martensitic stainless steel Grade AISI-420	2400		
1010420105	NEEDLE - 16 MM 3/8 CC SILK 4/0 SUTURE SIZE:4/0 16mm,3/8circle,cutting Wire diameter: 0.56+/-0.02mm Bore diameter: 0.28+/-0.01mm. Material:High Quality martensitic stainless steel Grade AISI-420.	30000		
1010420106	NEEDLE -1/2 CTC17MM SILK 4/0 Wire diameter: 0.56 (-0.0+0.02)mm. Bore diameter: 0.28 (-0.01+0.02) mm. Material:High Quality martensitic stainless steel Grade AISI-420	2400		
1010420107	NEEDLE -1/2 CRB 16MM SILK 4/0 Specifications: suture size 4/0 Wire diameter:0.54 +/-0.01mm Bore diameter: 0.28+/-0.01mm 16mm, 1/2 C, Round body Material :High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420108	NEEDLE-5/0 12 MM 3/8 CRB SILK 5/0 SILK SUTURE: 5/0 12mm, 3/8circle,round body Wire diameter: 0.46+/-0.02mm Bore diameter: 0.23+/-0.01mm Material: High Quality Martens tic stainless steel Grade AISI-420.	2400		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Specify HSN Code:			
1010420109	NEEDLE-5/0, 12 MM 3/8 CRC 5/0	12000		
	Wire diameter: 0.48 (-0.0+0.02)mm Bore diameter: 0.22 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420110	NEEDLE-3/8 CRB 16MM SILK 5/0	6000		
	Wire diameter: 0.48 (-0.0+0.02)mm Bore diameter: 0.22 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420111	NEEDLE-1/2 CTC,17MM SILK 5/0	1200		
	Wire diameter: 0.48 (-0.0+0.02)mm Bore diameter: 0.22 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420112	NEEDLE Straight RB 20MM SILK 5/0	1200		
	Wire diameter: 0.48 (-0.0+0.02)mm Bore diameter: 0.22 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420113	NEEDLE-3/8 CRB 20MM SILK 5/0	1200		
	Wire diameter: 0.48 (-0.0+0.02)mm Bore diameter: 0.22 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420114	needle-SILK 6/0,3/8 CRB 13mm	1200		
	Specification: Wire diameter: 0.38+/-0.01mm Bore diameter: 0.18+/-0.01mm Material :High Quality martensitic stainless steel Grade AISI-420			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
1010420115	NEEDLE - 60 MM 3/8 CC SILK 1 SUTURE SIZE:1 60mm,3/8circle,cutting 1. Bore Diameter (mm): 0.55+/-0.01mm 2. Wire diameter (mm): 1.20+/-0.05mm. Material :High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420116	NEEDLE-36MM, 3/8 CTC SILK-1 USP size: Silk 1 Wire diameter: 1+/-0.02mm Bore diameter: 0.55+/-0.01mm Material :High Quality martensitic stainless steel Grade AISI-420	1200		
1010420117	Needle-3/8 CRC, 12mm PP: 0 1. Wire diameter: 0.78+/-0.02mm 2. Bore diameter: 0.47+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420118	NEEDLE - 25 MM DOUBLE 1/2 CRB PP 3/0 Wire diameter: 0.61 (-0.0+0.02)mm Bore diameter: 0.28 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420119	NEEDLE 25MM, 3/8 CC PP 3/0 USP Size: Polypropylene 3/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.61+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420120	NEEDLE-17 MM 1/2CTC PP4/0 Wire diameter: 0.56(-0.0+0.02)mm Bore diameter: 0.25(-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420	1200		
1010420121	NEEDLE FOR PP SUTURE	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Specification: SUTURE SIZE:5/0 16mm,3/8circle,round body Wire diameter: 0.48+/-0.02mm Bore diameter: 0.19+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420122	NEEDLE-12 MM 3/8 CRB PP 5/0 Specifications: SUTURE SIZE:5/0 Wire diameter:0.48 +/-0.01mm Bore diameter: 0.19+/-0.01mm 12mm 3/8 C, Round body Material:High Quality martensitic stainless steel Grade AISI-420.	6000		
1010420123	NEEDLE-16MM, 3/8 CRC PP 5-0 SUTURE SIZE: PP 5-0 Wire diameter: 0.48+/-0.02mm Bore diameter: 0.19+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	6000		
1010420124	NEEDLE-3/8 CRC,12MM PP 6/0 Suture size-6/0 Bore diameter: 0.15+/-0.01mm Wire diameter: 0.35+/-0.01mm high quality martensitic grade stainless steel-Grade-AISI-420	1200		
1010420125	NEEDLE-PP 6/0,3/8 CRB,13MM Suture size-6/0 (monofilament) Wire diameter: 0.36+/-0.02mm Bore diameter: 0.15+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420126	NEEDLE-PP 6/0,3/8 CC,15MM PP Specification: Suture size-6/0 Wire diameter: 0.36+/-0.02mm Bore diameter: 0.15+/-0.01mm Material :High Quality martensitic stainless steel Grade AISI-420	2400		
1010420127	NEEDLE - 30 MM 1/2 CRB PP:0 SUTURE SIZE: 0 30mm,1/2circle,round body Wire diameter: 0.92+/-0.02mm Bore diameter: 0.45+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	24000		
1010420128	NEEDLE-25 MM 1/2 CTC PP:0	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	SUTURE SIZE:0 25mm,1/2circle,taper cut Wire diameter: 0.80+/-0.02mm Bore diameter: 0.45+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420130	NEEDLE-1/2 CRB 26MM PP-1/0	1200		
	Specifications: suture size-1/0 Wire diameter: 0.78 +/-0.01mm Bore diameter: 0.45+/-0.01mm 26mm 1/2 C, Round body Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420130	NEEDLE-1/2 CRB 26MM PP-1/0	1200		
	Wire diameter: 0.81 (-0.0+0.02)mm Bore diameter: 0.45 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420			
1010420131	NEEDLE-1/2CTC,30MM(HEAVY),PP1/0	2400		
	Specifications: suture size-1/0 Wire diameter: 1.0 +/-0.01mm Bore diameter: 0.45+/-0.01mm 30mm 1/2 C, Taper cutting, Heavy Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420132	NEEDLE 30MM, 1/2 CC PP USP 0	1200		
	USP Size: Polypropylene: 0 Bore diameter: 0.45+/-0.01mm Wire diameter: 0.92+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420133	NEEDLE 40MM, 1/2 CRB PP: 0	1200		
	USP Size: Polypropylene 0 Bore diameter: 0.45+/-0.01mm Wire diameter: 0.92+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420134	NEEDLE 40MM, 1/2 CC PP: 0	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	USP Size: Polypropylene 0 Bore diameter: 0.45+/-0.01mm Wire diameter: 0.92+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420135	NEEDLE 45MM 1/2 CRB PP 0	1200		
	USP Size: Polypropylene 0 Bore diameter: 0.45+/-0.01mm Wire diameter: 1+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420136	NEEDLE 45MM, 1/2 CC PP : 0	1200		
	USP Size: Polypropylene 0 Bore diameter: 0.45+/-0.01mm Wire diameter: 0.92+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420			
1010420137	NEEDLE - 30 MM 1/2 CRB PP2/0	18000		
	"Specification: Needle: 30mm, 1/2 circle round body Wire diameter: 0.80 +/- 0.01mm Bore diameter: 0.36 +/- 0.01mm Suture size: polypropylene USP 2/0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420138	NEEDLE-25 MM 1/2 CTC PP 2/0	3000		
	Needle: 25mm, 1/2 circle taper cutting Wire diameter: 0.78 +/- 0.01mm Bore diameter: 0.36 +/- 0.01mm Suture size: polypropylene USP 2/0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420139	NEEDLE-2/0, 25 MM, 1/2 CTC PP2/0	1200		
	Wire diameter: 0.81 (-0.0+0.02)mm Bore diameter: 0.36 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420			
1010420140	NEEDLE-1/2 CTC17MM PP-2/0	2400		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Wire diameter: 0.74 (-0.0+0.02)mm Bore diameter: 0.28 (-0.01+0.02)mm Material: High Quality martensitic stainless steel Grade AISI-420			
1010420141	NEEDLE-PP 3/0 25 MM 1/2 CRB	2400		
	NEEDLE-25MM, 1/2 CIRCLE ROUND BODY, POLYPROPYLENE- USP 3/0, Ribbed & Siliconised. 1. Wire diameter: 0.61 (-0.0+0.02)mm 2. Bore diameter: 0.28 (-0.01+0.02)mm 3. Suture material: Polypropylene, USP 3/0 (Monofilament) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420143	NEEDLE-20 MM 1/2 CRB PP 3/0	1200		
	Wire diameter: 0.58 (-0.0+0.02)mm Bore diameter: 0.28 (-0.01+0.02)mm Material: High Quality martensitic stainless steel Grade AISI-420			
1010420144	NEEDLE-3/0, 22 MM, 3/8 CC PP	3000		
	SUTURE SIZE: Polypropylene USP3/0, 22mm, 3/8circle, cutting Bore diameter: 0.27 to 0.29mm Wire diameter: 0.57 to 0.59mm Material: High Quality Martensitic stainless steel Grade AISI-420.			
1010420145	NEEDLE 1/2 CTC 17MM, PP 3/0	1200		
	Specifications: USP Size: 3/0 Wire diameter: 0.60+/-0.01mm Bore diameter: 0.28+/-0.01mm material; high quality martensitic stainless steel Grade; AISI-420			
1010420146	NEEDLE-PP4/0 16 MM 3/8 CC	1200		
	Wire diameter: 0.56 (-0.0+0.02)mm Bore diameter: 0.25 (-0.01+0.02)mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420147	NEEDLE-PP 4/0 16 MM 1/2 CRB	6000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	SUTURE SIZE:4/0 16mm,1/2circle,round body Wire diameter: 0.54+/-0.02mm Bore diameter: 0.25+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420148	NEEDLE -3/8 CRC,12MM,PP 4/0	1200		
	Specifications: suture size-4/0 Wire diameter: 0.54 +/-0.01mm Bore diameter: 0.25+/-0.01mm 12mm, 3/8 C, Reverse cutting Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420149	NEEDLE -1/2 CTC,25MM,PP 4/0	1200		
	Wire diameter: 0.56 (-0.0+0.02)mm Bore diameter: 0.25 (-0.01+0.02)mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420150	NEEDLE-16 MM 3/8 CRB	3000		
	Specifications: SUTURE SIZE:5/0 16mm,3/8circle,round body Wire diameter: 0.48+/-0.02mm Bore diameter: 0.19+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420151	NEEDLE-PP 5/0 12 MM 3/8 CRC	1200		
	SUTURE SIZE:5/0 12mm,3/8circle,reverse cutting Wire diameter: 0.50+/-0.01mm Bore diameter: 0.19+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420152	NEEDLE-3/8 CC,16MM PP 5/0	1200		
	Wire diameter: 0.48(-0.0+0.02)mm Bore diameter: 0.19+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420153	NEEDLE -PP:1 40 MM 1/2 CRB	24000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Specification: Item name: Needle: 40mm, ½ circle round body, heavy. -Ribbed &siliconised. 1.Wire diameter: 1.2+/-0.01mm 2. Bore diameter: 0.53+/-0.01mm 3. Suture material: Polypropylene, USP 1 (Monofilament) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420154	NEEDLE-30 MM 1/2 CRB PP:1	24000		
	SUTURE SIZE:1 30mm,1/2circle,round body(heavy) Wire diameter: 1.15+/-0.05mm Bore diameter: 0.53+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420			
1010420155	NEEDLE-PP:1 45 MM HEAVY, 1/2CRC	24000		
	Specifications: Wire diameter: 1.2+/-0.01mm Bore diameter: 0.53+/-0.01mm SUTURE SIZE: 1 45mm heavy,1/2circle rev cutting Material: High Quality Martensitic stainless steel Grade AISI-420.			
1010420156	NEEDLE-PP:1,1/2 CTC,30MM(HEAVY)	1200		
	Specifications: suture size-1(USP) Wire diameter: 1.12 +/-0.01mm Bore diameter: 0.53+/-0.01mm 30mm 1/2 C, Taper cutting, Heavy Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420157	NEEDLE 30MM 1/2 CC,MO PP: 1	1200		
	USP Size: Polypropylene 1 Bore diameter: 0.53+/-0.01mm Wire diameter: 1+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420158	NEEDLE 40MM 1/2 CC,MO: PP: 1	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	USP Size: Polypropylene 1 Bore diameter: 0.53+/-0.01mm Wire diameter: 1+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420159	NEEDLE 45MM 1/2 CC MO PP: 1	1200		
	USP Size: Polypropylene 1 Bore diameter: 0.53+/-0.01mm Wire diameter: 1.2+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420160	NEEDLE 45MM 1/2 CRB,MO:PP: 1	1200		
	NEEDLE 45MM 1/2 CIRCLE ROUND BODY MONOFILAMENT PP SUTURE USP 1 USP Size: Polypropylene 1 Bore diameter: 0.53+/-0.01mm Wire diameter: 1.2+/- 0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420161	NEEDLE NYLON 1/0, 1/2 C.R.B. 40MM	24000		
	"Item name: Needle: NEEDLE-40MM, 1/2 CIRCLE ROUND BODY, NYLON-USP 0, Siliconised& ribbed 1. Wire diameter: 1.00 (-0.0+0.025)mm 2. Bore diameter: 0.45 (-0.01+0.02)mm 3. Suture material: Polyamide Nylon, USP 0. (Monofilament) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420162	NEEDLE NYLON 1/0, STR C 60MM	1200		
	Specification: USP Size: 0 Wire diameter: 0.84+/-0.02mm Bore diameter: 0.45+/-0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420.			
1010420163	NEEDLE NYLON 2/0, 3/8 CIR.REV.CU.45MM	600000		
	USP size: 2-0, 45mm, 3/8 circle reverse cutting.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Bore Dia 0.36 +/- 0.01mm Wire dia : 0.99 to 1.00mm Material: High Quality martensitic Stainless Steel Grade AISI-420			
1010420164	NEEDLE 26MM, 3/8 CRC NYLON 2/0 USP Size: Nylon 2/0 Bore diameter: 0.36+/-0.01mm Wire diameter: 0.7+/-0.02mm Material:High Quality martensiticstainless steel Grade AISI-420.	1200		
1010420165	NEEDLE 16MM, 3/8 CRC NYLON 2/0 USP Size: Nylon 2/0 Bore diameter: 0.36+/-0.01mm Wire diameter: 0.61+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420166	NEEDLE NYLON 3/0, 3/8 CC G 16MM Monofilament Nylon Suture 1. Bore diameter (mm): 0.28+/-0.01mm 2. Wire diameter (mm): 0.60+/-0.05mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	1200		
1010420167	NEEDLE NYLON 3/0, 3/8 CIR.REV.CU. 26MM Monofilament Nylon Suture, Bore Dia 0.28 +/- 0.01mm wire dia: 0.71 to 0.72mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	300000		
1010420168	NEEDLE NYLON 3/0 , 3/8 CRC 24 MM Wire diameter: 0.61 (-0.0+0.02)mm Bore diameter: 0.28 (-0.01+0.02)mm Material: High Quality martensitic Stainless Steel Grade AISI-420.	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
1010420169	NEEDLE 45MM, 3/8 CRC NYLON 3/0	1200		
	USP Size: Nylon 3/0 Bore diameter: 0.28+/-0.01mm Wire diameter: 0.74+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420170	NEEDLE NYLON 4/0, 3/8 CC 16MM	75000		
	Specification: Wire diameter: 0.54+/-0.02mm Bore diameter: 0.25+/-0.01mm Suture size: Monofilament Nylon Suture, USP 4-0 Material: High Quality martensitic Stainless Steel Grade AISI-420.			
1010420171	NEEDLE 16MM, 3/8 CRC NYLON 4/0	1200		
	Suture size: Nylon 4/0 Wire diameter: 0.54+/-0.02mm Bore diameter: 0.25+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420172	NEEDLE 19MM, 3/8 CC NYLON 4/0	1200		
	NEEDLE 19MM, 3/8 CIRCLE CUTTING FOR NYLON 4/0 SUTURES Suture size: Nylon 4/0 Wire diameter: 0.54+/-0.02mm Bore diameter: 0.25+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420			
1010420173	NEEDLE NYLON 5/0, 3/8 CC 16MM	12000		
	Monofilament Nylon Suture, Wire dia : 0.46(-0.0+0.02)mm Bore Dia 0.19 ± 0.01mm Material: High Quality martensitic Stainless Steel Grade			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	AISI-420.			
1010420174	NEEDLE NYLON 5/0, 3/8 CIR.REV.CU 12MM	3000		
	Specification: Wire diameter: 0.46+/-0.02mm Bore diameter: 0.19+/-0.01mm Suture size: Monofilament Nylon 5-0 Material: High Quality martensitic Stainless Steel Grade AISI-420.			
1010420175	NEEDLE 13MM, 3/8 CRC NYLON 5/0	1200		
	Suture size: Nylon 5/0 Wire diameter: 0.46+/-0.02mm Bore diameter: 0.19+/-0.01mm Material: High Quality martensitic Stainless Steel Grade AISI-420			
1010420176	NEEDLE NYLON 1,1/2 CRC.45MM(Heavy)	72000		
	Monofilament Nylon Suture, 1.Bore Diameter (mm): 0.53+/-0.01mm 2. Wire diameter (mm): 1.20+/-0.05mm Material: High Quality martensitic Stainless Steel Grade AISI-420.			
1010420177	NEEDLE NYLON 1, 1/2 Cl. R.B. 40MM	6000		
	Monofilament Nylon Suture, Wire dia : 1.12-0.0+0.03mm Bore Dia 0.53-0.01+0.025mm Material: High Quality martensitic Stainless Steel Grade AISI-420.			
1010420178	Needle 40mm, 1/2 CRB.heavy,PGA:1	1200		
	Specification: USP SIze: 1 (Braided) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420179	Needle 40mm,1/2 CTC.heavy,PGA:1 Specification: USP Size: 1 (Braided) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420180	Needle 40mm, 1/2 CRC.PGA: 1 Specification: USP Size: PGA 1 (braided) Bore diameter: 0.55+/- 0.01MM Wire diameter: 1.2+/-0.02mm Material: High Quality martensitic stainless steel Grade AISI-420.	2400		
1010420181	Needle 50mm,1/2 CRC.heavy,PGA:1 Specification: USP Size: PGA 1 (Braided) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	24000		
1010420182	Needle 50mm, 1/2 CRB. heavy,PGA:1 Specification: USP Size: PGA 1 (Braided) Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420	1200		
1010420183	NEEDLE 45 MM, 3/8 CRB.PGA 1 Specification: USP Size: PGA 1 (Braided) Wire diameter: 1.00+/-0.02mm Bore diameter: 0.55+/-0.01mm	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420184	Needle 40mm, 1/2 CRC PGA 2/0 Specification: USP Size: 2/0 (Braided) Wire diameter: 0.84+/-0.02mm Bore diameter: 0.42+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420185	NEEDLE 30 mm, 1/2 CRB.PGA 2/0 Specification: USP Size: PGA 2-0 (Braided) Wire diameter: 0.80+/-0.02mm Bore diameter: 0.42+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420186	NEEDLE 40mm, 1/2 CRB.PGA 2/0 Specification: USP Size: PGA 2/0 (Braided) Wire diameter: 0.84+/-0.02mm Bore diameter: 0.42+/-0.02mm. Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420187	Needle 22mm, 3/8 CC.PGA: 3/0 Specification: USP Size: 3/0 (Braided) Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420188	Needle 20mm, 1/2 CCB.PGA: 3/0 Specification: USP Size: PGA 3/0 (Braided)	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Wire diameter: 0.62+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420189	Needle 16mm, 3/8 CC.PGA: 3/0 Specification: USP Size: PGA 3/0 (Braided) Wire diameter: 0.54+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420190	Needle 22mm, 1/2 CC.PGA: 3/0 Specification: USP Size: PGA 3/0 (Braided) Wire diameter: 0.70+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420191	Needle 12mm, 3/8 CRC.PGA: 3/0 Specification: USP Size: PGA 3/0 (Braided) Wire diameter: 0.54+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420192	Needle 17mm, 1/2 C.TA.CU.PGA: 3/0 Specification: USP Size: PGA 3/0 (Braided) Wire diameter: 0.60+/-0.02mm Bore diameter: 0.33+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420	1200		
1010420193	Needle 36mm, 1/2 C.TA.CU.PGA 3/0	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Specification: USP Size: PGA 3/0 (Braided) Wire diameter: 0.71+/-0.02mm Bore diameter: 0.33+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420194	Needle 20mm, 1/2 CRB.PGA: 4/0	1200		
	Specification: USP Size: 4/0 (Braided) Wire diameter: 0.54+/-0.02mm Bore diameter: 0.28+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420195	Needle 16mm, 1/2 CRB.PGA: 4/0	1200		
	Specification: USP Size: 4/0 (Braided) Wire diameter: 0.54+/-0.02mm Bore diameter: 0.28+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420196	Needle16mm, 5/8 CC PGA: 4/0	1200		
	Specification: USP Size: PGA 4/0 (Braided) Bore dia 0.28 +/-0.02mm Wire dia : 0.48 to 0.51mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420197	Needle 12mm, 3/8 CRCPGA: 4/0	1200		
	Specification: USP Size: PGA 4/0 (Braided) Wire diameter: 0.54+/-0.02mm Bore diameter: 0.28+/-0.01mm Material: High Quality martensitic stainless steel Grade			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	AISI-420.			
1010420198	NEEDLE 16mm, 1/2 CRB.PGA: 5/0 Specification: USP SIze:5/0 (Braided) Wire diameter: 0.48+/-0.02mm Bore diameter: 0.22+/-0.01mm Material:High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420199	Needle12mm, 3/8CRC.PGA: 5/0 Specification: USP Size: PGA 5/0 (Braided) Wire diameter: 0.48+/-0.02mm Bore diameter: 0.22+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420200	Needle 40mm, 1/2CRB.PGA: 0 Specification: USP SIze: 0 (Braided) Wire diameter: 1+/-0.02mm Bore diameter: 0.47+/-0.01mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420201	Needle 40mm, 1/2C.TA.CPGA: 0 Specification: USPSIze:PGA 0 (Braided) Wire diameter:1+/-0.02mm Borediameter:0.47+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAIS-420.	1200		
1010420202	Needle 36mm,1/2CRC(OS),PGA: 0 Specification:	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	USPSize:PGA 0 (Braided) Borediameter:0.47+/-0.01mm Wire diameter:1.2+/-0.02mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420203	NEEDLE 30 MM, 1/2 CRB.PGA: 0	1200		
	Specification: USPSize:PGA 0 (Braided) Borediameter:0.47+/-0.01mm. Wire diameter:0.84+/-0.02mm. Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420204	NEEDLE , 40MM 1/2CRB. HEAVY,PDO: 1	50000		
	Specification: Suture type:Polydioxanone(PDO) Monofilament sutureUSP1. Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420205	NEEDLE 50MM 1/2 CRB HEAVY PDO: 1.	1200		
	Specificaton: Suture type:Polydioxanone(PDO) MonofilamentUSP1. Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420206	NEEDLE 45MM 1/2CRB HEAVY PDO: 1	12000		
	Specification: SUTURE SIZE:USP1 Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	420.			
1010420207	NEEDLE 30MM, 1/2 CRB.PDO: 2-0 Specification: Suture size:USP2-0,monofilament poly(p-DIOXANONE). Wire diameter:0.78+/-0.02mm Borediameter:0.45+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	6000		
1010420208	NEEDLE 20MM 1/2 CRB.PDO: 3/0 Specification: SUTURE SIZE:USP3/0 Wire diameter:0.78+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	18000		
1010420209	NEEDLE 40MM 1/2 CRC-OS NEEDLE PDO Specification SUTURE SIZE:USP1 Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	9000		
1010420210	NEEDLE 40MM, 1/2 CRB. HEAVY, PDO: 0 Specification: Suture type:Polydioxanone(PDO) monofilament USP0. Wire diameter:1.12+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
1010420211	NEEDLE 40MM, 1/2CRB. HEAVY,PDO: 0.	1200		
	Specification: Suture type:Polydioxanone(PDO) monofilament USP0. Wire diameter:1.12+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420212	NEEDLE 30MM 1/2 CRB.PDO: 0	1200		
	Specification: SUTURE SIZE:USP0 Wire diameter:1+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420213	NEEDLE 48 MM 1/2 CRC.PDO:1	12000		
	Specification: SUTURE SIZE:USP1 Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420214	NEEDLE-40MM 1/2 C.TA.C.G HEAVY PDO:1	6000		
	Specification: Suture material:PDOmonofilament USP1 Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420215	NEEDLE PGA-PCL 2-0,26MM 1/2 CRB.	24000		
	Specification: Suture type:PGA-PCL[Poly			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	(glycolide-co-caprolactone] monofilament suture,USP2-0 Wire diameter:0.84+/-0.02mm Borediameter:0.45+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420216	NEEDLE PGA-PCL 2-0,26MM 3/8 CRB.	1200		
	Specification Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] monofilament suture,USP2-0. Wire diameter:0.84+/-0.02mm Borediameter:0.45+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420217	NEEDLE PGA-PCL 2/0 30MM, 1/2 CRB.	18000		
	Specification: SUTURE SIZE:USP2/0 Wire diameter:0.84+/-0.02mm Borediameter:0.45+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420218	NEEDLE PGA-PCL 2/0 30MM, 1/2 CRB	1200		
	Specification: Suture material:PGA-PCL MonofilamentUSP2-0 Wire diameter:0.84+/-0.02mm Borediameter:0.45+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420219	NEEDLE PGA-PCL3-0 25MM 3/8 CC.	1200		
	Specification: Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] MONOFILAMENTUSP3-0.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Wire diameter:0.84+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420220	NEEDLE PGA-PCL 3-0 26MM 1/2 CRB.	1200		
	Specification: Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] monofilament, USP3-0 Wire diameter:0.80+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420221	NEEDLE PGA-PCL3-0 26MM 3/8 CRB	1200		
	Specification: Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] monofilament sutureUSP3-0 Wire diameter:0.80+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420222	NEEDLE PGA-PCL 3/0 20MM, 1/2 CRB	1200		
	Specification: SUTURE SIZE:USP3/0 Wire diameter:0.78+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420223	NEEDLE PGA-PCL3/0 60MM, STR.CU.	1200		
	Specification: Suture material:PGA-PCL MonofilamentUSP3-0 Wire diameter:0.70+/-0.02mm			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420224	NEEDLE PGA-PCL 3/0 26 MM 1/2 CRB.	1200		
	Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] MONOFILAMENTUSP3-0. Wire diameter:0.62+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420225	NEEDLE PGA-PCL 3/0 24MM 3/8CRC USP 3-0	1200		
	Specification: Suture size :PGA-PCL3-0 (Monofilament) Wire diameter:0.62+/-0.02mm Borediameter:0.36+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420226	NEEDLE PGA-PCL 4-0 16MM 3/8 CRC.	6000		
	Specification: Suture size:PGA-PCL[Poly (glycolide-co-caprolactone)] MONOFILAMENTUSP4-0. Wire diameter:0.54+/-0.02mm Borediameter:0.28+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420227	NEEDLE PGA-PLA 4/0 20MM, 1/2 CRB.	12000		
	Specification: SUTURE SIZE:USP4/0 Wire diameter:0.70+/-0.02mm Borediameter:0.33+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420228	NEEDLE 20MM 1/2 CRB.PDO:4/0 Specification SUTURE SIZE:USP4/0 Wire diameter:0.62+/-0.02mm Borediameter:0.28+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	3000		
1010420229	NEEDLE16 MM 3/8 CRB.PGA-PCL 4-0 Suture type:PGA-PCL[Poly (glycolide-co-caprolactone)] MONOFILAMENTUSP4-0. Wire diameter:0.48+/-0.02mm Borediameter:0.28+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	1200		
1010420230	NEEDLE30MM 1/2 CRB.PGA-PCL 0 Specification: SUTURE SIZE:USP0 Wire diameter:1+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	1200		
1010420231	NEEDLE 40MM 1/2 C.T.A.C. HEAVY PDO:0 Specification: SUTURE SIZE:USP0 Wire diameter:1.12+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	12000		
1010420232	NEEDLE 30MM 1/2CRB HEAVY PDO:0 Specification: SUTURE SIZE:USP0 Wire diameter:1.12+/-0.02mm Borediameter:0.55+/-0.01mm	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420233	NEEDLE 40MM 1/2C.TA.C.PGA-PLA: 0 Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP0 Wire diameter:1+/-0.02mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	1200		
1010420234	NEEDLE 30MM, 1/2 CRB.PGA-PLA: 2/0. Specification: SUTURE SIZE: poly(glycolide-co-lactide)USP2/0 Wire diameter:0.80+/-0.02mm Borediameter:0.47+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	18000		
1010420235	NEEDLE 40MM 1/2 CRB.PGA-PLA: 2/0 Specification SUTURE SIZE: Poly(glycolide-co-lactide)USP2/0 Wire diameter:0.84+/-0.02mm Borediameter:0.47+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	18000		
1010420236	NEEDLE 40MM 1/2 CRC.PGA-PLA: 2/0 Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP2/0 Wire diameter:0.84+/-0.02mm Borediameter:0.47+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	1200		
1010420237	NEEDLE 20MM 1/2 CRB.PGA-PLA: 3/0 Specification: Needle:20mm,1/2circleroundbody	100000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Wire diameter:0.62 +/- 0.01mm Borediameter:0.36 +/- 0.01mm Suture size:poly(glycolide-co-lactide) USP3/0 Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420238	NEEDLE 22MM, 3/8 CC.PGA-PLA: 3-0	1200		
	"Specification: Needle:22mm,3/8circlecutting Wire diameter:0.62 +/- 0.01mm Borediameter:0.36 +/- 0.01mm Suture size:poly(glycolide-co-lactide) USP3/0 Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420239	NEEDLE 16MM 1/2 CRB.PGA-PLA: 5/0.	3000		
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP5/0 Wire diameter:0.48+/-0.02mm Borediameter:0.28+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420240	NEEDLE-40MM 1/2 CIR.REV.CU. OS	1200		
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP1 Wire diameter:1.2+/-0.01mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420241	NEEDLE 40MM, 1/2 CRB.PGA-PLA: 0	150000		
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP0 Wire diameter:1+/-0.01mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420242	NEEDLE 40MM 1/2 CRB.HEAVY PGA-PLA:1	1200		
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide)USP1 Wire diameter:1.2+/-0.01mm Borediameter:0.60+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420243	NEEDLE - 36MM 1/2 CRC.OS NEEDLE	60000		
	Specification: SUTURE SIZE: Ploy(glycolide-co-lactide)USP0 Wire diameter:1.2+/-0.01mm Borediameter:0.55+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420244	NEEDLE40MM 1/2 C.TA.CU.HEAVY PGA-PLA	24000		
	"Specification: Needle:40mm,1/2circletaper cutting, heavy Wire diameter:1.2 +/- 0.01mm Borediameter:0.60 +/- 0.01mm Suture size:poly(glycolide-co-lactide) USP1 Material:High Qualitymartensitic stainlesssteelGradeAISI-420			
1010420245	NEEDLE 30MM, 1/2 CRB. PGA-PLA: 0	24000		
	Wire diameter: 0.84+/-0.02mm Bore diameter: 0.53+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420			
1010420248	NEEDLE 40MM 1/2 C.TA.CU.HEAVY PGA-PLA	1200		
	Specification: SUTURE SIZE: Poly(glycolide-co-lactide),USP1 Wire diameter:1.2+/-0.02mm Borediameter:0.60+/-0.01mm Material:High Qualitystainlesssteel Grade300.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420249	NEEDLE SUTURE 36MM, 1/2 CRB.PGA-PLA:1	3000		
	Specification: Needle:36mm,1/2circleroundbody. Suture size:Polyglactin910, USP1 (Braided) Wire diameter:1+/-0.02mm Borediameter:0.57+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420251	NEEDLE 40MM 1/2 C.TA.CU.HEAVY PGA-PLA.	1200		
	Specification: Suture size:Polyglactin910, Braided, USP1. Wire diameter:1+/-0.01mm Borediameter:0.57+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420267	NEEDLE-30MM, 3/8 CIRCLE ROUND BODY (PGA)	1200		
	Specification: Suture size:PGA, USP2-0 Wire diameter:0.80 +/- 0.01mm Borediameter:0.42 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420268	NEEDLE-37MM, 1/2 CIRCLE REV CUT(PGA PLA)	9000		
	Specification: Suture size:PGA-PLA,USP2-0 Wire diameter:0.84 +/- 0.01mm Borediameter:0.47 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420			
1010420269	NEEDLE-37MM,1/2 CIRCLE ROUND BODY	1200		
	PGS-PLA)Specification: Suture size:PGA-PLA,USP2-0 Wire diameter:0.84 +/- 0.01mm Borediameter:0.47+/-			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420270	NEEDLE-26MM, 3/8 CIRCLE ROUND BODY (PGA)	1200		
	Specification: Suture size:PGA, USP3-0 Wire diameter:0.70 +/- 0.01mm Borediameter:0.33 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420271	NEEDLE-30MM, 3/8 CIRCLE ROUND BODY.	1200		
	Specification: Suture size:PGA, USP3-0 Wire diameter:0.70 +/- 0.01mm Borediameter:0.33 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420272	NEEDLE-37MM, 1/2 CIRCLE ROUND BODY	1200		
	(PGA PLA)Specification: Suture size:PGA-PLA,USP3-0 Wire diameter:0.70 +/- 0.01mm Borediameter:0.36 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420273	NEEDLE-37MM,1/2 CIRCLE REV-CUT (PGA PLA)	1200		
	Specification: Suture size:PGA-PLA,USP3-0 Wire diameter:0.70 +/- 0.01mm Borediameter:0.36 +/- 0.01mm Material:High QualitymartensiticstainlesssteelGradeAISI-420.			
1010420274	NEEDLE-13MM, 1/2 CIRCLE ROUND BODY	1200		
	Suture size:PGA, USP4-0 Wire diameter:0.54 +/- 0.01mm Borediameter:0.28 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	420.			
1010420276	NEEDLE 45MM 3/8 CIRCLE CUTTING PGA 1 SPECIFICATION: USPSIZE:PGA 1 Wire diameter:1.2mm Borediameter:0.55mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	1200		
1010420277	NEEDLE-36MM, 3/8 CIRCLE ROUND BODY (PGA) Specification: Suture size:PGA, USP0 Wire diameter:0.84 +/- 0.01mm Borediameter:0.47 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	1200		
1010420278	NEEDLE-36MM, 1/2 CIRCLE ROUND BODY (PGA) Specification: Suture size:PGA, USP1 Wire diameter:1.00 +/- 0.01mm Borediameter:0.55 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.	1200		
1010420279	NEEDLE-26MM, 1/2 CIRCLE TAPER CUT (PGA) Specification: Suture size:PGA, USP1 Wire diameter:0.84 +/- 0.01mm Borediameter:0.55 +/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420	1200		
1010420280	NEEDLE FOR PDO SUTURE-40MM, 1/2 C R B specification: NEEDLE :40MM, 1/2CIRCLE ROUNDBODY. Suture type:Polydioxanone(PDO) monofilament sutureUSP1. Wire diameter:1+/-0.01mm	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Borediameter:0.57+/-0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420281	NEEDLE-40MM,1/2 CIRCLE REV CUT (PGA-PLA)	3600		
	Specification: Suture size:PGA-PLA,USP0 Wire diameter:1.00 +/- 0.01mm Borediameter:0.55+/- 0.01mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420282	NEEDLE PGA-PLA 2-0 SUTURE 30MM,1/2 C, TC	9000		
	Specification: Needle:30mm,1/2circle taper cutting. Suture material:PGA-PLAUSP2-0 Wire diameter:0.80mm Borediameter:0.47mm Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420283	NEEDLE-50MM 1/2C REVERSE CUTTING, HEAVY.	1200		
	SPECIFICATION: NEEDLE:50MM, 1/2CIRCLE REVERSE CUTTING HEAVY. SUTURE:PGA-PLAUSP1 WIREDIAMETER:1.2+/-0.01MM BORE DIAMETER:0.60 MM Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420284	NEEDLE-30MM 1/2 C REVERSE CUTTING	1200		
	Specification: Item name:Needle:30mm,1/2circle reversecutting-siliconised. 1.Wire diameter:0.92+/-0.01mm 2. Borediameter:0.55+/-0.01mm 3. Suture material:PGA-PLA,USP0 (Braided) 4. Allneedlesmustbesiliconised 5. Material:High Qualitymartensitic			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	stainlesssteelGradeAISI-420.			
1010420285	NEEDLE-40MM 1/2 CIRCLE ROUND BODY	75000		
	NEEDLE-40MM 1/2CIRCLE ROUND BODY Specification: NEEDLE:40MM, 1/2 CIRCLE ROUNDBODY.SUTURE MATERIAL:PGA- PLAUSPSIZE 2. Wire diameter: 1.2MM Borediameter:0.72MM Material:High Qualitymartensitic stainless steelGradeAISI- 420			
1010420287	NEEDLE-30MM 1/2 CIRCLE ROUND BODY	1200		
	NEEDLE-30MM 1/2CIRCLE ROUNDBODYSPECIFICATION: NEEDLE:30MM, 1/2 CIRCLE ROUNDBODYSUTURE MATERIAL:PGA- PLAUSP2 BORE DIAMETER:0.72MM WIREDIAMETER:1MM Material: High Qualitymartensiticstainless steel GradeAISI-420			
1010420288	NEEDLE-40MM 1/2 CIRCLE TAPER CUTTING	24000		
	NEEDLE-40MM 1/2CIRCLE TAPER CUTTING Specification:Needle: 40mm,1/2 circletapercuttingSuturematerial: PGA-PLAUSP2 Wire diameter: 1.2mmBorediameter:0.72mm Material:High Qualitymartensitic stainless steelGradeAISI- 420			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
1010420289	NEEDLE-48MM 1/2 CIRCLE ROUND BODY	50000		
	NEEDLE-48MM 1/2CIRCLE ROUNDBODYSPECIFICATION: Needle:48mm,½ circleround body Suturematerial: PGA-PLAsutureUspsize 2 Wire diameter:1.3mm Borediameter: 0.72mm Material:High Quality martensitic stainlesssteelGrade AISI-420			
1010420292	NEEDLE- PGA PLA 2/0,5/8 TAPER POINT 27MM	1200		
	"Specifications: SUTURE SIZE:2/0 Wire diameter:0.71+/-0.01mm Borediameter:0.47+/-0.01mm 27mm, 5/8C, TaperPoint Material:High Qualitymartensitic stainlesssteelGradeAISI-420.			
1010420293	NEEDLE-30MM,1/2 CRB,HIDOX USP 3-0	6000		
	Specification: Needle:30mm,½ circleroundbody Wire diameter:0. 78mm Borediameter:0.36mm Suture mtl:PolydioxanoneUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420			
1010420294	NEEDLE-50MM,1/2 CRB,HEAVY,PGA-PLA USP 1	75000		
	Specification: Needle:50mm,½ circleroundbody, heavy. Wire diameter:1.2mm Borediameter:0.60mm Suture mtl:Polyglactin910 USP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
1010420295	NEEDLE-30MM,1/2 CRC, PGA USP 0	1200		
	Specification: Needle:30mm,1/2 circlereverse cutting Wire diameter:0. 84mm Borediameter:0.47mm Suture mtl:PolyglycolicacidUSP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420296	NEEDLE-22MM,1/2CIRCLE CUTTING,PGA-PLA3-0	24000		
	Specification: Needle:22mm,1/2circlecutting Wire diameter:0. 62mm Borediameter:0.36mm Suture mtl:Polyglactin910 USP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420297	NEEDLE-45MM,1/2 CIRCLE CUTTING,PP USP-0	1200		
	Specification: Needle:45mm,1/2circlecutting Wire diameter:1.1mm Borediameter:0.45mm Suture mtl:PolypropyleneUSP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420298	NEEDLE-16MM,3/8 CIRCLE RC,Silk USP 3-0	1200		
	Specification: Needle:16mm,3/8circlereverse cutting Wire diameter:0.54mm			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Borediameter:0.33mm Suture mtl:SilkUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420299	NEEDLE-26MM,3/8CIRCLE CUTTING,PGA-PLA2-0	6000		
	Specification: Needle:26mm,3/8circlecutting Wire diameter:0. 78mm Borediameter:0.47mm Suture mtl:Polyglactin910 USP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420300	NEEDLE-22MM,1/2CIRCLE CUTTING,PGA-PLA4-0	1200		
	Specification: Needle:22mm,1/2circlecutting Wire diameter:0. 54mm Borediameter:0.33mm Suture mtl:Polyglactin910 USP4-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420301	NEEDLE-30MM,1/2 CIRCLE CUTTING,SILK 2-0	1200		
	Specification: Needle:30mm,1/2circlecutting Wire diameter:0.78 +/- 0.01 mm Borediameter:0.42mm Suture mtl:SilkUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420302	NEEDLE-36MM,1/2CIRCLE TAPER CUT,PGA 2-0	6000		
	Specification: Needle:36mm,1/2circletaper cutting			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Wire diameter:0.78 +/- 0.01 mm Borediameter:0.42+/- 0.01 mm Suture mtl:PolyglycolicAcidUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420303	NEEDLE-36MM,1/2CIR. ROUND BODY,PGA 2-0	6000		
	Specification: Needle:36mm,1/2circleroundbody Wire diameter:0.78+/- 0.01 mm Borediameter:0.42+/- 0.01 mm Suture mtl:PolyglycolicAcidUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420304	NEEDLE-40MM,3/8CIR. ROUND BODY,PGA 1	1200		
	Specification: Needle:40mm,3/8circleroundbody Wire diameter:1.00+/- 0.01 mm Borediameter:0.55+/- 0.01 mm Suture mtl:PolyglycolicAcidUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420305	NEEDLE-25MM,1/2CIR. ROUND BODY,PP 2/0	6000		
	Needle: 25mm, 1/2 circle round body Wire diameter:0.78+/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:PolypropyleneUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420306	NEEDLE-25MM,3/8CIR. CUTTING,PP 3/0	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Specification: Needle:25mm,3/8circlecutting Wire diameter:0.61+/- 0.01 mm Borediameter:0.28+/- 0.01 mm Suture mtl:PolypropyleneUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420307	NEEDLE-40MM,1/2CIRCLE REVERS CUT,PGA 3-0	1200		
	Specification: Needle:40mm,1/2circlereverse cutting-ribbed&siliconised Wire diameter:0.78 +/- 0.01 mm Borediameter:0.33+/- 0.01 mm Suture mtl:PolyglycolicAcidUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420308	NEEDLE-25MM, 1/2 CRC, PGA-PLA USP 3-0	1200		
	Specification: Needle:25MM, 1/2CIRCLE REVERSECUTTING-ribbed& siliconised Wire diameter:0.62 +/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:PGA-PLAUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420309	NEEDLE-26MM, 3/8 CRC, HEAVY,PGA-PLA USP	1200		
	Specification: Needle:26MM, 3/8CIRCLE REVERSECUTTING, HEAVY-ribbed&siliconised Wire diameter:0.84 +/- 0.01 mm			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Borediameter:0.60+/- 0.01 mm Suture mtl:PGA-PLAUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420310	NEEDLE-26MM, 3/8 CIRCLE CUTTING, USP 3-0	1200		
	Specification: Needle:26mm,3/8circle cutting-ribbed&siliconised Wire diameter:0.62 +/- 0.01 mm Borediameter:0.33+/- 0.01 mm Suture mtl:PolyglycolicAcidUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420311	NEEDLE 30MM,3/8 CIRCLE CUTTING USP 3/0	1200		
	Specification: Needle:30MM, 3/8CIRCLE CUTTINGNEEDLE-ribbed& siliconised Wire diameter:0.66 (+0.02 mm) Borediameter:0.28 (-0.01+0.02mm) Suture mtl: NylonUSP3/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420312	NEEDLE 30MM,3/8 CIRCLE CUTTING USP 2/0	1200		
	Specification: Needle:30MM, 3/8CIRCLE CUTTINGNEEDLE-ribbed&siliconised Wire diameter:0.74 (+0.02mm) Borediameter:0.36 (-0.01+0.02mm) Suture mtl: NylonUSP2/0 Material:High Qualitymartensitic			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	stainlesssteelGrade AISI-420.			
1010420313	NEEDLE 13MM,3/8 CIRCLE TAPER CUT USP 6/0	1200		
	Specification: Needle:13MM, 3/8CIRCLE TAPER CUT-ribbed&siliconised Wire diameter:0.27 (+0.02 mm) Borediameter:0.15 (+/- 0.01mm) Suture mtl:PolypropeleneUSP6/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420314	NEEDLE-12MM, 1/2 CIRCLE CUTTING	1200		
	Specification: Needle:12MM, 1/2CIRCLE CUTTINGNEEDLE-ribbed&siliconised Wire diameter:0.46 (-0.0+0.02mm) Borediameter:0.22(-0.01+0.02mm) Suture mtl:PGA USP5-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420315	NEEDLE-30MM, 1/2C, REVERSE CUTTING	6000		
	Specification: Needle:30MM, 1/2CIRCLE REVERSECUTTING-ribbed&siliconised Wire diameter:0.74 (-0.0mm +0.02mm) Borediameter:0.36 (-0.01mm +0.02mm) Suture mtl:PolypropyleneUSP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
1010420316	NEEDLE-16MM, 1/2C, REVERSE CUTTING	3000		
	Specification: Needle:16MM, 1/2CIRCLE REVERSE CUTTING- ribbed&siliconised Wire diameter:0.56 (-0.0mm +0.02mm) Borediameter:0.28 (-0.01mm +0.02mm) Suture mtl:PolyglycolicacidUSP5-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420317	NEEDLE-20MM, 1/2C, REVERSE CUTTING	1200		
	Specification: Needle:20MM, 1/2CIRCLE REVERSE CUTTING-ribbed& siliconised Wire diameter:0.74 (-0.0mm +0.02mm) Borediameter:0.36 (-0.01mm +0.02mm) Suture mtl:PolydioxanoneUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420318	NEEDLE-30 MM 1/2CIRCLE CUTTING	1200		
	Specification: Needle:30MM, 1/2CIRCLE, CUTTING-ribbed&siliconised Wire diameter:0.71 (-0.0mm +0.02mm) Borediameter:0.33 (-0.01mm +0.02mm) Suture mtl:SilkUSP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420319	NEEDLE 13MM 3/8CIRCLE ROUND BODY	1200		
	Specification: Needle:13MM, 3/8CIRCLE, ROUNDBODY-ribbed&siliconised Wire diameter:0.24 (-0.0+0.02mm) Borediameter:0.18 (± 0.01mm) Suture mtl:SilkUSP6-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420320	NEEDLE 23MM 3/8CIRCLE REV.CUTTING HEAVY	1200		
	Specification: Needle:23MM, 3/8CIRCLE, REVERSECUTTING, HEAVY-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420321	NEEDLE 23MM 3/8CIRCLE REV.CUTTING HEAVY	1200		
	Specification: Needle:23MM, 3/8CIRCLE REVERSECUTTING, HEAVY-ribbed&siliconised Wire diameter:0.84 +/- 0.01 mm Borediameter:0.60+/- 0.01 mm Suture mtl:PGA-PLAUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420322	NEEDLE-26MM 3/8CIR. REVERSE CUTTING,2/0	1200		
	Specification: Needle:26MM, 3/8CIRCLE,			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	REVERSE CUTTING- ribbed& siliconised Wire diameter:0.71 (-0.0+0.02mm) Borediameter:0.36 (-0.01+0.02mm) Suture mtl:NYLON USP2-0 Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420323	NEEDLE-19MM 3/8CIR. REVERSE CUTTING,3/0	1200		
	Specification: Needle:19MM, 3/8CIRCLE, REVERSE CUTTING- ribbed& siliconised Wire diameter:0.61 (-0.0+0.02mm) Borediameter:0.28 (-0.01+0.02mm) Suture mtl:NYLON USP3/0 Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420324	NEEDLE-22MM 3/8CIR. REVERSE CUTTING,3/0	1200		
	Specification: Needle:22MM, 3/8CIRCLE, REVERSE CUTTING-ribbed& siliconised Wire diameter:0.61 (-0.0+0.02mm) Borediameter:0.28 (-0.01+0.02mm) Suture mtl: NYLON USP3/0 Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420325	NEEDLE-19MM 3/8CIR. REVERSE CUTTING,4/0	1200		
	Specification: Needle:19MM, 3/8CIRCLE, REVERSE CUTTING- ribbed& siliconised Wire diameter:0.56 (-0.0+0.02mm)			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Borediameter:0.25 (-0.01+0.02mm) Suture mtl: NYLON USP4/0 Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420326	NEEDLE-22MM 3/8CIR. REVERSE CUTTING,4/0	3000		
	Specification: Needle:22MM, 3/8CIRCLE, REVERSE CUTTING-ribbed&siliconised Wire diameter:0.56 (-0.0+0.02mm) Borediameter:0.25 (-0.01+0.02mm) Suture mtl: NYLON USP4/0 Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420327	NEEDLE-19MM 3/8CIR. REVERSE CUTTING,3/0	6000		
	Specification: Needle:19MM, 3/8CIRCLE, REVERSE CUTTING-ribbed&siliconised Wire diameter:0.61 (-0.0+0.02mm) Borediameter:0.33 (-0.01+0.02mm) Suture mtl: SILK USP3/0 Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420328	NEEDLE-19MM 3/8CIR. REVERSE CUTTING,4/0	6000		
	Specification: Needle:19MM, 3/8CIRCLE, REVERSE CUTTING- ribbed&siliconised Wire diameter:0.56 (-0.0+0.02mm) Borediameter:0.28 (-0.01+0.025mm) Suture mtl: SILK USP4/0 Material:High Quality martensitic stainless steel Grade AISI-420.			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420329	NEEDLE-23MM 1/2 CIR.REV.CUT,HEAVY,USP O	1200		
	Specification: NEEDLE-23MM 1/2CIRCLE REVERSE CUTTING, HEAVY, USP O-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420330	NEEDLE-23MM 1/2 CIR.REV.CUT,HEAVY,USP 1	1200		
	Specification: Needle:NEEDLE-23MM 1/2CIRCLE REVERSE CUTTING,HEAVY,USP 1- ribbed&siliconised Wire diameter:0.84 (-0.0+0.02mm) Borediameter:0.60 (-0.01+0.02mm) Suture mtl:PGA-PLAUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420331	NEEDLE-23MM 1/2 CIR.REV.CUT,HEAVY,USP O	75000		
	Specification: NEEDLE-23MM 1/2CIRCLE REVERSE CUTTING,HEAVY,USP O-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420332	NEEDLE-23MM 1/2 CIR.REV.CUT,HEAVY,USP 1	75000		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Specification: NEEDLE-23MM 1/2CIRCLE REVERSE CUTTING,HEAVY,USP 1-ribbed&siliconised Wire diameter:0.84 (-0.0+0.02mm) Borediameter:0.60 (-0.01+0.02mm) Suture mtl:PGA-PLAUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420333	NEEDLE-36MM 1/2 CIR. REVERSE CUTTING, 2/	75000		
	Specification: Needle:NEEDLE-36MM 1/2CIR. REVERSE CUTTING,2/0- ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.47 (-0.01+0.02mm) Suture mtl:Polyglactin910Rapid USP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420334	NEEDLE-36MM 1/2 CIR. ROUND BODY, 2/0	75000		
	Specification: Needle:NEEDLE-36MM 1/2CIR. ROUNDBODY,2/0-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.47 (-0.01+0.02mm) Suture mtl:Polyglactin910Rapid USP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420335	NEEDLE-20MM 3/8 CIR. ROUND	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	BODY, 4/0			
	Specification: Needle:NEEDLE-20MM 3/8CIR. ROUNDBODY,4/0-ribbed&siliconised Wire diameter:0.58 (-0.0+0.02mm) Borediameter:0.33 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP4-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420336	NEEDLE-45MM 3/8 CRB, PGA USP 1	6000		
	Specification: Needle:NEEDLE-45MM 3/8CIRCLE ROUNDBODY,USP1- ribbed&siliconised Wire diameter:0.86 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:PolyglycolicacidUSP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420			
1010420337	NEEDLE-40MM 1/2 CRB, PGA-PLA USP 0	1200		
	Specification: Needle:NEEDLE-40MM 1/2CIRCLE ROUNDBODY,USP0- ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420338	NEEDLE-40MM 1/2 CRB, HEAVY PGA-PLA USP1	1200		
	Specification: Needle:NEEDLE-40MM 1/2CIRCLE ROUNDBODY,HEAVY,USP 1-ribbed&siliconised Wire diameter:1.00 (-0.0+0.02mm)			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Borediameter:0.60 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420			
1010420339	NEEDLE-40MM 1/2 CTC, PGA-PLA USP0	6000		
	Specification: Needle:NEEDLE-40MM 1/2CIRCLE TAPERCUTTING,USP0-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420340	NEEDLE 40MM 1/2 CRC.PGA-PLA: 2/0	30000		
	Specification: Needle:NEEDLE-40MM 1/2CIRCLE REVERSECUTTING,USP 2/0-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.47 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420341	NEEDLE-40MM, 3/8CRC,NYLON1	1200		
	Specification: Needle:NEEDLE-40MM 3/8CIR. REVERSECUTTING,ribbed& siliconised Wire diameter:1 (-0.0+0.025mm) Borediameter:0.53 (-0.01+0.025mm) Suture mtl: Nylon1 Material:High Qualitymartensitic			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	stainlesssteelGrade AISI-420.			
1010420342	NEEDLE-40MM,1/2CRB,HEAVY,AISI 302,USP1	1500000		
	Specification: Item name:Needle:NEEDLE-40MM, 1/2CIRCLE ROUND BODY, HEAVY-USP 1- Siliconised. 1.Wire diameter:1.00 (-0.0+ 0.025)mm 2. Borediameter:0.60 (-0.01+ 0.025)mm 3. Suture material:PGA-PLA,USP1 (Braided) 4. Allneedlesmustbesiliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.			
1010420343	NEEDLE-40MM,1/2CTC,HEAVY,AISI 302,USP1	1000000		
	Specification: Item name:Needle:NEEDLE-40MM, 1/2CIRCLE TAPERCUTTING, HEAVY-USP 1 Siliconised. 1.Wire diameter:1.00 (-0.0+ 0.025)mm 2. Borediameter:0.60 (-0.01+ 0.025)mm 3. Suture material:PGA-PLA,USP1 (Braided) 4. Allneedlesmustbesiliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.			
1010420344	NEEDLE-45MM,1/2CRC,PGA-PLA USP 2	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Specification: Item name:Needle:NEEDLE-45MM, 1/2CIRCLE REVERSE CUTTING,-USP2 Ribbed &siliconised. 1.Wire diameter:1.20 (-0.0+0.03)mm 2. Borediameter:0.72 (-0.01+0.025)mm 3. Suture material:PGA-PLA,USP2 (Braided) 4. Allneedlesmustbesiliconised& ribbed. 5. Material:High Qualitymartensitic stainlessteelGrade AISI-420.			
1010420345	NEEDLE-19MM,1/2 CIR.CUT.,PGA- PLA USP 5/0	1200		
	Specification: Item name:Needle:NEEDLE-19MM, 1/2CIRCLE CUTTING-USP5/0 Ribbed &siliconised. 1.Wire diameter:0.56 (-0.0+0.02)mm 2. Borediameter:0.28 (-0.01+0.02)mm 3. Suture material:PGA-PLA,USP 5/0(Braided) 4. Allneedlesmustbesiliconised& ribbed. 5. Material:High Qualitymartensitic stainlessteelGrade AISI-420.			
1010420346	NEEDLE-30MM,1/2 CRB,AISI 302,PGA-PLA 3/0	12000		
	Specification: Item name:Needle:NEEDLE-30MM, 1/2CIRCLE ROUND BODY,USP 3-0 Siliconised&RIBBED. 1.Wire diameter:0.56 (-0.0+0.02)mm 2. Borediameter:0.36 (-0.01+ 0.02)mm 3. Suture material:PGA-PLA,USP 3-			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	0(Braided) 4. Allneedlesmustbesiliconised& ribbed. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.			
1010420347	NEEDLE-25MM,3/8 CIRCLE PRIME CUTTING	120000		
	Specification: Item name:Needle: NEEDLE-25MM,3/8 CIRCLE PRIME CUTTING, AISI302,USP3-0 Siliconised 1.Wire diameter:0.58 (-0.0+0.02)mm 2. Borediameter:0.36 (-0.01+ 0.02)mm 3. Suture material:PGA-PCL,USP 3-0 4. Allneedlesmustbesiliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.			
1010420348	NEEDLE-19MM, 3/8 CRC, PGA-PLA 3-0	3000		
	Needle:19MM, 3/8CIRCLE, REVERSECUTTING-ribbed&siliconised Wire diameter:0.61 (-0.0+0.02mm) Borediameter:0.36 (-0.01+0.02mm) Suture mtl: SILKUSP3/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420349	NEEDLE-40MM 1/2 CRB,AISI 302,PGLA USP 0	300000		
	Needle:NEEDLE-40MM 1/2CIRCLE ROUND BODY,USP0-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High QualityAustenitic stainlesssteelGrade AISI-302.			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420350	NEEDLE-40MM 1/2 CTC,AISI 302,PGLA USP0	120000		
	Needle:NEEDLE-40MM 1/2CIRCLE TAPERCUTTING,USP0-ribbed&siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High QualityAustenitic stainlesssteelGrade AISI-302.			
1010420351	NEEDLE 30MM, 1/2 CRB,AISI 302,PGLA 2/0.	240000		
	NEEDLE: 30MM, 1/2CIRCLE ROUND BODY- Ribbed&siliconised, PGA-PLA: 2/0. Wire diameter:0.74(-0.0+0.02)mm Borediameter:0.47(-0.01+0.02)mm Material:High QualityAustenitic stainlesssteelGrade AISI-302.			
1010420352	NEEDLE-40MM 1/2 CIR.REV.CU. OS, AISI-302	240000		
	Specification: NEEDLE:40MM, 1/2CIRCLE REVERSE CUTTING, OS NEEDLE-ribbed&siliconised, pga-plaUSP1 Wire diameter:1.2(-0.0+0.03)mm Borediameter:0.60(-0.01+0.025)mm Material:High QualityAustenitic stainlesssteelGrade AISI-302.			
1010420353	NEEDLE -60 MM 3/8 CC SILK 1 BD053	120000		
	Specification: Needle:60mm,3/8circle,cutting-ribbed &siliconised, silkusp 1 1. BoreDiameter(mm): 0.53(-0.01+0.025) 2. Wire diameter(mm): 1.20(-0.0+0.03)			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420354	NEEDLE-36MM 1/2 CRC,PGA-PLA USP 0	1200		
	Specification: Needle:NEEDLE-36MM 1/2CIRCLE REVERSECUTTING,USP0-ribbed &siliconised Wire diameter:0.81 (-0.0+0.02mm) Borediameter:0.55 (-0.01+0.02mm) Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420			
1010420355	NEEDLE-22MM 1/2 CRC, PGA-PLA USP2-0	1200		
	Specification: NEEDLE: 22MM, 1/2CIRCLE REVERSECUTTING- Ribbed&siliconised, PGA-PLA: 2/0. Wire diameter:0.71(-0.0+0.02)mm Borediameter:0.47(-0.01+0.02)mm Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420356	NEEDLE-40MM 1/2 CRC, PGA-PLA USP2	18000		
	Specification: Needle:NEEDLE-40MM 1/2CIRCLE REVERSECUTTING,USP2-ribbed &siliconised Wire diameter:1.2 (-0.0+0.03mm) Borediameter:0.72 (-0.01+0.025mm) Suture mtl:Polyglactin910 USP2 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420357	NEEDLE-22MM 1/2 CRC, PGA-PLA USP3-0	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Specification: NEEDLE: 22MM, 1/2CIRCLE REVERSE CUTTING-Ribbed& siliconised, PGA-PLA: 3/0. Wire diameter:0.62(-0.0+0.02)mm Borediameter:0.36(-0.01+0.02)mm Material:High Qualitymartensitic stainlesssteelGrade AISI-420			
1010420358	NEEDLE-60MM STRAIGHT CUTTING,USP 2-0	1200		
	Specification: NEEDLE: 60MM, STRAIGHT CUTTING-Ribbed&siliconised, NYLON2/0. Wire diameter:0.74(-0.0+0.02)mm Borediameter:0.36(-0.01+0.02)mm Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420359	NEEDLE-50MM 1/2 CRB, PGA 0	24000		
	Specification: Needle:NEEDLE-50MM 1/2CIRCLE ROUNDBODY,USP0-ribbed& siliconised Wire diameter:1 (-0.0+0.025)mm Borediameter:0.47 (-0.01+0.02)mm Suture mtl:PGA USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420360	NEEDLE-50MM, 1/2 CRC-PGA 2/0	1200		
	Specification: Needle:NEEDLE-50MM 1/2CIRCLE REVERSE CUTTING,USP 2/0-ribbed&siliconised Wire diameter:0.88 (-0.0+0.025mm)			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Borediameter:0.42 (-0.01+0.02mm) Suture mtl:PGA USP2/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420361	NEEDLE-50MM, 1/2 CRB-PGA 2/0	1200		
	Specification: Needle:NEEDLE-50MM 1/2CIRCLE ROUNDBODY,USP2/0-ribbed& siliconised Wire diameter:0.88 (-0.0+0.025)mm Borediameter:0.42 (-0.01+0.02)mm Suture mtl:PGA USP2/0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420			
1010420362	NEEDLE-50MM, 1/2 CRB, PGA-PLA 1,AISI 302	240000		
	"Item name: Needle: NEEDLE-50MM, 1/2CIRCLE ROUNDBODY,HEAVY-USP 1 Siliconised, AISI-302 1.Wire diameter:1.08 (-0.0+ 0.025)mm 2. Borediameter:0.60 (-0.01+ 0.025)mm 3. Suture material:PGA-PLA,USP1 (Braided) 4. Allneedlesmustberibbedand siliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.			
1010420363	NEEDLE-45MM 1/2 CTC, HEAVY, PET USP 2	75000		
	Item name: Needle: NEEDLE-45MM, 1/2CIRCLE TAPERCUTTING,HEAVY-USP 2 SiliconisedandRibbed. 1. Wire diameter:1.20 (-0.0+ 0.03)mm			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	2. Borediameter:0.65 (-0.01+ 0.025)mm 3. Suture material:Polyester,USP2 (Braided) 4. Allneedlesmustbesiliconised and ribbed. 5. Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420364	NEEDLE-26MM, 1/2 CC, PGA-PLA USP 3/0	1200		
	Specification: Item name:NEEDLE-26MM, ½ CIRCLE CUTTING,Siliconised. 1.Wire diameter:0.56 (-0.0+0.02)mm 2. Borediameter:0.36 (-0.01+ 0.02)mm 3. Suture material:PGA-PLA,USP3-0 (Braided) 4. Allneedlesmustbesiliconised& ribbed. 5. Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420365	NEEDLE 45MM,1/2CRB 4-0	1200		
	"Item name: Needle: NEEDLE-45 MM, 1/2CIRCLE ROUND BODY, USP4-0, AISI- 302 1.Wire diameter:0.81(-0.0+0.02)mm 2. Borediameter:0.28 (-0.01+ 0.02)mm 3. Suture material: PLA,USP 4-0 (Braided) 4. Allneedlesmustberibbedand siliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.			
1010420366	NEEDLE 45MM,1/2CRB 5-0	1200		
	Item name: Needle: NEEDLE-45 MM, 1/2CIRCLE ROUND BODY, USP5-0, AISI- 302			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	1.Wire diameter:0.74 (-0.0+0.02)mm 2. Borediameter:0.22 (-0.01+ 0.02)mm 3. Suture material: PLA,USP 5-0 (Braided) 4. Allneedlesmustberibbedand siliconised. 5. Material:High Qualityaustenitic stainlesssteelGrade AISI-302.			
1010420367	NEEDLE-40MM 1/2 CTC, PGA, USP 2	6000		
	Item name: Needle: NEEDLE-40MM, 1/2CIRCLE TAPERCUTTING-USP2 SiliconisedandRibbed. 1. Wire diameter:1.12 (-0.0+ 0.03)mm 2. Borediameter:0.60 (-0.01+ 0.025)mm 3. Suture material:PGA, USP2 (Braided) 4. Allneedlesmustbesiliconised and ribbed. 5. Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			
1010420368	NEEDLE-45 MM,3/8 CRC, NYLON-USP 0	75000		
	1.Wire diameter: 1(-0.0+0.025)mm 2. Bore diameter: 0.45(-0.01+0.02)mm 3. Suture material: Nylon USP 0 (Monofilament) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420			
1010420369	NEEDLE-48 MM,1/2 CTC, PGA-PLA USP2	1200		
	Specification: Item: 48 mm, 1/2 Circle Taper Cutting, Ribbed &siliconised. SUTURE SIZE: Polyglactin 910 USP 2			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Wire diameter: 1.20 (-0.0+0.03) mm Bore diameter: 0.72 (-0.01+0.025) mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420370	NEEDLE 48MM ½ CRC-PGA PLA 2	1200		
	Item name: Needle: NEEDLE-48mm, 1/2 circle round body. PGA-PLA USP 0 Siliconised and Ribbed. 1. Wire diameter: 1.12 (-0.0 + 0.03)mm 2. Bore diameter: 0.55 (-0.01 + 0.025)mm 3. Suture material: PGLA, USP 0(Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420371	NEEDLE 48MM ½ CRB-PGA PLA 0	1200		
	Item name: Needle: NEEDLE-48mm, 1/2 circle round body. PGA-PLA USP 0 Siliconised and Ribbed. 1. Wire diameter: 1.12 (-0.0 + 0.03)mm 2. Bore diameter: 0.55 (-0.01 + 0.025)mm 3. Suture material: PGLA, USP 0(Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420372	NEEDLE 45MM ½ CRC PGA PLA-1	1200		
	Item name: Needle: NEEDLE-45mm, 1/2 circle reverse cutting. PGA-PLA USP 1 Siliconised and Ribbed.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	1. Wire diameter: 1.12 (-0.0+0.03)mm 2. Bore diameter: 0.60 (-0.01+0.025)mm 3. Suture material: PGLA, USP 1 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AIST-420			
1010420373	NEEDLE 16MM ½ CRC PGA PLA-4-0	1200		
	Item name: Needle: NEEDLE-16mm, 1/2 circle reverse cutting. PGA-PLA USP 4/0 Siliconised and Ribbed. 1. Wire diameter: 0.56 (-0.0+0.02)mm 2. Bore diameter: 0.33 (-0.01+0.02)mm 3. Suture material: PGLA, USP 4/0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420			
1010420374	NEEDLE 31 MM ½ CRB PGA PLA-2-0	2400		
	Item name: Needle: NEEDLE-31mm, 1/2 circle round body. PGA-PLA USP 2/0 Siliconised and Ribbed. 1. Wire diameter: 0.80 (-0.0+0.02)mm 2. Bore diameter: 0.47 (-0.01+0.02)mm 3. Suture material: PGLA, USP 2/0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420			
1010420375	NEEDLE 31MM ½ CRC PGA PLA-2-0	12000		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Item name: Needle: NEEDLE-31mm, 1/2 circle reverse cutting. PGA-PLA USP 2/0 Siliconised and Ribbed. 1. Wire diameter: 0.80 (-0.0+0.02)mm 2. Bore diameter: 0.47 (-0.01+0.02)mm 3. Suture material: PGLA, USP 2/0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420			
1010420376	NEEDLE 26 MM ½ CRB PGA PLA 3-0	12000		
	Item name: Needle: NEEDLE-26mm, 1/2 circle round body. PGA-PLA USP 3/0 Siliconised and Ribbed. 1. Wire diameter: 0.66 (-0.0+0.02)mm 2. Bore diameter: 0.36 (-0.01+0.02)mm 3. Suture material: PGLA, USP 3/0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steelgrade AIST-420			
1010420377	NEEDLE 50 MM ½ CRC PGA PLA 0	1200		
	Item name: Needle: NEEDLE-50mm, 1/2 circle reverse cutting. PGA-PLA USP 0 Siliconised and Ribbed. 1. Wire diameter:1.12 (-0.0+0.03)mm 2. Bore diameter: 0.55 (-0.01+0.025)mm 3. Suture material: PGLA, USP 0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	stainless steelgrade AIST-420			
1010420378	NEEDLE 45MM ½ CRB PGA PLA -1 Item name: Needle: NEEDLE-45mm, 1/2 circle round body. PGA-PLA USP 1 Siliconised and Ribbed. 1. Wire diameter: 1.12 (-0.0+0.03)mm 2. Bore diameter: 0.60 (-0.01+0.025)mm 3. Suture material: PGLA, USP 1 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420	1200		
1010420379	NEEDLE-35 MM,3/8 CRC,RAPID 0,B55W1 "Needle: NEEDLE-35MM 3/8 CIRCLE REVERSE CUTTING, USP 0-ribbed & siliconised Wire diameter: 1.00 (-0.0+0.025mm) Bore diameter: 0.55 (-0.01+0.02mm) Suture mtl: PGA-PLA Rapid USP 0 Material: High Quality martensitic stainless steel grade AIST-420	1200		
1010420380	NEEDLE-23 MM,1/2 CTC,PDO, USP 0 "Needle: NEEDLE-23MM 1/2 CIRCLE TAPER CUTTING, USP 0-ribbed & siliconised Wire diameter: 0.81 (-0.0+0.02mm) Bore diameter: 0.55 (-0.01+0.02mm) Suture mtl: Polydioxanone USP 0 Material: High Quality martensitic stainless steel grade AIST-420	1200		
1010420381	NEEDLE-55MM,1/2 CTC,HEAVY,POLYESTER- 5 Needle: NEEDLE-55MM 1/2 CIRCLE TAPER CUTTING- POLYESTER USP	60000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	5-ribbed & siliconised Wire diameter: 1.3 (-0.0+0.03mm) Bore diameter: 0.90 (-0.01+0.02mm) Suture mtl: Polyester USP 5 Material: High Quality martensitic stainless steel grade AIST-420			
1010420382	NEEDLE-90 MM, 3/8 CCC,NYLON USP 1	1200		
	"Item name: Needle: NEEDLE-90 mm, 3/8 Circle Conventional Cutting,Nylon USP 1 Siliconised and Ribbed. 1. Wire diameter: 1.20 (-0.0+0.03)mm 2. Bore diameter: 0.53 (-0.01+0.025)mm 3. Suture material: Nylon, USP 1 (Monofilament) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420			
1010420383	NEEDLE-65MM,1/2CRB,PGA-1	1200		
	NEEDLE FOR PGA SUTURE(BRAIDED) 65 MM, 1/2 CIRCLE ROUND BODY SIZE:1 (65mm,1/2circle,round body(heavy))Wire diameter: 1.2+/-0.02mm Bore diameter: 0.55+/-0.02mm Material:High Quality martensitic stainless steel Grade AISI-420.			
1010420384	NEEDLE-30MM,1/2CRB,PGA-4/0	1200		
	Needle 30 MM,1/2 CIRCLE ROUND BODY Suture size:4/0 PGA Material:high quality Martensitic grade stainless steel			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Grade;AISI-420 CORROSION RESISTANT,Free from dirt,grease,fibresetc Needle type:Drilled end suitable for above said applications Bore dia 0.28 +/-0.02mm Wire dia : 0.48 to 0.51mm			
1010420385	NEEDLE-30MM,1/2CRC,NYLON 4/0	1200		
	Needle 30 MM ,1/2 CIRCLE REVERSE CUTTING Wire diameter: 0.54+/-0.02mm Bore diameter: 0.25+/-0.01mm Suture size: Monofilament Nylon Suture, USP 4-0 Material: High Quality martensitic Stainless Steel Grade AIST-420			
1010420386	NEEDLE-48MM,1/2 CRB,PGAPLA2-0	24000		
	"Item name: Needle: NEEDLE- 48MM, 1/2 CRB,PGLA 2-0 Siliconised and Ribbed. 1. Wire diameter: 1.12 (-0.0+0.03) mm 2. Bore diameter: 0.47 (-0.01+0.02) mm 3. Suture material: PGLA, USP 2-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AIST-420			
1010420387	NEEDLE-16MM,1/2 CRB,PGAPLA3-0	12000		
	Item name: Needle: NEEDLE- 16 MM, 1/2 CRB,PGLA 3-0 Siliconised and Ribbed. 1. Wire diameter: 0.61 (-0.0+0.02) mm 2. Bore diameter: 0.36 (-0.01+0.02) mm			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	3. Suture material: PGLA, USP 3-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420388	NEEDLE- 45MM, 1/2 CRB,PGA3-0	60000		
	Item name: Needle: NEEDLE- 45MM, 1/2 CRB,PGA3-0 Siliconised and Ribbed. 1. Wire diameter: 0.84 (-0.0+0.025) mm 2. Bore diameter: 0.33 (-0.01+0.02) mm 3. Suture material: PGA, USP 3-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel Grade AIST-420			
1010420389	NEEDLE-35MM, 3/8 CRC,HISIL 2-0	3000		
	tem name: Needle: NEEDLE-35MM, 3/8 CIRCLE REVERSE CUTTING- USP 2-0 Siliconised and Ribbed. 1. Wire diameter: 0.84 (-0.0+0.025)mm 2. Bore diameter: 0.42 (-0.01+0.02)mm 3. Suture material: Silk, USP 2-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420			
1010420390	NEEDLE-35MM,3/8 CRC,HISIL 0	3000		
	Item name: Needle: NEEDLE-35MM, 3/8 CIRCLE REVERSE CUTTING- USP 0 Siliconised and Ribbed. 1. Wire diameter: 0.92 (-0.0+0.025)mm			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	2. Bore diameter: 0.47 (-0.01+0.02)mm 3. Suture material: Silk, USP 0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality martensitic stainless steel grade AIST-420			
1010420391	NEEDLE 30MM-1/2 CC(PGCL) 2/0	6000		
	Specification: Needle: 30mm, 1/2 circle cutting Wire diameter: 0.84(-0.0+0.025)mm Bore diameter: 0.45(-0.01+0.02)mm Suture size: PGA-PCL, USP 2-0 Material: High Quality martensitic stainless steel Grde AIST-420			
1010420392	NEEDLE 30MM-1/2 CRC(PGCL) 2/0	6000		
	Specification: Needle: 30mm, 1/2 circle reverse cutting Wire diameter: 0.84(-0.0+0.025)mm Bore diameter: 0.45(-0.01+0.02)mm Suture size: PGA-PCL, USP 2-0 Material: High Quality martensitic stainless steel Grade AIST-420			
1010420393	NEEDLE 20MM-1/2 CC(PGCL) 4/0	2400		
	Specification: Needle: 20mm, 1/2 circle cutting Wire diameter: 0.66(-0.0+0.02)mm Bore diameter: 0.28(-0.01+0.02)mm Suture size: PGA-PCL, USP 4-0 Material: High Quality martensitic stainless steel Grde AIST-420			
1010420394	NEEDLE 30MM-1/2CRB(PGCL)3/0	6000		
	Specification: Needle: 30mm, 1/2 circle round body Wire diameter: 0.74(-0.0+0.02)mm Bore diameter: 0.36(-0.01+0.02)mm			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Suture size: PGA-PCL, USP 3-0 Material: High Quality martensitic stainless steel GradeAIST-420			
1010420396	NEEDLE 20MM-1/2 CRC(PGCL) 4/0	6000		
	Specification: Needle: 20mm, 1/2 circle reverse cutting Wire diameter: 0.66(-0.0+0.02)mm Bore diameter: 0.28(-0.01+0.02)mm Suture size: PGA-PCL, USP 4-0 Material: High Quality martensitic stainless steel GradeAIST-420			
1010420397	NEEDLE 30MM-1/2 CRB(PGCL) 4/0	6000		
	pecification: Needle: 30mm, 1/2 circle round body Wire diameter: 0.71(-0.0+0.02)mm Bore diameter: 0.28(-0.01+0.02)mm Suture size: PGA-PCL, USP 4-0 Material: High Quality martensitic stainless steel Grade AIST-420			
1010420398	NEEDLE 45MM-1/2 CRB(HIDOX)	1200		
	Specification: Needle: 45mm, 1/2 circle round body Wire diameter: 1.12 (-0.0+0.03)mm Bore diameter: 0.60 (-0.01+0.025)mm Suture size: PDO USP 1 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420399	NEEDLE 45MM-1/2 CRB(HIDOX) 2	12000		
	pecification: Needle: 45mm, 1/2 circle round body Wire diameter: 1.12 (-0.0+0.03)mm Bore diameter: 0.72 (-0.01+0.025)mm Suture size: PDO USP 2 Material: High Quality martensitic stainless steel Grade AISI-420.			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420400	NEEDLE 20MM-1/2 CRB(HIDOX)2/0 Specification: Needle: 20mm, 1/2 circle round body Wire diameter: 0.71(-0.0+0.02)mm Bore diameter: 0.45(-0.01+0.02)mm Suture size: PDO USP 2/0 Material: High Quality martensitic stainless steel Grade AISI-420.	6000		
1010420401	NEEDLE 26MM-1/2CRC(PP)2/0 Needle: NEEDLE-26MM 1/2 CIRCLE REVERSE CUTTING, USP 2-0-ribbed &siliconised Wire diameter: 0.71 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: Polypropylene USP 2-0 Material: High Quality martensitic stainless steel Grade AISI-420.	6000		
1010420402	NEEDLE 22MM-1/2CRB(PP)3/0 Needle: NEEDLE-22MM 1/2 CIRCLE ROUND BODY, USP 3-0-ribbed &siliconised Wire diameter: 0.61 (-0.0+0.02mm) Bore diameter: 0.28 (-0.01+0.02mm) Suture mtl: Polypropylene USP 3-0 Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420403	NEEDLE 22MM-1/2CRC(PP)3/0 Needle: NEEDLE-22MM 1/2 CIRCLE REVERSE CUTTING, USP 3-0-ribbed &siliconised Wire diameter: 0.61 (-0.0+0.02mm) Bore diameter: 0.28 (-0.01+0.02mm)	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Suture mtl: Polydioxanone USP 3-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420404	NEEDLE 20MM-1/2 CRC(HIDOX)3/0 E-20MM 1/2 CIRCLE REVERSE CUTTING, USP 3-0-ribbed &siliconised Wire diameter: 0.61 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: Polydioxanone USP 3-0 Material: High Quality martensitic stainless steel Grade AISI-420.	6000		
1010420405	NEEDLE 16MM-1/2 CRC(HIDOX)4/0 Needle: NEEDLE-16MM 1/2 CIRCLE REVERSE CUTTING, USP 4-0-ribbed &siliconised Wire diameter: 0.46 (-0.0+0.02mm) Bore diameter: 0.28 (-0.01+0.02mm) Suture mtl: Polydioxanone USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.	3000		
1010420406	NEEDLE 12MM-1/2 CRC(RAPID)5/0 Needle: NEEDLE-12MM 1/2 CIRCLE REVERSE CUTTING, USP 5-0-ribbed &siliconised Wire diameter: 0.38 (-0.0+0.02mm) Bore diameter: 0.28 (-0.01+0.02mm) Suture mtl: PGA-PLA RAPID USP 5-0 Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420407	NEEDLE 13MM-1/2CRC(PP)5/0 Needle: NEEDLE-13MM 1/2 CIRCLE REVERSE CUTTING, USP 5-0-ribbed &siliconised	18000		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Wire diameter: 0.38 (-0.0+0.02mm) Bore diameter: 0.19 (-0.01+0.02mm) Suture mtl: Polypropylene USP 5-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420408	NEEDLE 31MM-1/2CRB(PP)2/0	6000		
	Needle: NEEDLE-31MM 1/2 CIRCLE ROUND BODY, USP 2-0-ribbed &siliconised Wire diameter: 0.74 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: Polypropylene USP 2-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420409	NEEDLE 26MM-1/2CRB(PP)2/0	6000		
	Needle: NEEDLE-26MM 1/2 CIRCLE ROUND BODY, USP 2-0-ribbed &siliconised Wire diameter: 0.71 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: Polypropylene USP2-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420410	NEEDLE 16MM-1/2 CRB(PP)4/0	6000		
	Needle: NEEDLE-16MM 1/2 CIRCLE ROUND BODY, USP 4-0-ribbed &siliconised Wire diameter: 0.56 (-0.0+0.02mm) Bore diameter: 0.25 (-0.01+0.02mm) Suture mtl: Polypropylene USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420411	NEEDLE 13MM-1/2CRB(PP)5/0	12000		
	Needle: NEEDLE-13MM 1/2 CIRCLE			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	ROUND BODY, USP 5-0-ribbed &siliconised Wire diameter: 0.38 (-0.0+0.02mm) Bore diameter: 0.19 (-0.01+0.01mm) Suture mtl: Polypropylene USP 5-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420412	NEEDLE 30MM-1/2CRC(PGCL)3/0	1200		
	Needle: NEEDLE-30MM 1/2 CIRCLE REVERSE CUTTING, USP 3-0-ribbed &siliconised Wire diameter: 0.71 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: Polyglecaprone25 USP 3-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420413	NEEDLE 25MM-1/2CRC (PP)4/0	1200		
	Needle: NEEDLE-25MM 1/2 CIRCLE REVERSE CUTTING, USP 4-0-ribbed &siliconised Wire diameter: 0.51 (-0.0+0.02mm) Bore diameter: 0.25 (-0.01+0.02mm) Suture mtl: Polypropylene USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420414	NEEDLE 25 MM-1/2CRB(PP)4/0	6000		
	Needle: NEEDLE-25MM 1/2 CIRCLE ROUND BODY, USP 4-0-ribbed &siliconised Wire diameter: 0.51 (-0.0+0.02mm) Bore diameter: 0.25 (-0.01+0.02mm) Suture mtl: Polypropylene USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010420415	NEEDLE 26MM-1/2CRC(PGLA)3/0 Needle: NEEDLE-26MM 1/2 CIRCLE REVERSE CUTTING, USP 3-0-ribbed &siliconised Wire diameter: 0.71 (-0.0+0.02mm) Bore diameter: 0.36 (-0.01+0.02mm) Suture mtl: PGA-PLA USP 3-0 Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420416	NEEDLE- 26MM, 1/2 C TROCAR,PGLA2-0 Specification: SUTURE SIZE: Poly(glycolide-co-lactide) USP 2-0 Wire diameter: 0.84 (-0.0+0.025)mm Bore diameter: 0.47 (-0.01+0.02)mm Material: High Quality martensitic stainless steel Grade AISI-420.	24000		
1010420417	NEEDLE-40MM ,1/2C TROCAR,PGLA1 Specification: SUTURE SIZE: Poly(glycolide-co-lactide) USP 1 Wire diameter: 1.00 (-0.0+0.025)mm Bore diameter: 0.60 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	24000		
1010420418	NEEDLE-30MM,1/2C TROCAR,PGLA0 Specification: SUTURE SIZE: Poly(glycolide-co-lactide) USP 0 Wire diameter: 0.92 (-0.0+0.025)mm Bore diameter: 0.55 (-0.01+0.025)mm Material: High Quality martensitic	24000		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	stainless steel Grade AISI-420.			
1010420419	NEEDLE-40MM,1/2CTC,SHORT TIP,SS302,USP1	100000		
	""Specification: Item name: Needle: NEEDLE-40MM, 1/2 CIRCLE TAPER CUTTING, HEAVY-USP 1 with short taper cutting edge 1.Wire diameter: 1.00 (-0.0 + 0.025)mm 2. Bore diameter: 0.60 (-0.01 + 0.025)mm 3. Suture material: PGA-PLA, USP 1 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel G			
1010420420	NEEDLE-35MM,1/2CTC,SS302,USP2-0	1200		
	pecification: Item name: Needle: NEEDLE-35MM, 1/2 CIRCLE TAPER CUTTING-USP 2-0 1.Wire diameter: 0.88 (-0.0 + 0.02)mm 2. Bore diameter: 0.47 (-0.01 + 0.025)mm 3. Suture material: PGA-PLA, USP 2-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-302.			
1010420421	NEEDLE-31MM,1/2CRB,PGLA-0	6000		
	Specification: Item name: Needle: NEEDLE-31MM, 1/2 CIRCLE ROUND BODY-USP 1-0			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	1.Wire diameter: 0.84 (-0.0 + 0.025)mm 2. Bore diameter: 0.55 (-0.01 + 0.025)mm 3. Suture material: PGA-PLA, USP 1-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420422	NEEDLE-17MM,1/2CRB,PGLA5-0	1200		
	Specification: Item name: Needle: NEEDLE-17MM, 1/2 CIRCLE ROUND BODY -USP 5-0 1.Wire diameter: 0.51 (-0.0 + 0.02)mm 2. Bore diameter: 0.28 (-0.01 + 0.02)mm 3. Suture material: PGA-PLA, USP 5-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420423	NEEDLE-20MM,3/8CC,PGLA4-0	2400		
	Specification: Item name: Needle: NEEDLE-20MM, 3/8 CIRCLE CUTTING-USP 4-0 1.Wire diameter: 0.58 (-0.0 + 0.02)mm 2. Bore diameter: 0.33(-0.01 + 0.02)mm 3. Suture material: PGA-PLA, USP 4-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420424	NEEDLE-36MM,1/2CRB,PGLA-0	1200		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	<p>Specification: Item name: Needle: NEEDLE-36MM, 1/2 CIRCLE ROUND BODY -USP 1-0 1.Wire diameter: 1.00(-0.0 + 0.025)mm 2. Bore diameter: 0.55 (-0.01 + 0.025)mm 3. Suture material: PGA-PLA, USP 1-0 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.</p>			
1010420425	NEEDLE-40MM,1/2CC,PGLA-2	3000		
	<p>Specification: Item name: Needle: NEEDLE-40MM, 1/2 CIRCLE CUTTING-USP 2 1.Wire diameter: 1.20 (-0.0 + 0.03)mm 2. Bore diameter: 0.72 (-0.01 + 0.025)mm 3. Suture material: PGA-PLA, USP 2 (Braided) 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.</p>			
1010420426	NEEDLE-19MM,3/8CC,PGLA3-0	1200		
	<p>Specification: Item name: Needle: NEEDLE-19MM, 3/8 CIRCLE CUTTING-USP 3-0 1.Wire diameter: 0.62 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PGA-PLA, USP 3-0 (Braided) 4. All needles must be siliconised and ribbed.</p>			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420427	NEEDLE-40MM,1/2CRB,HILON2-0	12000		
	pecification: Item name: Needle: NEEDLE-40MM, 1/2 CIRCLE ROUND BODY-USP 2-0 1.Wire diameter: 1.00 (-0.0 + 0.025)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 2-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420428	NEEDLE-40MM,1/2CRC,HILON2-0	12000		
	Specification: Item name: Needle: NEEDLE-40 MM, 1/2 CIRCLE REVERSE CUTTING-USP 2-0 1.Wire diameter: 1.00 (-0.0 + 0.025)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 2-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420429	NEEDLE-36MM,1/2CRB,HILON3-0	6000		
	Specification: Item name: Needle: NEEDLE-36MM, 1/2 CIRCLE ROUND BODY-USP 3-0 1.Wire diameter: 0.86 (-0.0 + 0.025)mm 2. Bore diameter: 0.28 (-0.01 + 0.02)mm			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	3. Suture material: Nylon, USP 3-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420430	NEEDLE-36MM,1/2CRC,HILON3-0	6000		
	Specification: Item name: Needle: NEEDLE-36MM, 1/2 CIRCLE REVERSE CUTTING- USP 3-0 1.Wire diameter: 0.86 (-0.0 + 0.025)mm 2. Bore diameter: 0.28 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 3-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420431	NEEDLE-20MM,1/2CRB,HILON4-0	6000		
	Specification: Item name: Needle: NEEDLE-20MM, 1/2 CIRCLE ROUND BODY -USP 4-0 1.Wire diameter: 0.56 (-0.0 + 0.02)mm 2. Bore diameter: 0.25 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 4-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420432	NEEDLE-20MM,1/2CRC,HILON4-0	1200		
	Specification: Item name: Needle: NEEDLE-20MM, 1/2 CIRCLE REVERSE CUTTING- USP 4-0			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	1.Wire diameter: 0.56 (-0.0 + 0.02)mm 2. Bore diameter: 0.25 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 4-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420433	NEEDLE-20MM,1/2CRB,HILON5-0	2400		
	Specification: Item name: Needle: NEEDLE-20 MM, 1/2 CIRCLE ROUND BODY -USP 5-0 1.Wire diameter: 0.51 (-0.0 + 0.02)mm 2. Bore diameter: 0.19 (+/-0.01)mm 3. Suture material: Nylon, USP 5-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420434	NEEDLE-20MM,1/2CRC,HILON5-0	6000		
	Specification: Item name: Needle: NEEDLE-20 MM, 1/2 CIRCLE REVERSE CUTTING- USP 5-0 1.Wire diameter: 0.51 (-0.0 + 0.02)mm 2. Bore diameter: 0.19 (+/-0.01)mm 3. Suture material: Nylon, USP 5-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420435	NEEDLE-38MM,1/2CRC,HILON2-0	1200		
	Specification: Item name: Needle: NEEDLE-38MM, 1/2 CIRCLE REVERSE CUTTING-			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	USP 2-0 1.Wire diameter: 1.00 (-0.0 + 0.025)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 2-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.			
1010420436	NEEDLE-38MM,1/2CRC,HILON1-0 Specification: Item name: Needle: NEEDLE-38MM, 1/2 CIRCLE REVERSE CUTTING-USP 1-0 1.Wire diameter: 1.00 (-0.0 + 0.025)mm 2. Bore diameter: 0.45 (-0.01 + 0.02)mm 3. Suture material: Nylon, USP 1-0 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AIST-420	1200		
1010420437	NEEDLE-38MM,1/2CRC,HILON1 Specification: Item name: Needle: NEEDLE-38MM, 1/2 CIRCLE REVERSE CUTTING-USP 1 1.Wire diameter: 1.12 (-0.0 + 0.025)mm 2. Bore diameter: 0.53 (-0.01 + 0.025)mm 3. Suture material: Nylon, USP 1 4. All needles must be siliconised and ribbed. 5. Material: High Quality austenitic stainless steel Grade AISI-420.	1200		
1010420438	NEEDLE-40MM,1/2CRC,HILON2	2400		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Specification: Item: 40mm, 1/2 Circle Reverse Cutting, Ribbed & siliconised. SUTURE SIZE: Polyamide Nylon USP 2 Wire diameter: 1.12 (-0.0+0.03) mm Bore diameter: 0.60 (-0.01+0.025) mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420439	NEEDLE-45MM,1/2CRB,HILON2	1200		
	Specification: Item: 45MM, 1/2 Circle Round Body, Ribbed & siliconised. SUTURE SIZE: Polyamide Nylon USP 2 Wire diameter: 1.20 (-0.0+0.03) mm Bore diameter: 0.60 (-0.01+0.025) mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420440	NEEDLE-45MM,1/2CRC,HILON2	1200		
	Specification: Item: 45MM, 1/2 Circle Reverse cutting, Ribbed & siliconised. SUTURE SIZE: Polyamide Nylon USP 2 Wire diameter: 1.20 (-0.0+0.03) mm Bore diameter: 0.60 (-0.01+0.025) mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420441	NEEDLE-45MM,1/2CRC,PGLA -0	6000		
	Specification: Item: 45mm, 1/2 circle Reverse cutting. SUTURE SIZE: Polyglactin 910 USP 0 Wire diameter: 1.00 (-0.0+0.025) mm Bore diameter: 0.55 (-0.01+0.025)mm Material: High Quality martensitic			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	stainless steel Grade AISI-420.			
1010420442	NEEDLE-30MM,1/2CRB,PGCL-0	1200		
	Specification: Item: 30mm, ½ circle round body. SUTURE SIZE: Polyglecaprone 25 USP 0 Wire diameter: 0.92 (-0.0+0.025) mm Bore diameter: 0.55 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420443	NEEDLE-36MM,1/2CRB,HILON2-0	3000		
	Specification: Item: 36mm, ½ circle round body. SUTURE SIZE: Nylon USP 2-0 Wire diameter: 0.88 (-0.0+0.025) mm Bore diameter: 0.36 (-0.01+0.02)mm Material: High Quality martensitic stainless steel Grade AIST-420			
1010420444	NEEDLE-36MM,1/2CRC,HILON2-0	3000		
	Specification: Item: 36mm, ½ circle reverse cutting. SUTURE SIZE: Nylon USP 2-0 Wire diameter: 0.88 (-0.0+0.025) mm Bore diameter: 0.36 (-0.01+0.02)mm Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420448	NEEDLE-48MM-1/2 CRB,RAPID-1	6000		
	Specification: Item: 48mm, ½ circle Round Body. SUTURE SIZE: Polyglactin 910 Undyed USP 1 Wire diameter: 1.20 (-0.0+0.025) mm Bore diameter: 0.60 (-0.01+0.025)mm Material: High Quality martensitic			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	stainless steel Grade AISI-420.			
1010420449	NEEDLE-48MM-1/2 CRC,RAPID-1 Specification: Item: 48mm, 1/2 circle Reverse cutting. SUTURE SIZE: Polyglactin 910 Undyed USP 1 Wire diameter: 1.20 (-0.0+0.025) mm Bore diameter: 0.60 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	3000		
1010420450	NEEDLE-48MM-1/2 CRB,PGCL-1 Specification: Item: 48mm, 1/2 circle Round body. SUTURE SIZE: Polyglecaprone 25 Undyed USP 1 Wire diameter: 1.20 (-0.0+0.025) mm Bore diameter: 0.60 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	1200		
1010420451	NEEDLE-48MM-1/2 CRB,PGCL-2 Specification: Item: 48mm, 1/2 circle Round body. SUTURE SIZE: Polyglecaprone 25 Undyed USP 2 Wire diameter: 1.20 (-0.0+0.025) mm Bore diameter: 0.72 (-0.01+0.025)mm Material: High Quality martensitic stainless steel Grade AISI-420.	3000		
1010420452	NEEDLE_3/8 CIRCLE RC 22MM PGLA 4-0 ""Specification: Needle: 22mm, 3/8 circle reverse cutting	1200		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.

तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Wire diameter: 0.61 (-0.0+0.02mm) Bore diameter: 0.33mm (-0.01+0.02) Suture mtl: Polyglactin 910 USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420453	NEEDLE 13MM 1/2 CRB VISI BLACK PP 5-0	1200		
	Specification: Needle: NEEDLE 13MM1/2 CRB VISI BLACK Wire diameter: 0.46mm (-0.0+0.02) Bore diameter: 0.19mm (-0.01+0.02) Suture mtl: Polyglactin 910 USP 2-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010420454	NEEDLE 30MM-3/8CRC PA2-0	1200		
	Specification: Needle: 30MM-3/8 circle reverse cutting Wire diameter: 0.99mm (-0.0+0.02) Bore diameter: 0.36mm (-0.01+0.02) Suture mtl: Polyamide USP 2-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010430008	36MM, 3/8CIRCLE REVERSE CUTTING	1200		
	Specification: Needle:36MM, 3/8CIRCLE REVERSECUTTING-ribbed& siliconised Wire diameter:1.00 +/- 0.01 mm Borediameter:0.60+/- 0.01 mm Suture mtl:Polyglactin910 USP1 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
1010430009	36MM,1/2CIRCLE TAPER CUTTING Specification: Needle:36MM,1/2CIRCLE TAPER CUTTING-ribbed&siliconised Wire diameter:0.84 +/- 0.01 mm Borediameter:0.55+/- 0.01 mm Suture mtl:Polyglactin910 USP0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	3000		
1010430010	35MM,1/2CIRCLE TAPER CUTTING Specification: Needle:35MM,1/2CIRCLE TAPER CUTTING-ribbed&siliconised Wire diameter:0.80 +/- 0.01 mm Borediameter:0.47+/- 0.01 mm Suture mtl:Polyglactin910 USP2-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420	75000		
1010430011	16MM,3/8CIRCLE CUTTING Specification: Needle:16MM,3/8CIRCLE CUTTING-ribbed&siliconised Wire diameter:0.54 +/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:Polyglactin910 USP3-0 Material:High Qualitymartensitic stainlesssteelGrade AISI-420.	6000		
1010430012	22MM, 1/2CIRCLE CUTTING Specification: Needle:22MM, 1/2CIRCLE CUTTING-ribbed&siliconised Wire diameter:0.62 +/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:Polyglactin910 USP3-0 Material:High Qualitymartensitic	6000		

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	stainlesssteelGrade AISI-420.			
1010430013	26MM, 3/8CIRCLE REVERSE CUTTING	24000		
	Specification: Needle:26MM, 3/8CIRCLE REVERSE CUTTING-ribbed& siliconised Wire diameter:0.62 +/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:Polyglactin910 USP3-0 Material:High Quality martensitic stainlesssteelGrade AISI-420.			
1010430014	26MM, 3/8CIRCLE ROUND BODY	12000		
	Specification: Needle:26MM, 3/8CIRCLE ROUND BODY-ribbed&siliconised Wire diameter:0.62 +/- 0.01 mm Borediameter:0.36+/- 0.01 mm Suture mtl:Polyglactin910 USP3-0 Material:High Quality martensitic stainlesssteelGrade AISI-420.			
1010430015	16MM, 3/8CIRCLE CUTTING	24000		
	Specification: Needle: 16MM, 3/8CIRCLE CUTTING-ribbed &siliconised Wire diameter: 0.48 +/- 0.01 mm Bore diameter: 0.33+/- 0.01 mm Suture mtl: Polyglactin 910 USP 4-0 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010430017	45MM, 1/2CIRCLE, ROUND BODY	3000		
	Specification: Needle: 45MM, 1/2CIRCLE ROUND BODY-ribbed &siliconised Wire diameter: 1.2 (-0.0mm +0.03mm) Bore diameter: 0.60 (-0.01 mm +0.02mm)			

एचएलएल लाइफकेयर लिमिटेड /HLL LIFECARE LIMITED
(भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
		Unit: pcs	(Specify Currency)	
	Suture mtl: Polyglycolicacid USP 2 Material: High Quality martensitic stainless steel Grade AISI-420.			
1010430028	NEEDLE 45MM,1/2 CRB PGA1 Specification: Item name: Needle: NEEDLE-45MM, 1/2 CIRCLE ROUND BODY, PGA-USP 1 Ribbed &siliconised. 1.Wire diameter: 1.12(-0.0 + 0.03)mm 2. Bore diameter: 0.55(-0.01 + 0.025)mm 3. Suture material: PGA USP 1 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	60000		
1010430029	NEEDLE 45MM,1/2 CRC PGA2-0 Specification: Item name: Needle: NEEDLE-45MM, 1/2 CIRCLE REVERSE CUTTING, PGA-USP 2-0 Ribbed &siliconised. 1.Wire diameter: 0.92(-0.0+0.025)mm 2. Bore diameter: 0.42(-0.01+0.02)mm 3. Suture material: PGA USP 2-0 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.	24000		
1010430030	NEEDLE 45MM,1/2 CRB PGA2-0 Specification: Item name: Needle: NEEDLE-45MM, 1/2 CIRCLE ROUND BODY, PGA-USP 2-0 Ribbed &siliconised. 1.Wire diameter: 0.92(-0.0+0.025)mm 2. Bore diameter: 0.42(-0.01+0.02)mm 3. Suture material: PGA USP 2-0	60000		

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	(Braided) 4. All needles must be siliconised & ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010430031	NEEDLE 48MM,1/2 CRB PGA1	1200		
	Specification: Item name: Needle: NEEDLE-48MM, 1/2 CIRCLE ROUND BODY, PGA-USP 1 Ribbed & siliconised. 1. Wire diameter: 1.12(-0.0 + 0.03)mm 2. Bore diameter: 0.55(-0.01 + 0.025)mm 3. Suture material: PGA USP 1 (Braided) 4. All needles must be siliconised & ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010430032	NEEDLE 48MM,1/2 CRB PGA2	1200		
	Specification: Item name: Needle: NEEDLE-48MM, 1/2 CIRCLE ROUND BODY, PGA-USP 2 Ribbed & siliconised. 1. Wire diameter: 1.2(-0.0+0.03)mm 2. Bore diameter: 0.60(-0.01 + 0.025)mm 3. Suture material: PGA USP 2 (Braided) 4. All needles must be siliconised & ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010430033	NEEDLE 48MM,3/8 CRC PGA2	1200		
	Specification: Item name: Needle: NEEDLE-48MM, 3/8 CIRCLE REVERSE CUTTING, PGA-USP 2 Ribbed & siliconised. 1. Wire diameter: 1.2(-0.0+0.03)mm			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	2. Bore diameter: 0.60(-0.01 + 0.025)mm 3. Suture material: PGA USP 2 (Braided) 4. All needles must be siliconised&ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010430035	NEEDLE-45MM,1/2CRB,PGLA-2	60000		
	Specification: Item name: Needle: NEEDLE-45MM, 1/2 CIRCLE ROUND BODY,-USP 2 Ribbed &siliconised. 1.Wire diameter: 1.20 (-0.0+0.03)mm 2. Bore diameter: 0.72 (-0.01+0.025)mm 3. Suture material: PGA-PLA, USP 2 (Braided) 4. All needles must be siliconised&ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
1010430037	NEEDLE 26MM-1/2CRC(PGLA)3/0	1200		
	Specification: Item name:NEEDLE-26MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.74 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PGLA Braided 4. All needles must be siliconised&ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
1010430041	NEEDLE 16MM-1/2 CRC(HIDOX)4/0	1200		
	Specification:			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	Item name:NEEDLE-16 MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.56 (-0.0 + 0.02)mm 2. Bore diameter: 0.28 (-0.01 + 0.02)mm 3. Suture material: PDO Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
1010430042	NEEDLE 12MM-1/2 CRC(RAPID)5/0	1200		
	Specification: Item name:NEEDLE-12 MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.48 (-0.0 + 0.02)mm 2. Bore diameter: 0.28 (-0.01 + 0.02)mm 3. Suture material: PGLA RAPID Braided 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
1010430043	NEEDLE 13MM-1/2CRC(PP)5/0	1200		
	Specification: Item name:NEEDLE-13 MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.48 (-0.0 + 0.02)mm 2. Bore diameter: 0.19 (-0.01 + 0.02)mm 3. Suture material: PP Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
1010430044	NEEDLE 31MM-1/2CRB(PP)2/0	1200		
	Specification: Item name:NEEDLE-31MM, 1/2			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	CIRCLE ROUND BODY, Siliconised. 1.Wire diameter: 0.81 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PP Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
1010430045	NEEDLE 26MM-1/2CRB(PP)2/0	1200		
	Specification: Item name:NEEDLE-26MM, 1/2 CIRCLE ROUND BODY, Siliconised. 1.Wire diameter: 0.74 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PP Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
1010430048	NEEDLE 30MM-1/2CRC(PGCL)3/0	1200		
	Specification: Item name:NEEDLE-30MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.81 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PGCL Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
1010430049	NEEDLE 25MM-1/2CRC (PP)4/0	1200		
	Specification: Item name:NEEDLE-25MM, 1/2 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.60 (-0.0 + 0.02)mm			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materials@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	2. Bore diameter: 0.25 (-0.01 + 0.02)mm 3. Suture material: PP Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
1010430050	NEEDLE 25 MM-1/2CRB(PP)4/0	1200		
	Specification: Item name:NEEDLE-25MM, 1/2 CIRCLE ROUND BODY, Siliconised. 1.Wire diameter: 0.60 (-0.0 + 0.02)mm 2. Bore diameter: 0.25 (-0.01 + 0.02)mm 3. Suture material: PP Monofilament 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
1010430051	NEEDLE-3/8 CIRCLE RC 26MM PGLA2-0	1200		
	Specification: Item name:NEEDLE-26MM, 3/8 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.71 (-0.0 + 0.02)mm 2. Bore diameter: 0.47 (-0.01 + 0.02)mm 3. Suture material: PGA-PLA,USP 2-0 (Braided) 4. All needles must be siliconised& ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
1010430052	NEEDLE-3/8 CIRCLE RC 22MM PGLA3-0	1200		
	Specification: Item name:NEEDLE-22M, 3/8 CIRCLE REVERSE CUTTING, Siliconised. 1.Wire diameter: 0.62 (-0.0 + 0.02)mm			

एचएलएल लाइफ़केयर लिमिटेड /HLL LIFECARE LIMITED
 (भारत सरकार का उद्यम) / (A GOVT.OF INDIA ENTERPRISE)
आक्कुलम फैक्टरी , श्रीकार्यम पी .ओ / AKKULAM FACTORY,SREEKARIYAM P.O.
तिरुवन्नतपुरम -695 017 / THIRUVANANTHAPURAM-695 017

फोन नं : 0471 2442642 फाक्स नं : 0471 2441383 / PHONE NO:0471 2442642; FAX NO:0471 2441383;
 EMAIL: materialsft@lifecarehll.com

TENDER NO. HLL/AFT/MTLS/SUT/NEEDLES/2018-20

Dated 16.02.2018

Material Number	Material Description	Quantity	Rate / Unit	Amount
	Specify HSN Code:	Unit: pcs	(Specify Currency)	
	2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PGA-PLA, USP 3-0 (Braided) 4. All needles must be siliconised & ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420			
10104203499	NEEDLE-26MM, 1/2 CC, PGA-PLA USP 3/0	1200		
	Specification: Item name:NEEDLE-26MM, 1/2 CIRCLE CUTTING, Siliconised. 1.Wire diameter: 0.56 (-0.0 + 0.02)mm 2. Bore diameter: 0.36 (-0.01 + 0.02)mm 3. Suture material: PGA-PLA, USP 3-0 (Braided) 4. All needles must be siliconised & ribbed. 5. Material: High Quality martensitic stainless steel Grade AISI-420.			
PLACE: DATE:	NAME & SIGNATURE OF THE APPLICANT (WITH OFFICE SEAL)			

Minimum time required to start the supply after receipt of letter of indent/supply orders }

Validity : **One Year (April 2018 to March 2019)**

Certified that the rate quoted will hold good for one year during which period no upward revision will be sought.

PLACE:
DATE:

NAME & SIGNATURE OF THE APPLICANT
(WITH OFFICE SEAL)