

nne pharmaplan®	<b>User Requirement Specifications</b>				
	<b>Equipment/System</b>	Ultra Filtration System			
	<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01	
	<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07	

# User Requirement Specifications

## Ultra Filtration System

PROCESS CODE	AREA	EQUIPMENT CODE	QTY(NOS)
D	Diphtheria	D-UFS 01	1

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	<b>User Requirement Specifications</b>				
	<b>Equipment/System</b>	Ultra Filtration System			
	<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01	
	<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07	

### Table of Contents

<b>1.0</b>	<b>APPROVAL SIGNATURE.....</b>	<b>4</b>
<b>2.0</b>	<b>EQUIPMENT DESCRIPTION.....</b>	<b>5</b>
2.1	OPERATING CONDITIONS: .....	5
2.2	SYSTEM SPECIFICATIONS.....	5
2.3	VESSEL SPECIFICATIONS .....	6
2.4	VESSEL DESIGN SPECIFICATION.....	6
<b>3.0</b>	<b>PROCESS DESCRIPTION.....</b>	<b>10</b>
3.1	INPUT & CHARGING METHOD .....	10
3.2	BRIEF PROCESS STEPS .....	10
3.3	OUTPUT & DISCHARGING METHOD.....	10
<b>4.0</b>	<b>PRODUCTIVITY REQUIREMENT .....</b>	<b>10</b>
4.1	DESIRED/ SUGGESTED CAPACITY.....	10
4.2	STANDARD BATCH SIZE .....	10
4.3	CHANGE OVER TIME.....	10
4.4	OTHER (IF ANY) .....	10
<b>5.0</b>	<b>CONTAINMENT .....</b>	<b>11</b>
<b>6.0</b>	<b>GMP REQUIREMENTS .....</b>	<b>11</b>
6.1	PROCESS CONTROL.....	11
6.2	FAILURE MODE DETECTION.....	11
6.3	IN –PROCESS CONTROL.....	12
6.4	LEVEL OF INSTRUMENTATION.....	12
6.5	BATCH DATA DISPLAY AND RECORD PRINTING.....	12
6.6	GMP REQUIREMENTS (OTHERS).....	12
6.7	SPECIFIC REQUIREMENTS.....	13
<b>7.0</b>	<b>CONSTRAINTS .....</b>	<b>14</b>
7.1	EQUIPMENT LOCATION AND AVAILABLE SPACE .....	14
7.2	AVAILABLE UTILITY .....	14
<b>8.0</b>	<b>ABBREVIATION .....</b>	<b>15</b>

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	<b>User Requirement Specifications</b>				
	<b>Equipment/System</b>	Ultra Filtration System			
	<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01	
	<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07	

### URS Annexure List

URS Annex No.	Detail
1.	Layout showing the location of the Ultrafiltration System in the concentration room
2.	Tentative P& ID for Ultrafiltration System

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	<b>User Requirement Specifications</b>				
	<b>Equipment/System</b>	Ultra Filtration System			
	<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01	
	<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07	

### 1.0 APPROVAL SIGNATURE

This document is prepared by the Process, Validation and GMP Compliance team of "NNE Pharmaplan India" for the project "Revival of DPT Vaccines manufacturing Facility" (**Project number:-110831**) of Pasteur Institute of India, Coonoor under the authority of their Project Manager. Hence, this document before being effective shall be approved by the QA team of Pasteur Institute of India, and authorized by the appropriate Project Authority.

Prepared by		
Name/ Designation	Signature	Date
<b>Ms. Shilpa Rao</b> Sr. Project Engineer-Biotech NNE Pharmaplan India Ltd.		

Checked by		
Name/ Designation	Signature	Date
<b>Mr. K. Sridhar Babu</b> Assistant Manager- Validation & GMP Compliance NNE Pharmaplan India Ltd.		

Approved by		
Name/ Designation	Signature	Date
<b>Mr. Vikas Katial</b> GM and Head-COC Vaccines NNE Pharmaplan India Ltd.		
<b>HLL Lifecare Limited</b>		
<b>Pasteur Institute of India</b>		

Authorized by		
Name/ Designation	Signature	Date
<b>Project Authority</b> Pasteur Institute of India		

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	<b>User Requirement Specifications</b>				
	<b>Equipment/System</b>	Ultra Filtration System			
	<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01	
	<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07	

### 2.0 EQUIPMENT DESCRIPTION

The equipment described by this URS is an “**Ultra Filtration System**” (**Quantity- 1NO.**). This system will be used for the concentration of Diphtheria Toxoid (non-live material) after detoxification of the toxin (live material).

#### 2.1 Operating Conditions:

- Pressure bar :0 - 4 bar
- Flow range: vendor to specify
- Temperature range: 0 - 40 °C during process
- Surface area of the membrane: 5 m<sup>2</sup> with a provision to increase the filter area to 10m<sup>2</sup> (size of modules)
- Minimum working Volume: 0.01% of Total Volume

#### 2.2 System Specifications

**TABLE 1**

S. No.	Description	Purpose	MOC	Capacity/Size
1.	System frame	To hold cassette holder	SS304	Vendor to specify
2.	Cassette holder	To hold filter cassettes	SS316L	5 m <sup>2</sup> with a provision to increase the filter area to 10m <sup>2</sup> .
3.	Recirculation pump	To recirculate the permeate and retentate	Sanitary type-SS316L	Vendor to specify according to membrane area
4.	Flow path	For the circulation of feed, permeate and retentate	SS316L	Vendor to specify
5.	Pressure transmitter in the feed, permeate and retentate lines	To measure pressure differential (ΔP) and Trans membrane pressure (TMP)	Sanitary type with SS316L diaphragm	NA
6.	Surface Finish	System should be Internally Electro polished and passivated Ra≤0.6 μm, according to ASME BPE guidelines(2009)		
		Externally Mechanically polished and passivated up to Ra<1.2 μm		
		Stainless steel piping interior Ra≤ 0.6 μm, , according to ASME BPE guidelines (2009)		

System should be provided with rotary lobe pump proven to handle shear sensitive cell with suitable flow rate and pressure with variable speed control and auto cut off of the pump at dry condition of the tank.

All pressurized parts of the system should be piped and connected via sanitary connections

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®

### User Requirement Specifications

<b>Equipment/System</b>	Ultra Filtration System		
<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01
<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07



## 2.3 Vessel Specifications

**TABLE 2**

S. No.	Description	Purpose	MOC
1.	Shell	Cylindrical	SS316L
2.	Top closure	Flat Lid	SS316L
3.	Bottom closure	Torispherical dish	SS316L
4.	Spray Ball	For the cleaning of the interior of the vessel.	SS 316L
5.	Cladding	Cladding is welded around the insulation (mineral wool)	SS304

## 2.4 Vessel design specification

**TABLE 3**

SI.NO	Description	Specification
1.	Maximum working volume	150L
2.	Quantity	1 No
3.	Working temperature range	25 °C-134 °C
4.	Surface Finish	Internally Electro polished Ra ≤ 0.6 μm, according to ASME BPE
		Externally Mechanically polished up to Ra <1.2μm mirror finish.
		Stainless steel piping interior Ra≤ 0.6μm, according to ASME BPE

### 2.4.a The Chassis mounted system (with castor wheel) comprising of One Ultrafiltration unit, Semi-Automatic System and feed tank. The following main features:

**A. Dosing unit for feed:** The feed (from Nalgene bottles of 50 L) shall be fed into the feed tank of the filtration unit using the feed pump.

#### **B. Air Filters**

- Inlet Air filters:
  - Reusable and Sterilizable SS housing with 0.2/0.22 micron sterile filter (Code 7) with manual diaphragm valve, which is to be sterilized along with vessel.

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	User Requirement Specifications				
	Equipment/System	Ultra Filtration System			
	Identification #	D-UFS 01	Document#	URS/D-UFS 01	
	Effective Date	2013-12-18	Revision#	07	

**C. Pressure Indication:**

Pressure of the vessel

- Compound Pressure gauge for vessel
- Rupture disc, to release the excess pressure in the vessel

**D. Flush bottom valve:** It should be zero dead leg type valve attached directly to the bottom of the vessel. The diaphragm shall be of PTFE type.

**E. Feed Line:**

- Pressure transmitter

**F. Permeate Line:**

- Pressure transmitter
- Electromagnetic Flow Measuring System for measuring the flow of conductive fluids in process applications
- Sampling valve

**G. Retentate Line:**

- Pressure transmitter
- Electromagnetic Flow Measuring System for measuring the flow of conductive fluids in process applications
- Retentate Sampling
- Control Valve for  $\Delta P$  and TMP

**H. CIP(Clean in Place):**

- Manual CIP of the system must be made possible

**I. General characteristics of the Ultrafiltration membrane:**

- Filter area of 5 m<sup>2</sup> with a provision to increase the filter area to 10m<sup>2</sup>
- Membrane with 30 KD Molecular Weight cut off.
- Membranes to be CIPable only
- Hydrophilic in nature.
- High velocity and high particulate level capability.
- Maximum containment of hazardous fluids.
- Low hold-up volume
- Low protein binding.
- Good compatibility with most of the cleaning, sanitizing, depyrogenation and storage agents
- Void free composite PES membrane with 0.5M caustic compatibility.

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	User Requirement Specifications				
	Equipment/System	Ultra Filtration System			
	Identification #	D-UFS 01	Document#	URS/D-UFS 01	
	Effective Date	2013-12-18	Revision#	07	

- Capable of withstanding 100 psi forward pressure at 25degC.
- Should meet USP class VI biological tests invivo

**Note:** membrane made up of equivalent polymer with above characteristics is optimal for use

**J. Additional requirements:**

Calibration of measuring instruments according to international standards, full-loop calibration is required for GMP and quality relevant instruments. The Supplier has to provide calibration protocols and guidelines for writing SOPs for recalibration.

**K. Controller:** PLC Based Controller (Semi- Automatic operation) with a 10” industrial touch screen large HMI (Displaying data trends as Graphs, synoptic view of running parameters etc).

**L.** The HMI shall be touch screen type (Provision for manual operation also to be provided)

- HMI screen size shall be of 10” with resolution of 1200 x 800 pixels
- Human machine interface must be used to enter the process details, which should appear in the print out.
- All critical alarms
- All critical parameters & interlocks
- Addition of the buffer
- All Recipes/ sequences (process, CIP, transfer etc)
- P&ID of the vessel along with instrumentation details
- Login details
- HMI screen shots shall be available

**M. Nozzles Schedule :**

**Top Head Plate (design of shell to be provided by the vendor)**

- GMP type Spray ball assembly with 360o spray (design of the same shall be submitted by the vendor)
- Sterilizing grade hydrophobic vent filter (0.2/0.22µ filter) with SS housing and manual diaphragm valve with connection to drain line
- Rupture disc – NA connector with connection the drain line
- Port for Light/Sight Glass – Bolted with gasket.
- Port for Spare port- TC clamps with gasket
- Port for Pressure gauge
- Port for inlet of retentate with “J” type nozzle
- Port for level sensor with accuracy of ± approx. 0.1% of the total range

**Upper wall side:**

- Port for the addition of feed/ buffer with peristaltic pump
- Permeate recirculation port with “J” type nozzle

**Lower wall/Bottom connections:**

- Port for temperature transmitter

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	<b>User Requirement Specifications</b>				
	<b>Equipment/System</b>	Ultra Filtration System			
	<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01	
	<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07	

- Port for conductivity meter

**N. All points of the IRS except the below mentioned would be applicable for the equipment**

- 4.1.10, 4.1.11, 4.1.13, 4.1.17
- FDA Guidance for industry, CE Certification,
- ANSI/NSF 49-2008, ISO 14664, ISO 8362

**Note:**

I.	This Technical Specification is the basis for an inquiry to a vendor and therefore the basis for the vendor's proposal.
II.	The vendor is asked to state in "REMARKS" column with "yes" if the described requirement will be completely fulfilled and with "no" in case the requirement will not or cannot be fulfilled with the proposed equipment. In case of a deviation a comment must be inserted or enclosed as a separate annexure by referring to the respective URS specification number.
III.	The vendor must clearly comment each item of the Technical Specification. The comments must be in English language. If extra cost for options become necessary the item must be clearly stated.
IV.	In case that the requirement includes a question or request or information from the vendor, the answer / information should be stated in the "REMARKS" column.
V.	The final version of this document including the vendor's comments will become basis of a potential purchase order or contract.
VI.	The Technical Specification serves to define a summary of all vendor's requirements concerning scope of delivery and services.
VII.	The vendor is responsible for technically unobjectionable function of the equipment. This TS is not intended to dictate a technical design to the vendor. If agreed upon with the vendor, the vendor can apply his practically proven design.
VIII.	Special Instruction a. If no comments against any specification shall be considered as "NO" and b. If there is no reply / comments against the complete URS by the vendor then it shall be treated as unresponsive / technically non-compliant and rejected.
IX.	All the instruments and controls mentioned in the URS(s) are expected to be standard supply and part of your standard equipment model. In case of any deviation or redundancy or additional scope of supply is noticed, vendor is required to obtain clarification from HLL before submitting the quotes.
X.	The makes requested are standard international makes. In case of any deviation, vendor to seek clarification from HLL before submitting the offers.
XI.	Refer document Installation Requirement Specification and Specific Instructions with URS; NPI_110831_IRS_PII_01
XII.	Refer Tender document with URS; NPI/110831/EQP/TD/03

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	<b>User Requirement Specifications</b>				
	<b>Equipment/System</b>	Ultra Filtration System			
	<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01	
	<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07	

Specifications	Remarks
<b>3.0 PROCESS DESCRIPTION</b>	
<b>3.1 Input &amp; Charging method</b>	
3.1.1 The feed (from Nalgene bottles of 50 L) shall be fed into the feed tank of the filtration unit using the feed pump.	
<b>3.2 Brief Process Steps</b>	
<p><b>3.2.1 The equipment will be used for concentration of toxoid</b></p> <p>a) The toxoid shall be directly fed into feed tank with peristaltic pump.</p> <p>b) The flow-rate of the toxoid into the system and the feed pump shall be controlled by a control panel.</p> <p><b>(Basic Requirement: Concentration factor: 100X)</b></p> <p>The required number of filtration modules shall be installed as per the required filtration area in single or in multiple membrane cassettes.</p> <p>c) Permeate and retentate will be recirculated into the vessel until the equilibrium is achieved.</p> <p>d) After achieving the equilibrium, permeate is collected separately and retentate shall be recirculated in the vessel</p> <p>e) System to be designed to process simultaneously product feed inlet and concentration operation.</p> <p>f) A separate provision to be made for product recovery to flush the module.</p>	
<b>3.3 Output &amp; Discharging method</b>	
a) After diafiltration, the retentate (50 L) shall be collected in the 100 L (WV) mobile pressure vessel which will serve as feed tank to sterile filtration system.	
<b>4.0 PRODUCTIVITY REQUIREMENT</b>	
<b>4.1 Desired/ suggested capacity</b>	
a) Ultrafiltration Poly Ether Sulphone membrane to handle toxoid fermentation broth with low protein binding.	
b) Feed Vessel: 5L (Minimum) and 150L (Maximum working Volume)	
<b>4.2 Standard batch size</b>	
Infeed: 16 nalgene bottles of 50L (G.V.)	
<b>4.3 Change over time</b>	
Not Applicable	
<b>4.4 Other (If any)</b>	
Not Applicable	

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	<b>User Requirement Specifications</b>				
	<b>Equipment/System</b>	Ultra Filtration System			
	<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01	
	<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07	

Specifications	Remarks
<b>5.0 CONTAINMENT</b>	
Not Applicable	
<b>6.0 GMP REQUIREMENTS</b>	
<b>6.1 Process control</b>	
6.1.1 The Ultrafiltration unit should essentially have the necessary provisions for adjustment / control of the following critical process parameters:	
6.1.2 Following parameters shall be controlled by the equipment	
a) Pressure	
b) Trans membrane pressure (TMP), differential pressure ( $\Delta P$ )	
c) Temperature of the product	
d) Duration of the cycle	
e) The drain valve position and control.	
f) Flow of the permeate	
g) Parameters during CIP	
h) Conductivity	
i) NWP stabilization and measurement time	
j) Variable frequency drive	
k) Level of the product	
6.1.3 Following conditions need only notification to operator for procedural control:	
l) Emergency stop activated.	
m) Power failure.	
n) Malfunction of sensors of temperature, pressure, flow and conductivity	
<b>6.2 Failure mode detection</b>	
6.2.1 <b>Equipment shall be capable to detect the following failure, notify the operator with alarm and shutdown the process:</b>	
a) Pressure level low / high	
b) Alarm in case of not reached pre-set value. (TMP, $\Delta P$ )	
c) Alarm is activated in case Temperature is out of range	
<b>File Name</b>	NPI_110831_EQP_URS_B1(D)-UFS 01.docx
<b>Page No.</b>	Page 11 of 17

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	<b>User Requirement Specifications</b>				
	<b>Equipment/System</b>	Ultra Filtration System			
	<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01	
	<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07	

Specifications	Remarks
----------------	---------

d) Emergency stop activated	
-----------------------------	--

<b>6.3 In –Process control</b>	
--------------------------------	--

6.3.1 TMP, ΔP	
---------------	--

6.3.2 Temperature measurement and control of the product.	
---	--

6.3.3 Flow measurement on the feed inlet line	
---	--

6.3.4 Measurement and control of fluid level in the tank.	
---	--

<b>6.4 Level of instrumentation</b>	
-------------------------------------	--

Sufficient and suitable instrumentation for the process, safety and productivity control as indicated in the following table:

Parameters	Purpose	Type of control and Instrumentation
Temperature	Monitor and control the temperature	RTD sensor and temperature indicator & controller on the tank
Pressure	Monitor and control the pressure(TMP&Δ P)	Pressure gauge for vessel and Pressure transmitter for the skid
Flow rate	Monitor the rate of flow of permeate	Mass Flow rate indicator and flow meter
Conductivity	To measure the conductivity during CIP	Conductivity meter
Level of the volume	To maintain the correct volume of the product	With the accuracy of 0.1% of the completed range.

<b>6.5 Batch data display and record printing</b>	
---	--

Refer IRS (Installation Requirement Specification and Specific Instruction) Non editable data shall be available / transferred to USB Drive for printing the batch report, alarm log. Real time online printing shall be available for batch report	
---	--

<b>6.6 GMP requirements (Others)</b>	
--------------------------------------	--

6.6.1 Equipment design must be designed for aseptic processing.	
---	--

6.6.2 All process relevant wiring has to be executed in fail safe manner.	
---	--

6.6.3 All parts of the machine exposed in A/C area must be resistant to standard disinfectants or vendor shall provide the name of specific disinfectants.	
--	--

--	--	--	--

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	User Requirement Specifications				
	Equipment/System	Ultra Filtration System			
	Identification #	D-UFS 01	Document#	URS/D-UFS 01	
	Effective Date	2013-12-18	Revision#	07	

Specifications	Remarks		
<b>6.7 Specific requirements</b>			
6.7.1 In general the equipment has to be designed in a way to get easy and quick access to all necessary maintenance points e. g. Motors, etc.			
6.7.2 All ports should be attached with Sanitary Tri clamps			
6.7.3 All diaphragm valves to be sterile type			
6.7.4 Make of level sensor. ( <b>vendor to specify with technical data sheet</b> )			
6.7.5 Automated Temperature control during clarification process ( <b>tolerance limit: ±0.2 degree Celsius</b> )			
6.7.6 <b><u>Pump specification: Sterile Sanitary design</u></b> <b>CIP</b> : Yes <b>Surface finish</b> :Ra <0.6 µm (electropolished) <b>MOC</b> : SS316L <b>Seal</b> : Single mechanical seal <b>Elastomers</b> : EPDM - FDA/US			
6.7.7 Nozzle shell shall be seamless.			
6.7.8 Nozzle connection to be Triclover.			
6.7.9 Nozzles, adaptors, instrument shall comply with ASME BPE compliant.			
6.7.10 Total motor drive assembly with SS304 cover with TEFC eff 1.			
6.7.11 Flexible hose with TC end of 2m length to be provided for connections between the Utility header and the system			
6.7.12 Flexible hose with TC end of 2m length to be provided for connections between system and drain point.			
6.7.13 Performance Requirements: Vendor to demonstrate the following during FAT/SAT <ul style="list-style-type: none"> <li>• Sterility of the complete system</li> <li>• Flux- to be demonstrated with model solution</li> <li>• CIP- effectiveness to be demonstrated and spray ball coverage.</li> <li>• Temperature Control along with the level accuracy to be demonstrated.</li> </ul>			
6.7.14 Design Considerations: <ul style="list-style-type: none"> <li>• Vessel working Pressure: -1 to 3 bar(g)</li> <li>• Vessel working Temperature: 0°C to 134°C.</li> <li>• Vessel design Pressure: 4 bar or (<b>vendor to specify</b>)</li> <li>• Vessel design Temperature: 0-150<sup>0</sup> C or (<b>vendor to specify</b>)</li> <li>• Design pressure for safety release valve: 4.9 bar or (<b>vendor to specify</b>)</li> </ul>			
6.7.15 Module holder (process scale holder), all pipelines and connections, which come			
File Name	NPI_110831_EQP_URS_B1(D)-UFS 01.docx	Page No.	Page 13 of 17

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	User Requirement Specifications				
	Equipment/System	Ultra Filtration System			
	Identification #	D-UFS 01	Document#	URS/D-UFS 01	
	Effective Date	2013-12-18	Revision#	07	

Specifications	Remarks
----------------	---------

in contact with the product should be made SS 316 L and electro polished- To accommodate all mentioned makes of cassettes.	
6.7.16 Vendor shall provide the FRL (Filter, regulator, lubricator), automatic valve assembly and air pressure switch for instrument air. Connections to automatic diaphragm valve shall be in vendor scope.	
6.7.17 Instrumentation and control for Semi- Automatic operation of the unit.	
6.7.18 Vendor shall provide special tools for maintenance of the equipment	
6.7.19 The Vendor shall ensure maintenance parts availability for a minimum of 15 months from delivery.	
6.7.20 The filtration system should have provision for filter integrity tests system, for filter integrity test done pre and post filtration.	

7.0 CONSTRAINTS	
-----------------	--

7.1 Equipment location and available space	
--	--

<p>This equipment will be installed in the area of <b>DPT Vaccine Manufacturing Facility, PII, Coonoor.</b></p> <p><b>Equipment Location:</b>  <u>Toxoid Concentration Room (B1G060)</u>                      Block: Diphtheria Block                      Floor: Ground Floor                      Room Size: 4800 mm x 6510 mm                      False Ceiling: 3000 mm</p> <p><b>Physical condition of the rooms:</b></p> <ol style="list-style-type: none"> <li>1. Room will be non-hazardous</li> <li>2. Class: EU Class "C"</li> <li>3. Differential Pressure: 35Pa Absolute</li> <li>4. Temperature maintained: 22±2 °C</li> <li>5. Relative Humidity: &lt;60% Rh</li> </ol> <p>The equipment location is indicated in the relevant block of the layout enclosed as <b>URS Annex-1.</b></p>	
---	--

7.2 Available Utility	
-----------------------	--

7.2.1 Electricity: 3 ph, 415V AC, 50Hz	
7.2.2 Water for Injection: 80-85° C (point of use cooler provided)	
7.2.3 Pure Steam: 3 bar	
7.2.4 Plant steam: 3 - 3.5bar g	
7.2.5 Compressed air: 8 - 10 bar g	

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®	User Requirement Specifications				
	Equipment/System	Ultra Filtration System			
	Identification #	D-UFS 01	Document#	URS/D-UFS 01	
	Effective Date	2013-12-18	Revision#	07	

Specifications	Remarks
7.2.6 Chilled Water Supply: 6 - 7°C,	
7.2.7 <b><i>Vendor need to take care the utility requirement as desired for the system. As per the system if changes required in the utility pressure vendor should consider within the package.</i></b>	

### 8.0 ABBREVIATION

Abbreviation	Definition
ΔP	Differential Pressure
°C	Degree Centigrade
db	Decibel
HMI	Human Machine Interface
MOC	Material Of Construction
NPI	NNE Pharmaplan India Ltd
NWP	Nominal Water Permeability
PID	Proportional Integral Derivative
PII	Pasteur Institute of India
PLC	Programmable Logic Controller
RPM	Revolutions Per Minute
SS	Stainless steel
TMP	Trans membrane Pressure
UFS	Ultra Filtration System

### REVISION INDEX

Revision	Date	Reason for Revision
00	12.12.2011	1 <sup>st</sup> draft for client's review
01	18.05.2012	Client's Comment Incorporated
02	2012-10-22	Format changed as per HLL requirement
03	2013-01-23	HLL comments incorporated, received during the workshop dated 22 <sup>nd</sup> and 23 <sup>rd</sup> January 2013
04	2013-02-21	PIIC comments incorporated, received on 2013-03-18
05	2013-05-14	Format internally revised
06	2013-11-11	Updated as per Telecon dated 2013-11-08
07	2013-12-18	Updated as per commented URS received from HLL on e-mail dtd 2013-12-17

# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®


### User Requirement Specifications

<b>Equipment/System</b>	Ultra Filtration System		
<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01
<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07



### URS Annexure 1: LAYOUT OF DIPHTHERIA BLOCK

#### Room No: B1G060



# HLL LIFECARE LIMITED, CHENNAI

## Revival of DPT Vaccine Manufacturing Facility, PII, Coonoor

nne pharmaplan®

### User Requirement Specifications

<b>Equipment/System</b>	Ultra Filtration System		
<b>Identification #</b>	D-UFS 01	<b>Document#</b>	URS/D-UFS 01
<b>Effective Date</b>	2013-12-18	<b>Revision#</b>	07



### List of Preferred Make of components

SL.NO	DESCRIPTION	MAKE
<b>A</b>	<b>INSTRUMENTATION</b>	
1	PLC	Allen Bradley/ Siemens
2	Operator Interface/HMI	Allen Bradley/ Siemens
3	Temperature transmitter	Radix/ Yokogawa/Emerson
4	Temperature sensor	NEGELE/E&H
5	pH sensor	METTLER TOLEDO/E&H
6	Pressure transmitter	Wika /Dwyer/labom
7	Pressure regulator	FESTO
8	Flow meter	E&H / Khrono Marshall/ Rosemount
9	Temperature indicator	Radix/ Wika/ Waaree instruments
10	Steam trap	Spirax Marshall
11	Rupture Disc	Fike
12	Printer	Canon/Epsilon/HP
<b>B</b>	<b>MECHANICAL</b>	
13	Pressure gauges	WIKA/Denver/Negele
14	Vent filter cartridge	Sartorius/PALL/Millipore
15	Filter housing	Sartorius/ PALL/Millipore
16	Spray ball	HAKE/LECHLER
17	Diaphragm valve(Manual)	GEMU/ITT/Novaseptic
18	Ball valve(Manual)	Modentic/Alfa laval
19	Flush bottom valve	Novaseptic/GEMU
20	Sampling Valve	Novaseptic/GEMU
21	Flow switch	E&H/ Wika/Emerson
22	Recirculation Pump	Johnson/ Alfa Laval
<b>C</b>	<b>PNEUMATIC</b>	
23	Diaphragm valve(Automatic)	GEMU/ Novaseptic
24	Angle Seat valve	GEMU/Spirax/Sarco
<b>D</b>	<b>ELECTRICAL</b>	
25	Lamp	PAPENMEIER