

**Amendment No. 21****Date: 20.07.2015****Subject: Amendment no. 21 to the Tender Enquiry Document****Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY-II/06/14-15 dated 12.01.2015 and subsequent amendments published thereafter.**

The following changes in technical specifications are made further to the tender document and subsequent amendments published.

**Section I**  
**Notice Inviting Tenders(NIT)****Due to certain technical reasons in the e-tendering portal, the bidders are requested to submit a hardcopy of their technical bids and all related documents (uploaded in the portal) EXCEPT for the Price bid before the aforesaid “Closing date & time for physical submission of bids”.**

In view of above, the due date for submission of bids for the following schedules are revised as given below.

Sl. No	Sch. No	e-TenderRef.No (Event No.)	Brief Description
1	2	3000000087	C-Arm for ERCP
2	4	3000000089	Argon Plasma Coagulation
3	7	3000000092	Color Doppler Echocardiography System With 3d Facility
4	10	3000000095	Mammography with CR System
5	11	3000000096	800mA X-Ray unit with Single Detector (U/C Arm)
6	12	3000000097	800mA X-Ray unit with Ceiling Mount column and Dual Detector
7	20	3000000105	VATS Equipment

**Revised dates for schedules referred above:**

Sl.	Description	Schedule
c	Closing date & time for submission of <b>tender fee and EMD in physical form</b>	<b>28-July-2015,1700 hrs (IST)</b> Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
d	i) Closing date & time for submission of online bids	<b>30-July-2015, 1300 hrs IST</b>
	ii) Closing date & time for physical submission of bids (Price bid shall only be uploaded online; All other technical and commercial documents uploaded in the e-portal are to be submitted)	<b>31-July-2015, 1300 hrs IST</b>
e	Time and date of opening of bids	<b>31-July-2015, 1400 hrs IST</b>

**Note:**

**(i) If EMD is submitted in the form of BG, then the validity of the BG should be at least 165 days from the date of tender opening, i.e, upto 12.01.2016.**

**(ii) Tender fee( Rs.5,000/-) and EMD ( As applicable) should be deposited in the Tender Box located at HLL Lifecare Limited, Procurement and Consultancy Division, B-14 A, Sector-62, Noida-201307, Uttar Pradesh on or before 31-July-2015,1700 hrs (IST) . Submission beyond stipulated date & time would result in REJECTION of BID .**

**Section VII**  
**Technical Specifications**

**Schedule No. 4**  
**Argon Plasma Coagulation System**

**1. Existing Para:** HF power limitations: 350-400 Watt or more for cut, 120W or more for coagulation with option of change in steps

**Read as:** HF power limitations: 350-400 Watt or more for **monopolar cut**, 120W or more for **both monopolar and bipolar coagulation** with option of change in steps

**2. Existing Para 6: Irrigation pump for endoscopic procedures:**

Specifications:

- Option of tissue washing activation either by using the footswitch or directly at the unit
- Variable rate of water flow with option of selection
- Maximum flow rate (with compatible tubing set) 500 ml/ min or more
- Compactable tubing set for use with Olympus endoscopes

**Read as:** Irrigation pump for endoscopic procedures **with following specification:**

Specifications:

- Option of tissue washing activation either by using the footswitch or directly at the unit
- Variable rate of water flow with option of selection
- Maximum flow rate (with compatible tubing set) 500 ml/ min or more
- Compactable tubing set for use with Olympus endoscopes

**(OR)**

**Alternatively Non-sticky Bipolar forceps may be offered**

**Schedule No. 7**  
**Color Doppler Echocardiography System with 3D Facility**

**1. Existing Para 3.3 (After Amendment):** Adult & Pediatric Trans thoracic Cardiac, TEE (Adult - 3D and 2D capability) & Pediatric 3D or 2D capability) and Vascular Probes to be supplied which should be latest generation wide band transducers.

**Read as:** Adult & Pediatric Trans thoracic Cardiac, TEE (**Adult - 3D or 2D capability**) & (Pediatric 3D or 2D capability) and Vascular Probes to be supplied which should be latest generation wide band transducers.

- 2. Existing Para 4.5 (After Amendment):** Multi plane TEE Probe for Adult (**3D and 2D capability**) and Pediatric (3D or 2D capability) echocardiography —01 each.

**Read as:** Multi plane TEE Probe for Adult (**3D or 2D capability**) and Pediatric (3D or 2D capability) echocardiography —01 each.

- 3. Existing Para 4.5.a (After Amendment): 4D Volume Probe** (1-4 MHz) has to be offered — 01No.

**Read as: 4D Volume or Matrix Probe** (1-4 MHz) has to be offered — 01No.

**All other terms and conditions of the tender enquiry remain unaltered.**