

**Amendment No. 1****Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-III/32/GNS/18-19 dated 20/06/2018**

The following changes are being incorporated in the above referred Tender Enquiry Document.

**Section VII  
Technical Specification****Schedule No. 1  
Multipurpose OT Table (3000002936)**

<b>Para</b>	<b>TENDER SPECIFICATION</b>	<b>READ AS</b>
3.f	The table top can be moved cranially and caudally on its base.	The table top can be moved cranially and caudally <b>or longitudinal slide of min 250mm</b>
3.2.a	Height: 685-1100mm with 50 -70 mm mattress.	Height: <b>690-1080mm</b> with <b>50 -80</b> mm mattress.
3.2.b	Side tilt: +/-20degrees	Side tilt: <b>minimum +/-15</b> degrees
3.2.c	Back section adjustment: - 40 degrees to 70 degrees or more	Back section adjustment: - 40 degrees to <b>60</b> degrees or more
3.2.e	Trendelenburg: 30 degree	Trendelenburg: <b>minimum 25</b> degree
3.2.f	Anti trendelenburg: 30 degree	Anti trendelenburg: <b>minimum 25</b> degree
3.2.j	Cranial and caudal traversing 200 – 300 mm	Cranial and caudal traversing <b>or longitudinal movement of minimum 250 mm</b>
4.2	All accessories including the following should be quoted. The specific accessories and their quantity will be decided on the basis of actual requirement:-	All accessories including the following should be quoted. The specific accessories and their quantity will be decided on the basis of actual requirement. <b>All the accessories should be from the same manufacturer (except gel pad) and same quality standards as asked for table (para 7.1)</b>
4.2.g.ii	Gel prone positioner	Gel prone positioner <b>(feet)</b>
4.2.h	Optional accessories (Price of each item should be mentioned separately A. X-Ray Top for 9" Urology extension B. Metal Drain pan C. Power lift stirrup set with side rail clamp D. Foot control E. Split leg pair F. Reverse Trendelburg restraint strap G. Light weight transfer board.	<b>Additional accessories</b> (Price of each item should be mentioned separately A. X-Ray Top for 9" Urology extension B. Metal Drain pan C. Power lift stirrup set with side rail clamp D. Split leg pair E. Reverse Trendelburg restraint strap F. Light weight transfer board.
7.1	Should be US-FDA or European CE with 4 digit notified body number or Declaration of conformity along with ISO 13485 from notified body for the quoted model	Should be US-FDA or European CE with 4 digit notified body number or Declaration of conformity along with ISO 13485 from notified body <b>or BIS certified</b> for the quoted model

Para	TENDER SPECIFICATION	READ AS
		<b>Added para:</b> Physical demonstration of the quoted model must be taken

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

**Note:**

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**