

**AMENDMENT NO. I**

**AMENDMENT TO THE PRE BID MOM OF THE TENDER FOR SUPPLY, INSTALLATION, COMMISSIONING AND VALIDATION OF LAB EQUIPMENT (PHASE-IV) AT IVC, CHENGALPATTU**

**REF: TENDER NO: NPI-120310-EQP-S1-22 DATED 25-11-2017**


With reference to the tender for Supply, Installation, Commissioning and Validation of Lab Equipment (Phase-IV) at IVC, Chengalpattu, the following amendment may please be noted:

S. NO	CLAUSE OF PRE-BID MINUTES	AS PER PRE-BID MOM	AMENDED AS
A.VII	<b>Air Sampler</b> Specific revision in the datasheet	3.12. Airflow – 200 – 300 Litres /min	3.12. Airflow – <b>100 – 300</b> Litres /min
A. XIII	<b>Deep Freezer Low Temperature</b> Specific revision in the datasheet	Vertical Type – 250 L capacity -1 No (Quantity Reduced)	<b>REMOVED FROM LIST (Equipment Not Required)</b>
		Horizontal Type – 425 L capacity - Qty is same as 3 Nos	<b>Will remain same</b>
<b>Annexure-1</b>	<b>Revised List of Equipment and EMD (For Schedule No: 14)</b>	Deep Freezer Low Horizontal Type – 425 L capacity (Qty :2 Nos) – EMD: 21,000	<b>Deep Freezer Low Horizontal Type – 425 L capacity (Qty: 3 Nos) – EMD : 31,500</b>
<b>Annexure - A</b>	<b>Pre-Bid MOM – Weighing Balance</b> <b>Equipment Id -B1-Hep-WGB 06</b>	Equipment Id -B1-Hep-WGB 06 - Capacity - 40 mg - 1000g, Qty - 1 No, Readability -10 mg	Equipment Id -B1-Hep-WGB 06 - Capacity - 1g -1000g, Qty - 1 No, Readability -1 mg ,USP standard. Printer Req- 1No

	<b>Equipment Id -B1-Hep-WGB</b>	Equipment Id -B1-Hep-WGB 07 - Capacity - 20 mg - 410g, Qty - 1 No,Readability -1 mg	Equipment Id -B1-Hep-WGB 07 - Capacity - 100 mg - 400g, Qty - 1 No, Printer Req- 1No Readability -0.1 mg .USP standard. Printer Req- 1No
	<b>Equipment Id -F4-WGB 01,02</b>	Equipment Id -F4-WGB 01,02 - Capacity - 40 mg - 1000g, Qty - 2 No,Readability -10 mg	Equipment Id -F4-WGB 01,02 - Capacity - 100 mg -400g, Qty - 2 No,Readability -0.1 mg .USP standard. Printer Req- 1No
	<b>Equipment Id -B1-Hep-WGB 06</b>	Equipment Id -R1-WGB 01- Capacity - 40 mg -1000g, Qty - 1No,Readability -10 mg	Equipment Id -R1-WGB 01- Capacity - 100 mg -400g, Qty - 1No,Readability -0.01 mg .USP standard. Printer Req- 1No

All other terms and condition will remain as per the Tender No: NPI-120310-EQP-S1-22 DATED 25-11-2017 and Minutes of the Pre-Bid Meeting dated 04-12-2017 to 07-12-2017

For HLL Biotech Limited



Handwritten signature and date: 21/12/17

Authorized Signatory