

HLL pharmaplant	User Requirement Specifications			 <small>HLL BIOTECH LIMITED Subsidiary of HLL Group, India 9th Floor, HLL Building, Chennai</small>	
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

User Requirement Specifications Ceiling Suspended Laminar Air Flow

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

ne pharmaplan	User Requirement Specifications			 <small>HLL BIOTECH LIMITED Subsidiary of HLL Lifesciences Limited 95, Government of India Estate</small>	
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

URS Annexure List:

URS Annex No.	Detail
1	Excel sheet showing room location, quantity and dimension details

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

ne pharmaplan'	User Requirement Specifications			 <small>HLL BIOTECH LIMITED CHENGALPATTU INTEGRATED VACCINES COMPLEX</small>	
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

Table of Contents

1	APPROVAL SIGNATURE	4
2	EQUIPMENT DESCRIPTION	6
3	PROCESS DESCRIPTION	8
	3.1 INPUT & CHARGING METHOD	8
	3.2 BRIEF PROCESS STEPS.....	8
	3.3 OUTPUT & DISCHARGING METHOD	8
4	PRODUCTIVITY REQUIREMENT	8
	4.1 DESIRED/ SUGGESTED CAPACITY	8
	4.2 STANDARD BATCH SIZE	8
	4.3 CHANGE OVER TIME	8
	4.4 OTHER PRODUCTIVITY REQUIREMENT	8
5	CONTAINMENT	8
6	GMP REQUIREMENTS	8
	6.1 PROCESS CONTROL	8
	6.2 IN –PROCESS CONTROL	8
	6.3 LEVEL OF INSTRUMENTATION.....	9
	6.4 BATCH DATA DISPLAY AND RECORD PRINTING	9
	6.5 TECHNICAL SPECIFICATION	10
	6.6 MATERIAL OF CONSTRUCTION	10
	6.7 SPECIFIC EQUIPMENT REQUIREMENT	10
	6.8 OTHER REQUIREMENT	11
	6.9 REGULATORY GUIDELINES / STANDARDS	12
	6.10 SAFETY REQUIREMENTS.....	12
	6.11 DOCUMENTS.....	12
7	CONSTRAINTS	13
	7.1 EQUIPMENT LOCATION AND AVAILABLE SPACE	13
	7.2 UTILITY	14
8	ABBREVIATION	14
9	REVISION INDEX	14


HLL BIOTECH LIMITED, CHENNAI



INTEGRATED VACCINES COMPLEX, CHENGALPATTU

User Requirement Specifications	User Requirement Specifications				
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

1 APPROVAL SIGNATURE

This document is prepared by the Process, Validation and GMP Compliance team of "NNE Pharmaplan India" for the project "Integrated Vaccine Complex, Chengalpattu, Chennai" (project number: 120310) of HLL BIOTECH LIMITED (Chennai) under the authority of their Project Manager. Hence, this document before being effective shall be reviewed by the QA team of HBL, approved by Team lead and authorized by the appropriate Project authority.

NNE Pharmaplan India Limited			
Name	Designation	Signature	Date
Prepared by			
Mr. Manohar M S	Process Engineer		
Checked by			
Mr. Yogesha M J	Process Engineer		
Approved by			
Mr. Vikas Katial	GM & Head-COC Vaccines		

HLL Biotech Limited			
Name	Designation	Signature	Date
Reviewed by			
User Department (Viral Vaccine Formulation Block) G. L. ARSHAD PUNNARAO	DY. Manager - BVP		
User Department (Viral Vaccine Formulation Rabies) P. NAVIN KUMAR	DY. MANAGER - VIRAL VACCINE		
User Department (Rabies Vaccine Formulation Masles) K. RADHAKRISHNAN	DY. MANAGER - PRODUCTION		
User Department (Rabies Vaccine) S. SATHESH	MANAGER - VIRAL VACCINE		
User Department (Water Purification) K. RADHAKRISHNAN	DY. MANAGER - PRODUCTION		
SANJEEV KUMAR R (Quality control block)	MANAGER		
User Department (MBB Block-Hep B) C. H. DABHI	Assistant Manager		
User Department (MBB Block-H1N1) S. G. BURESH	SI. MANAGER		
User Department (BCG block) Polina Sreenath	Deputy Manager		
HARSHAD DABHI Project / Engineering Department	DY. MANAGER (HVAC)		
Approved By			

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

ne pharmaplan	User Requirement Specifications				
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

2 EQUIPMENT DESCRIPTION

A Ceiling Suspended Laminar air flow unit is used to carry out aseptic operations.

All points of the IRS except the below mentioned would be applicable for the equipment

- 4.1.11
- FDA Guidance for industry- Documentation for sterilization Process Validation
- ASME, ANSI / NSF 49-2008, EN12469
- ISO 14364, ISO 8362
- 5.2.7, 5.2.8

Note:

I.	This Technical Specification is the basis for an inquiry to a vendor and therefore the basis for the vendor's proposal.
II.	The vendor is asked to state in "REMARKS" column with "yes" if the described requirement will be completely fulfilled and with "no" in case the requirement will not or cannot be fulfilled with the proposed equipment. In case of any deviation a comment must be inserted or enclosed as a separate annexure by referring to the respective URS specification number.
III.	The vendor must clearly comment each item of the Technical Specification. The comments must be in English language. If extra cost for necessary options becomes necessary the item must be clearly stated.
IV.	In case that the requirement includes a question or request or an information from the vendor, the answer / information should be stated in the "REMARKS" column.
V.	The final version of this document including the vendor's comments will become basis of a potential purchase order or contract.
VI.	The Technical Specification serves to define a summary of all vendor's requirements concerning scope of delivery and services.
VII.	The vendor is responsible for technically unobjectionable function of the equipment. This TS is not intended to dictate a technical design to the vendor. If agreed upon with the vendor, the vendor can apply his practically proven design.
VIII.	Special Instruction a. If no comments against any specification should be considered as "NO" and b. If there is no reply / comments against the complete URS by the vendor then it should be treated as unresponsive / technically non-compliant and rejected.
IX.	All the instruments and controls mentioned in the URS(s) are expected to be standard supply and part of your standard equipment model. In case of any deviation or redundancy or additional scope of supply is noticed, vendor is required to obtain clarification from HBL.

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

ne pharmaplan	User Requirement Specifications			 <small>HLL BIOTECH LIMITED A Public Limited Company Registered in India</small>	
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

	before submitting the quotes.
X.	The makes requested are standard international makes. In case of any deviation, vendor to seek clarification from HBL before submitting the offers.
XI.	Refer document Installation Requirement Specification and Specific Instructions with URS; NPI-120310-IRS-S1-01
XII.	Refer Tender document with URS; NPI-120310-EQP-S1-TD-14

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

HLL Pharmaplan	User Requirement Specifications				
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

Specifications		Remarks
----------------	--	---------

3 PROCESS DESCRIPTION

3.1 Input & Charging method

A LAF is used to Carry out aseptic operations.

3.2 Brief Process Steps

3.2.1 Blower in the LAF should be switched on for prior to using the LAF.

3.2.2 LAF's use vertical laminar airflow to create a barrier to product & Environment.

3.3 Output & Discharging method

Not Applicable

4 PRODUCTIVITY REQUIREMENT

4.1 Desired/ suggested capacity

Refer URS annexure - 1

4.2 Standard batch size

Not Applicable

4.3 Change Over Time

Not Applicable

4.4 Other Productivity Requirement

Continuous 24x7 operation required

5 CONTAINMENT

Not applicable

6 GMP REQUIREMENTS

6.1 Process control

Not Applicable

6.2 In -Process control

Not applicable

--	--

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

ne pharmaplan	User Requirement Specifications				
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

6.3 Level of instrumentation

Sufficient and suitable instrumentation for the process, safety and productivity control as indicated in the following table:

Type of control	Purpose	Instrumentation
Differential Pressure	Monitor the pressure drop of the HEPA (H-14) filter w.r.t the ambient pressure	Microprocessor control to provide current filter condition
Air velocity	To measure the laminarity of air	Air velocity sensing device with continuous digital display on the LED
Speed controller	To maintain the stable motor voltage and airflow despite building voltage fluctuations	Microprocessor based speed controller
UV light with Hour meter	UV radiation	with digital hour meter On/Off automation (Timer) for exposure time with Manual Switch.

6.4 Batch data display and record printing

Not Applicable

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

HLL Pharmaplan	User Requirement Specifications				
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

6.5 Technical Specification

6.5.1	Model	cGMP	
6.5.2	Type	Grade A (ISO 5)	
6.5.3	Inner Area (mm)	Refer URS Annexure - 1	
6.5.4	Air flow pattern	Vertical Laminar Air Flow type a) Uni-directional Laminar Air Flow type b) Double Stage filtration	
6.5.5	Pre- Filter	EU EN779 Class G3 Efficiency 95% down to 5 μ	
6.5.6	UV light	a) UV light emitting lamp with Hour meter The minimum operate able life of UV Lamp should be 8000 hr & light emitted shall be short-wave UV radiation with a peak at 253.7 nm (UV-C) for germicidal action. The hour meter should be able to be manually re-set. b) UV light with digital Hour Meter	
6.5.7	HEPA Filter	EU EN 1822 Class H14 / ISO 29463 Class 45H Efficiency >>99.995% down to 0.3 μ	
6.5.8	Air flow rate	0.45 ±20 % m/s	
6.5.9	Quantity	Refer Annexure 1	
6.5.10	Electrical Requirement	220-230 V, 50 Hz , Single phase	

6.6 Material of Construction

6.6.1	Body Construction	SS 304, min 240 grit	
6.6.2	Grills	SS 304, min 240 grit	
6.6.3	Coving	SS in built	
6.6.4	Curtains	Anti-static PVC curtains (Except for filling and sealing line ceiling suspended LAFs)	
6.6.5	MOC Fan	Aluminium or SS 304	
6.6.6	All welds shall be ground finish		

6.7 Specific Equipment requirement

6.7.1	LAF shall comply ISO Class 5 (Grade A as per WHO)		
6.7.2	Pre-filters should be easily accessible for periodic cleaning.		
6.7.3	Sleeving for accommodating the pre filters.		
6.7.4	Side access panel for final filters and blowers.		
6.7.5	HEPA filter shall have an efficiency of 99.995 % when tested with POA.		
6.7.6	1 No. magnehelic gauge to be provided to monitor the differential pressure across the HEPA filter. If the number of modules more than one than each module shall have separate magnehelic gauge.		

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

Inne pharmanplaplan	User Requirement Specifications				
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

6.7.7	Blower System shall be balanced for vibration free operation and noise level.	
6.7.8	Blower shall be permanently lubricated	
6.7.9	Individual blower motor shall be provided on each standard HEPA filter as per the LAF area	
6.7.10	Motor and electrical devices designed for working in normal operating conditions.	
6.7.11	Preferred make for Motor and Blower: Crompton Greaves/ ABB/ GE/ Siemens/EBM-PAPST/ebm-Nadi / Zeil-abegg/ Kruger/ Nicotra	
6.7.12	Light level shall be minimum 650 Lux. Lighting with fluorescent lamps installed in such a way as to generate no turbulence in air flow.	
6.7.13	Soft touch controls for blower, light shall be provided along color coded indicator (different for each application). It shall be unit mounted.	
6.7.14	Ergonomic design to access electrical connections and instrumentation behind the control panel	
6.7.15	PVC curtains shall be provided around the Ceiling suspended LAF to maintain the air flow pattern. Length of the curtains shall be till 300-400mm above the ground floor. (Except for filling and sealing line ceiling suspended LAFs)	
6.7.16	Vendor shall specify the sizes and quantity of HEPA filter and pre filter in each Ceiling suspended LAF as per the sizes	
6.7.17	Detailed Installation and fixing details shall be provided	
6.7.18	Approved makes for filter : Camfil Farr/ AAF/ Freudenberg	
6.7.19	Approved makes for Magnehelic Gauge : Dwyer	

6.8 Other requirement

6.8.1	Cleaning shall be done manually.	
6.8.2	All bolts, nuts shall be of dome type of SS304 material	
6.8.3	Vendor to give code numbers for each component	
6.8.4	There shall be no crevices, edges, cracks, pits so as to avoid dust accumulation	
6.8.5	In general the equipment has to be designed in a way to get easy and quick access to all necessary maintenance points e. g. motors, filters, etc.	
6.8.6	The design shall be maintenance friendly for the ease of replacement of filters	
6.8.7	All parts of the machine exposed in classified area must be resistant to standard disinfectants or vendor shall provide the name of specific disinfectants	
6.8.8	The heat given off by the unit must be stated (inside the room).	

File Name	NPI_120310_EQP_URS_CLF_01	Start Date	16-06-2015	Page No.	Page 11 of 14
-----------	---------------------------	------------	------------	----------	---------------

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

HLL Biopharmaplan	User Requirement Specifications				
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

6.8.9	<p>Failure mode detection</p> <p>A. Equipment shall be capable to detect the following failure, notify the operator with alarm ,Visual indication:</p> <p>a) Blower motor overload.</p> <p>b) Emergency stop activated.</p> <p>B. Following condition need only notification to operator for procedural control:</p> <p>a) Differential pressure across HEPA filter should provide constant indication of pressure drops and a visual alarm signal of high and low pressure in the filter</p>	
-------	---	--

6.8.10	<p>The following tests to be conducted at the site during qualification:</p> <ol style="list-style-type: none"> 1. Air velocity test 2. Filter integrity test 3. Flow visualization test (Videography) 4. Non-viable particle count 5. Recovery test 6. Lux level 7. Sound level 	
--------	---	--

6.9 Regulatory guidelines / standards

6.9.1	ISO Class 5 14644 – 1 (For Cleanliness Class)	
6.9.2	ISO 14644 – 3 (For HEPA filter integrity testing & Velocity testing)	
6.9.3	DIN EN 1822 and ISO 29463 (Filter Class)	
6.9.4	DIN EN779 (2012) for Filter Efficiency	
6.9.5	IEST-RP-CC001.3, CC007.1, CC034.1 (Filtration)	
6.9.6	IEC 61010-1 (Electrical safety)	
6.9.7	EU-GMP-Guideline Part 1, Annexes 1, 11 & 15	
6.9.8	Schedule M of Indian Drugs and Cosmetics Act	
6.9.9	Code of Federal Regulations (CFR) 21, Part 210: cGMP in Manufacturing, Processing, Packing and Holding of Drugs	

6.10 Safety requirements

6.10.1	Noise level should not be more than 65 decibels.	
6.10.2	Appropriate closure of all the rotating parts.	

6.11 Documents

6.11.1	Vendor to submit detailed fabrication drawing for approval before fabrication.	
6.11.2	Phase 1: Post ordering and prefabrication stage of the equipment	
6.11.2.1	Functional design specification	
6.11.2.2	Equipment descriptions	
6.11.2.3	Equipment operation steps	
6.11.2.4	List of failure indications and interlocks (as applicable)	

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

HLL Biotech Limited	User Requirement Specifications				
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

6.11.2.5	Critical list of major component, devices and instruments with their specific functions, specs and data sheets.	
6.11.2.6	GA/ Schematic diagram of the equipment	
6.11.2.7	DQ specification as per the User approved format.	
6.11.2.8	IQ specification as per the User approved format.	
6.11.2.9	OQ specification as per the User approved format.	
6.11.3	Phase - 2	
6.11.3.1	Vendor shall provide the FAT protocol at least 1 month in advance of the date of FAT, for the approval by the user.	
6.11.3.2	System shall be inspected and tested (FAT) at the Vendor's site in the presence of user's representative before delivery.	
6.11.4	Phase - 3	
6.11.4.1	Vendor shall provide the following documents in the delivery package in minimum 2 sets. The delivery package shall reach the site of user at least 15 days before the delivery equipment's for the engineering check of the documents.	
6.11.4.2	Shipping checklist.	
6.11.4.3	Operation and maintenance manuals; preventive maintenance instruction & schedule for equipment major component as well as the operating system. Control system operation manual. Cleaning procedures to be provided.	
6.11.4.4	Operation and maintenance manuals for the bought out items (as applicable).	
6.11.4.5	Drawings: Electrical, instrumentation, final GA drawing etc.	
6.11.4.6	Spare and/ or change parts list with ordering information.	
6.11.4.7	MOC certificates	
6.11.4.8	Calibration certificates of critical instruments with respect to the traceable national reference standard instrument and their calibration procedure.	
6.11.4.9	Guaranty/ warranty certificate for each equipment and major bought-out items and comprehensive 1 years warranty from the date of installation.	
6.11.4.10	Types of Lubricant and Lubrication instructions. Food grade certificates.	
6.11.4.11	The Vendor shall provide start-up services through successful completion of the site acceptance test. The site acceptance test will be a repeat of the factory integration test performed at the Vendor's facility.	

7 CONSTRAINTS

7.1 Equipment location and available space

Refer annexure 1

HLL BIOTECH LIMITED, CHENNAI

INTEGRATED VACCINES COMPLEX, CHENGALPATTU

NNE Pharmaplan	User Requirement Specifications				
	Equipment/System	Ceiling Suspended Laminar Air Flow Unit			
	Identification #:	-	Document No:		URS/CLF-01
	Effective Date:		Revision No:		01

7.2 Utility

a) Electricity: Single Phase (220 V) & 3 phase (420 - 440 V) (Report Requirement)

8 ABBREVIATION

Abbreviation	Definition
CFC	Chlorofluorocarbon
CFR	Code of Federal Regulations
DQ	Design Qualification
FDA	Food and Drug Administration
GMP	Good Manufacturing Practice
HBL	HLL Biotech Ltd
HMI	Human Machine Interface
IQ	Installation Qualification
ISO	International Standards Organization
LAF	Laminar Air Flow
MOC	Material Of Construction
NA	Not Applicable
NPI	NNE Pharmaplan India Limited
OQ	Operational Qualification
PAO	Poly Alpha Olefin
PQ	Performance Qualification
QA	Quality Assurance
RH	Relative Humidity
SS	Stainless steel
TBD	To be discussed
UPS	Uninterrupted Power Supply
URS	User Requirement Specifications

9 REVISION INDEX

Revision	Date	Reason for revision
00	16-06-2015	First Draft for Client's Review
01	14-07-2015	Updated as per HBL comments dated 10-07 2015

HLL BIOTECH LIMITED,							
INTEGRATED VACCINES COMPLEX, CHENGALPATTU							
Document Name: URS Annexure-1: Ceiling Suspended LAF List							
Document number: NPI_120310_EQP_URS_CLF_01							
Date / Revision: 15-07-2015 / 00							
Sl. No.	Room Number	Room Name	Equipment Code	Description	Quantity	Room Height, mm	Remark
BACTERIAL VACCINES FORMULATION BLOCK							
1	F2G009	Prep. Room	F2-CLF-01	HEPA Filter Area (3.0 m x 1.2 m)	1	3000	
2	F2G011	Sterile Buffer	F2-CLF-02	HEPA Filter Area (3.8 m x 1.8 m)	1	3000	
3	F2G024	Disinfection preparation	F2-CLF-03	HEPA Filter Area (1.5 mx 1.2 m)	1	3000	
4	F2G028	Blending	F2-CLF-04	HEPA Filter Area (1.6 m x 1.5 m)	1	4000	
5	F2G030	Buffer & solution Preparation	F2-CLF-05	HEPA Filter Area (2.0 m x 1.0 m)	1	3500	
6	F2G044	Filling Room	F2-CLF-06, 07	HEPA Filter Area (4.5 m x 2.7 m), (1.8 m x 2.5 m)	2	3000	
VIRAL VACCINES FORMULATION BLOCK							
7	F1G036	Blending / Formulation Room	F1-CLF-01	HEPA Filter Area (1.5 m x 2.0 m)	1	4000	
8	F1G037	Sterile Material Receiving	F1-CLF-02, 03	HEPA Filter Area (1.8 m x 2.9 m), (1.8 m x 2.3 m)	2	3000	
9	F1G043	Vial Filling & Capping	F1-CLF-04, 05, 06	HEPA Filter Area (3.3 m x 2.3 m), (1.8 m x 5.2 m), (2.0 m x 2.0 m)	3	3000	
10	F1G081	Sterile Material Receiving	F1-CLF-07	HEPA Filter Area (1.5 m x 1.8 m)	1	3000	
11	F1G081	Sterile Material Receiving	F1-CLF-08	HEPA Filter Area (1.0 m x 1.8 m)	1	3000	
12	F1G086	Vial Filling	F1-CLF-09, 10, 11	HEPA Filter Area (2.5 m x 3.4 m), (5.1 m x 1.6 m), (1.8 m x 1.8 m)	3	3000	
13	F1G028	Media Preparation	F1-CLF-12	HEPA Filter Area (1.8 x 1.0) m	1	3500	
14	F1G028	Media Filtration	F1-CLF-13	HEPA Filter Area (1.8 x 1.0) m	1	3500	
15	F1G080	Blending and Formulation Room	F1-CLF-14	HEPA Filter Area (1.5 m x 2.0 m)	1	4000	
HIB(MBB) BLOCK							
16	B1G120	Media Preparation Room	B1-CLF-01	HEPA Filter Area (1.5 mtr x 1.0 mtr)	1	3000	
17	B1G137	Sterile Material Unloading Room	B1-CLF-02, 03	HEPA Filter Area (3.0 m x 2.5 m) (3.3 m X 2.5 m)	2	3000	
18	B1G145	Dis-Infecant Preparation Room	B1-CLF-04	HEPA Filter Area (1.5 mtr x 1.2 mtr)	1	3000	
HEPATITS-B(MBB) BLOCK							
19	B1G043	Purification & Chromatography Room	B1-CLF-05	HEPA Filter Area (1.5 mtr x 1.2 mtr)	1	3000	
20	B1G045	Sterile Material Unloading Room	B1-CLF-06	HEPA Filter Area (10.0 mtr x 2.5 mtr)	1	3000	
21	B1G048	Dis-Infecant Preparation Room	B1-CLF-07	HEPA Filter Area (1.5 mtr x 1.2 mtr)	1	3000	
22	B1G041	Aerosol -Preparation Room	B1-CLF-08	HEPA Filter Area (1.2 mtr x 1.2 mtr)	1	3000	
RABIES BULK BLOCK							
23	B4G033	Sterilized Material Unloading	B4-CLF 01, 02	HEPA Filter Area (24 x 1.4) m. (for DHS) ; (1.2 x 2.4) m (for autoclave)	2	3000	
24	B4G036	Cell Culture	B4-CLF 03	HEPA Filter Area (1.650 x 1.25) m	1	2700	
25	B4G057	Purification	B4-CLF 04	HEPA Filter Area (2.0 x 1.5) m	1	3000	
26	B4G046	Virus Propagation	B4-CLF 05	HEPA Filter Area (1.650 x 1.25)	1	2700	
BCG BLOCK							
27	F4G041	Media Bottle Storage	F4-CLF 01	HEPA Coverage Area (1.2 m X 1.2 m)	1	3000	
28	F4G035	Sterile Unloading	F4-CLF 02, 03	HEPA Coverage Area (1.2 m X 1.2 m)	2	3000	
29	F4G040 & F4G040A	Filling & Lyo Loading	F4-CLF 04, 05, 06, 07	HEPA Filter Area (2.5 m x 3.4 m), (5.1 m x 1.6 m), (1.8 m x 1.8 m), (2.2 m x 5.0 m)	4	3000	
QC BLOCK							
30	Q1S041	Sterile Filtration Area	Q1-CLF 01	HEPA Coverage Area (1.2 x 0.6) m.	1	2400	
WAREHOUSE BLOCK							
31	W1G036	Autoclave Room	W1-CLF 01	HEPA Coverage Area (1.2 x 0.6) m.	1	3000	