

**Amendment No. 3****Date: 23/09/2014****Subject: Amendment to the Tender Enquiry Document****Ref: NIT No.: HLL/PCD/GNCTD/10/GBPH/14-15 dated 31/07/2014 read with its Amendment no.1 and 2 dated 06.09.2014 and 15.09.2014 respectively.**

The following changes have been incorporated in the referred NIT against Tender ID: 2014\_HFWD\_61971\_14 only.

**Section - I****NOTICE INVITING TENDERS (NIT)****Existing Schedule of closing and opening of tender:**

Sl. no.	Tender ID	Description	Qty.	Tender Fees (Rs.)	EMD Amount (Rs.)	Date & time of Prebid meeting	Date & time of closing of online tender	Closing date & time for submission of physical Tender	Date & time of opening of tender
14	2014_HFWD_61971_14	Fully Automated Endoscope Reprocessor (Endowasher)	1	500	30,000	12-08-14	25-09-14	26-09-14	26-09-14
						11:00 AM	6:00 PM	1:00 PM	2:00 PM

**Amended Schedule of closing and opening of tender:**

Sl. no.	Tender ID	Description	Qty.	Tender Fees (Rs.)	EMD Amount (Rs.)	Date & time of Prebid meeting	Date & time of closing of online tender	Closing date & time for submission of physical Tender	Date & time of opening of tender
14	2014_HFWD_61971_14	Fully Automated Endoscope Reprocessor (Endowasher)	1	500	30,000	12-08-14	08-10-14	09-10-14	09-10-14
						11:00 AM	6:00 PM	1:00 PM	2:00 PM

**SECTION VII**  
**Technical Specification**

**Item No. 14**  
**Fully Automated Endoscope Reprocessor (Endowasher)**

- 1. Existing:- Para 9:-** To include supply of all necessary chemicals/reagents for 1 month usage.

**Read as:- Para 9:-** To include supply of all necessary chemicals/reagents for 1 month usage.  
**(Approx. Cidex-60 litres, Low Foam Enzymatic detergent -03 litres, 70% Ethyl Alcohol-03 litres)**

**Added Para: -**

The bidder should provide and install in all respect a 100 ltr/hour RO system with a 500 ltrs overhead tank and a motor pump in order to fulfil the pre-requisites and service requirements for the Endowasher.

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

**Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**