

AMENDMENT NO 03

REF: EXPRESSION OF INTEREST (EOI) for registration and empanelment of Manufacturers / Suppliers for the supply of Medical and imaging laboratory consumables / Reagents for HLL Lifecare Ltd Diagnostics centers across India vide EOI no: HLL/CHO/SD/HCS/2019-20/EOI 001 dated 07.06.2019 read with its amendment HLL/CHO/SD/HCS/2019-20/EOI 001 Amendment 01dated 20/06/2019 and Amendment 02 dated 13/08/2019.

Amendment 03 – Additional items added to the list of consumables given as per annexure 03 of EOI document and as per amendment 01. The details of the new consumables and conditions as given below:

SL NO	NAME OF ITEM	UOM	DETAILED SPECIFICATION	REMARKS	PACKSIZE (To be filled by the bidder)	Whether the product is available YES/NO (To be filled by the bidder)	Manufacturer name (To be filled by the bidder)
M	ADDITIONAL ITEMS						
	POINT OF CARE CUVETTES AND STRIPS						
27	Micro Cuvette with dry Reagent - Heomoglobin	NOS	Disposable microcuvette containing dry reagent microcuvette should be made of polystyrene plastic and contains no active Ingredients	Should hold 10 µL of blood sample			
28	Micro Cuvette with dry Reagent - Glucose	NOS	Disposable microcuvette containing dry reagent Microcuvette : The microcuvette to be made of polystyrene plastic and contain the following reagents: cuvette (Mutarotase, Glucose Dehydrogenase and Diaphorase NAD (Nicotinamide-	Sample Volume : 5 µL			

			Adenine Dinucleotide), Polypropylene glycol, MTT (Methylthiazolyldiphenyl Tetrazolium), Saponin, Ammonium Chloride Sodium Fluoride).				
29	Micro Cuvette with dry Reagent - Albumin	NOS	Disposable microcuvette containing dry reagent Microcuvette : The microcuvette is made of polystyrene plastic. The distance between the walls of the optical eye is 0,5 mm. Contain rabbit antihuman albumin ingredients antibody (polyclonal), PEG, Tris/HCl-buffer, polymer and non-reactive .	Should hold 18 µL of urine sample			
30	Micro Cuvette with dry Reagent - WBC	NOS	Disposable microcuvette containing dry reagent Microcuvette : Microcuvette to be made of polystyrene plastic and contain the following reagents: cuvette (Methylene Blue, Saponin, Surfonyl 465, Triton X-100).	Sample Volume : 10µL			
31	Cartridge with reagent – HbA1c	NOS	Disposable cartridge containing reagent. Cartridge: Chamber 1 (reagent solution) contains: 25 v/v % boronate affinity bead, surfactant, 0.1 w/v % nonreactive ingredients buffer. Chamber 2 (rinsing solution) contains: surfactant, 0.1 w/v % nonreactive ingredients, buffer solution) contains: surfactant, 0.1 w/v % nonreactive ingredients, buffer.	Should hold 4 µL of blood sample Room temperature storage for 18 months			

GENERAL CONDITIONS (ITEMS – M 27 to M 31)

1. ALL the items (Cuvette / Cartridge) should have a factory calibrated analyzer to be placed as POC machine for particular operation.
2. The product should have quality certificates such as CE marking and the facility should be ISO 13485 certified.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of submission of bids