

AMENDMENT No.3

Date: 16/07/2014

Subject: Amendment to the tender Enquiry Document**Ref: Tender Enquiry No.: HLL/PCD/PATIALA/02/14-15 dated 27/05/2014**

The following amendments are being incorporated in the referred tender enquiry document.

SECTION-I**NOTICE INVITING TENDERS (NIT)**

Sl. No.	Description	Existing Schedule	Revised Schedule
i.	Dates for document download of tender enquiry documents	29.05.2014 to 19.07.2014	29.05.2014 to 29.07.2014
v.	Closing date & time for receipt of Tender in e-Portal	21.07.2014; 1800hrs IST	29.07.2014 ; 1800hrs IST
vi.	Closing date & time for receipt of physical documents as stated in para 7 below	22.07.2014; 1400hrs IST	30.07.2014 ; 1400hrs IST
vii.	Time and date of opening of Techno – Commercial tenders in e-Portal	22.07.2014; 1430hrs IST.	30.07.2014 ; 1430hrs IST.

SECTION-VII**Technical Specifications****Item no. 3****Computerized cardiopulmonary exercise testing system with treadmill for humans**

Existing: Para 6. The system should have a bidirectional volume sensor with the following specifications:-

- (i) Volume: 0 to 10 lit.
- (ii) Accuracy: 50 ml or 2%
- (iii) Resolution: 3 ml
- (iv) Flow: 0 to 15 l/s

Read as : Para 6. The system should have a bidirectional volume sensor with the following specifications:-

- (i) **Volume: 0 to 8 lit. or more**
- (ii) Accuracy: 50 ml or 2%
- (iii) Resolution: 3 ml
- (iv) **Flow: 0 to 15 l/s or more**

Item no. 13

1.5 TESLA MAGNETIC RESONANCE IMAGING SYSTEMS

Existing: Added Para- Quantification of fat in the liver

Read as: Quantification of fat in the liver (**optional**)

Existing: Para 8.n. Suitable coils for Proton MR spectroscopy for brain, muscle, cardiac and liver spectroscopy. (Price should be offered separately for coils and software if available)

Read as : Suitable coils for Proton MR spectroscopy for brain, muscle, cardiac and liver spectroscopy.

Existing : Para 8. G Dedicated coil with at least 32 channels for peripheral angiography application.

Read As: Para 8 g **Suitable** coil with at least 32 channels for peripheral angiography application

Item no. 18

Neonatal Ventilator

Existing: Para 10 Should be compatible with the Nitric oxide delivery system Sensor NOx (Viasys Healthcare, CA, USA) and the necessary connectors/adaptors/tubings and any other accessories required should be provided.

Read as: PARA 10 - DELETED

All other terms and conditions of the tender enquiry remain unaltered.

Note:

- 1. Bidders who have already prepared their EMD in form of bank guarantee as per the requirement of the tender enquiry scheduled for closing/opening on 21.07.2014 /22.07.2014 may submit their tender with the same bank guarantee without any further extension of its validity to meet the revised schedule.**
- 2. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.**