

Amendment No. 3

16.02.19

Sub: Amendment to the Bidding Document

Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/39/18-19 dated 09.01.2019 read with its Amendment no. 01 & 02 dated 02.02.19 & 11.02.19 respectively.

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

**SECTION - VII
TECHNICAL SPECIFICATION AND GENERAL POINTS**

A. TECHNICAL SPECIFICATION:

Item No:- 4 (Rfx/Event no. 3000003732)

Rigid Cysto-Resectoscope

Reference to the Bidding Document	Specification/Clause of the bidding documents	Amended as
Amendment No. 1; Sl. no. 02; Pg.no. 02	ESU with 300W or more output power. ESU should be of same make of cysto-resectoscope	ESU with 300W or more output power. ESU should be of same make of cysto-resectoscope. If same make ESU is not available, then bidder should quote any reputed make ESU along with a Declaration from OEM of Cysto-resectoscope stating "ESU is not manufactured by us". Bidder should give demonstration of complete system in case requested by committee

Item No:- 5 (Rfx/Event no. 300003733)**Micro Vascular Instruments Set**

Reference to the Bidding Document	Specification/Clause of the bidding documents	Amended as
Amendment No. 1; Sl. no. 01; Pg.no. 02	Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.	Compatible sterilization container with thermo lock (or similar) lock system & bacterial (or microbial) filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.

Item. No:- 6 (Rfx/Event no. 300003734)**Major Vascular Instruments Set**

Reference to the Bidding Document	Specification/Clause of the bidding documents	Amended as
Amendment No. 1; Sl. no. 05; Pg.no. 04	Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.	Compatible sterilization container with thermo lock (or similar) lock system & bacterial (or microbial) filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.

Item No:- 7 (Rfx/Event no. 300003735)**Thoracic Vascular Instruments Set**

Reference to the Bidding Document	Specification/Clause of the bidding documents	Amended as
Amendment No. 1; Sl. no. 02; Pg.no. 04	Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.	Compatible sterilization container with thermo lock (or similar) lock system & bacterial (or microbial) filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.

Item No:- 8 (Rfx/Event no. 300003736)**Plastic Surgical Instruments Set**

Reference to the Bidding Document	Specification/Clause of the bidding documents	Amended as
Amendment No. 1; Sl. no. 02; Pg.no. 05	Compatible sterilization container with thermo lock system & bacterial filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.	Compatible sterilization container with thermo lock (or similar) lock system & bacterial (or microbial) filter for instruments set along with SS wire basket/SS perforated basket having silicone supports for holding instruments, should be supplied in required nos. per each instruments set so as to accommodate all the instruments for sterilization.

Note:

Technical Compliance Sheet in HLL e-bid portal 'for RFX / Event number 300003732, 300003733, 300003734, 300003735 & 300003736' has been replaced by a new version i.e. 'Ver1.1'. The same need to be downloaded for submission.

All other contents of the Bidding Document including terms & conditions remain unaltered.