

13-12-2017

Amendment No. 5**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-III/19/MOT/17-18 dated 06-10-2017 read with its Amendment No. 1 dated 14-11-2017, Amendment No. 2 dated 27-11-2017, Amendment No. 3 dated 06-12-2017 and Amendment No. 4 dated 07-12-2017****Section VII
Technical Specification**

Sl. No.	Para/Page Ref.	Tender Specification	Read As
1	Page 48	Bidder Responsibility	Added Para (under Bidder responsibility) Para q: Bidder must have a satisfactory installation of complete MOT as asked in tender and demo may be taken for the same.
2	7	Touch Screen Control Panel 20" or more	Added Para: (Under Touch Screen Control Panel 20" or more): Para 7.12 : The Control Panel should be able to display the Isolation Panel Alarm Conditions alongwith MGPS Alarms
3	14.1 f	The Pendant Service Heads should be modular with minimum 1200mm head. The heads should be capable of accepting a range of shelves, and infusion poles or other accessories. The pendant Heads should support the range of Physiological Monitor Mounting Solutions.	Para 14.1f: The Pendant Service Heads should be modular with minimum 800mm head. The heads should be capable of accepting a range of shelves, infusion poles, electrical switch/sokets, gas outlets other accessories as asked in tender. The Pendant Heads should support the range of Physiological Monitor Mounting Solutions.
4	14.2 f	The Pendant Service Heads should be modular with minimum 1200mm head. The heads should be capable of accepting a range of shelves, and infusion poles or other accessories. The Pendant Heads should support the range of Physiological Monitor Mounting Solutions.	Para 14.2f: The Pendant Service Heads should be modular with minimum 800mm head. The heads should be capable of accepting a range of shelves, infusion poles, electrical switch/sokets, gas outlets other accessories as asked in tender. The Pendant Heads should support the range of Physiological Monitor Mounting Solutions.

Sl. No.	Para/Page Ref.	Tender Specification	Read As
5	18	Isolation Panel System	Added Para: (Under Isolation Panel System): Para 18.6 : The IPS should be able to integrate with HIS/BMS and Surgeon Control Panel as standard
6	19	Para 19.5: Bidder shall offer UPS from make – APC/TATA Liebert/Delta/Hitachi/Consul Neowatt	Para 19.5: Bidder shall offer UPS from make – APC/ TATA Liebert/ Delta /Hitachi/ Consul Neowatt / UNILINE/3EM
7	22.1 p.	Surgical Light System Should be compliant with relevant European CE /US FDA standards	Para 22.1.p: Surgical Light System Should be European CE with 4digit notified body/US FDA certified or Declaration of Conformity for quoted model with ISO 13485 issued by 4digit notified body.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

Bidders are advised to ensure that the revised price bid format is used for submission of their price offer with the same file name only as in the system. Any other file name will not be accepted by the system while submitting their bid online.