

## TENDER ENQUIRY NOTICE

HLL/ID/Trssr-Med Equip/2014-15 /

20.12.2014

Name of Work : Supply of 1 no: of External pacemaker for the setting up of emergency and critical care department at Thrissur Medical College.

Limited tenders are invited on behalf of Principal, Thrissur Medical College, Govt. of Kerala for the Supply of 1 no: of External pacemaker for the setting up of emergency and critical care department at Thrissur Medical College. The specification for the External pacemaker is enclosed herewith.

### **Address to which the offers are to be submitted:**

#### **Deputy Vice President (Technical)**

HLL Lifecare Limited

“ADARSH”, TC - 6/1718 (1)

Vettamukku, Thirumala P O

Thiruvananthapuram - 695 006

Ph : 0471 - 2365872/73

Fax: 0471 - 2368144

### **Submission of Offer**

The offer shall be submitted in single bid system. The offer should be submitted in a sealed envelope superscribing “Supply of 1 no: of External pacemaker for the setting up of emergency and critical care department at Thrissur Medical College.”

Last date and time of Receipt of offers : 30.12.2014 11.00 am.

Date and time of Opening of Offers : 30.12.2014 11.30 am

### **Payment terms**

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

a) **On delivery:**



### **Other Conditions**

1. Warranty should be for minimum 2 years from the date of acceptance of External pacemaker and the Warranty has to be given in the name of client.

### **Performance Security**

2. Within Fifteen (15) days from date of the issue of purchase order the successful bidder, shall furnish performance security for an amount equal to ten percent (10%) of the total value of the contract, valid up to sixty (60) days after the date of completion of all contractual obligations by the supplier, including the warranty obligations.
3. Demo of equipments and sample of instruments should be arranged when asked for
4. Earnest Money of Rs. 5,000/- has to be submitted along with the tender in the form of a Demand Draft/ Fixed Deposit Receipt (FDR) of a scheduled bank issued in favour of HLL Lifecare Limited, Thiruvananthapuram which should be placed in a separate sealed cover marked "Earnest Money" shall be submitted along with the tender.
5. The rate quoted should be inclusive of all taxes, duties, transport, loading and unloading up to consignee site.
6. The tenderer shall submit the compliance for technical specifications. The make and Model should also be attached along with the bid.
7. The original Literature or the relevant part of the user/ service manual should be attached as proof.
8. The offer shall be valid for 3 months.
9. The invoice/ bills shall be raised in the name of "**Principal, Thrissur Medical College a/c HLL Lifecare Ltd**" and the same shall be submitted to the HLL Engineer in charge.
10. TIN no. of HLL should not be mentioned in any of the documents as HLL is acting on behalf of the client in the capacity of consultant alone.
11. The completed works shall be handed over to the client after due verification by HLL Engineer in Charge.
12. The bidders should seal and sign the Tender Enquiry Notice and should be submitted along with the offer.

**Deputy Vice President (Tech)**

**HLL Lifecare Limited**

<b>1</b>	<b>EXTERNAL PACEMAKER</b>
<b>1</b>	<b>Technical Specifications</b>
1.1	Should be able to use for asynchronous and synchronous modes
1.2	Should be a dual chamber model
1.3	Base pacing rate should be 30-200 ppm
1.4	Modes : DDD,DOO,DDI,DVI,AAI,AOO,VVI,VOO
1.5	Output amplitude for atrial should be 0.1 to 20 mA and Ventricular should be 0.1 to 25 mA
1.6	AV interval should be 20 to 300 ms
1.7	Atricular sensitivity should be 0.4 to 10 Mv
1.8	Ventricular sensitivity should be 0.8 to 20 mV
1.9	Should work with 9V alkaline or lithium battery
1.1	Should be supplied with all standard accessories at free of cost
<b>4</b>	<b>System Configuration Accessories, spares and consumables</b>
4.1	System as specified-
4.2	10 disposables kits should be provided with equipment
4.3	Consumables should be available for atleast 5 years after the sale of machine.
4.4	All consumables required for installation and standardization of system to be given free of cost.
<b>2</b>	<b>Standards, Safety and Training</b>
2.1	Should have the ISO certification and the copy of the same should be enclosed along with the technical bid.
2.2	The quoted model should have FDA/CE/BIS certificate and copy of the same should be enclosed along with the technical bid.
<b>3</b>	<b>Documentation</b>
3.1	Two numbers of complete User/Technical/Maintenance manuals to be supplied in English .
3.2	Certificate of calibration and inspection from factory to be supplied during delivery of the equipment.
3.3	Warranty & CMC as per tender terms.
3.4	Must submit atleast 3 user list and performance report within last 5 years from major hospitals.
3.5	Must submit atleast 2 nos of latest purchase order of the quoted model dated within 1 year along with the <b>price bid</b> .