

Minutes of the Prebid Meeting held at HLL Bio Medical Office Thycaud on 05/01/2015 at 11.00 am for the Supply of Furniture and Medical Equipment for the Strengthening & Upgradation of Govt: Medical College ,Kozhikode -Phase II.

Ref: HLL/ID/14/66

FURNITURE

Plain glass for table top

Existing Item:

Sl No:	Item	Qty
9	Plain glass for table top Supply and fixing of 10mm glass cut to size and edge finished smoothly for table top use.	230

Read As:

Deleted

MEDICAL EQUIPMENT

Polygraph -16 Channel

Item no:27 Technical Specification

27	POLYGRAPH -16 CHANNEL
27.1	Specification
	No of Channels : 16
	Data Acquisition and analysis Software
	Apparatus for recording and calculating HRV and blood pressure Variability temperature,
	Transducers and softwares for recording and analyzing plethysmography, GSR, Skin temperature,Continuous real-time beat-to-beat blood pressure, Non Invasive Cardiac Out Put, respiration, phonocardiogram and carotid pulse.
	21 inch TFT monitor
	160 GB storage facility and 1GB RAM for the computer
	High resolution, high speed colour laser printer
	Bio Harness ( transmitter / recorder) device with transmit range up to 100m, , memory capacity 480 hours, 250 Hz sampling rate, radio band frequency

**To be read as**

<b>27</b>	<b>POLYGRAPH -16 CHANNEL</b>
<b>27.1</b>	<b>Specification</b>
	No of Channels : 16
	Data Acquisition and analysis Software
	Apparatus for recording and calculating HRV and blood pressure Variability temperature,
	Transducers and softwares for recording and analyzing plethysmography, GSR, Skin temperature, Continuous real-time beat-to-beat blood pressure, Non Invasive Cardiac Out Put, respiration, phonocardiogram and carotid pulse.
	21 inch TFT monitor
	160 GB storage facility and 1GB RAM for the computer
	High resolution, high speed colour laser printer

**The date and time of Submission and Opening of the tender is also amended as follows:**

<b>Sl. No.</b>	<b>Original</b>	<b>Amended</b>
1	Last date and time of Submission of Tender document - 12.01.2015 at 11.00 am	Last date and time of Submission of Tender document - 17.01.2015 at 11.00 am
2	Last date and time of Opening of Technical bid - 12.01.2015 at 11.30 am	Last date and time of Opening of Technical bid - 17.01.2015 at 11.30 am

**The supplier should give CMC for 5 years for the Equipments, after the warranty period of two years .Cost of CMC will not be considered for price comparison purpose.**

**The pre bid meeting minutes shall form part of the tender document and all the bidders are requested to seal and sign the pre bid meeting minutes and should be submitted along with the tender.**