

19-09-2017

Amendment No. 3**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-III/11/ANST/17-18 dated 31-07-2017 read along with Amendment No. 01 dated 22-08-2017 and Amendment No. 02 dated 09-09-2017**

The following changes are being incorporated in the above referred Tender Enquiry Document

**Section VII
Technical Specification**

SCH 01. 12 Channel ECG Machine		
Para	TENDER SPECIFICATION	READ AS
1	Twelve channel 6" or more LCD display for all 12 leads along with on screen details.	Twelve channel 5.7" or more LCD display for all 12 leads along with on screen details.
SCH 02. Blood and Fluid Warmer		
Para	TENDER SPECIFICATION	READ AS
1	Should be able to warm fluid /blood at a temperature range of 37-43 deg C.	Should be able to warm fluid /blood at a temperature range of 37-41 deg C.
SCH 04. High-end Monitor for ICU with CNS		
Para	TENDER SPECIFICATION	READ AS
14	All Monitors should be able to communicate with each other and can display other patient monitor data without the need of central monitor	Deleted (Note: monitor to monitor over view facility already mentioned in para 19, hence deleted from here.)
SCH 10. Ventilator-High End (I.C.U)		
Para	TENDER SPECIFICATION	READ AS
	BOQ	BOQ
	Proximal flow sensor 1 no	Proximal flow sensor 1 no (With necessary hardware and software required in the machine.)
	External Compressor - 1 no	External Compressor - 1 no (if applicable)
SCH 21. Defibrillator with CPR monitoring and TC pacing		
Para	TENDER SPECIFICATION	READ AS
14	Li-Ion smart battery – 1 No (100 shocks at full charge)	Battery – 1 No (100 shocks at full charge)

SCH 22. Defibrillator with ECG Monitor		
Para	TENDER SPECIFICATION	READ AS
4.2	Patient Cable - 02	Deleted (Note: ECG cable already asked in para 4.6, hence deleted from here)

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**