

Amendment No.3

Date: 14.05.2018

Subject: Amendment no. 3 to the Tender Enquiry Document

Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY/AIIMS-II/27/18-19 dated 09.04.2018

The following changes are being incorporated in the above referred Tender Enquiry Document.

Section VII
Technical Specifications

Schedule No. 02 - ADVANCED ELECTRO SURGICAL UNIT WITH VESSEL SEALING (Rfx No. 3000002681)		
Para	CURRENT TENDER SPECIFICATION	READ AS
16.iii.a	Pure cut for clean, precise cut in general surgery having maximum power of atleast 400W	Pure cut/Auto Cut for clean, precise cut in general surgery having maximum power of atleast 300W
16.iii.b	Blend mode for cutting with homeostasis having maximum power of 200 W.	Blend/ Dery Cut mode for cutting with homeostasis having maximum power of 200 W.
17.iv.a	Maryland Jaw Vessel Sealer 5mm 34 to 37 cm for lap surgery 2 Units reusable or 50 Nos. disposable	Maryland/Curved Jaw Vessel Sealer 5mm 34 to 37 cm for lap surgery 2 Units reusable or 50 Nos. disposable
17.iv.b	Maryland Jaw Vessel Sealer 5mm 20 to 23cm for open surgery - 2 Units reusable or 50 Nos. disposable.	Maryland /Bent/ curved Jaw Vessel Sealer 5mm 20 to 23cm for open surgery - 2 Units reusable or 50 Nos. disposable.

Schedule No. 06 Laproscopy Set (Rfx No. 3000002685)		
Para	CURRENT TENDER SPECIFICATION	READ AS
A.6	Video Outputs: DVI, S-Video, Component, Composite	Video Outputs: DVI / HD-SDI /Equivalent HD
B.1	Full High Definition LED Color Monitor 26" Flat Panel, medical grade	Full High Definition Medical Grade LED/LCD Color Monitor 26" or more Flat Panel Monitor
B.3	Composite, S-Video, Component, and DVI inputs	DVI / HD-SDI /Equivalent HD
C.1	Red, Green & Blue LED based cold light source	LED based cold light source
C.2	Color temperature should be around 6400K	Color temperature should be 5800k or more
C.4	Intuitive simple user interface with LCD touch screen with stand-by mode	Deleted
E.3	Ability to transfer captured still and video images to external media using USB and DVD recorder	Ability to transfer captured still and video images to external media using USB
E.7	Video input and output should be HDMI, DVI, S-Video, BNC	Video input and output should be DVI / HD-SDI /Equivalent HD

All other terms and conditions of the tender enquiry remain unaltered.