

Amendment No. 1

Date: 24/12/2015

Sub: Amendment to the Tender Enquiry Document

Ref: NIT No.: HLL/PCD/CDSCO-05/15-16 Dated 10.12.2015

The following changes have been incorporated in the referred NIT.

Section – VII

Technical Specification

Item No. 01

HPLC Quaternary Gradient with UV-VIS Detector with Auto sampler

1. **Existing Specification :**

MULTI WAVELENGTH UV-VIS: Multi wavelength Measurement at four points.

Read as:

Dual Wavelength UV-VIS detector.

2. **Existing Specification :**

**THERMOSTATED COLUMN COMPARTMENT WITH COOLING FACILITY:
WORKABLE TEMPERATURE RANGE - 5 to 80 degree C**

Read as:

**THERMOSTATED COLUMN COMPARTMENT WITH COOLING FACILITY:
WORKABLE TEMPERATURE RANGE - 5 to 85 degree C**

3. **Existing Specification :**

AUTOSAMPLER: Replicate Injection per vial 1-50

Read as:

AUTOSAMPLER: Replicate Injection per vial **1-30**

4. **Existing Specification :**

AUTOSAMPLER: Sample Capacity 100 x 2 ml vials.

Read as:

AUTOSAMPLER: Sample capacity **should be 100 x 2 ml vials or more.**

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids