

**Amendment No. 07**

Date: 06/07/2018

Sub: Amendment No.07 to the Tender Enquiry Document

Ref: (i) Tender No: HITES/PCD/AIIMS-BBSR/01/18-19 dated 20.04.2018 and subsequent amendments published for the referred tender.

**Section I**  
**Notice Inviting Tenders(NIT)**The schedule for submission of tenders for Sch.01, (RFx No. 3000002824), 128-Slice PET/ CT Scanner against the referred NIT is revised as below:**For:- Sch.01, (RFx No. 3000002824), 128-Slice PET/ CT Scanner**

Sl. No.	Description	Schedule
d.	Closing date & time for submission of online bids	10.07.2018, 01:00 PM
c.	Closing date & time for submission of <b>tender processing fee and EMD in physical form*</b>	10.07.2018, 02:00 PM
e.	Time and date of opening of online bids	10.07.2018, 02:30 PM

Read As:

Sl. No.	Description	Schedule
c	Closing date & time for submission of online bids	<b>18.07.2018, 01:00 PM</b>
d	Closing date & time for submission of <b>tender processing fee and EMD in physical form*</b>	<b>18.07.2018, 02:00 PM</b>
e	Time and date of opening of online bids	<b>18.07.2018, 02:30 PM</b>

**Note: If EMD is submitted in the form of BG, then the validity of the BG should be at least 165 days from the date of tender opening.****Section VII**  
**Technical Specifications****Schedule No. 01**  
**128-SLICE PET - CT SCANNER**

Para	Existing specification	Amended As
6	Workstation (Data Acquisition, Processing & Management).	Workstation (Data Acquisition, Processing & Management). <b>Five workstations with concurrent users for all applications in PET and CT.</b>

Para	Existing specification	Amended As
6.c	The processing workstation should be high performance CPU with multi tasking operating system having minimum of 8 GB RAM, 2.53 GHz processor speed, minimum 1GB graphic card, 2TB or more SCSI hard drive (if less, another HDD maybe included), Optical Mouse, Key-board and high resolution flat panel dual view LCD monitor of minimum 19" size with minimum resolution of 1280 x 1024. It should also have CD and DVD combo drive preferably with writer facility.	The processing workstation should be high performance CPU with multi tasking operating system having minimum of <b>4 GB RAM</b> , 2.5 GHz processor speed, minimum 1GB graphic card, 2TB or more SCSI hard drive (if less, another HDD maybe included), Optical Mouse, Key-board and high resolution flat panel dual view LCD monitor of minimum 19" size with minimum resolution of 1280 x 1024. It should also have CD and DVD combo drive preferably with writer facility.

**Note:**

i) The bidders are requested to upload their techno-commercial compliance in the revised excel sheet uploaded in e-tender portal having file name: Sch.01, TC 3000002824 v2.0, based on the revised technical specifications.

ii) Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids

**All other terms and conditions of the tender enquiry remain unaltered**