

Amendment No. 3**Date: 12/08/2014****Sub: Amendment to the tender Enquiry Document****Ref: Tender Enquiry No.: HLL/PCD/Rohtak/01/14-15 dated 24/06/2014
Amendment No. 1 dated 25-07-2014
Amendment No. 2 dated 31-07-2014**

The following changes are incorporated into technical specifications of the referred tender enquiry.

SECTION-I**NOTICE INVITING TENDERS (NIT)****i) Existing:**

Sl. No.	Name of Equipment	Total Qty.	EMD Amount (Rs.)
22	Endobronchial Ultrasound System (EBUS) with Endoscope Reprocessor Unit	1	2,00,000

Read as:

Sl. No.	Name of Equipment	Total Qty.	EMD Amount (Rs.)
22	Endobronchial Ultrasound System (EBUS)	1	2,00,000

ii) Existing:

Sl. No.	Description	Schedule
i.	Dates of sale of tender enquiry documents	20.06.2014 to 13.08.2014, 1000 hrs to 1600 hrs IST
vi.	Closing date & time for receipt of Tender	14.08.2014, 1400 hrs IST
vii.	Time and date of opening of Techno – Commercial tenders	14.08.2014, 1430 hrs IST

Amended as:

Sl. No.	Description	Schedule
i.	Dates of sale of tender enquiry documents	20.06.2014 to 22.08.2014 , 1000 hrs to 1600 hrs IST
vi.	Closing date & time for receipt of Tender	25.08.2014 , 1400 hrs IST
vii.	Time and date of opening of Techno – Commercial tenders	25.08.2014 , 1430 hrs IST

**SECTION - VI
LIST OF REQUIREMENTS**

Part I**i) Existing:**

Sl. No.	Name of Equipment	Departments	Total	Warranty Period (in years)	CMC Period (in years)
		Pulmonary & Critical Care			
22	Endobronchial Ultrasound System (EBUS) with Endoscope Reprocessor Unit	1	1	2	6

Read as:

Sl. No.	Name of Equipment	Departments	Total	Warranty Period (in years)	CMC Period (in years)
		Pulmonary & Critical Care			
22	Endobronchial Ultrasound System (EBUS)	1	1	2	6

**Section – VII
Technical Specifications**

**Item sl. no. 19
Endoscope & Colonoscope**

- 1. Existing Para:-** Compatible 100 – 300 Watts Xenon Light Source with 2 extra Xenon bulbs.

Read as : Compatible 100 – 300 Watts Xenon Light Source with 2 extra Xenon bulbs **or LED based light source.**

Item sl. no. 22

Endo-Bronchial Ultrasound System (EBUS)

1. **Existing Para vi** : Endoscope Reprocessor (Automatic with Ultrasonic Cleaning provision)
Read as Para : Deleted
2. **Existing Para 1.4** : Rigid distal width:Probe not more than 6.5 x 7.0 mm Optic not more than 7.5mm
Read as Para : Distal end : Probe not more than 6.5 x 7.0 mm Optic not more than 7.5mm
3. **Existing Para 2.4** : Total Length – 1850 mm or less
Read as Para : Deleted
4. **Existing Para 3.2** : 300 W xenon lamp with a spare bulb
Read as Para : 100 - 300 W xenon lamp with a spare bulb
5. **Existing Para 5** : Digital Ultrasound Scanner
Read as Para : Digital Ultrasound Scanner / Colour Doppler

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of submission of bids