

Response To Pre-Bid Queries (Pre-Bid date: 24.04.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/14/18-19 Dated: 13.04.2018

Schedule No. 08 - Video Cystoscope(Rfx/Event number 3000002784)

Sl. No	Reference to the bidding document	Specification/Clause of the bidding documents	Name of firms	Queries/Clarification asked by firms	Recommendations/replies by the Committee
1	Page 68 Para 1	Instrument channel: 5.5mm or more	Olympus	Kindly amend instrument channel as 6.5 Fr or more. Remarks: As 6.5 mm is typographical error	Amended as: Instrument channel: 6.5 Fr or more
2	Page 68 Para 3	Light source: Should have following technical specifications/features - 1 no			Deleted
3	Page 68 Para 3	Separate/combined module from video processor			Deleted
4	Page 68 Para 3	Must be latest and high end model from bidder			Deleted
5	Page 68 Para 3	Powerful 300 Watt Xenon Lamp with backup lamp of Halogen/LED	Karl Storz	The back up lamp of Halogen/LED is a very company specific and usually it is not required with powerful xenon 300 watt light source therefore kindly delete back up lamp of Halogen/LED.	Deleted
6	Page 68 Para 3	Lamp life should not be less than 500 hrs (approx) one spare 300W xenon lamp should be supplied	Olympus	Kindly amend it as lamp life should not be less than 500 hrs (approx), one spare 300W xenon lamp should be supplied - In case of LED no spare lamp is required	Deleted
7	Page 68 Para 3	Should have inbuilt air pump	Olympus	Kindly delete as not applicable for Cystoscopy. Remark: Inbuilt Air pump is required in GI Endoscopy	Deleted
8			Karl Storz	Kindly delete inbuilt air pump as it is not required in cystoscopy.	
9	Page 68 Para 4	Video Processor & Light source - 1 no.	Karl Storz	Kindly delete light source as you have already asked the same in point 3.	No Change
10	Page 68 Para 4.1	Outputs - RGB, Y/C, VBS Composite, XGA & DV simultaneous	Karl Storz	Kindly add that video processor should be also HD with resolution of 1920x1080P with DVI output.	Amended as: Outputs - RGB, Y/C, VBS Composite, XGA & DVI simultaneous.
11	Page 68 Para 4.6	Light source - Combined or separate LED with emergency backup facility.	Karl Storz	Delete	Amended as: Combined or separate LED/Xenon light source.
12	Page 68 Para 4.7	Air pump - Inbuilt air pump with minimum two variable air flow control.	Olympus	Kindly delete as not applicable for Cystoscopy	Deleted
13	Page 68 Para 4.8	Lamp can be turned on/off without turning off the equipment.	Karl Storz	Delete	Deleted
14	Page 68 Para 4.9	Electronic magnification up to 1.5X by a touch of scope - note switches	Karl Storz	Delete	Deleted
15	Page 68 Para 4.10	One spare LED Lamp should be supplied.	Olympus	Kindly delete as Led spare lamp is not available	Amended as: 2 nos. spare xenon lamps to be offered in case

Sl. No.	Reference to the bidding document	Specification/Clause of the bidding documents	Name of firms	Queries/Clarification asked by firms	Recommendations/replies by the Committee
17			Karl Storz	Delete as its covered under point 3 of light source.	of xenon light source
18	Page 69 Para 5	5. Monitor - 1 no.			
19	Page 69 Para 5	15 inches or more LCD/LED HD Monitor of Medical Grade. It should be mountable on trolley.	Karl Storz	Kindly amend this to "19 inches or more LCD/LED HD Monitor of Medical Grade. It should be mountable on trolley".	No Change
20	Page 69 Para 6	Computer with Software - 1 no.			
21	Page 69 Para 6	Should be supplied with suitable computer system with facility for recording images and video	Karl Storz	Kindly add "Colour Laser Printer" along with table for computer system.	Amended as: Should be supplied with suitable computer system with facility for recording images and video. Color laser printer and trolley to mount computer system & printer to be supplied
22		BOQ for Video Cystoscope	Karl Storz		
23	BOQ sl.no. 2	Light source as per technical specifications - 1 no.	Karl Storz	should be optionally integrated	Deleted
24	BOQ sl.no. 3	Video Processor & Light source as per technical specifications - 1 no.	Karl Storz	should be optionally integrated	No Change
25	BOQ sl.no. 4	Monitor as per technical specifications	Karl Storz	should be optionally integrated	No Change
26	BOQ sl.no. 5	Computer with Software as per technical specifications - 1 no.	Karl Storz	should be optionally integrated	No Change
27	BOQ sl.no. 6	Trolley as per technical specifications- 1 no.	Karl Storz	can be of other manufacturer	No Change

gm

Savilur

Hy/Geo

Hy/Geo

Hy