

Equipment Specification Data Sheet

Equipment Name: Karl fisher autotitrator
(Volumetric)

Document No.: DS-KFA 01




Revision: 00

Project No.: 120310






Project Name: Integrated Vaccines
Complex, Chengalpattu


Block Code	Block Name	Identification No.	Capacity	Quantity
Q1	Admin, QA & QC	Q1-KFA 01	-	1


NNE Pharmaplan India Limited


Name	Designation	Signature	Date
Prepared by			
Mr. Tushar Shinde	Engineer - Process		11/02/2016
Checked by			
Mr. Yogesha M J	Process Engineer		4/02/2016
Approved by			
Dr. Harshad Mali	Lead - Process Engineer		11/02/2016

HLL Biotech Limited

Name	Designation	Signature	Date
Reviewed by			
SANJEEV KUMAR R	Manager		15/02/2016
VIGNESHWARAN V	DM - PROJECTS		15/2/16
Approved by			
V. SIVAGANESH	Manager		15/02/16
D. SURESH BABU	DM - Q.C. & Q.A		17.02.2016
Authorized by			
RAMAN K. R.	C.E.O.		19.2.16

Equipment Specification Data Sheet		
HLL Biotech Limited, Chennai		
nne pharmaplan®	INTEGRATED VACCINES COMPLEX, CHENGALPATTU	
	Equipment Name	Karl fisher autotitrator (Volumetric)
	Project #	120310
	Document #	DS-KFA 01
		
1	Process requirements	
1.1	Karl fisher autotitrator colourimetric titration is a classic titration method in analytical chemistry that uses coulometric titration to determine trace amounts of water in sample.	
2	Equipment ID	
2.1	Q1- KFA 01	
3	Technical Specification	
3.1	Model	Vendor to specify
3.2	Titration Type	Karl Fischer Volumetric
3.3	Determination Range	100 ppm to 100%
3.4	Choice of end point indication	bi- volumetric or bi- amperometric
3.5	Burette Resolution	1/10000 of Burette cylinder volume
3.6	Titration Dimension	Vendor to specify
3.7	Ambient Temperature	+5 to +40° C
4	Material of Construction	
4.1	Main body	vendor to specify
5	Specific Equipment requirement	
5.1	Should have the facility to STOP the titration by Delay time or the novel "drift stop", where titration stops when a preset drift value is reached	
5.2	The equipment should not have a separate method for drift determination and the drift compensation should be automatic	
5.3	10 ml Burette complete set to be offered along with the system. The burette should have a zero dead volume or close to it and preferably a positive displacement type for safety in case of spillage	
5.4	Automatic recognition of cylinder volume should be built into the instrument preferably with a unique serial number	
5.5	The instrument should be equipped with an intelligent burette capable of storing the complete information on the Titrant like the Titer history, Value, and serial number.	
5.6	Result can be displayed or printed in a choice of units as %, ppm, g/L, mg/L, mol/L, g, mg, ml, mg/pc, no unit & unit programmable	
5.7	Possibility to connect host of accessories for automation	
5.8	A suitable stirrer along with a pump for handling the solvent and the waste should be provided with the system	
5.9	The instrument shall have the facility to measure pH of a solution using a suitable electrode if required.	
5.10	The instrument shall have the facility to do parallel titration if required	
5.11	The instrument should be compatible with an autosampler, for automation of sample analysis if required in the future	
5.12	Latest windows based software should be provided	
6	Other requirement	
6.1	Training and demo for user on operation and cleaning to be provided	
7	Regulatory aspects	
7.1	CE certification	
8	Safety requirements	
8.1	Proper earthing is necessary	
8.2	Appropriate closure of all parts.	

Equipment Specification Data Sheet		
HLL Biotech Limited, Chennai		
nne pharmaplan®	INTEGRATED VACCINES COMPLEX, CHENGALPATTU	
	Equipment Name	Karl fisher autotitrator (Volumetric)
	Project #	120310
	Document #	DS-KFA 01
		
9	Documents	
	Following documents, but not limited to these, are expected from the vendor as part of the supply package in hard copy as well as editable electronic file	
9.1	IOQ documents	
9.2	Operation and maintenance manuals shall be provided along with IOQ documents during installation at site	
9.3	Warranty Letter for Minimum 1 year from the date of supply.	
9.4	Vendor should provide list of standard spare parts with ordering information.	
9.5	NPL traceable calibration certificates and calibration procedures	
9.6	Vendor should provide list of change parts (if applicable) with ordering information	
10	Timelines	
10.1	Not Applicable	
11	Preferred list of Makes	
11.1	Metrohm, Mettler toledo	

Equipment Specification Data Sheet					
HLL Biotech Limited, Chennai					
nne pharmaplan®	INTEGRATED VACCINES COMPLEX, CHENGALPATTU				 HLL BIOTECH LIMITED <small>Established by HLL Biotech Limited in Government of India's Enterprise</small>
	Equipment Name	Karl fisher autotitrator (Volumetric)			
	Project #	120310			
	Document #	DS-KFA 01			
TABLE NO: 1					
Equipment ID	Block Name	Room Name	Room No	Room dimension in mm	Room height in mm
Q1- KFA 01	QC	Chemical Lab	Q1S022	-	-
Table-2: Change Log					
Date	Name	Revision	Section	Change/Comment	
09-09-2015	Tushar Shinde	00	-	New document	
Table-3: Annexure					
Not applicable					

