

## **Amendment No. 1**

**Date: 28/12/2015**

**Sub: Amendment to the Tender Enquiry Document**

**Ref: NIT No.: HLL/PCD/CDSCO-04/15-16 Dated 18.12.2015**

The following is the specification for the item at sl. no. 28 of TED.

### **Section – VII**

#### **Technical Specification**

##### **Item no-28**

##### **Laboratory Spot Extractor**

Laboratory Spot Extractor with high performance, durability, easily movable and excellent performance.

**Design Structure** : Wall Mounted / Ceiling Mounted

#### **Technical Specification:**

Pipe : Dia 75 mm Aluminum Extruded Anodized.  
Fittings : Polypropylene.  
Dome : Dia 385 mm Polypropylene.  
: Inbuilt Damper and exhaust .  
: Made up of anti-corrosive and chemical resistant material.  
: All 'O' Ring made of Nitrile Rubber.  
: should cover about 5 feet working area  
: Protects both user and workplace safely from harmful fumes

#### **Standard, safety and training**

1. Should be FDA/UL/CE/BIS approved product.
2. Manufacturer should have ISO 9001 certificate for quality standards.
3. On site comprehensive training for lab staff and support services till customers satisfaction with the system.
4. **Installation testing:** supplier of the instrument must provide free installation, commissioning and testing.
5. Validation document should supply by vendor etc IQ, OQ, PQ.

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

**Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**