

Amendment No. 7

Date: 09.02.2016

Sub: Amendment to the Tender Enquiry Document

Ref: NIT No.: HLL/PCD/PMSSY-II/NAGPUR/13/15-16 dated 06/11/2015

The schedule for submission of tenders against the referred NIT has been revised as follows:

Sl. No	e- Tender Ref. No (Event No.)	Item Name	Qty.
2	3000000640	Centralized Medical Gas Pipeline system	1

SECTION I**NOTICE INVITING e-TENDERS (NIT)**

Sl. No	Description	Schedule
d	i) Closing date & time for submission of online bids	23-02-2016, 1800 hrs IST
	ii) Closing date & time for tender fee, EMD and bid documents in physical form (Price bid shall only be uploaded online; All other technical and commercial documents uploaded in the e-portal are to be submitted)	24-02-2016, 1200 hrs IST Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
e	Time and date of opening of bids	24-02-2016, 1500 hrs IST

TECHNICAL AMENDMENT**Schedule No. 2****Centralized Medical Gas Pipeline System****1. Existing para as per Amendment no.2:****MEDICAL AIR SYSTEM (Imported):**

Medical quality air shall be delivered at a nominal pressure of 400 kPa (4 bar) gauge for supply of the hospital medical air system. The medical air plant shall deliver both medical and surgical air; with a minimum total flow rate of 2000 l/min +/- 10%.

Read as:

Medical quality air shall be delivered at a nominal pressure of 400 kPa (4 bar) gauge for supply of the hospital medical air system. The medical air plant shall deliver medical air with a minimum total flow rate of 2000 l/min +/- 10%.

2. Existing Para: as per Amendment no. 3

It must consist of the following – 1 no. suction regulator and 2 nos. 3000-4000ml polysulfone/ polycarbonate collection jar and both to be mounted on a trolley

Read as:

It must consist of the following – 1 no. suction regulator and 2 nos. 1800ml or more polysulfone/ polycarbonate collection jar and both to be mounted on a trolley

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids