

AMENDMENT NO.2 Dated 22.05.2020 for change in specification of goggles and latex gloves in tender Ref No: TENDER NO: HLL/SD/RBD/2019-20/TENDER/15 DT:05.03.2020 WHICH HAS NOW BEEN CLOSED, FRESH REQUEST FOR PROPOSAL FOR THE SUPPLY OF PERSONAL PROTECTIVE EQUIPMENTS - GOGGLES, N95 FACE MASK WITH / WITHOUT EXPIRATORY VALVE AND NITRILE GLOVES floated on 11 May 2020.

The following amendment has been incorporated to the bid document for the above tender:

Specification

1. Goggles

1. With transparent glasses, Zero power, well fitting, covered from all sides with elastic band/or adjustable holder.
2. Good seal with the skin of the face.
3. Flexible frame to easily fit all face contours without too much pressure.
4. Covers the eyes and surrounding areas and accommodates for prescription glasses.
5. Fog and scratch resistant.
6. Adjustable band to secure firmly so as not to become loose during clinical activity.
7. Indirect venting to reduce fogging.
8. May be re-usable (provided appropriate arrangements for decontamination are in place) or disposable.
9. Quality compliant with the below standards, or equivalent:
 - a. EU standard directive 86/686/EEC, EN 166/2002
 - b. ANSI/SEA Z87.1-2010
 - c. IS 5983-1980, reaffirmed in 2018

MAY BE READ AS

1. Goggles

1. Made of transparent scratch resistant material, Zero optical power, well fitting, covered from all sides with elastic band/or adjustable holder.
2. Good seal with the skin of the face.
3. Flexible frame to easily fit all face contours without too much pressure.
4. Covers the eyes and surrounding areas and accommodates for prescription glasses.
5. Fog and scratch resistant.
6. Adjustable band to secure firmly so as not to become loose during clinical activity.
7. Indirect venting to reduce fogging.
8. May be re-usable (provided appropriate arrangements for decontamination are in place) or disposable.
9. Quality compliant with the below standards, or equivalent:
 - d. EU standard directive 86/686/EEC, EN 166/2002
 - e. ANSI/SEA Z87.1-2010
 - f. IS 5983-1980, reaffirmed in 2018

2. Gloves in pairs of Size 6.5 & 7.0 each

- Nitrile
- Non-sterile
- Powder free
- Outer gloves preferably reach mid-forearm (minimum 280mm total length)
- Different sizes in pairs of 6.5 & 7.0 each
- Quality compliant with the below standards, or equivalent:
 - a. EU standard directive 93/42/EEC Class I, EN 455
 - b. EU standard directive 89/686/EEC category III, EN 374
 - c. ANSI/SEA 105-2011
 - d. ASTM D6319-10
- Packing – The Product should be supplied in boxes of 100 pieces (The gloves of size 6.5 & 7.0 shall be individually packed).

OR

Latex Medical Examination Gloves

- Ambidextrous shape – fits either hand

- Size: Medium and Large in pairs
- Single use
- Non Sterile
- Should conform to IS 15354 Part I 2018 or EN455 or ISO 11193 -1 of 2008.
- Expiry: minimum 3 years from date of manufacture
- Label on primary packaging – name and/ or trademark of manufacturer, manufacturer's product reference, expiry date by year and month

MAY BE READ AS

2. Gloves in pairs of Size 6.5 & 7.0 each

- Nitrile
- Non-sterile
- Powder free
- Outer gloves preferably reach mid-forearm (minimum 280mm total length)
- Different sizes in pairs of 6.5 & 7.0 each
- Quality compliant with the below standards, or equivalent:
 - a. EU standard directive 93/42/EEC Class I, EN 455
 - b. EU standard directive 89/686/EEC category III, EN 374
 - c. ANSI/SEA 105-2011
 - d. ASTM D6319-10
- Packing – The Product should be supplied in boxes of 100 pieces (The gloves of size 6.5 & 7.0 shall be individually packed).

OR

Latex Medical Examination Gloves

- Ambidextrous shape – fits either hand
- Size: Medium and Large in pairs
- Length: minimum 230mm total length
- Disposable powder free latex gloves
- Single use
- Non Sterile
- Should conform to IS 15354 Part I 2018 or EN455 or ISO 11193 -1 of 2008.

- Expiry: minimum 3 years from date of manufacture
- Label on primary packaging – name and/ or trademark of manufacturer, manufacturer's product reference, expiry date by year and month

All other relevant specifications and clauses of the tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

**Vice President
Sourcing Division
HLL Lifecare Ltd.
HLL Bhavan, Poojappura,
Thiruvananthapuram.**

Offers may be submitted to:
sdcovidep.hll@lifecarehll.com / sdcovidep.hll@gmail.com
Contact: 0471 2354949 Ext: 323 (9.15 AM to 5.45 PM)