

Clarification Regarding Technical Specifications (Part I)

Date: 14/05/2014

Subject: Clarification regarding Technical Specifications

**Ref: (i) Tender Enquiry No.: HLL/PCD/PMSSY/AIIMS-II/12/13-14 dated 03/03/2014.
(ii) Amendment No.1 dated 15/04/2014.
(iii) Amendment No.2 dated 03/05/2014.**

The following clarifications regarding technical specifications are made further to those referred above.

Section – VII
Technical Specifications

Schedule No. 1

ETO Sterilizer

1. Capacity: Our specification for capacity is 200 litres or more. Conversion from cu.ft to litres is as per standard conversion. For information, 1 cu.ft is 28.317 litres approximately.
2. Five or more operating temperatures in a range of 33 deg C - 55 deg C.
3. Requirement is 5 nos. ETO steriliser, capacity of each 200 litres or more
4. Turn key is related to installation and commissioning of ETO sterilisers only.
5. Furniture's and fixtures to be supplied as required for functioning of ETO sterilisers
6. Appropriate pollution control device for safe disposal of E.O like catalytic converter or equivalent technology gas disposal management to be as per local pollution control norms. Apparently, it seems there is no local pollution control norm for ETO steriliser.
7. Compliance is as per International norms.

All other terms and conditions of the tender enquiry remain unaltered.