



DATA SHEET				
HLL BIOTECH LIMITED, CHENNAI				
nne pharmaplan®		INTEGRATED VACCINES COMPLEX, CHENGALPATTU		 <div>HLL BIOTECH LIMITED (Subsidiary of HLL Lifecare Limited) (A Government of India Enterprise)</div>
		BUFFER VESSEL		
		PROJECT NO:	120310	
		EQUIPMENT ID:	F1-BUV 01	
		DOCUMENT NO:	DS/F1-BUV 01	
1	Process Requirement			
1.1	Buffer vessel having formulated vaccine shall be used in filling area and the same shall be attached to the manifold of filling machine.			
2	Technical Specifications			
2.1	Model	cGMP		
2.2	Type	Buffer vessel for filling machine		
2.3	Capacity	12 L Geometric volume (10 L working volume)		
2.4	Quantity	1 no.		
2.5	Design	Shell	Cylindrical	(Vessel Quality Certification- ASME BPE)
		Top dish	Flat type	
		Bottom dish	Torispherical type	
3	Material Of Construction			
3.1	Shell	SS 316L		
	Top dish	SS 316L		
	Bottom dish	SS 316L		
3.2	Gaskets/ O-rings	PTFE / PVDF / VITON.		
3.3	Surface finish	Internally Electro polished Ra≤0.6 microns, conforming to SFC4, according to ASME BPE(2009).		
		Externally mechanically polished up to ≤1.2 Ra		
		Flexible hosing shall be used (FDA approved material) for connections between manifold of the filling line and buffer vessel.		
3.4	Nozzles or ports	SS 316 L		
3.5	Nuts - bolts (internal, external, insulation support)	SS 316 L		
4	Specific Requirements			
4.1	Nozzles Schedule :			
	Top Flat Head			
	• Top lid will be of TC connection with minimum of 4 inches diameter.			
	• 1no. Port for product inlet with S2S connection (TC clamps with gasket and Serrated nipple) for bulk addition.			
	• 1no. Port for vent filter with TC connection.			
	• 1no. Port for product recirculation S2S connection (TC clamps with gasket and Serrated nipple)			
	• 1no. Port for load cell.			
	• 1no. spare port.			
4.2	Bottom Torispherical dish			
	• Vessel should have bottom discharge through actuated flush bottom valve and S2S connection for connecting the manifold of filling machine.			
	• Fabrication of equipment and accessories should comply with the latest GMP standards.			
	• There should not be any sharp edges/Corners, crevices, pin holes in the process wetted parts of the equipment.			
	• All welded joints, internal or external, shall be buffed and smooth for easy cleanability.			
4.3	All nozzle pipes shall be seamless, unless otherwise specified.			
4.4	Flexible hosing/ tube shall be considered for transferring of the product, MOC: Silicone platinum cured.			
4.5	Top lid shall be openable / flanged type.			
4.6	Vessel should have one number autoclaveable disc filter.			
4.7	The vessel shall be autoclaveable.			
4.8	Vendor to ensure 100 % drainability of the product.			
4.9	GA drawing of the vessel and arrangement of the vessel should be provided by the vendor during the techno-commercial quotation.			
5	Documents Required			

DATA SHEET						
HLL BIOTECH LIMITED, CHENNAI						
nne pharmaplan®	INTEGRATED VACCINES COMPLEX, CHENGALPATTU				 HLL BIOTECH LIMITED <small>(Subsidiary of HLL Lifecare Limited) (A Government of India Enterprise)</small>	
	BUFFER VESSEL					
	PROJECT NO:	120310				
	EQUIPMENT ID:	F1-BUV 01				
	DOCUMENT NO:	DS/F1-BUV 01				
5.1	DQ Document					
5.2	MOC certificates					
5.3	Test Certificate					
5.4	List of MAKE with certificate (to be used during fabrication of this unit)					
	NOTE: Accurate size and technical specification need to be mentioned by the vendor.					
	AFI Approved for Enquiry AFO Approved for Ordering					
0	26.08.2013	MJY	SDBB	<input type="checkbox"/>	<input type="checkbox"/>	
Rev	Date	Completed By	Checked By	AFI	AFO	Sheet 1/1