

27-04-2018

Amendment No. 1**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PMSSY-III/26/CAD/18-19 dated 10/04/2018**

The following changes are being incorporated in the above referred Tender Enquiry Document.

**Section VII
Technical Specification****Schedule No. 03
Ambulatory BP Monitor**

Para	TENDER SPECIFICATION	READ AS
Para 6	Clock Display : 24 hour (1997-2096 and auto leap year setting)	Deleted
Para 10	Dimensions : 2.8"W x 1.1"H x 3.9"D 72mm W x 27mm H x 100mm D	Deleted
Para 11	Weight: 0.5lbs (215g)	Weight: <250gm

**Schedule No. 05
Colour Doppler 4D**

Para	TENDER SPECIFICATION	READ AS
Para 9.1	Scanning Parameters The system should have minimum 65,000 digital system processing channels.	Scanning Parameters The system should have minimum 1,50,000 digital system processing channels.
		Added Para: System should have High Resolution LCD/LED Touch Screen Control for ease of operation.

All other contents of the tender enquiry including terms & conditions remain unaltered.**Note:**

- i. **Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**