

Amendment No. 5

Date: 18.01.2016

Sub: Amendment to the Tender Enquiry Document**Ref: NIT No.: HLL/PCD/PMSSY-II/NAGPUR/11/15-16 dated 06/11/2015****SECTION I**
NOTICE INVITING e-TENDERS (NIT)

The following line item from the tender enquiry ref. no HLL/PCD/PMSSY-II/NAGPUR/11/15-16 dated 06/11/2015 has been deleted.

Sl. No	e- Tender Ref. No (Event No.)	Item Name	Qty	EMD (Rs.)
3	3000000643	Ortho drill & saw system	2	60,000/-

TECHNICAL SPECIFICATIONS**Schedule No. 1**
C Arm 3D/CT for Ortho

- Existing Para Image Intensifier:** Open interface for 3D surgical navigation. Compatible with Nav-3 (Stryker) for fluoro **and CT navigation**
Read As: Isocentric C Arm with Open interface for 3D surgical navigation. Compatible with Nav-3 (Stryker) for fluoro
- Existing Para X Ray Tube Anode:** 120vac (Maximum output): 1.4kW
Read As: Deleted
- Existing Para X Ray Tube Anode:** 220vac (Maximum output): 1.4kW
Read As: Maximum output: 2kW or more
- Existing Para X Ray Tube Anode:** Heat capacity, hu (X Ray Tube) 50000 anode 1200000 unit
Read As: Heat capacity, hu (X Ray Tube) 45000 anode 1200000 unit
- Existing Para Image Processing and storage:** Frame Integration (Image Processing and Storage): 0-32 Frame
Read As: Frame Integration (Image Processing and Storage): 0-30 Frame
- Existing Para Image Processing and storage:** Max No. of Images: Up to 900 on HD; CD burner standard
Read As: Max No. of Images: Up to 10,000 on HD; CD burner standard
- Existing Para Image Processing and storage:** Image matrix size: 512 X 444
Read As: Image matrix size: 1024 X 1024

- 8. Existing Para C Arm:** Free Space, cm(in): 78(31)
Read As: Free Space: **75 cm or more**
- 9. Existing Para C Arm:** H x W x D of C Arm Frame: 183x80x214.6
Read As: Deleted
- 10. Existing Para C Arm:** Weight: 350kg, 147kg Monitor and cart
Read As: Deleted
- 11. Existing Para Generator:** Power Rating: 1.4 kW
Read As: Power Rating: **2 kW or more**
- 12. Existing Para Generator:** Pulse Per Second(Monitor): 2.00
Read As: Pulse Per Second (Monitor): **8.00**

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids