

AMENDMENT NO.1 Dated 18.June.2017

Ref Tender No: HLL/SD/CHO/2017-18/TENDER/003

The following amendments have been incorporated to the bid document for the above Tender for supply of films, contrast and DICOM printing on paper for HLL diagnostic imaging centres across India

SI No	Original Tender Detail:	Amended tender details:
1.	<p>ANNEXURE 1</p> <p>SI NO:4</p> <p>Laser Film</p> <p>Suitable Dry Chemsitry film printer with following specification shall be provided free of cost:</p> <ol style="list-style-type: none"> 1.Resolution: 16 bits/ 600 dpi or more 2. With minimum three ports. 3.Support Multiple Film Sizes: one of which must be 17"x14" 4. DICOM Compatible (Attach conformance statement). 	<p>ANNEXURE 1</p> <p>SI NO:4</p> <p>Laser Film</p> <p>Suitable Dry Chemsitry laser FILM PRINTER with following specification shall be provided free of cost:</p> <ol style="list-style-type: none"> 1.Resolution: 16 bits/ 500 dpi or more Minimum One tray, Support Multiple Film Sizes: one of which must be 14"x17". (adaptable for film size of 10 X 8, 12 X 10 and 17 X 14) 3. DICOM Compatible (Attach conformance statement).
2.	<p>Terms and Condition</p> <p>5. Validity: The quoted rates must be valid for a period for 3 Years from the date of closing of the tender. The overall offer for the assignment and bidder(s) quoted price shall remain unchanged during the period of validity. If the bidder quoted the validity shorter than the</p>	<p>Terms and Condition</p> <p>5. Validity: The quoted rates must be valid for a period for 3 Years from the date of closing of the tender. Validity may be further extended upto 2 years based on mutual consent. The overall offer for the assignment and bidder(s) quoted price shall remain unchanged during the period of validity. If</p>

AMENDMENT NO.1 Dated 18.June.2017

	<p>required period, the same will be treated as unresponsive and it may be rejected.</p> <p>In case the tenderer withdraws, modifies or change his offer during the validity period, bid is liable to be rejected and the earnest money deposit shall be forfeited without assigning any reason thereof. The tenderer should also be ready to extend the validity, if required, without changing any terms, conditions etc. of their original tender.</p>	<p>the bidder quoted the validity shorter than the required period, the same will be treated as unresponsive and it may be rejected.</p> <p>In case the tenderer withdraws, modifies or change his offer during the validity period, bid is liable to be rejected and the earnest money deposit shall be forfeited without assigning any reason thereof. The tenderer should also be ready to extend the validity, if required, without changing any terms, conditions etc. of their original tender.</p>
<p>3.</p>	<p>ANNEXURE 1</p> <p>1.B Approximate requirement of films, Contrast and Dicom Paper for Hindlabs imaging centre (Indicative).</p> <p>SL NO 4 Laser film</p> <p>Approximate Quantity required per Annum (Indicative)</p> <p>9,72,500 Numbers</p>	<p>ANNEXURE 1</p> <p>1.B Approximate requirement of films, Contrast and Dicom Paper for Hindlabs imaging centre (Indicative).</p> <p>SL NO 4 Laser film</p> <p>Approximate Quantity required per Annum (Indicative)</p> <p>9,72,500 Numbers</p> <p>The annual requirement of Film quantity - 95 % of total quantity is 17 X 14 size and 5% other size (10 X 8 and 12 X10)</p>

AMENDMENT NO.1 Dated 18.June.2017

4. Amended tender details

The below mentioned point shall be applicable for the MRI contrast Medicine, CT Intravenous Contrast Medicine and CT oral Contrast Medicine:

Supplier shall submit three years market standing certificate for MRI contrast Medicine, CT Intravenous Contrast Medicine, CT oral Contrast Medicine.

All other terms and conditions in the original Tender Notice and tender documents remain same and unchanged.

For HLL Lifecare Ltd.

Senior Manager (SD)