

27-07-2017

Amendment No. 05

Sub: Amendment to the referred tender enquiry

Ref.: Tender Enquiry HITES/PCD/PMSSY-III/07/CTVS/2017-18 dated 08-06-2017 read with Amendment No. 01 dated 03-07-2017, Amendment No. 02 dated 10-07-2017, Amendment No. 03 dated 20-07-2017 & Amendment No. 04 dated 25-07-2017

The following changes have been incorporated in the technical specifications of the referred tender enquiry.

**Section VII
Technical Specification**

SCH 02. IABP (Intra-Aortic Balloon Pump)		
Para	Existing Specification	Read As
3.1	Pneumatics: Drive system: Dual head reciprocating compressor Drive gas- Helium (Available with disposable canister or refillable cylinder. Pumping Volume: 0.5 cc-50 cc Counter pulsation rate: 40-200 pulsations per minute	Pneumatics: Drive system: Dual head reciprocating compressor or stepper motor driven bellows Drive gas- Helium (Available with disposable canister or refillable cylinder. Pumping Volume: 0.5 cc-50 cc Counter pulsation rate: 40-200 pulsations per minute

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. **Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**