

**Prebid queries/clarifications asked by bidders and its recommendations/replies (ref: prebid meeting dated:09/08/2018)**

NIB Ref: HITES/PCD/NCI-AIIMS/25/18-19

Schedule No. 1- Rfx/Event Number - 3000003197 - Mobile OT Light - 9 No.

Sl. No.	Ref. to the Bidding Document	Specification/Clauses in the Bidding Document	Bidder	Queries/clarifications asked by Bidders	Recommendations/ replies by the Committee
1	Page 46,point1	Should be portable Cool and homogeneous mobile light with LED technology	Karl stroz	Kindly add: Ra = 96 or more and R9 = 95 or more, Color Temperature 4200K	<b>Added Para:</b> Ra=95 or more, Ri=95 or more
2	Page 46,point2	Preferably the light should have less number of LEDs with low power consumption.	Karl stroz	Kindly add: Not more than 8w.	<b>No Change</b>
3	Page 46,point3	There should be single dome equipped with not more than 20 LEDs and minimum 50,000 lux (measured at 1 meter distance).	Karl stroz	Kindly amend this as "There should be single dome equipped with not more than 5 LEDs and maximum 65,000 lux (measured at 1 meter distance). Variable 30000 lux to 65000 lux.	<b>To be Amended as:</b> There should be single dome equipped with <b>not more than 26 LEDs</b> and minimum 50,000 lux (measured at 1 meter distance).
4			Getinge	There should be single dome equipped with not more than 26 LEDs and minimum 50,000 lux (measured at 1 meter distance).  We suggest more LEDs result in better Light shadow resolution, we have 26 LEDs in our Light Head.	
5	Page 46,point4	Total power consumption should not be more than 30 W +/- 5%.	Karl stroz	Kindly amend this as "Total power consumption should not be more than 8W +/- 5%".	<b>No Change</b>
6	Page 46,point6	Light field diameter (D50) not less than 80 mm & Depth of illumination not less than 150 mm	Karl stroz	Kindly amend this as "Light field diameter (D10) not less than 140 mm & Depth of illumination not less than 1250 mm"	<b>To be Amended as:</b> Light field diameter (D50) not less than 80 cm & Depth of illumination not less than <b>125 cm</b>
7	Page 46,point7	LED's average life of not less than 50000 Hours.	Karl stroz	Kindly amend this as "LED's average life of not less than 60000 Hours."	<b>No Change</b>
8	Page 46,point 8	Inbuilt 30 min battery backup should be offered.	Getinge	to be deleted As all the OT's are UPS equipped so battery backup not required. Also additional battery will make the unit cumbersome.	<b>Deleted</b>
9	Page 46,point9	All components should be CE or FDA or BIS certified with 4 digit notified body number	Karl stroz	Kindly amend this as "All components should be European CE or US FDA or BIS certified with 4 digit notified body number. Certificate of compliance to MDD 93/42/EEC 2007 Medical Device Directive should be submitted".	<b>No Change</b>
10		Suggested by	Getinge	Colour Temperature (K) 4,300 Colour Rendering Index (CRI) (Ra) 95 Radiant energy < 3.9(mW/m <sup>2</sup> .lx)  These are the necessary required parameter for Mobile OT light.	<b>No Change</b>

Schedule No. 2- Rfx/Event Number - 3000003198 -Ultrasonic Bone Scalpel - Qty 3 No.

Sl. No.	Ref. to the Bidding Document	Specification/Clauses in the Bidding Document	Bidder	Queries/clarifications asked by Bidders	Recommendations/ replies by the Committee
1	Page 47 ,point 2	System should be able to handle all types of cortical bones (D1, D2, D3 types).	Cure Surgical	System should be able to handle all types of cortical bones (D1,D2,D3 types ). Ultrasonic system provide non- abrasive ,compression cut with minimul bone loss in combination not more than 0.5 - 1.0 mm thickness blade.	No Change
2	Page 47 ,point 5	Hand piece with LED mounted for illumination of surgical field	Cure Surgical	Delete	No Change
3	Page 47 ,point 6	The standard configuration of the system should have a set of surgical angled tips, angled Right & Left tips (accessible to all surgical areas) for all kind of interventions with facility for non-linear transverse oscillations.	Cure Surgical	The standard configuration of the system should have a set of Bone scalpel blade blunt tips , 8-10 mm ,18-20mm ,18-20mm unilateral seration. 2mm shaver & 4 mm shaver Right angled (accessible to all surgical areas) for all kind of interventions with facility for linear oscillations for effortless disscetion of hard cortical bone.	No Change
4	Page 47 ,point 9	Should be supplied with 10 irrigation sets.	Cure Surgical	The system should have integrated in line irrigation which should run through the hand piece and through the blade.	No Change
5	Page 47 BOQ point 12	The system should be qucted with the following accessories a. Osteotomy saw blades – 05 Nos. b. Anglec Right and left surgical tips – 05 Nos. c. Diamond finishing blade of differen: shapes – 05 Nos. d. Irrigation sets - 10 Nos	Cure Surgical	a)Osteotomy saw blades / 18-20mm blade blunt tip with unilateral saretion -05 Nos b)Angled Right and left Tips /Shaver Tips /Shaver qty -05 Nos . c)Diamond finishing blade/ shaver - 05 nos . D)irrigation sets :- 10 Nos e) 8-10mm blade blunt tips - 5 nos.	No Change

*JA*

*Shard*

*See*

*in*

*BA*

## Schedule No. 7- Rfx/Event Number - 3000003202, laproscopy Instrument Set Qty-6

Sl. No.	Ref. to the Bidding Document	Specification/Clauses in the Bidding Document	Bidder	Queries/clarifications asked by Bidders	Recommendations/ replies by the Committee
1	page 53 point a)	Single piece instruments:(Qty 2 nos. each	Karl stroz	we must emphasize that all instruments should be 3 part dismantal for better cleaning, and durability of instruments.	<b>To be Amended as:</b> Single or <b>three piece</b> instruments: (Qty 2 nos. each
2	page 53 point a)	Single piece instruments:(Qty 2 nos. each	DSS Imagetech Pvt Ltd	Kindly amend it as all the hand instrumentsshould be of length of <b>33-36 cm</b>	
3	page 53 point a.1)	Dissecting and grasping forcecs double action, <b>dolphin nose</b> , size 5 mm, length 36 cm			<b>To be Amended as:</b> Dissecting and grasping forceps double action, dolphin nose, size 5 mm, length <b>33-36 cm</b>
4	page 53 point a.2)	Dissecting and grasping forceps with, double action size 5 mm, length 36 cm			<b>To be Amended as:</b> Dissecting and grasping forceps with, double action size 5 mm, length <b>33-36 cm</b>
5	page 53 point a.3)	<b>REDDICK-OLSEN</b> dissecting and grasping forceps, double action, atraumatic, fenestrated, size 5 mm, length 36 cm			<b>To be Amended as:</b> REDDICK-OLSEN dissecting and grasping forceps, double action, atraumatic, fenestrated, size 5 mm, length <b>33-36 cm</b>
6	page 53 point a.4)	<b>BABCOCK</b> grasping forceps double action, atraumatic, fenestrated, size 5 mm, length 36 cm			<b>To be Amended as:</b> BABCOCK grasping forceps double action, atraumatic, fenestrated, size 5 mm, length <b>33-36 cm</b>
7	page 53 point a.5)	Grasping forceps double action, atraumatic, <b>fenestrated, with multiple teeth</b> , size 5 mm, length 36 cm			<b>To be Amended as:</b> Grasping forceps double action, atraumatic, fenestrated, with multiple teeth, size 5 mm, length <b>33-36 cm</b>
8	page 53 point a.6)	Grasping forceps double action, <b>spoon-shaped, with multiple teeth</b> , size 5 mm, length 36 cm	DSS Imagetech Pvt Ltd	Kindly amend it as Grasping forceps double action,, with multiple teeth, size 5 mm, length 33-36 cm	<b>To be Amended as:</b> Grasping forceps double action, spoon-shaped or similar, with multiple teeth, size 5 mm, length <b>33-36 cm</b>
9	page 53 point a.7)	<b>DeBAKEY</b> grasping forceps double action, atraumatic, size 5 mm, length 36 cm			<b>To be Amended as:</b> DeBAKEY grasping forceps double action, atraumatic, size 5 mm, length <b>33-36 cm</b>

NIB Ref: HITES/PCD/NCI-AIIMS/25/18-19

Schedule No. 7- Rfx/Event Number - 3000003202, laproscopy Instrument Set Qty-6

10	page 53 point a.8)	Grasping forceps double action with wavy serration, size 5 mm, length 36 cm			<b>To be Amended as:</b> Grasping forceps double action with wavy serration, size 5 mm, length <b>33-36 cm</b>
11	page 53 point a.9)	Grasping forceps double action, atraumatic, hollow jaws, size 5 mm, length 36 cm			<b>To be Amended as:</b> Grasping forceps double action, atraumatic, hollow jaws, size 5 mm, length <b>33-36 cm</b>
12	page 53 point a.10)	<b>Bowel Grasper</b> double action, atraumatic, fenestrated, size 5 mm, length 36 cm			<b>To be Amended as:</b> Bowel Grasper double action, atraumatic, fenestrated, size 5 mm, length <b>33-36 cm</b>
13	Page 54 point b)	General Specifications for 3 piece instruments (Qty 2 nos. each) Working length = 36cm for all instruments	Pridex	Working Length 33 – 36 CM for Instruments  Standard working length in Laparoscopy is <b>33 -36 cm</b> . Even Needle holder asked at point number G is of 33 cm	<b>Addressed at below para</b>
14	b.1	KELLY Dissecting and Grasping Forceps, with connector pin for unipolar coagulation, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, long, double action, plastic handle without ratchet.			<b>To be Amended as:</b> KELLY Dissecting and Grasping Forceps, with connector pin for unipolar coagulation, rotating, with connector pin for unipolar coagulation, size 5 mm, length <b>33-36 cm</b> , long, double action, plastic handle without ratchet
15	b.2	CROCE-OLMI Grasping Forceps, single action with connector pin for unipolar coagulation, rotating, size 5 mm, length 36 cm, long, plastic handle without ratchet.			<b>To be Amended as:</b> CROCE-OLMI Grasping Forceps, single action with connector pin for unipolar coagulation, rotating, size 5 mm, length <b>33-36 cm</b> , long, plastic handle without ratchet
16	b.3	Bowel Grasper, rotating, double action jaws, with connector pin for unipolar coagulation, fenestrated, size 5 mm, length 36 cm, long, plastic handle without ratchet			<b>To be Amended as:</b> Bowel Grasper, rotating, double action jaws, with connector pin for unipolar coagulation, fenestrated, size 5 mm, length <b>33-36 cm</b> , long, plastic handle without ratchet

*(Handwritten signatures and initials in blue ink)*

NIB Ref: HITES/PCD/NCI-AIIMS/25/18-19

Schedule No. 7- Rfx/Event Number - 3000003202, laproscopy Instrument Set Qty-6

17	b.4	Right Angled Dissecting and Grasping Forceps, double action jaws, with connector pin for unipolar coagulation, fenestrated, rotating, size 5 mm, length 36 cm, long, plastic handle without ratchet			<b>To be Amended as:</b> Right Angled Dissecting and Grasping Forceps, double action jaws, with connector pin for unipolar coagulation, fenestrated, rotating, size 5 mm, length <b>33-36</b> cm, long, plastic handle without ratchet
18	b.5	MANHES Grasping Forceps, single action, Tiger Jaws, 2X4 teeth, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, long, plastic handle with Manhes type ratchet			<b>To be Amended as:</b> MANHES Grasping Forceps, single action, Tiger Jaws, 2X4 teeth, rotating, with connector pin for unipolar coagulation, size 5 mm, length <b>33-36</b> cm, long, plastic handle with Manhes type ratchet
19	b.6	Anvil Graspers, double action, with connector pin for unipolar coagulation, rotating, double action jaws, fenestrated, size 5 mm, length 35 cm, long, Plastic handle without ratchet.			<b>To be Amended as:</b> Anvil Graspers, double action, with connector pin for unipolar coagulation, rotating, double action jaws, fenestrated, size 5 mm, length <b>33-36</b> cm, long, Plastic handle without ratchet
20	b.7	MANHES Dissecting Grasping Forceps, single action, with connector pin for unipolar coagulation, duck-bill jaws length 15mm, blunt, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, long, plastic handle without ratchet			<b>To be Amended as:</b> MANHES Dissecting Grasping Forceps, single action, with connector pin for unipolar coagulation, duck-bill jaws length 15mm, blunt, rotating, with connector pin for unipolar coagulation, size 5 mm, length <b>33-36</b> cm, long, plastic handle without ratchet
21	b.8	CLAW Grasping Forceps, single action, with ratchet, 2X3 teeth, single action jaw, rotating, size 10 mm, length 35 cm, long.			<b>To be Amended as:</b> CLAW Grasping Forceps, single action, with ratchet, 2X3 teeth, single action jaw, rotating, size 10 mm, length <b>33-36</b> cm, long.

Schedule No. 7- Rfx/Event Number - 3000003202, laparoscopy Instrument Set Qty-6

22	b.9	METZENBAUM Scissors, double action, curved, rotating, dismantling, with connector pin for unipolar coagulation, with luer-lock connection for cleaning, size 5 mm, length 36 cm, long plastic handle without ratchet.			<b>To be Amended as:</b> METZENBAUM Scissors, double action, curved, rotating, dismantling, with connector pin for unipolar coagulation, with luer-lock connection for cleaning, size 5 mm, length <b>33-36 cm</b> , long, plastic handle without ratchet.
23	b.10	Hook Scissors, single action, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, long plastic handle without ratchet.			<b>To be Amended as:</b> Hook Scissors, single action, rotating, with connector pin for unipolar coagulation, size 5 mm, length <b>33-36 cm</b> , long plastic handle without ratchet.
24	b.11	Dissecting-Electrode L-tip, size 5 mm, length 36 cm, with connector pin for unipolar coagulation			<b>To be Amended as:</b> Dissecting-Electrode L-tip, size 5 mm, length <b>33-36 cm</b> , with connector pin for unipolar coagulation
25	b.14	CUSCHIERI Retractor, size 10 mm, length 36cm			<b>To be Amended as:</b> CUSCHIERI Retractor, size 10 mm, length <b>33-36cm</b>
26	b.15	Grasping forceps, with especially fine atraumatic serrated, fenestrated rotating 5mm size, length 36cm with plastic handle with out ratchets, single action jaw			<b>To be Amended as:</b> Grasping forceps, with especially fine atraumatic serrated, fenestrated rotating 5mm size, length <b>33-36cm</b> with plastic handle with out ratchets, single action jaw
27			DSS Imagetech Pvt Ltd	Kindly amend it as ALL the hand instruments should be of length of <b>33-36 cm</b>	<b>Already addressed</b>
28	Page 54 point b).16	ROTHERBERG Bowl grasper, rotating 5mm size, length 36cm with plastic handle without ratchets	Pridex	Bowel grasper, rotating 5mm size, length 33 - 36cm with plastic handle without ratchets  ROTHERBERG – is company specific. Standard working length in Laparoscopy is 33 -36 cm	<b>To be Amended as:</b> <b>ROTHERBERG or similar Bowl grasper</b> , rotating 5mm size, length <b>33-36cm</b> with plastic handle without ratchets

NIB Ref: HITES/PCD/NCI-AIIMS/25/18-19

Schedule No. 7- Rfx/Event Number - 300003202, laparoscopy Instrument Set Qty-6

29	Page 54 point b).17	MOURET dissecting and grasping forceps, slender, atraumatic fenestrated rotating 5mm size, length 36cm with plastic handle with out ratchets	Pridex	Fissecting and grasping forceps, slender, Atraumatic fenestrated rotating 5mm size, length 36cm with plastic handle with out ratchets.  MOURET – is company specific. Standard working length in Laparoscopy is 33 -36 cm	To be Amended as: <b>MOURET or similar</b> Dissecting and grasping forceps, slender, atraumatic fenestrated rotating 5mm size, length <b>33-36cm</b> with plastic handle with out ratchets
30	Page 54 point b).18	SCHNEIDER lymph node grasping forceps, atraumatic rotating 5mm size, length 36cm with plastic handle with out ratchets	Pridex	Lymph node grasping forceps, atraumatic rotating 5mm size, length 36cm with plastic handle with out ratchets .  SCHNEIDER – is company specific. Standard working length in Laparoscopy is 33 -36 cm	To be Amended as: <b>SCHNEIDER or similar</b> Lymph node grasping forceps, atraumatic rotating 5mm size, length <b>33-36cm</b> with plastic handle with out ratchets
31	Page 54 point b).19	DUVAL grasping forceps with reinforced bayonet coupling for use with outer sheath	Pridex	Grasping forceps with reinforced bayonet coupling for use with outer sheath  DUVAL – is company specific. Standard working length in Laparoscopy is 33 -36 cm	To be Amended as: <b>DUVAL or similar</b> Grasping forceps with reinforced bayonet coupling for use with outer sheath
32	Page 54 point b).23	Hemolock Clip applicator large clips 10mmx33cm, jaw size 15mm to fit interlockable clips	Karl storz	It is company specific, therefore request you to kindly amend this as "Hemolock <b>type</b> Clip applicator large clips 10mmx33cm, jaw size 15mm to fit interlockable clips".	To be Amended as: <b>Hemolock or Similar</b> Clip applicator large clips 10mmx33cm, jaw size 15mm to fit interlockable clips
33	Page 54 point b).24	Hemolock Clip applicator large clips 10mmx25cm,jaw size 15mm to fit interlockable clips	Karl storz	It is company specific, therefore request you to kindly amend this as "Hemolock <b>type</b> Clip applicator large clips 10mmx33cm, jaw size 15mm to fit interlockable clips".	To be Amended as: <b>Hemolock or Similar</b> Clip applicator large clips 10mmx33cm, jaw size 15mm to fit interlockable clips
34	Page 55 point b).25	MULTIFIRE laparoscopic clip applicator small 5mm, length 37cm	Pridex	Laparoscopic clip applicator small 5mm, length 37cm  MULTIFIRE is company specific – Standard working length in Laparoscopy is 33 – 36 cm	To be Amended as: Reusable Laparoscopic Clip Applicator 33-36cm for use with small clips - 2 nos.
35			Karl storz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	
36	Page 55 point b).26	MULTIFIRE laparoscopic clip applicator small 10mm, length 37cm	Pridex	Laparoscopic clip applicator small 10mm, length 37cm  MULTIFIRE is company specific – Standard working length in Laparoscopy is 33 – 36 cm	To be Amended as: Reusable Laparoscopic Clip Applicator 33-36cm for use with medium size clips - 2 nos.
37			Karl storz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	

NIB Ref: HITES/PCD/NCI-AIIMS/25/18-19

Schedule No. 7- Rfx/Event Number - 300003202, laparoscopy Instrument Set Qty-6

38	Page 55 point b).27	MULTIFIRE laparoscopic clip applicator medium 5mm, length 37cm	Pridex	Laparoscopic clip applicator medium 5mm, length 37cm  MULTIFIRE is company specific – Standard working length in Laparoscopy is 33 – 36 cm	<b>To be Amended as:</b> Reusable Laparoscopic Clip Applicator 33-36cm for use with medium large size clips - 2 nos.
39			Karl storz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	
40	Page 55 point b).28	MULTIFIRE laparoscopic clip applicator medium 10mm, length 37cm	Pridex	Laparoscopic clip applicator medium 10mm, length 37cm  MULTIFIRE is company specific – Standard working length in Laparoscopy is 33 – 36 cm	<b>To be Amended as:</b> Reusable Laparoscopic Clip Applicator 33-36cm for use with large size clips - 2 nos.
41			karl Stroz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	
42	Page 55 point b).29	MULTIFIRE laparoscopic clip applicator large 5mm, length 37cm	Pridex	Laparoscopic clip applicator large 5mm, length 37cm  MULTIFIRE is company specific – Standard working length in Laparoscopy is 33 – 36 cm	<b>To be Amended as:</b> Clips small, medium, medium large & large - 100 nos. each size for use clip applicators
43			Karl Stroz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	
44	Page 55 point b).30	MULTIFIRE laparoscopic clip applicator large 10mm, length 37cm	Pridex	Laparoscopic clip applicator large 10mm, length 37cm  MULTIFIRE is company specific – Standard working length in Laparoscopy is 33 – 36 cm	<b>Deleted</b>
45			Karl Stroz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	
46	Page 55 point b).31	MULTIFIRE laparoscopic clip applicator small 5mm, length 26cm	Pridex	Laparoscopic clip applicator small 5mm, length 26cm  MULTIFIRE is company specific – Standard working length in Pead. Laparoscopy is 20 – 26 cm	<b>Deleted</b>
47			Karl Stroz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	
48	Page 55 point b).32	MULTIFIRE laparoscopic clip applicator small 10mm, length 26cm	Pridex	Laparoscopic clip applicator small 10mm, length 26cm  MULTIFIRE is company specific – Standard working length in Pead. Laparoscopy is 20 – 26 cm	<b>Deleted</b>



NIB Ref: HITES/PCD/NCI-AIIMS/25/18-19

Schedule No. 7- Rfx/Event Number - 300003202, laproscopy Instrument Set Qty-6

49			Karl Stroz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	
50	Page 55 point b).33	MULTIFIRE laparoscopic clip applicator Medium 5mm, length 26cm	Pridex	Laparoscopic clip applicator Medium 5mm, length 26cm  MULTIFIRE is company specific – Standard working length in Pead. Laparoscopy is 20 – 26 cm	Deleted
51			Karl Stroz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	
52	Page 55 point b).34	MULTIFIRE laparoscopic clip applicator Medium 10mm, length 26cm	Pridex	Laparoscopic clip applicator Medium 10mm, length 26cm  MULTIFIRE is company specific – Standard working length in Pead. Laparoscopy is 20 – 26 cm	Deleted
53			Karl Stroz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	
54	Page 55 point b).35	MULTIFIRE laparoscopic clip applicator Large 5mm, length 26cm	Pridex	Laparoscopic clip applicator Large 5mm, length 26cm  MULTIFIRE is company specific – Standard working length in Pead. Laparoscopy is 20 – 26 cm	Deleted
55			Karl Stroz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	
56	Page 55 point b).36	MULTIFIRE laparoscopic clip applicator Large 10mm, length 26cm	Pridex	Laparoscopic clip applicator Large 10mm, length 26cm  MULTIFIRE is company specific – Standard working length in Pead. Laparoscopy is 20 – 26 cm	Deleted
57			Karl Stroz	Kindly specify multifire and whether single use or reusable as it is manufacutred by specific company.	

NIB Ref: HITES/PCD/NCI-AIIMS/25/18-19

Schedule No. 7- Rfx/Event Number - 3000003202, laproscopy Instrument Set Qty-6

58	Page 55 point b).37	LEROY H retractor, long jaws with silicon pads on the distal tip double action jaw, jaw open in parallel, autoclavable, ratchet engage with handle open 10mm, 36cm length includes a. Metal handle-with disintegrable ratchet b. Outer sheet- insulated c. H-retractor – insert d. Silicon Pad package of 6	Pridex	LEROY H retractor, long jaws with silicon pads on the distal tip double action jaw, jaw open in parallel, autoclavable, ratchet engage with handle open 10mm, 36cm length includes a. Metal handle-with disintegrable ratchet b. Outer sheet- insulated c. H-retractor – insert d. Silicon Pad package of 6  OR Equivalent retractor  LEROY is company specific, hence it should be omitted or Equivalent retractor should be requested	<b>To be amended as:</b> <b>LEROY H or Similar Retractor</b> , long jaws with silicon pads on the distal tip double action jaw, jaw open in parallel, autoclavable, ratchet engage with handle open 10mm, 36cm length includes a. Metal handle-with disintegrable ratchet b. Outer sheet- insulated c. H-retractor – insert <b>or similar</b> d. Silicon Pad package of 6
59	Page 55 point b).38	LEROY articulating H-retractor, long jaws, double action jaw, jaws open in parallel, bendable 90 deg, autoclavable, 10mm36cm length	Pridex	LEROY articulating H-retractor, long jaws, double action jaw, jaws open in parallel, bendable 90 deg, autoclavable, 10mm 36cm length OR Equivalent Retractor  LEROY is company specific, hence it should be omitted or Equivalent retractor should be requested	<b>To be amended as:</b> <b>LEROY or Similar Articulating H-retractor</b> , long jaws, double action jaw, jaws open in parallel, bendable 90 deg, autoclavable, 10mm36cm length
60	Page 55 point b).39	Handle attachment autoclavable, for use with LEROY retractors for secure hold	Pridex	Handle attachment autoclavable, for use with LEROY retractors for secure hold OR Handle for retractor  LEROY is company specific, hence it should be omitted or Equivalent retractor should be requested	<b>To be amended as:</b> Handle attachment autoclavable, for use with retractors for secure hold
61	Page 55 point b).40	BERCH facial closure instrument for subcutaneous ligature of trocar incision size 2.8cm length 17cm	Pridex	Facial closure instrument for subcutaneous ligature of trocar incision size 2.8cm length 17cm  BERCH is company specific	<b>To be amended as:</b> Facial closure instrument for subcutaneous ligature of trocar incision size 2.8cm length 17cm
62	Page 56.	Trocar & Cannula: Qty 10 nos. each	Pridex		

NIB Ref: HITES/PCD/NCI-AIIMS/25/18-19

Schedule No. 7- Rfx/Event Number - 3000003202, laproscopy Instrument Set Q-6

63	Page 56 point 1	Cannula size: 13.5 mm diameter; should have multifunctional valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. Trocar should have pyramidal tip with pin holes near the tip for safety outlet of CO2 gas. The working length of the cannula should be 115 mm.	Pridex	Cannula size: <b>12.5 to 13.5 mm</b> diameter; should have multifunctional valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. Trocar should have pyramidal tip with pin holes near the tip for safety outlet of CO2 gas. The working length of the cannula should be 115 mm.  Should be changed to 12.5 to 13.5, As all asked instruments are either 10 mm or 5 mm.	<p><b>To be amended as:</b> Cannula size: <b>12.5mm-13.5mm</b> diameter; should have multifunctional valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. Trocar should have pyramidal tip with pin holes near the tip for safety outlet of CO2 gas. The working length of the cannula should be 115 mm.</p>
64			DSS Imagetech Pvt Ltd	kindly amend it as <b>13mm</b> diameter; should have multifunctional valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. Trocar should have pyramidal tip. The working length of the cannula should be 80mm Or more.	
65	Page 56 point 2	Cannula size: 11 mm diameter; should have multifunctional valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. Trocar should have pyramidal tip with pin holes near the tip for safety outlet of CO2 gas. The working length of the cannula should be 105 mm.	DSS Imagetech Pvt Ltd	kindly amend it as 11mm diameter; should have multifunctional valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. Trocar should have pyramidal tip. The working length of the cannula should be <b>105mm -110mm</b> .	<p><b>To be amended as:</b> Cannula size: 11 mm diameter; should have multifunctional valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. Trocar should have pyramidal tip with pin holes near the tip for safety outlet of CO2 gas. The working length of the cannula should be <b>105 mm+/-5%</b></p>

Schedule No. 7- Rfx/Event Number - 3000003202, laparoscopy Instrument Set Qty-6

66	Page 56 point 3	Cannula size: 6mm diameter; should have multifunctional valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. Trocar should have pyramidal tip with pin holes near the tip for safety outlet of CO2 gas. The working length of the cannula should be 105 mm.	DSS Imagetech Pvt Ltd	kindly amend it as 5.5- 6mm diameter; should have multifunctional valve/lip seal valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. Trocar should have pyramidal tip. The working length of the cannula should be <b>105mm -110mm.</b>	<b>To be amended as:</b> Cannula size: <b>5.5-6mm</b> diameter; should have multifunctional valve to prevent damage of sharp instruments and tip lens while passing through the cannula valve. It should have stopcock for CO2 insufflation. Trocar should have pyramidal tip with pin holes near the tip for safety outlet of CO2 gas. The working length of the cannula should be <b>105 mm+/-5%</b>
67	Page 56 point 4	Double reducer, 13.5/10 & 13.5/5mm	Pridex	Double reducer, <b>12.5 - 13.5/10 &amp; 13.5/5mm</b>  Should be changed to 12.5 to 13.5, As all asked instruments are either 10 mm or 5 mm.	<b>No Change</b>
68	Page 57 ,point d.1	Veress Pneumoperitoneum needle with spring loaded blunt inner cannula, LUER-Lock, working length 10cm	DSS magetech Pvt Ltd	Kindly amend it as Veress Pneumoperitoneum needle with spring loaded blunt inner cannula, LUER-Lock, working length <b>8-10cm.</b>	<b>To be amended as:</b> Veress Pneumoperitoneum needle with spring loaded blunt inner cannula, LUER-Lock, working length <b>8-10cm</b>
69	Page 57 ,point d.2	Veress Pneumoperitoneum needle with spring loaded blunt inner cannula, LUER-Lock, working length 13 cm	DSS magetech Pvt Ltd	Kindly amend it as Veress Pneumoperitoneum needle with spring loaded blunt inner cannula, LUER-Lock, working length <b>12-13cm.</b>	<b>To be amended as:</b> Veress Pneumoperitoneum needle with spring loaded blunt inner cannula, LUER-Lock, working length <b>12-13 cm</b>
70	Page 57 point h.	<b>Laparoscopic Bipolar Forceps – Qty 2nos. each</b>	Pridex	<b>Laparoscopic Bipolar Forceps – Qty 2nos. each</b>	
71	Page 57 point h.1	Three piece laparoscopic Bipolar Coagulating Grasping Forceps, CLERMONT- FERRAND MODEL, Wide jaws, with connector pin for bipolar coagulation, 360 degree rotational sheeth, size 5 mm, length 36 cm, long, single action jaws, ergonomic Plastic handle with larger contact area, can be dismantled with the press of a button.	Pridex	Three piece laparoscopic Bipolar Coagulating Grasping Forceps, Wide jaws, with connector pin for bipolar coagulation, 360 degree rotational sheeth, size 5 mm, length 33 – 36 cm, long, single action jaws, ergonomic Plastic handle with larger contact area, can be dismantled with the press of a button  It's Company specific. Standard working length in Laparoscopy is 33 -36 cm	<b>To be amended as:</b> Three piece laparoscopic Bipolar Coagulating Grasping Forceps, CLERMONT-FERRAND MODEL, Wide jaws, with connector pin for bipolar coagulation, 360 degree rotational sheeth, size 5 mm, length <b>33-36 cm</b> , long, single action jaws, ergonomic Plastic handle with larger contact area, can be dismantled with the press of a button.

NIB Ref: HITES/PCD/NCI-AIIMS/25/18-19

Schedule No. 7- Rfx/Event Number - 3000003202, laparoscopy Instrument Set Qty-5

72			DSS Imagetech Pvt Ltd	kindly amend it as three piece laparoscopic Bipolar Coagulating Grasping Forceps, CLERMONT- FERRAND MODEL, Wide jaws, with connector pin for bipolar coagulation, 360 degree rotational sheeth, size 5 mm, length <b>33-36 cm</b> , long, single action jaws, ergonomic Plastic handle with larger contact area, can be dismantled with the press of a button.	
73	Page 57 point h.2	Three piece laparoscopic Bipolar KELLY Dissecting & Grasping Forceps, CLERMONT- FERRAND MODEL, Wide jaws, with connector pin for bipolar coagulation, 360 degree rotational sheeth, size 5 mm, length 36 cm, long, single action jaws, ergonomic Plastic handle with larger contact area, can be dismantled with the press of a button.	Pridex	Three piece laparoscopic Bipolar KELLY Dissecting & Grasping Forceps, Wide jaws, with connector pin for bipolar coagulation, 360 degree rotational sheeth, size 5 mm, length 33 - 36 cm, long, single action jaws, ergonomic Plastic handle with larger contact area, can be dismantled with the press of a button  It's Company specific. Standard working length in Laparoscopy is 33 -36 cm	<b>To be amended as:</b> Three piece laparoscopic Bipolar KELLY Dissecting & Grasping Forceps, CLERMONT- FERRAND MODEL, Wide jaws, with connector pin for bipolar coagulation, 360 degree rotational sheeth, size 5 mm, length <b>33-36 cm</b> , long, single action jaws, ergonomic Plastic handle with larger contact area, can be dismantled with the press of a button.
74			DSS Imagetech Pvt Ltd	kindly amend it as three piece laparoscopic Bipolar KELLY Dissecting & Grasping Forceps, CLERMONT- FERRAND MODEL, Wide jaws, with connector pin for bipolar coagulation, 360 degree rotational sheeth, size 5 mm, length <b>33-36 cm</b> , long, single action jaws, ergonomic Plastic handle with larger contact area, can be dismantled with the press of a button.	
1	Page 53 Point 8	Bidder should quote for all the Instruments of the set. At least 30% of the quoted instruments should be from same manufacturer.	M/s DSS Imagetech Pvt. Ltd.	Kindly amend it as Bidder should quote for all the Instruments of the set. At least 70% of the quoted Instruments should be from the same manufacturer.	<b>Amended as:</b> Bidder should quote for all the Instruments of the set. <b>At least 70% of the quoted instruments should be from same manufacturer.</b>

Response to Commercial Queries

2	Page 60 Point B.3 'For Laprosopy Instrument Set'	The bidder should have successfully executed at least 3 contracts for Surgical Instruments to any reputed hospital during last Five years from the date of Bid Opening.	M/s DSS Imagetech Pvt. Ltd.	Kindly amend it as the tenderers quoting as authorized representative of the manufacturer <b>should have executed at least one contract in the last five years from the date of tender opening of medical equipment anywhere in India</b> of the same manufacturer, similar to the major parameters. Similar to the PMSSY Terms & Conditions.	<b>To be amended as:</b> The bidder should have successfully executed at least <b>01 (one) contract</b> for <b>similar Surgical Instruments</b> to any reputed hospital during last Five years from the date of Bid Opening.
---	--	---	-----------------------------------	---	---

Schedule No. 7- Rfx/Event Number - 3000003202, laparoscopy Instrument Set Qty-6

3	Page 44 Part II	Supply, Installation and Commissioning to be completed within 90 days from the date of NOA or date of opening of LC or date of approval of layout drawing (in case applicable), whichever is later.	M/s. KARL STORZ ENDOSCOPY INDIA PVT. LTD.	Kindly amend this to "Supply to be completed within 90 days from the date of NOA or date of opening of LC" and installation to be done with in 45days of delivery of goods/date of site readiness.	<b>To be amended as:</b>  i) Supply to be completed within <b>75 days</b> from the date of NOA or date of opening of LC, whichever is later.  ii) Installation and commissioning shall be done within <b>15 days</b> of receipt of the stores/ goods at site or handing over the site for installation, whichever is later.
4	Page 37	Point 21. Terms and Mode of Payment 21.1 Payment TermsB) Payment for Imported Goods (M&E): a) On Shipment: 75% of the net FCA, CIP price (i.e. FCA/CIP price less Indian Agency commission) of the goods despatch by Sea/Air shall be paid through irrevocable, non-transferable Letter of Credit (LC) opened in favour of the supplier in a bank in his country and upon submission of documents specified hereunder: vii) Inspection certificate issued by the Purchaser's Inspector; (if applicable) viii) Any other document(s) as and if required in terms of the contract.	M/s. KARL STORZ ENDOSCOPY INDIA PVT. LTD.	Kindly clarify for points vii & viii or request you to kindly delete these points.	- For Point no. vii- Please refer GCC Clause no. 8.8 at Pg. no. 32 of the bidding document.  - Point no. viii is self explanatory
5	Page 44 and 58 'For Laparoscopy Instrument Set'	As per page 44, the warranty is 1year and CAMC not applicable As per page 58, warranty is 5years and CAMC also required.	M/s. KARL STORZ ENDOSCOPY INDIA PVT. LTD.	Kindly clarify for warranty & CAMC. Generally for instrument, warranty is 1year and CAMC is not required.	Please refer GCC Clause no. 15.5 at Pg. no. 34 and SCC Clause at Pg. no. 43 of the bidding document, which are self explanatory.