

# Techno-commercial Responsiveness & Non Responsiveness details – JULY 2016

Procurement & Consultancy Division: HITES – HLL Lifecare Limited

For Tender Ref No. HLL/PCD/PMSSY-II/NAGPUR/15/15-16 dated 29.01.2016							
Item No.	Name of Equipment	Name of Bidder	Commercial Responsiveness	Reason for Commercial Non-Responsiveness	Technical Responsiveness	Reason for Technical Non-Responsiveness	Overall Responsiveness
1	Digital X-ray Machine	M/s Carestream Health India Pvt. Ltd.	Responsive	N/A	Responsive	N/A	Responsive
		M/s Siemens Limited	Found technically Non- Responsive hence not evaluated further		Non-Responsive	Supplier did not confirm the following: 1KV step adjustment of KV settings, Table top material and make and model of monitor	Non-Responsive