

**AMENDMENT NO: 04 Dated 01.11.2022**

**AMENDMENT TITLE:**

Change in Technical Specification.

**REF NO:**HLL/SD/HCS/2022-23/Tender 18 dated 28<sup>th</sup> September 2022

E-TENDER FOR SUPPLY OF FULLY AUTOMATIC IMMUNOASSAY ANALYSER - HIGH THROUGHPUT FOR HLL LIFECARE DIAGNOSTIC LABORATORIES LOCATED AT VARIOUS LOCATION PAN INDIA ON REAGENT RENTAL BASIS

The following amendment has been incorporated to the bid document for the above tender;

In ANNEXURE 04, TECHNICAL SPECIFICATION AND PARAMETERS

<b>Sr. No.</b>	<b>Existing Technical Specifications as per Tender</b>	<b>Amendment</b>
6	Test Throughput should be more than 100- 200 samples per hour	Test Throughput should be more than 100- 200 tests per hour
14	Sample Volume: 1 to 250 µl per test, depending on assay protocol.	Sample volume: 10-250 ul per test or higher depending on assay protocol.

All relevant clauses of the tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

VICE PRESIDENT (SOURCING DIVISION)  
HLL LIFECARE LTD  
HLL BHAVAN, POOJAPURA  
THIRUVANANTHAPURAM