

25-04-2019

Amendment No. 7

Sub: Amendment to the referred tender enquiry

Ref.: Tender Enquiry HITES/PCD/AIIMS-IV/14/MGPS/18-19 dated 14/02/2019 read with Amendment No.01, 02, 03, 04, 05 & 06 Dated-19.3.2019 , 27.03.2019 ,29.03.2019 06.04.2019 , 09.04.2019 & 22.04.2019 respectively.

The following changes are being incorporated in the above referred Tender Enquiry Document.

SECTION - II
GENERAL INSTRUCTIONS TO TENDERERS (GIT)

For:

C. PREPARATION OF TENDERS

11. Documents comprising the e-Tender

A) Details of Technical Tender (Un priced Tender)

Bidders shall furnish the following information along with technical tender (in pdf format):

- xxi) A tenderer quoting imported goods located within India shall produce documentary evidence of the goods having been imported and already located within India (i.e. Bills of Entry for the quoted items and a self-declaration confirming that the quoted items were imported for the purpose of storage in bidder warehouse and for further sale), along with their techno-commercial bid.

Read As:

C. PREPARATION OF TENDERS

11. Documents comprising the e-Tender

B) Details of Technical Tender (Un priced Tender)

Bidders shall furnish the following information along with technical tender (in pdf format):

- xxii) A Tenderer quoting imported items in INR will have to submit necessary documents like
- a) Declaration confirming that the quoted items would be imported for the intended project (Tender reference number is to be quoted) only.
 - b) Bidder must submit Bill of Entry in the name of the project at the time of supply along with third party inspection certificate (from TUV/SGS/Lloyd/ Bureau Veritas) from the OEM country. The documents should clearly state the name of the intended Project (Tender reference and/or Project Name) along with the quantities.

SECTION - IV

GENERAL CONDITIONS OF CONTRACT (GCC)

For:

21. Terms and Mode of Payment

21.1 Payment Terms

Payment shall be made subject to recoveries, if any, by way of liquidated damages or any other charges as per terms & conditions of contract in the following manner.

A) Payment for domestic goods or goods of foreign origin located within India.

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

a) On delivery:

Seventy percent (70%) payment of the delivered goods price shall be paid on receipt of goods in good condition and upon the submission of the following documents:

- (iv) Inspection certificate issued by the nominated Inspection agency, if any.

Read As:

21. Terms and Mode of Payment

21.1 Payment Terms

Payment shall be made subject to recoveries, if any, by way of liquidated damages or any other charges as per terms & conditions of contract in the following manner.

A) Payment for domestic goods or goods of foreign origin offered in INR.

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

(v) On delivery:

Seventy percent (70%) payment of the delivered goods price shall be paid on receipt of goods in good condition and upon the submission of the following documents:

- (iv) For Domestic goods: Inspection certificate issued by the nominated Inspection agency, if any.
For goods of foreign origin offered in INR: Third Party Inspection certificate issued by TUV/ SGS/ Lloyd / Bureau Veritas from the OEM country along with bill of entry in the name of intended project (Tender reference And/or Project Name to be mentioned).

Section – VII
Technical Specifications

Tender Page & Para	TENDER SPECIFICATION	READ AS
Page 58 Para 5	Ward Vacuum Units	
Page 58 Para 5.2	Digital Suction Regulator: Suction regulator should be supplied with a safety jar, including and antibacterial filter and an anti-overflow safety device. Should have wide membrane continuous suction controller	Suction regulator(Digital/Analogue): Suction regulator should be supplied with a safety jar, including and antibacterial filter and an anti-overflow safety device. Should have wide membrane continuous suction controller. In case of digital suction regulator, battery should be replaced by the bidder during warranty & CMC period
Page 59 Para 6	Theatre Vacuum unit for OT	
Page 59 Para 6.1	Digital Suction Regulator and 2nos. 1500ml or more polysulfone/ polycarbonate collection jar and both to be mounted on a trolley.	Suction Regulator (Digital/Analogue) and 2nos. 1500ml or more polysulfone/ polycarbonate collection jar and both to be mounted on a trolley. In case of digital suction regulator, battery should be replaced by the bidder during warranty & CMC period
Page 59 Para 7	AGSS (Anesthetic Gas Scavenging System) Plant	
Page 59	The package should consist of rotary vane/claw type vacuum pumps (Dry/Oil less only) , a control panel, and mounted on a common base frame.	The package should consist of duplex rotary vane/claw type vacuum pumps or blower as applicable to standards , a control panel with automatic changeover , and mounted on a common base frame.
Page 59	AGSS pump: AGSS pump shall operate completely dry . Each pump should be completely air cooled and have absolutely no water requirements. The suitable wiring from OTs to AGSS plant for remote control/suitable reservoir (as applicable) is the responsibility of the bidder.	AGSS pump: Each pump/blower should be completely cooled and have absolutely no water requirements. The suitable wiring from OTs to AGSS plant for remote control/suitable reservoir (as applicable) is the responsibility of the bidder.
Pg 61 Para 11.1	Master Alarm (Digital)	

Tender Page & Para	TENDER SPECIFICATION	READ AS
Pg 61	Bidder shall be responsible for all cabling from local alarm panels(OTS & ICUs) to master alarm panel	Bidder shall be responsible for all cabling from local alarm panels(OTS & ICUs) to master alarm panel. Bidder should quote unit rate per meter interconnection cost and 1000m will be considered for ranking purpose. Payment will be made at actuals
		Added para in BOQ
		Interconnection from Local alarm panels to master alarm panel - qty 1000m

Section – IX Qualification Criteria

Added Para under Clause No 7:

Order copy along with installation certificates carried out by the bidder for the same / similar equipment (of the same manufacturer as quoted in the tender BOQ) should be submitted for at least one successful installation in India in last five years from the date of tender opening for all the items for which **Exclusive Manufacturer's Authorisation** has been asked in the tender.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. **Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**