

Amendment No.1

Sub: Amendment to Tender Enquiry Document.

Ref: NIT No. HLL/PCD/NEIFM-05/15-16 Dated 16.10.2015

The following amendments are issued with respect to the above Tender Enquiry Document.

SECTION-I NOTICE INVITING TENDERS (NIT)

Existing:

(1) **Tender No.: HLL/PCD/NEIFM-05/15-16**

Sl. No.	Description	Schedule
c	Closing date & time for submission of tender fee and EMD in physical form	20.11.2015, 1700 Hrs. (IST) Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
e	Closing date & time for submission of online bids	22.11.2015, 1700 Hrs. (IST)
f	Time, date and venue of e-tender opening of online bids	23.11.2015,1400 Hrs , HLL Life care Limited, Procurement & Consultancy Services Division B-14 A, Sector-62, Noida-201 307

Read As:

(1) **Tender No.: HLL/PCD/NEIFM-05/15-16**

Sl. No.	Description	Schedule
c	Closing date & time for submission of tender fee and EMD in physical form	01.12.2015, 1700 Hrs. (IST) Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
e	Closing date & time for submission of online bids	02.12.2015, 1700 Hrs. (IST)
f	Time, date and venue of e-tender opening of online bids	03.12.2015, 1400 Hrs , HLL Life care Limited, Procurement & Consultancy Services Division B-14 A, Sector-62, Noida-201 307

Section – VII
Technical Specification

Schedule-1
DIGITAL BRIX METER

Existing:

2. Refractive Index 0.00001
Sugar (°Brix) 0.01Accuracy
3. Refractive Index ± 0.00002
Sugar (°Brix) ± 0.02
6. Sucrose (°Brix) 5 – 80 °C

Read As:

- 2. Refractive Index 0.0001**
Sugar (°Brix) **0.1Accuracy**
- 3. Refractive Index ± 0.0002**
Sugar (°Brix) **± 0.1**
- 6. Sucrose (°Brix) 5 – 50 °C or more**

Schedule-7
TLC SETUP WITH PLATES

Added Para: Scanner for quantitative analysis-(Price to be quoted separately).

Schedule-8
500 mA HIGH FREQUENCY X-RAY UNIT

Existing:

- 2) X-Ray Tube & Collimator & Ceiling Suspension:

Read As:

- 2) X-Ray Tube, Collimator & Floor to ceiling column stand:

Note: All other contents of the Tender Enquiry Document including terms and conditions of the tender enquiry remain unchanged.