

AMENDMENT – I TO MINUTES OF THE MEETING

TENDER FOR SUPPLY, INSTALLATION, COMMISSIONING AND VALIDATION OF LAB EQUIPMENTS AT PII, COONOR


Document No. : NPI/110831/EQP/TD/15 Dated: 26.02.2016
Venue : HLL Biotech Limited, Chennai
Pre-bid Date : 07.03.2016 – 10.03.2016
Project : Revival of DPT Vaccines manufacturing facility
Issued on : 22nd March'2016
Issued from : Head - Procurement - HBL

Agenda

- | | |
|----|---|
| 1. | Pre-bid Meeting for Lab equipment's at PII, Conoor. |
|----|---|

Tender for Supply , Installation, commissioning and Validation of Lab Equipment's at PII, Coonoor					
A Discussion on Tender Enquiry Document: NPI/110831/EQP/ TD/15					
S. No. Clarifications on URSS					
Schedule No 4. Weighing balance (Addition/Clarification)					
1.	<ul style="list-style-type: none"> a. E1 class (1mg to 5g – 16 Nos.) with NABL accreditation certification -1 set b. F1 class (1mg to 200g – 23 Nos.) with NABL accreditation certification -1 set c. F1 Class individual weights (500g-1 No., 1 Kg-1No., 2 Kg -2 Nos., 5 Kg -1 No., 10 Kg -1 No., 20 Kg-2 Nos.) with NABL accreditation certification -1set d. MOC of all the weights mentioned above must be SS 316 L and the boxes for the weights must be compactable to the clean room application. e. The Batch Report must have a header, Footer for signature, Batch details with date and time along with the weights. 				
Schedule No 19. pH Meter					
	<table border="1"> <thead> <tr> <th>DS point</th> <th>Point modified as</th> </tr> </thead> <tbody> <tr> <td>1. 3.1Glass electrode The glass electrode must be made from Borosilicate glass</td> <td>3.1 Electrode type The electrode must be gel filled type Separate temperature probe to be provided</td> </tr> </tbody> </table>	DS point	Point modified as	1. 3.1Glass electrode The glass electrode must be made from Borosilicate glass	3.1 Electrode type The electrode must be gel filled type Separate temperature probe to be provided
DS point	Point modified as				
1. 3.1Glass electrode The glass electrode must be made from Borosilicate glass	3.1 Electrode type The electrode must be gel filled type Separate temperature probe to be provided				

For HLL Biotech Limited


22/3/16.

Head - Procurement