

Amendment No. 1

Date: 25/02/2016

Sub: Amendment to the Tender Enquiry Document**Ref: NIT No.: HLL/PCD/CDSCO-03-RT/15-16 Dated 02.02.2016**

The following changes are being incorporated in the referred tender enquiry.

**SECTION I
NOTICE INVITING TENDERS (NIT)****FOR:**

Sl. No.	Description	Schedule
c	Closing date & time for submission of tender fee and EMD in physical form	26.02.2016, 1500 hrs (IST) Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
d	Closing date & time for submission of online bids	26.02.2016, 1400 hrs (IST)
e	Time and date of opening of online bids	26.02.2016, 1600 hrs (IST)

READ AS:

Sl. No.	Description	Schedule
c	Closing date & time for submission of tender fee and EMD in physical form	04.03.2016, 1500 hrs (IST) Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
d	Closing date & time for submission of online bids	04.03.2016, 1400 hrs (IST)
e	Time and date of opening of online bids	04.03.2016, 1600 hrs (IST)

Section – VII**Technical Specification****Near Infra Red (NIR) Analyzer****1. Existing Specification:**

Para: Built in Library - The instrument should have minimum 1000 Reference spectra of drugs and must have a provision of development of user reference library that can be used as Reference library for overlay comparison.

Read as:

Built in Library - The instrument should have **the ability to store reference library of minimum 750 spectra drugs based on samples tested at the site.**

2. Existing Specification:

Para: Mode of use - The Instrument can measure directly through plastic drum liners and glass containers permitting analysis without opening the packaging.

Read as:

Mode of use - Instrument can identify the sample over the transparent package and glass containers permitting analysis without contaminating the sample.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids