

16-11-2019

Amendment No. 4

Ref.: Tender Enquiry: HITES/PCD/MP/CLINICAL/RC-05/19-20 dated 26.09.2019 and subsequent amendment(s).

Sub: Amendment No. 4 to the referred tender enquiry.

The following changes are being incorporated in the above referred Tender Enquiry Document

**SECTION-VII
TECHNICAL SPECIFICATIONS**

Tender Page & Para	Existing specification	Read As
Sch. 1 Laparoscopy set with hand instrument set		
Page 46		
•	Laprosopic right angled liga clip applicator -1no. (from a reputed make)	Deleted
Sch.7 Low Temperature Plasma Sterilizer 50 litre capacity		
Page 58		
8	Deleted	Added Para:- Lumen sterilization claims should be validated and endorsed by USFDA/510K / European CE/ BIS
Sch.8 Urodynamic machine with Uroflometry		
Page 60		
17	The Urodynamic should be supplied with 3D Bladder scanner with following specifications: - It Should be supplied by manufacturer of Urodynamic system only. - 3D volume calculations – PVR saved into Urodynamic application automatically. - Manual bladder correction mode. - It Should be Real-time, save, easy and non-invasive bladder scanning - Auto location and auto tracking of the bladder in real-time. - Full-page reports in JPEG and Dicom	Deleted
Sch.9 C-ARM IMAGE INTENSIFIER		
Page 60		
3	C) X-RAY GENERATOR: Fluoro& Rad. Kv 40 to 120 KV.	Fluoro & Rad. Kv 40 to 110 KV. Or more
Page 62		
•	Pre-Processing Features • Pulse fluoroscopy facility with Frame rate up to 30fps for fluoro& cinemodes. • DICOM 3.0 version.	Pulse fluoroscopy facility with Frame rate up to 25fps or more for fluoro Deleted
Sch.12 Upper & Lower Urinary tract instruments		
Page 66		
	Grasping Forceps for large stone fragments, double action jaws, 5 Fr, Flexible/rigid, length 55-60 cms	Grasping Forceps for stone fragments, double/single action jaws compatible with Supplied URS
	System should have inbuilt vessel sealing capability or should be upgraded to vessel sealing in near future.	System should have inbuilt vessel sealing capability

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.
- ii. If EMD is submitted in the form of BG, then the validity of the BG should be at least 165 days from the date of tender opening