

Amendment No. 5**Date: 23/07/2014****Subject: Amendment to the Tender Enquiry Document****Ref: NIT ref.: HLL/PCD/GNCTD/05/GBPH/14-15 dated 04/06/2014****Tender ID: 2014_HFWD_58409_4 and 2014_HFWD_58409_6**

The following changes have been incorporated in the referred tender.

**SECTION I
Notice Inviting Tender****Existing:**

Sl. no.	Tender ID	Description	Qty	Tender Fees (Rs.)	EMD Amount (Rs.)	Date & time of Prebid meeting	Date & time of closing of online tender	Closing date & time for submission of physical Tender	Date & time of opening of tender
4	2014_HFWD_58409_4	High Speed Pneumatic Craniotome and Drill System	1	500	38,000	09-06-14	24-07-14	25-07-14	25-07-14
						11:00 AM	6:00 PM	1:00 PM	2:00 PM
6	2014_HFWD_58409_6	High Speed Electrical Craniotome and Drill System	1	1,000	80,000	09-06-14	24-07-14	25-07-14	25-07-14
						11:00 AM	6:00 PM	1:00 PM	2:00 PM

Amended as:

Sl. no.	Tender ID	Description	Qty	Tender Fees (Rs.)	EMD Amount (Rs.)	Date & time of Prebid meeting	Date & time of closing of online tender	Closing date & time for submission of physical Tender	Date & time of opening of tender
4	2014_HFWD_58409_4	High Speed Pneumatic Craniotome and Drill System	1	500	38,000	09-06-14	30-07-14	31-07-14	31-07-14
						11:00 AM	6:00 PM	1:00 PM	2:00 PM
6	2014_HFWD_58409_6	High Speed Electrical Craniotome and Drill System	1	1,000	80,000	09-06-14	30-07-14	31-07-14	31-07-14
						11:00 AM	6:00 PM	1:00 PM	2:00 PM

**SECTION VII
Technical Specification****Item no.4****High speed pneumatic craniotome and drill system**

- Existing:** The motor should run in forward, reverse and oscillating mode

Amended as: The motor should run in forward mode

Item no.6
High speed electrical craniotome and drill system

1. **Existing:** Should have an electrically powered hand-piece for rapid, controlled tissue removal with a single entry point, designed specifically for nucleotomy or hard tumors

Amended as: DELETED

2. **Existing:** Should be dedicated hand piece for Recip Saw speed (14000cpm to 17000cpm,) Oscillating saw speed (18000 cpm to 20000cpm) and sagittal saw speed (17000 to 20000cpm)

Amended as: DELETED

3. **Existing:** Blade tip should rotate 360 degree independently of shaft.

Amended as: DELETED

4. **Existing:** Should have multiple option consisting of both straight and curved blades.

Amended as: DELETED

5. **Existing:** Blade tip should have designed to protect against annular penetration

Amended as: DELETED

6. **Existing:** System should have inbuilt irrigation and suction facility.

Amended as: DELETED

7. **Existing:** Straight and angled blade with dia of 4.0mm and 4.5mm should be quoted (2 pieces each)

Amended as: DELETED

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

1. **Bidders who have already prepared their EMD in form of bank guarantee as per the requirement of the tender enquiry for submission of bid on 14/07/2014 may submit their tender with the same bank guarantee without any further extension of its validity to meet the revised schedule of submission for the above items only.**
2. **Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.**