

Amendment No: 10 Issued on dated 16/05/2015

Ref: Tender Enquiry No: HLL/PCD/RIMS-04/14-15 and Pre-Bid meeting held on 02.02.2015 at HLL Lifecare Ltd., B-14A, Sector-62, Noida - 201 307

The following amendments are issued to the above Tender Enquiry Document with respect to only for **Item no.3** as below:

Item No.	Description	Qty.(Nos)	EMD (In Rs)
3	C Arm (System Local)	3	72,000

Section – VII

Technical Specification

Item no-3: C-ARM (SYSTEM LOCAL)

1. Existing Specification :

Generator: Microprocessor controlled generator with 2.0Kw or more with integrated beam filters to reduce patient skin radiation dose.

Read as:

Generator: Microprocessor controlled generator **with 1.4 kW or more** with integrated beam filters to reduce patient skin radiation dose.

2. Existing Specification :

X ray mode (kV & mA range): Fluoroscopy should not exceed 5 mA

Read as:

X ray mode (kV & mA range): Fluoroscopy **should not exceed 9 mA.**

3. Existing Specification:

Image Processing: (a) Archival memory CD/DVD mode.

Read as:

Image Processing: (a) Archival memory CD/DVD mode or through USB. (If through USB, desktop computer of latest configuration i3 or above with 4GB or more and 500GB HDD, CD/DVD with software should be supplied as standard.)

4. Added para :

- 1. System should be European CE or USFDA certified.**
- 2. System should be AERB type approved & ELORA registered.**

All other contents of the Tender Enquiry Document including terms and conditions of the tender enquiry remain unchanged.