

AMENDMENT NO: 04 Dated 01.11.2022

AMENDMENT TITLE:

Change in technical Specification

REF NO: HLL/SD/HCS/2022-23/Tender 19 dated 28th September 2022

E-TENDER FOR SUPPLY OF FULLY AUTOMATIC IMMUNOASSAY ANALYSER - LOW THROUGHPUT FOR HLL LIFECARE DIAGNOSTIC LABORATORIES LOCATED AT VARIOUS LOCATION PAN INDIA ON REAGENT RENTAL BASIS

The following amendment has been incorporated to the bid document for the above tender;

In ANNEXURE 04, TECHNICAL SPECIFICATION AND PARAMETERS

Sr. No.	Existing Technical Specifications as per Tender	Proposed amendment
13	Sample Volume: 1 to 250 µl per test, depending on assay protocol.	Sample Volume: 10 to 150 µl per test or higher, depending on assay protocol.

All relevant clauses of the tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

VICE PRESIDENT (SOURCING DIVISION)
HLL LIFECARE LTD
HLL BHAVAN, POOJAPURA
THIRUVANANTHAPURAM