

**Amendment No. 1**

17.12.2019

**Sub: Amendment to the Bidding Document****Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/44/19-20 dated 05.12.2019.**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

**SECTION - VII****TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item no. 33 (Tender ID: 2019\_HLL\_37262\_33)****Biological X-Ray based Blood Irradiator**

<b>Tender Page &amp; Para</b>	<b>Existing Specification</b>	<b>Amended as</b>
Page 87, Para 6	Centre dose rate should be between 2.5-5 Gys per min	Centre dose rate should be <b>2.5 Gy per minute or more.</b>
Page 87, Para 10	The system may include a positioning function for beam and specimen alignment.	<b>Deleted</b>
Page 87, Para 15	The system MUST be accompanied by a calibrated dosimeter, for dose and dose rate measurements within the irradiation chamber. The firm will also have to supply 1000 radiation tags along with any other quality control requirements including dose mapping in every 6 months till the life cycle of the equipment. The rates for the same must be quoted in the tender for 200,000 radiation tags for fixing of rates for 10 years.	The system MUST be accompanied by a calibrated dosimeter, for dose and dose rate measurements <b>with</b> the irradiation chamber. The firm will also have to supply 1000 radiation tags along with any other quality control requirements including dose mapping in every 6 months till the life cycle of the equipment. The rates for the same must be quoted in the tender for 200,000 radiation tags for fixing of rates for 10 years.

**B. GENERAL POINTS:**

***(Added Para for all the tendered items under the referred NIB where-ever applicable)***

**Para 7.**

IQ, OQ, PQ or instrument performance verification (IPV) should be provided by the bidder for all the quoted equipment (where applicable).

**Important Note:**

Technical Compliance Sheet for tender ID- **2019\_HLL\_37262\_33** 'Ver1.1' under this NIB is being uploaded under Corrigendum on CPPP. This will supersede the previous version i.e. 'Ver1.0' and the same need to be downloaded for necessary submission.

**All other contents of the Bidding Document including terms & conditions remain unaltered.**