

Amendment No. 6

Date: 28/03/2016

Sub: Amendment to the Tender Enquiry Document

Ref: NIT No.: HLL/PCD/PMSSY-II/NAGPUR/15/15-16 dated 29.01.2016 for Digital X-Ray Machine

SECTION I NOTICE INVITING e-TENDERS (NIT)

The schedule for submission of tenders against the referred NIT has been revised as follows.

Sl.	Description	Schedule
	i) Closing date & time for submission of online bids	08-April -2016, 1200 hrs IST
d	ii) Closing date & time for tender fee, EMD and bid documents in physical form (Price bid shall only be uploaded online; All other technical and commercial documents uploaded in the e-portal are to be submitted)	08- April -2016, 1400 hrs IST Bidders have to submit Original Bank Instruments viz. DD/BC/BG of tender fee and EMD within the above mentioned date and time
e	Time and date of opening of bids	08- April -2016, 1500 hrs IST

Note:

(i) If EMD is submitted in the form of BG, then the validity of the BG should be at least 165 days from the date of tender opening, i.e., upto 20 September 2016.

(ii) Tender fee (Rs. 5,000/-) and EMD (As applicable) should be deposited in the Tender Box located at **HLL Lifecare Limited, Procurement and Consultancy Division, B-14 A, Sector-62, Noida-201307, Uttar Pradesh** on or before **08-April-2016, 1200 hrs IST**. Submission beyond stipulated date & time would result in **REJECTION** of BID.

SECTION – VII TECHNICAL SPECIFICATIONS

Schedule No. 1

SPECIFICATION FOR X-RAY UNIT WITH DIGITAL FLAT PANEL DETECTOR

1. Existing Para 2.II (After Amendment): Detector Panel should be made of amorphous Silicon with CsI

Read As: Detector Panel should be made of amorphous Silicon with **Gadox**

Existing Para 2.VIII: DQE at 0lp/mm should be at least 65% or more

Read As: DQE at 0lp/mm should be at least **27% or more**