

## **Amendment No. 3**

**Date: 30/11/2015**

**Subject: Amendment to the Tender Enquiry Document**

**Ref: Tender Enquiry No.: HLL/HITES/PCD/RC-ME/01/2015 dated 06/10/2015 read with amendment no. 1 & 2 dated 21.11.2015 & 23.11.2015 respectively.**

Part of the technical specification for item at sl. no. 1 (i.e. radiant warmer) was missing in the existing Tender Enquiry Document due to typographical error. The same has been incorporated as under.

<b>SECTION VII Technical Specification</b>
--

### **Group: Special Neonatal Care Unit**

#### **Item No. 01 Radiant Warmer**

#### **Added Para**

##### **2. Technical Characteristics**

##### **2.1 Technical characteristics (specific to this type of device)**

1. It should be microcontroller based radiant warmer with manual and servo options.
2. It should have facility to display skin set, skin observed temperature in degree C and heat power separately.
3. Should have user friendly touch panel control.
4. It should have ceramic or quartz infrared or calrod heater.
5. It should have audiovisual alarm facility for overheating beyond set temperature range.
6. It should have alarm facility for patient temperature less than or greater than the required temperature i.e. above or below the set range. Machine should sense the skin probe failure and cut of the heater.
7. Warmer head should be rotatable in different direction, so as to allow taking X-ray.
8. It should have alarm for probe failure, power failure, system failure and heater failure.

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

#### **Note:**

**Prospective bidders are advised to check the website regularly prior to the closing date and time of online submission of tenders.**

*-- End --*