

Amendment No. 1

AMENDMENT TO THE PREBID MEETING MINUTES FOR SUPPLY,
INSTALLATION COMMISSIONING AND VALIDATION OF ULTRA
FILTRATION SYSTEMS AND STERILE FILTRATION SYSTEMS AT
PASTEUR INSTITUTE OF INDIA, COONOOR

Ref: Tender no. NPI/110831/EQP/ TD/04 dtd 30.06.2014

With reference to the tender for the supply, installation, commissioning and Validation of Microfiltration system at Pasteur institute of India, Coonoor, the following amendment may please be noted:

FOR:

**Ultrafiltration System -NPI_110831_EQP_URS_B1(D)-UFS 01 and
NPI_110831_EQP_URS_B2-UFS 01**

2.1 Operating Conditions:

- Surface area of the membrane: 0.5 / 0.7m² with a provision to increase the filter area to 7 m²/14m² (size of modules)
- Minimum working Volume: 0.5 % of Total Volume

READ AS:

**Ultrafiltration System -NPI_110831_EQP_URS_B1(D)-UFS 01 and
NPI_110831_EQP_URS_B2-UFS 01**

2.1 Operating Conditions:

- Surface area of the membrane: 0.5 / 0.7m² with a provision to increase the filter area to 7 m²/14m² (size of modules)
- Minimum working Volume: 0.5 % of Total Volume
- **SIP of the ultrafiltration system to be provided and it should have necessary arrangements.**

All other terms and conditions of the tender document and prebid meeting minutes remain the same.