

AMENDMENT NO.3 Dated 26.05.2020 for change in specification of N95 Mask in tender Ref No: TENDER NO: HLL/SD/RBD/2019-20/TENDER/15 DT:05.03.2020 WHICH HAS NOW BEEN CLOSED, FRESH REQUEST FOR PROPOSAL FOR THE SUPPLY OF PERSONAL PROTECTIVE EQUIPMENTS - GOGGLES, N95 FACE MASK WITH / WITHOUT EXPIRATORY VALVE AND NITRILE GLOVES floated on 11 May 2020.

The following amendment has been incorporated to the bid document for the above tender:

Specification

N-95 Masks (With / Without valve)

- Shape that will not collapse easily
- High filtration efficiency
- Good breathability, with expiratory valve
- Quality compliant with standards for surgical N95 respirator:
 - a. NIOSH N95, EN 149 FFP2, or equivalent
- Fluid resistance: minimum 80 mmhg pressure based on ASTM F1862, ISO 22609, or equivalent.
- Quality Compliant with standards for particulate respirator that can be worn with full face shield

MAY BE READ AS

N-95 Masks (With / Without valve)

- Shape that will not collapse easily
- High filtration efficiency
- Good breathability, with expiratory valve
- Quality compliant with standards for surgical N95 respirator:
 - a. NIOSH N95, EN 149 FFP2, BIS or equivalent
- Fluid resistance: minimum 80 mmhg pressure based on ASTM F1862, ISO 22609, or equivalent.
- Quality Compliant with standards for particulate respirator that can be worn with full face shield

All other relevant specifications and clauses of the tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

**Vice President
Sourcing Division
HLL Lifecare Ltd.
HLL Bhavan, Poojappura,
Thiruvananthapuram.**

Offers may be submitted to:
sdcovidep.hll@lifecarehll.com / sdcovidep.hll@gmail.com
Contact: 0471 2354949 Ext: 323 (9.15 AM to 5.45 PM)