

**Amendment No. 3**

**Date: 10/10/2014**

**Sub: Amendment to Tender Enquiry Document with respect to Tender IDs: 2014\_HFWD\_65918\_1 and 2014\_HFWD\_65918\_47.**

**Ref: NIT No.: HLL/PCD/GNCTD/14/BSAH/14-15 dated 09.09.2014 read with its amendment no. 1 & 2 dated 01/10/2014 & 04/10/2014 respectively.**

The following changes have been incorporated in the referred NIT.

**Section – VII  
Technical Specifications**

**Sl. No. 1  
AUTOMATED ANALYZER**

**Existing:**

**Para 2.2:** Should be capable of undertaking 160-200 tests/hr involving fixed time, end point and kinetic chemistry

**Amended as:**

Should be capable of undertaking **at least 200 tests/hr** involving fixed time, endpoint and kinetic chemistry

**Sl. No. 47  
MICROSCOPES, OIL IMMERSION**

**Existing:**

**Para 5: Focusing:** Revolving nosepiece Quintuple/Quintuple

**Amended as:**

**Focusing:** Revolving nosepiece **Quadruple/Quintuple**

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

**Note:**

Prospective bidders are advised

1. to ensure the validity of your EMD as per amendment no. 2 dated 04.10.2014.
2. to check the website regularly prior to the closing date and time of online submission of tenders.