

**Amendment No. 4**

15.06.18

**Sub: Amendment to the Bidding Document**

**Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/14/18-19 dated 13.04.2018 read with its Amendment no. 01, 02 & 03 dated 11.05.18, 28.05.18 & 08.06.2018 respectively.**

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

**SECTION - I****NOTICE INVITING BIDS (NIB)**

<b>Description</b>	<b>Existing</b>	<b>Amended as</b>
Last date and time of online submission of tender	19.06.2018 at 12:00 noon	<b>05.07.2018</b> at 12:00 noon
Last date and time of physical submission of EMD, Tender processing Fee, any other document specified in the Bidding Document	19.06.2018 at 2:00 pm	<b>05.07.2018</b> at 2:00 pm
Date of tender Opening	19.06.2018 at 2:30 pm	<b>05.07.2018</b> at 2:30 pm

**SECTION - VII  
TECHNICAL SPECIFICATION AND GENERAL POINTS**

**A. TECHNICAL SPECIFICATION:****Item No. 8 (Rfx/Event number 3000002784)****Video Cystoscope**

<b>Sl. No</b>	<b>Reference to the bidding document</b>	<b>Existing Specification/ Clause of the bidding documents</b>	<b>Amended as</b>
1	Sl. no. 1, page 23, Amendment No.2	<b>HD</b> Video Cysto-Urethroscope	Video Cysto Urethroscope
2	Sl. no. 3, page 23, Amendment No.2	Outputs - <b>RGB, Y/C, VBS Composite, XGA</b> & DVI simultaneous.	Outputs-Digital: DVI/HDMI/SDI/HDSDI
3	Sl. no. 4.3 page 68,	It should have structure and <b>edge enhancement</b> option for	It should have structure enhancement for better image

	Original Biding Document	better image quality	quality
4	Sl. no. 4.4 page 68, Original Biding Document	It should have various iris control option for better light distribution	It should have various iris control option/ <b>brightness control</b> for better light distribution.
5		<b>BOQ for Video Cystoscope</b>	
	Sl. no. 1 page 69, Original Biding Document	<b>HD</b> Video Cysto-urethroscope as per technical specifications	Video Cysto-urethroscope as per technical specifications

**Item No. 11 (Rfx/Event number 3000002787)**  
**Hysteroscope System**

<b>Sl. No</b>	<b>Reference to the bidding document/ Amendment</b>	<b>Existing Specification/ Clause of the bidding documents</b>	<b>Amended as</b>
1	Sl. no. 04, page 24, Amendment no. 2	Added Para:	<p><b>Following Para added:</b>  ESU with 300W or more (of any reputed make) output power, Should have Monopolar cut (pure, blend) &amp; coagulation (soft, forced, spray), Bipolar Cut (cut, Saline cut) Coagulation (soft coag, auto coag, saline coag), Should be compatible with all monopolar &amp; bipolar hysteroscope instruments and resectoscopes  Should be European CE or USFDA certified</p> <p>Should be supplied with footswitch for all applications requested, disposable monopolar pen-20, bipolar forceps straight 3 nos. &amp; bayonet 3 nos., disposable patient plate adult - 50 nos., disposable patient plate pediatric - 10 nos.</p> <p>All compatible accessories for therapeutic &amp; diagnostic hysteroscope work should be supplied along with ESU unit</p>
2		Added item in BOQ	BOQ Sl.no.17: Electro Surgical Unit as per technical specifications - 1 no.
3		Added item in BOQ	BOQ Sl.no.18: disposable monopolar pen-20
4		Added item in BOQ	BOQ Sl.no.19: bipolar forceps straight 3 nos. & bayonet 3 nos.

5		Added item in BOQ	BOQ Sl.no.20: disposable patient plate adult - 50 nos.
6		Added item in BOQ	BOQ Sl.no.21: disposable patient plate pediatric - 10 nos.

**Item No. 15 (Rfx/Event number 300002791)  
Rigid Cysto-Resectoscope**

Sl. No	Reference to the bidding document	Existing Specification/ Clause of the bidding documents	Amended as:
1		Added Para:	<p><b>Following Para added:</b> ESU with 300W or more (of any reputed make) output power, Should have Monopolar cut (pure, blend) &amp; coagulation (soft, forced, spray), Bipolar Cut (cut, Saline cut) Coagulation (soft coag, auto coag, saline coag), Should be compatible with all monopolar &amp; bipolar hysteroscope instruments and resectoscopes Should be European CE or USFDA certified</p> <p>Should be supplied with footswitch for all applications requested, disposable monopolar pen-20, bipolar forceps straight 3 nos. &amp; bayonet 3 nos., disposable patient plate adult - 50 nos., disposable patient plate pediatric - 10 nos.</p> <p>All compatible accessories for therapeutic &amp; diagnostic hysteroscope work should be supplied along with ESU unit</p>
2		Added item in BOQ	BOQ Sl.no.29: Electro Surgical Unit as per technical specifications - 1 no.
3		Added item in BOQ	BOQ Sl.no.30: disposable monopolar pen-20
4		Added item in BOQ	BOQ Sl.no.31: bipolar forceps straight 3 nos. & bayonet 3 nos.
5		Added item in BOQ	BOQ Sl.no.32: disposable patient plate adult - 50 nos.
6		Added item in BOQ	BOQ Sl.no.33: disposable patient plate pediatric - 10 nos.

**All other contents of the Bidding Document including terms & conditions remain unaltered.**